

Evaluation of Calibration Performance of a Low-cost Particulate Matter Sensor Using Collocated and Distant NO₂

¹, Kabseok Ko¹, Seokheon Cho², and Ramesh R. Rao²

¹Department of Electronics Engineering, Kangwon National University, Chuncheon, 24341, Korea

²Qualcomm Institute, University of California, San Diego (UCSD), La Jolla, CA, 92093, USA

Correspondence: Seokheon Cho (justinshcho@gmail.com)

Abstract. Low-cost optical particle sensors have the potential to supplement existing particulate matter (PM) monitoring systems to provide high spatial and temporal resolution. However, low-cost PM sensors have often shown questionable performance under various ambient conditions. Temperature, relative humidity (RH), and particle composition have been identified as factors that directly affect the performance of low-cost PM sensors. This study investigated if NO₂, which creates PM_{2.5} by chemical reactions in the atmosphere, can be used to improve the calibration performance of low-cost PM_{2.5} sensors. To this end, we evaluated the PurpleAir PA-II, called PA-II, a popular air monitoring system that utilizes two low-cost PM sensors that is frequently deployed near air quality monitoring sites of the Environmental Protection Agency (EPA). We selected a single location where 14 PA-II units have operated for more than two years since July 2017. Based on the operating periods of the PA-II units, we then chose the period of Jan. 2018 to Dec. 2019 for study. Among the 14 units, a single unit containing more than 23 months of measurement data with a high correlation between the unit's two PMS sensors was selected for analysis. Daily and hourly PM_{2.5} measurement data from the PA-II unit and a BAM 1020 instrument, respectively, were compared using the federal reference method (FRM), and a per-month analysis was conducted against the BAM-1020 using hourly PM_{2.5} data. In the per-month analysis, three key features, temperature, relative humidity (RH), and NO₂, were considered. The NO₂, called collocated NO₂, was collected from the reliable instrument collocated with the PA-II unit. The per-month analysis showed the PA-II unit had a good correlation (coefficient of determination, $R^2 > 0.819$) with the BAM-1020 during the months of Nov., Dec., and Jan. in both 2018 and 2019, but their correlation intensity was moderate during other months, such as July and Sep. 2018, and Aug., Sep., and Oct. 2019. NO₂ was shown to be a key factor in increasing the value of R^2 in the months when moderate correlation based on only PM_{2.5} was achieved. This study calibrated a PA-II unit using multiple linear regression (MLR) and random forest (RF) methods based on the same three features used in the analysis studies as well as their multiplicative terms. The addition of NO₂ had a much larger effect than that of RH when both PM_{2.5} and temperature were considered for calibration in both models. When NO₂, temperature, and relative humidity were considered, the MLR method achieved similar calibration performance to the RF method. In addressing the feasibility of utilizing distant NO₂ measurements for calibration in lieu of collocated data, the study highlights the effectiveness of distant NO₂ when correlated strongly with collocated measurements. This finding offers a practical solution for situations where obtaining collocated NO₂ data proves challenging or costly. We assessed the performance of different PA-II units to determine their efficacy. Our investigation reveals a significant enhancement in calibration performance across different PA-II units upon integrating NO₂. Importantly, this improvement

remains consistent even when employing models trained with different PA-II units within the same location. Overall, this investigation emphasizes the significance of NO₂ in improving calibration for low-cost PM_{2.5} sensors and presents insights into leveraging distant NO₂ measurements as a viable alternative for calibration in the absence of collocated data.

30 1 Introduction

Recently, attention has been paid to particulate matter (PM), which not only has adverse effects on visibility but also can impact human health by contributing to conditions such as cardiovascular disease, asthma, and lung cancer (Liu et al., 2018, 2013). PM that is less than 2.5 μm in diameter, referred to as PM_{2.5}, can penetrate the lungs and may thus increase the risk to human health. Globally, the estimated number of adult deaths attributable to PM_{2.5} exposure is over 0.67, 1.6, and 2.1 million for lung cancer, 35 cardiopulmonary disease, and all causes, respectively (Evans et al., 2013). To minimize the harmful effects, many countries regulate daily and annual PM_{2.5} concentrations by monitoring PM_{2.5} levels at air quality monitoring stations. The monitoring stations use instruments based on Federal Reference Methods (FRMs) or Federal Equivalent Methods (FEMs), which promote high precision and accuracy. The U.S. Environmental Protection Agency (EPA) approves both FRMs and FEMs as official designations for measuring ambient concentrations. Furthermore, the U.S. EPA carries out various cooperative programs, 40 including those on ambient monitoring methods and technologies, with many other countries in the world. These instruments can provide high-quality measurements of PM_{2.5} concentrations at the installed locations and nearby surroundings. However, these instruments are sparsely distributed due to the high cost of the equipment (ten thousand to tens of thousands of US dollars), so they cannot provide spatial variability. In other words, traditional monitoring stations frequently provide air quality data with poor spatio-temporal resolution, due to the limited number of high quality instruments.

45 As a cost-effective approach for a dense monitoring network, many stakeholders and researchers have turned to low-cost PM sensors that use a light scattering technique for measurement. In addition to low cost, these sensors have the advantages of low energy consumption and high sampling frequency, and they are easy to deploy and operate compared to traditional monitoring networks. Thus, low-cost PM sensors have been deployed in several communities to measure and report local air quality information (Jiao et al., 2016; PurpleAir, 2018).

50 However, low-cost PM sensors are not suitable for regulatory purposes because the data reported can be questionable in terms of accuracy, precision, and reliability. In worst-case scenarios, low-cost sensors report no meaningful data at all. Because manufacturers provide limited information on sensors' performance, some studies have been conducted to evaluate the performance of a variety of low-cost sensor models by comparing them with high-cost instruments in laboratory and outdoor ambient environments (Alvarado et al., 2015; Johnson et al., 2018; Wang et al., 2015; Holstius et al., 2014; Austin et al., 2015; 55 Gao et al., 2015; Kelly et al., 2017; Mukherjee et al., 2017; Sousan et al., 2016; Feinberg et al., 2018; Crilley et al., 2018; Badura et al., 2018; Budde et al., 2018; Liu et al., 2019; Cavaliere et al., 2015; Kelly et al., 2017; Zheng et al., 2018). Most sensors showed good performance under laboratory tests where the sensors measured, known concentrations of particles, such as polystyrene latex, in a chamber. On the other hand, under ambient conditions, the performance of low cost sensors varied depending on the sensor model and its deployed location. Some PM sensor units have inconsistent precision between units of

60 the same model (Feenstra et al., 2019; Feinberg et al., 2018), while other PM monitors, including the PurpleAir PA-II, have shown good precision (Barkjohn et al., 2020; Pawar and Sinha, 2020; Mailings et al., 2020). Field evaluations of PurpleAir PA-II units collocated with FEM instruments for approximately two months have shown good correlation with the FEM instruments (SCAQMD, 2017c). Furthermore, it was shown that PMS5003 sensors, which are used in PurpleAir PA-II monitors, have a good correlation with the FEM monitors (Kelly et al., 2017; Sayahi et al., 2019). However, the sensors still require
65 calibration for better performance before use in ambient conditions.

Several studies have developed calibration models for low-cost PM sensors based on the following approaches: simple linear regression (Zheng et al., 2018), multiple linear regression (Zimmerman et al., 2018), random forest (Zimmerman et al., 2018), and neural networks (Si et al., 2020). Moreover, to improve calibration performance, several studies have identified other factors in addition to $PM_{2.5}$ concentration that can affect the performance of low-cost sensors. These typical factors
70 include temperature, relative humidity, and particle properties (composition and size distribution) (Holstius et al., 2014; Gao et al., 2015; Kelly et al., 2017). In particular, some low-cost PM sensors have been shown to excessively overestimate $PM_{2.5}$ concentrations under high relative humidity conditions (Jayaratne et al., 2018). The reason for this overestimation is that some aerosols can uptake water via hygroscopy. To solve this problem, several correction models have been proposed, such as a correction model based on the κ -Köhler theory (Crilley et al., 2018, 2020), multiple linear regression (Barkjohn et al., 2021;
75 Nilson et al., 2022), and generalized additive model (Hua et al., 2021). Analysis of direct factors, such as temperature, relative humidity, and particle composition, can enhance the performance of low-cost sensors. In addition to these direct factors, we examine the impact of precursor gas NO_2 , acting as a source of $PM_{2.5}$ emissions, on calibration performance in low-cost $PM_{2.5}$ sensors. In general, $PM_{2.5}$ arises by secondary formation from a chemical reaction between precursor gases, such as NO_2 , in the atmosphere some distance downwind from the original emission source (Hodan et al., 2004). This study aims
80 to identify the significance of the precursor NO_2 and evaluate its potential for improving the performance of low-cost $PM_{2.5}$ sensors. To this end, we considered two machine learning methods, Multiple Linear Regression (MLR) and Random Forest (RF), for calibration models using various feature vectors, including temperature, relative humidity, and NO_2 . The trained MLR and RF models were evaluated on the test set, and their performance was compared. From an implementable perspective on NO_2 data, we investigated the feasibility of using data from distant NO_2 regulatory instruments due to the questionable data
85 quality of low-cost NO_2 sensors. The results of our study showed that incorporating distant NO_2 , in addition to temperature and relative humidity, into RF models yields lower errors than RF models that only include temperature and relative humidity.

2 Methods

2.1 Measurement data

2.1.1 PurpleAir PA-II Units

90 The PurpleAir PA-II Outdoor air quality monitor was developed for measuring particulate matter of various sizes. PA-II units can measure various particulate matter as well as temperature, relative humidity, and barometric pressure. PurpleAir also devel-

oped a crowdsourcing platform to share publicly gathered PM measurements obtained from all PA units. From the PurpleAir website (<https://www.purpleair.com/map>), we can observe and download data reported by all installed PA units.

95 A PA-II unit includes two identical PMS 5003 sensors. The PMS 5003 sensors based on a light scattering principle measure concentrations of $PM_{1.0}$, $PM_{2.5}$, and PM_{10} in real-time. By counting the number of particles per each diameter range that flows through a fan at a rate of 0.1L/min. Based on number of particles counted per diameter, each sensor estimates $PM_{1.0}$, $PM_{2.5}$, and PM_{10} concentrations and then averages the concentrations every 80 s¹. The PA-II unit sends the averaged concentrations obtained from two PMS sensors (A and B) to the PurpleAir server without storing the data in the unit itself. The PA-II unit does not calibrate the data, which implies it just collects the measured data.

100 The PurpleAir website provides the following information about all PA-II units via a JSON formatted file: a name, a unique ID, a latitude, a longitude, and an installation date. Each PA-II unit has two unique IDs for each of its PMS sensors A and B.

2.1.2 Air quality measurement data from EPA

Outdoor air quality data collected from across the U.S. is publicly available through the U.S. Environmental Protection Agency (EPA) website (<https://epa.gov/outdoor-air-quality-data>). Monitoring ambient air quality for purposes of determining compli-
105 ance with the U.S. National Ambient Air Quality Standards (NAAQSs) requires the use of either FRMs or FEMs. FRM and FEM instruments are accepted for methods for monitoring the NAAQS pollutants, such as particulate matters ($PM_{2.5}$ and PM_{10}), NO_2 , SO_2 , O_3 , and CO . Hourly measurements of $PM_{2.5}$ and PM_{10} , as well as other pollutants such as NO_2 , SO_2 , O_3 , and CO , obtained from FEM and non-FEM instruments can be downloaded via the EPA's application programming inter-
110 face (<https://aqs.epa.gov/data/api>) (U.S. EPA, 2011). Daily measurements of $PM_{2.5}$ obtained from an FRM instrument are also available.

2.1.3 Selection of PA-II units and reference monitoring sites

To investigate the performance of a PA-II unit itself and evaluate its calibration, we focused on PA-II units that are installed close to an EPA monitoring site (i.e. reference site) that provides reliable hourly $PM_{2.5}$ concentrations. We use the location information of the PA-II units and reference monitors to find PA-II and reference monitor pairs that are located less than 100 m
115 from each other (Wallace et al., 2021). Among the identified pairs, we selected a monitoring site, located at Rubidoux, CA, that has 14 PA-II units as pairs and can measure other pollutants such as NO_2 on an hourly basis. The monitoring site is identified by a state code of 06, a county code of 065, and a site number of 8001 (i.e., 06-065-8001). This monitoring site is located in an urban residential area within the south coast air basin at an elevation of 248 m. Air pollutants from the Los Angeles and coastal areas are transported to this air basin, which is known to have poor ventilation and may experience air stagnation during the
120 early evening and early morning periods. Local air pollution includes NO_x from diesel trucks, since the city of Jurupa Valley, which includes the community of Rubidoux, is a main transportation corridor for diesel trucks serving three air cargo terminals and the ports of Los Angeles and Long Beach.

¹After May 30, 2019, the averaging time is changed from 80 s to 120 s.

Table 1 describes information about the 14 PA-II units, such as their IDs, location (latitude and longitude), sensor name, start time of measurement, end time of measurement, and non-operating months². While we present the ID for only PMS sensor A of each PA-II unit, the ID of PMS sensor B is the ID of PMS sensor A plus 1. The geographic information on 14 PA-II units and the monitoring site is shown in Figure S1. Distances between PA-II units and the monitoring site are shown in Table S1. The minimum and maximum distance between a PA-II unit and the monitoring site is less than 10 m and 100 m, respectively.

Based on the non-operating months of the PA-II units found, we selected an appropriate period of sample data from Jan. 2018 to Dec. 2019 (24 months). Among the 14 identified PA-II units, we chose several that had more than 23 months of valid measurement data during the period selected for study. The selected units are RIVR_Co-loc2, 3, 5, 6, 7, and 8, which we call PA-II 2, 3, 5, 6, 7, and 8, respectively.

Before using PM_{2.5} data from the PA-II units, we checked the unit's data quality. We calculated the correlation among the selected PA-II units, considering both PMS 5003 sensors for each PA-II unit for the correlation analysis. Since these PA-II units are closely located, PM_{2.5} data should be highly correlated. Figure 1 shows the correlation results for all PMS 5003 sensors included in the PA-II units. The numbers on each axis represent the number of the selected PA-II units. Boxes to the left and right of each number indicate PMS sensors A and B for its corresponding PA-II unit, respectively. The PMS sensor A of PA-II unit 2, PMS sensors A and B of PA-II unit 5, and PMS sensor A of PA-II unit 6 all have a poor correlation with other PMS sensors. In addition, sensor A of PA-II unit 3 has slightly poor correlation with other sensors. Based on these results, we selected PA-II units 7 and 8.

2.1.4 Data preprocessing of PA-II units

The PA-II units selected for study are long-term installations, i.e., they have been in operation for more than two years. Therefore, PA-II units may have abnormal data due to failure and aging drift, so data quality control is required before calibrating the PA-II units. The quality control (QC) measure has been shown to be important for developing correction models of PA-II units (Barkjohn et al., 2021). They performed a QC measure by obtaining daily PM_{2.5} measurement data, but we applied the QC measure to obtain hourly PM_{2.5} measurement data. The QC measure has the following 3 steps: i) data from both channels A and B was removed when either channel A or B had a missing value, ii) data with abnormal temperature or relative humidity values was removed, and iii) data from channels A and B were compared. In the first step, when we calculate 1-hour averages of PM_{2.5} measurements generated with 2 min (or 80 sec) intervals, we remove the 1-hour average if the number of PM_{2.5} measurements is less than 27 (or 40). We considered two different measurement intervals for a PA-II unit because its old interval had been 80 sec until May 30, 2019. Its current interval is 2 min. After calculating 1-hour average data, we removed all data points for the 1-hour interval, where either sensor A or B had a missing value. The second step deals with temperature and RH data. PA-II units occasionally report extremely high or low values of temperature and relative humidity that are inaccurate. Therefore, we removed the data points whose corresponding time interval contained unrealistic measurement of temperature or relative humidity. In this study, the acceptable ranges of temperature and RH are (0 °F, 200 °F) and (0%, 100%), respectively. Once the unacceptable data points were removed, we calculated the 1-hour average for temperature and RH. The last step was

²We define non-operating month as the month, when the number of days without the measurement data is larger than 10 days.

to compare results sensors A and B in a PA unit to check data consistency. To do this, we used symmetric percentage error (SPE) as follows:

$$SPE = \frac{2(|PM_{2.5}^A| - |PM_{2.5}^B|)}{|PM_{2.5}^A + PM_{2.5}^B|}, \quad (1)$$

where $PM_{2.5}^A$ and $PM_{2.5}^B$ are hourly averaged $PM_{2.5}$ concentrations from sensors A and B in the same PA-II unit, respectively.

160 We removed the relevant data points with SPE larger than 0.61, which is 2 standard deviation. This value of SPE threshold has been used for 24-hr average $PM_{2.5}$ concentrations (Barkjohn et al. (2021)), but we use it here for 1-hour averaged $PM_{2.5}$ concentrations. The number of data points processed for each pre-processing step in PA-II 7 is summarized in Table S2.

The period of valid measurement data collected from the PA-II units we selected is 24 months, such as from Jan. 2018 to Dec. 2019. The measurement data in the years 2018 and 2019 from the two-year dataset were used for training and testing
165 for our calibration models, respectively. The reason why we split the two-year dataset at a 1:1 ratio is that $PM_{2.5}$ as well as the other environmental parameters, such as temperature and relative humidity, which we considered for calibration models, have a seasonal pattern. Also, we used whole-year dataset for training, to learn the relationship between PA-II and regulatory measurement over seasonality and thus enhance the performance of the calibration models over all 4 seasons.

2.2 Instrument intercomparisons

170 The monitoring site we considered has an FRM instrument and a BAM-1020 instrument with the parameter of 88502. These instruments produce daily and hourly $PM_{2.5}$ measurement data, respectively. Since we measure the PA-II units at intervals much shorter than a full day, it is much more reasonable to compare the $PM_{2.5}$ measurement of PA-II units with that of a BAM-1020 instrument with a shorter measurement interval, rather than that of an FRM instrument for evaluating the accurate calibration performance of PA-II units. However, we face the limitation that a BAM-1020 instrument can be classified as a non-
175 FEM-compliant device. Therefore, our approach for analyzing PA-II units to appropriately resolve these issues is as follows: we compared the BAM-1020 instrument's readings with daily $PM_{2.5}$ concentrations collected from an FRM instrument to ensure the BAM-1020 provides an acceptable level of performance as an FRM instrument, which is enough to assess the calibration performance of PA-II units. According to this affirmative observation, the BAM-1020 instrument can be used to evaluate the calibration performance of low-cost $PM_{2.5}$ sensors by comparing its readings with hourly $PM_{2.5}$ measurement data of PA-II
180 units.

We compared daily and hourly $PM_{2.5}$ measurement data obtained from an FRM and BAM-1020 instruments and PA-II 7 unit. Table 3 shows summary statistics of daily and hourly $PM_{2.5}$ measurement data from FRM and BAM instruments and PA-II 7³. These data suggest that a BAM-1020 instrument using non-FEM methods compares well to the statistics achieved with the FRM method. However, the measurements are not enough to evaluate how similar the performance of the BAM-1020
185 is to that of the FRM instrument. Hence, this study compared the performance of two instruments using a linear fitting scheme.

³A PMS 5003 sensor that collects $PM_{2.5}$ concentrations from within a PA-II unit exhibits a maximum consistency error of $\pm 10 \mu\text{g}/\text{m}^3$ at $0\text{-}100 \mu\text{g}/\text{m}^3$ and $\pm 10\%$ at $100\text{-}500 \mu\text{g}/\text{m}^3$. The sensor reports $PM_{2.5}$ concentrations as integer values on a per-second basis. A PA-II unit generates readings of its own $PM_{2.5}$ concentrations by averaging its 1-second $PM_{2.5}$ concentrations over 80 (or 120) seconds.

Figure 2 shows the calibration performance using linear regression. The R^2 , slope, and intercept are 0.896, 0.923, and 0.741, respectively. Also, the value of RMSE is $2.211 \mu\text{g}/\text{m}^3$. The BAM-1020 is close to an FEM instrument with the parameter of 88101. In order for the BAM-1020 to attain the 88101 code in terms of performance, the following conditions must be satisfied: R^2 is larger than 0.9, a slope is larger than 0.9 and less than 1.1, and an absolute value of the intercept is less than 190 2.0. Slope and intercept are satisfied with the requirement, while R^2 does not meet the condition very slightly. Nonetheless, the BAM-1020 instrument provides an acceptable level of performance to evaluate the calibration performance of PA-II units on an hourly basis.

Compared to the FRM and BAM-1020 instruments, the PA-II 7 unit overestimates the maximum daily $\text{PM}_{2.5}$ concentrations. Additionally, the mean daily $\text{PM}_{2.5}$ concentration from the PA-II 7 unit was higher than that of the FRM and BAM-1020 195 instruments. These results show that the PA-II unit has a good correlation (r) with the FRM instrument for the two-year period of interest, since its value is very close to 1. However, a comparison of metrics from the FRM instrument and the PA-II 7 unit did not correlate as favorably.

Next, we compared the PA-II unit's hourly $\text{PM}_{2.5}$ data with that of the BAM-1020 instrument over the course of the same two-year period. We did not consider the FRM instrument for exploring hourly $\text{PM}_{2.5}$ measurement data, since it only produces 200 daily concentrations. The PA-II unit's maximum hourly $\text{PM}_{2.5}$ measurement was almost twice that of the BAM-1020. In other words, the PA-II unit overestimates hourly $\text{PM}_{2.5}$ concentrations. Figure 3 shows the comparison of $\text{PM}_{2.5}$ measurement data obtained from the BAM-1020 and the selected PA-II 7 unit, as well as temperature and relative humidity measured from the selected PA-II 7 unit during winter season (from Dec. 2018 to Feb. 2019). The PA-II 7 unit showed a similar trend of $\text{PM}_{2.5}$ concentration measurements to that of the BAM-1020 instrument, but it generally overestimated hourly $\text{PM}_{2.5}$ concentrations 205 more often than the BAM-1020.

In addition, we compared the hourly $\text{PM}_{2.5}$ concentrations of the PA-II unit with that of the BAM-1020 instrument in terms of RMSE, MSE, MAE, and r . The results are as follows: RMSE of $6.194 \mu\text{g}/\text{m}^3$, MSE of $38.369 \mu\text{g}/\text{m}^3$, MAE of $7.919 \mu\text{g}/\text{m}^3$, and r of 0.876. The PA-II unit had a good correlation with the BAM-1020 instrument based on r . However, other metrics, such as RMSE, MSE, and MAE, did not correlate well.

210 2.3 Feature selection for calibration models

Temperature and relative humidity have been identified in previous studies as key factors for effective calibration. In particular, relative humidity has been shown to affect low-cost PM sensors under high relative humidity conditions. Furthermore, few papers have considered NO_2 in calibration models (Hua et al., 2021) because NO_2 , which is known to be a precursor to the formation of $\text{PM}_{2.5}$ through chemical reactions in the atmosphere, may indirectly affect $\text{PM}_{2.5}$ concentrations. Therefore, we 215 investigated the suitability of temperature, relative humidity, and NO_2 for the calibration of the PA-II 7 unit.

To identify the independent variables relevant for calibration, we conducted a correlation analysis involving $\text{PM}_{2.5}$ measurements from BAM-1020 and PA-II 7 unit readings, as well as temperature and relative humidity data, spanning a two-year period. The results are illustrated in Figure S2. The highest correlation was observed between $\text{PM}_{2.5}$ from BAM-1020 and

PA-II 7 unit, followed by NO₂ measurements. Subsequently, relative humidity and temperature exhibited the next level of correlation. As a result, we have identified temperature, relative humidity, and NO₂ as the selected candidate features.

To explore the potential for enhancing the calibration performance of low-cost PM sensors using temperature, relative humidity, and NO₂ as features, we conducted linear fitting. Before considering temperature, relative humidity, and NO₂, we evaluate the monthly performance based on hourly PM_{2.5} data from the PA-II 7 unit compared to the BAM-1020 instrument. Table 2 shows the value of R^2 , RMSE, and MAE of hourly PM_{2.5} measurement data from the PA-II 7 unit compared to that of the BAM-1020 instrument and the corresponding slope and intercept of each optimal linear fitting. During the months of Nov., Dec., and Jan., the PA-II unit is shown to have a high correlation, R^2 of 0.813 to 0.936, with the BAM-1020 instrument. This result is supported by the field evaluation of PA-II units taken by the Air Quality Sensor Performance Evaluation Center (AQ-SPEC) during the period of Dec. 2016 - Jan. 2017, which showed the value of R^2 as being 0.868 to 0.921 when the PA-II units were compared with the FEM. (Sayahi et al., 2019) showed that PMS sensors have a high correlation with tapered element oscillating microbalances (TEOM) instruments during the winter season by providing R^2 of 0.866 to 0.892. That is, the hourly PM_{2.5} measurement data from PA-II units seem to be highly correlated with that of FEM instruments during the months of November, December, and January, which implies the PM_{2.5} measurement performance of PA-II is reliable, especially during winter seasons. These months have different slopes and intercepts; for example, Jan. 2018 has a slope of 0.502 and an intercept of 3.898, while Jan. 2019 has 0.397 and 1.961, respectively.

On the other hand, the PA-II 7 unit has a correlation lower than 0.6 for months of Jul. and Sep. 2018 as well as Aug., Sep., and Oct. 2019. These months, except Sep. 2019, have larger RMSE values compared to other months over the two-year period, which need to be calibrated.

For multiple features, such as temperature, relative humidity, and NO₂, we used an MLR approach for regression analysis of PA-II units compared to the BAM-1020 instrument. A per-month analysis was conducted based on hourly PM_{2.5} measurements from the PA-II 7 unit under several feature vectors, such as (PM_{2.5}), (PM_{2.5}, T), (PM_{2.5}, RH), (PM_{2.5}, NO₂), (PM_{2.5}, T, RH), and (PM_{2.5}, T, NO₂), where T and RH represent temperature and relative humidity, respectively. For notational simplicity, we defined the above feature vectors (PM_{2.5}), (PM_{2.5}, T), (PM_{2.5}, RH), (PM_{2.5}, NO₂), (PM_{2.5}, T, RH), and (PM_{2.5}, T, NO₂) as 1, 2, 3, 4, 5, and 6, respectively. Figure 4 shows the R^2 and RMSE results of multiple linear regression for selected months with the above varying feature vectors. We considered feature vector 1 as a baseline for comparison among other feature vectors. On Jan. 2018, feature vector 5, referring to temperature and relative humidity, had little effect on the regression performance of R^2 and RMSE. The amount of R^2 increase by feature vector 5 from the baseline was around 0.001, and the amount of RMSE decrease was 0.038 $\mu\text{g}/\text{m}^3$. In the case of feature vector 6, including NO₂ instead of RH, R^2 increased from the baseline by 0.015, while RMSE was improved by 0.518 $\mu\text{g}/\text{m}^3$. Similarly, for Apr. 2018, R^2 (or RMSE) for feature vector 5 increased (or decreased) by 0.01 (or 0.072 $\mu\text{g}/\text{m}^3$) compared to its baseline. R^2 and RMSE for feature vector 6 increase by 0.05 and decrease by 0.52 $\mu\text{g}/\text{m}^3$ from the baseline, respectively. For regressions in Aug. and Sep. 2019, an increase in R^2 was larger than 0.17 when feature vector 6 was considered, but it was less than 0.07 when feature vector 5 was considered. These remarkable results suggest that NO₂ is generally a key factor that can improve the performance of PA-II units over a year, even though

the enhancement by NO₂ does not meet the values of 0.7 of R^2 and 3.5 $\mu\text{g}/\text{m}^3$ of RMSE during certain months, such as July 2018, August 2019, October 2019.

255 2.4 Calibration methods

A per-month analysis with a combination of features, including T, RH, and NO₂, showed an effect on calibration for the PA-II unit. However, it is challenging to use the per-month linear fitting result to calibrate PA-II units because each month has a different slope and intercept defined for the linear fitting. Moreover, their values exhibit a change over the years. For example, notably, the linear fitting result in Apr. 2018 exhibited a higher RMSE than the fitting result in Apr. 2019. On the contrary, the
260 calibration performance in Aug. 2018 was worse than that in Aug. 2019.

We used a machine learning approach to develop a calibration model, employing two machine learning algorithms, such as multiple linear regression (MLR) and random forest (RF). For both calibration methods, we considered various combinations of features, including PM_{2.5} measured from a PA-II unit, temperature, relative humidity, NO₂, and their multiplicative interaction terms.

265 2.4.1 Multiple linear regression (MLR)

An MLR method can be expressed as follows:

$$\hat{y} = \beta_0 + \beta_1 x_1 + \cdots + \beta_n x_n, \quad (2)$$

where \hat{y} represents a response, n is the number of predictor variables, β_i for $i = 0, 1, \dots, n$ are regression coefficients, and x_i for $i = 1, 2, \dots, n$ represent predictor variables (called features). Using a linear equation with multiple variables, we investigated
270 the relationship between features and a response.

All features in an MLR method should be independent. However, many studies have considered PM_{2.5}, temperature, and RH, which are not independent (Magi et al. (2019); Mailings et al. (2020)). Some studies have introduced multiplicative interaction terms (i.e., PM_{2.5} × RH) to exploit interdependence between features (Barkjohn et al. (2021)). We also consider multiplicative interaction terms in this study.

275 We use PM_{2.5} concentrations obtained from a reference monitor as the response. As predictor variables, we consider multiple features, such as PM_{2.5} measurement data from a PA-II unit, temperature, relative humidity, NO₂, and their multiplicative interaction terms (i.e., PM_{2.5} × RH, T × RH, PM_{2.5} × RH × T).

2.4.2 Random forest (RF)

An RF is an ensemble of K regression trees. Each regression tree is trained with a bootstrap sample of an original training
280 dataset. The output of an RF is the aggregation of regression trees, i.e., averaging estimates over all trees. Each regression tree is grown by selecting random m features among M input features at each possible split. The best cut is calculated at the randomly chosen features. Optimal cuts can be achieved using the Classification and Regression Trees split criterion (CART),

which compares the variance of the uncut node and one of all possible cuts along m directions. Every tree is fully grown with these splits (Breiman, 2001).

285 2.5 Performance evaluation metrics

In this study, we examined the root mean square error (RMSE), mean squared error (MSE), mean absolute error (MAE), and Pearson correlation coefficient r between daily $PM_{2.5}$ data from the FRM instrument and that from the PA-II units. In the cases of the RMSE, MSE, and MAE, the lower its value is, the better the performance or the lower the difference in measurement data between the FRM instrument and the PA-II units. The Pearson correlation coefficient is a metric measuring a linear correlation
 290 between two variables. It is a number between -1 and 1 that measures the strength and direction of their relationship. As the coefficient approaches an absolute value of 1, the values of measurement data from the FRM instrument and the PA-II units become more similar. These performance metrics are expressed as follows:

$$RMSE = \sqrt{\frac{1}{n} \sum_{i=1}^n (x_i - y_i)^2}, \quad (3)$$

$$MSE = \frac{1}{n} \sum_{i=1}^n (x_i - y_i)^2, \quad (4)$$

$$295 \quad MAE = \frac{1}{n} \sum_{i=1}^n |x_i - y_i|, \quad (5)$$

where x_i represents 1-hour averaged (24-hour period) sensor $PM_{2.5}$ concentrations for the i th hour (day) ($\mu\text{g}/\text{m}^3$), y_i represents 1-hour averaged (24-hour period) FRM or BAM-1020 $PM_{2.5}$ concentrations for the i th hour (day) ($\mu\text{g}/\text{m}^3$), and n is the number of data points.

3 Results and discussions

300 3.1 Calibration performance

The two-year dataset was divided into training and test sets at a 1:1 ratio, meaning the measurement data in the years 2018 and 2019 were used for training and testing, respectively. We used the training set to learn calibration models based on MLR and RF, and then used the test set to evaluate the calibration performance in terms of RMSE, MAE, and R^2 . A calibration performance for the PA-II 7 unit using MLR and RF methods was compared with several features, including temperature,
 305 relative humidity, and NO_2 , as well as their multiplicative terms.

3.1.1 MLR-based calibration model

Recently, calibration methods have employed multiplicative interaction terms, such as $PM_{2.5} \times \text{RH}$ and $T \times \text{RH}$. In our MLR models, we considered both additive and multiplicative interaction terms. The additive terms in our models include raw PurpleAir $PM_{2.5}$, T, RH, and NO_2 . We considered multiplicative interaction terms that involve less than four additive terms when

310 NO₂ was not included (i.e., we consider PM_{2.5}×T×RH), and less than three additive terms when NO₂ is included. There are 95 combinations of features. Out of 95 combinations tested, only 52 combinations had a p-value of less than 0.05. Of those, we select 21 combinations, among 52 combinations, by increasing the number of additive terms and the number of multiplicative interaction terms and identifying the combinations with the lowest RMSE among the same numbers of additive terms and multiplicative interaction terms. The selected combinations were shown in Table 4.

315 The calibration results of the PA-II 7 unit for test datasets using the MLR method with 21 selected combinations are presented in Table 5. Multicollinearity is a known issue with MLR models, as it can cause instability. One common method to diagnose this issue is to use the variance inflation factor (VIF) test for multicollinearity (Mansfield and Helms , 1982). Out of the 21 combinations tested, most VIF values were less than 5, indicating the absence of collinearity issues.

When a single additive term, such as T or RH, was applied, the RMSE values for two combinations, #2 and #3, improved by
320 more than 0.208 µg/m³, compared to considering only PM_{2.5}. The inclusion of an additive RH term in an MLR yielded a lower error than an additive T term did, since both RMSE and MAE for combination #3 were less than those for combination #2. The MLR model with PM_{2.5}, the single additive term with RH, and its multiplicative interaction term with PM_{2.5} yielded similar RMSE and MAE to the MLR model using PM_{2.5} and two meteorological variables, such as T and RH, as demonstrated by the results of combinations #4 and #5. When we considered two meteorological variables and incorporated four multiplicative
325 interaction terms, such as PM_{2.5}×T, PM_{2.5}×RH, and T×RH, the MLR model resulted in the lowest error with an RMSE of 4.151 µg/m³ and an MAE of 3.023 µg/m³, compared to all combinations generated from PM_{2.5}, T, RH, and their multiplicative terms.

The MLR model of combination #10 with PM_{2.5} and NO₂ had an RMSE of 4.424 µg/m³, which was lower than that of the MLR model with only PM_{2.5}, whose RMSE was 4.513 µg/m³, but larger than that of combination #2 with a single
330 environmental variable and an RMSE of 4.305 µg/m³. This implies that the addition of a single multiplicative term in that model has no performance enhancement. However, when the additive term T is incorporated into an MLR model with PM_{2.5} and NO₂, an RMSE of 3.997 µg/m³ can be achieved, which is lower than the values of all combination cases, not including NO₂, i.e., combinations #1 to #9. Coefficients of PM_{2.5}, T, and NO₂ in the MLR model, including T and NO₂, were around 0.446, 0.110, and 0.112, respectively. The temperature had more impact on error than relative humidity when considering NO₂.
335 Considering both temperature and relative humidity together with NO₂ may cause a non-zero correlation of relative humidity with other factors due to a p-value of 0.083. When some multiplicative terms were additionally integrated into T, RH, and NO₂, the MLR calibration models passed a p-value test. The model based on combination #18 with four additive terms, i.e., PM_{2.5}, T, RH, and NO₂, and multiplicative interaction terms, including PM_{2.5}×RH and T×RH, achieved the lowest RMSE of 3.912 µg/m³. Considering multiplicative terms with T and RH had little effect on calibration performance as shown in the results of
340 combinations #15, #19, and #20. From these results, we conclude that considering NO₂ together with meteorological variables and their multiplicative terms or a single variable, such as temperature, can improve the calibration performance of PA-II units.

3.1.2 RF-based calibration model

This study validated performance of RF-based calibration for PA-II units with 95 combinations of predictor variables mentioned in the previous subsection. An RF was implemented using the scikit-learn package in Python. An RF has several hyperparameters, such as `n_estimators`, `max_depth`, `min_samples_leaf`, and `max_features`, that need to be set for the best performance over each combination of features. For this study, the hyperparameters were tuned with a random search method by 5-fold cross-validation based on the training set. For a random search, the number of trees (`n_estimators`) was set to 10, 20, 50, 100, 200, and 400. The range of `max_depth` was set to 2, 4, 6, 8, 10, 16, and None. The range of `min_samples_leaf` was set to 1, 2, 3, 4, and 5. The range of `min_samples_split` was set to 2, 3, 5, 7, and 10. The range of `max_features` was set to None.

We selected 22 combinations according to the above mentioned method. The selected combinations were listed in Table 6. Table 7 summarizes calibration results, including R^2 , RMSE, and MAE of test sets for PA-II units using the RF method with the selected combinations of features.

Like the MLR method, the RF method showed better performance on the training set than on the test set. Some combinations had larger RMSE differences than $0.6 \mu\text{g}/\text{m}^3$ between training and test sets, while others have differences smaller than $0.4 \mu\text{g}/\text{m}^3$. We note that some combinations with multiplicative terms showed significant RMSE differences between two datasets, which might have occurred because of overfitting the training dataset. Nonetheless, the RF models with the other combinations had lower RMSE than the model using only $\text{PM}_{2.5}$. Considering a single environmental variable together with $\text{PM}_{2.5}$ improved the calibration performance in terms of values of RMSE and MAE compared to the RF model with only $\text{PM}_{2.5}$. Specifically, RH had more significant impact on the performance enhancement of the RF calibration model than T as seen in the results of combinations #2 and #3. Including the additional multiplicative term of $\text{PM}_{2.5} \times \text{RH}$ had an insignificant effect on RMSE compared to the RF model with $\text{PM}_{2.5}$ and RH. Both meteorological variables together, i.e., combination #5, yielded lower RMSE in the training set compared to the RF model with $\text{PM}_{2.5}$ and RH, i.e., combination #3, but similar RMSE in test set. In contrast to MLR models, more than one multiplicative term, i.e., combinations #6 to #9, bring about insignificant difference in RMSE compared to considering a single meteorological variable. When we analyze calibration methods without NO_2 , the RF model with $\text{PM}_{2.5}$, T, and RH improved RMSE by $0.117 \mu\text{g}/\text{m}^3$, compared to the best MLR model.

Utilizing NO_2 on RF models had different effects on calibration performance, depending on the combinations of predictor variables. The RF model of combination #10 with the additional NO_2 term resulted in an RMSE of $4.434 \mu\text{g}/\text{m}^3$, which was little improvement compared to combination #1 with only $\text{PM}_{2.5}$ and an RMSE of $4.439 \mu\text{g}/\text{m}^3$. The RF model with $\text{PM}_{2.5}$ and NO_2 had a larger RMSE than the MLR model with the same features, but the difference was not significant, it did not show enough performance improvement to warrant adding the multiplicative term of $\text{PM}_{2.5} \times \text{NO}_2$ from combination #10. Adding single or two meteorological variables to RF models of combinations #12 and #16 lead to remarkable performance enhancement over combination #10, with RH, RMSE decreasing by $0.462 \mu\text{g}/\text{m}^3$. Furthermore, RMSE dropped an additional $0.130 \mu\text{g}/\text{m}^3$ when T was added as an additional feature. The combinations consisting of one or more multiplicative interaction terms resulted in either an insignificant improvement or a slight decline in the performance in terms of RMSE and MAE when compared with

375 combination #16 consisting of $PM_{2.5}$, T, RH, and NO_2 . In other words, there is no need to consider multiplicative interaction terms when we use the RF model, because there is no outstanding performance improvement.

As with the MLR method, it was shown that including NO_2 as a consideration in RF methods can improve calibration performance. Moreover, by integrating two additional variables, such as T and RH, even better calibration performance can be achieved.

380 The RF method was shown to have a better performance than the MLR method when NO_2 was not considered. From the viewpoint of RMSE, the best performance from MLR and RF methods was $4.151 \mu\text{g}/\text{m}^3$ and $4.014 \mu\text{g}/\text{m}^3$, respectively. However, when we consider NO_2 , the best MLR model is not significantly different from the best RF model. For instance, the RMSE values from the best MLR and RF models were $3.912 \mu\text{g}/\text{m}^3$ and $3.840 \mu\text{g}/\text{m}^3$, respectively. Their corresponding R^2 values differ slightly, since their gap is only 0.008. Nonetheless, the MAE of $2.777 \mu\text{g}/\text{m}^3$ achieved from the best MLR
385 is lower than that achieved by the best RF, which is $2.831 \mu\text{g}/\text{m}^3$. From these results, we conclude that better calibration can be obtained by considering NO_2 additionally. Furthermore, when NO_2 is considered, the MLR model can enhance calibration performance without the need for an RF model.

3.2 Effect of distant NO_2 on calibration performance

In the previous subsections, it was demonstrated that including NO_2 as a consideration can effectively improve the calibration
390 performance of PA-II units. However, it is not always feasible to have an NO_2 instrument with high accuracy collocated with a low-cost PM sensor. Instead, an alternative approach is to collocate a low-cost NO_2 sensor with a PA-II unit, but this approach is hindered by the unreliability of NO_2 sensors. To address this issue, we investigated the usefulness of using data from distant NO_2 instruments installed with PA-II units for the calibration algorithm.

We selected two monitoring sites that measure NO_2 near the Rubidoux site. Two monitoring sites identified were 06-065-
395 8005 and 06-071-0027. The distances between the two monitoring sites and the Rubidoux site are 7.05 km and 18.87 km, respectively. The correlations of NO_2 measurements obtained from the Rubidoux site with that of 06-065-8005 and 06-071-0027 were 0.895 and 0.621, respectively. The site 06-065-8005 had NO_2 measurements that are much more highly correlated with the Rubidoux site compared with those from the site 06-071-0027. This result can occur when the distance from the Rubidoux site to the site 06-065-8005 is shorter than it is to the site 06-071-0027.

400 To evaluate the usefulness of distant NO_2 measurements on the calibration of a low-cost PM sensor, we used NO_2 data measured from monitoring sites near the PA-II 7 unit as a test dataset, rather than data from the collocated Rubidoux site. When we trained calibration models with the measurements from the PA-II 7 unit over 2018, we used highly accurate NO_2 concentrations measured by FEM instruments at the Rubidoux site. Subsequently, to verify the trained calibration models, we utilized a separate test dataset featuring distant NO_2 measurements taken by FEM instruments at sites 06-065-8005 and
405 06-071-0027. We considered this scenario to evaluate our proposed calibration models, previously trained with collocated NO_2 concentrations and distant NO_2 concentrations, when collocated NO_2 measurements cannot be collected.

Table 8 shows calibration performance using MLR and RF methods with NO_2 collected from the air quality monitoring sites near the PA-II unit. In the case of MLR methods used with 06-065-8005 data, the difference in RMSE between NO_2 data

obtained from a collocated NO₂ instrument, called collocated NO₂, and a distant NO₂ instrument, called distant NO₂, was less than 0.06 µg/m³ for every selected combination defined in previous two subsections for the MLR and RF methods. All MLR models using distant NO₂, except combinations #10 and #11, yielded lower errors than all MLR models without NO₂ as shown in Table 5. For example, the worst RMSE of the MLR methods using distant NO₂ data (except combinations #10 and #11) was 4.018 µg/m³, while the best RMSE without NO₂ was 4.151 µg/m³. Like RMSE, other metrics, such as R^2 and MAE, also showed a calibration performance enhancement for these combinations with distant NO₂.

When we used an MLR algorithm with NO₂ data, the result of the calibration performance for the monitoring site 06-071-0027 showed a new aspect from that of 06-065-8005. All MLR methods using distant NO₂ data from site 06-071-0027 had a higher RMSE than the MLR algorithm was based on data that did not include NO₂ data from the collocated Rubidoux instrument, which had an RMSE of 4.513 µg/m³ as shown in Table 5. This result can be explained by comparing correlation of NO₂ measured from the Rubidoux site with measurements from site 06-065-8005 as well as site 06-071-0027. The NO₂ correlation between Rubidoux measurements and site 06-065-8005 was 0.895, while the correlation with site 06-071-0027 was 0.621. These results shows that 06-065-8005 data is much more correlated with the Rubidoux site in terms of NO₂.

In the case of RF models, the use of the distant NO₂ data from site 06-065-8005 increased RMSE compared to using collocated NO₂ data, but not significantly, since the maximum gap of RMSE values for all feature vectors considered was just 0.060 µg/m³. Similar to the MLR method, all RF models referring to distant NO₂ from site 06-065-8005, except combinations #11, resulted in a better calibration performance than was seen in combination #1 without NO₂ which had an RMSE of 4.439 µg/m³ shown in Table 7. Other metrics, such as R^2 and MAE, also showed a calibration performance improvement. In the case of RF models using data from site 06-071-0027, calibration performance for each combination was degraded compared to the corresponding combination using collocated NO₂, which had similar results of the MLR model. As we explained previously, the higher the correlation of NO₂ measurements from the Rubidoux site with measurements from sites 06-065-8005 and 06-071-0027, the better the calibration performance of the RF model; that is, all combinations with distant NO₂ from 06-065-8005 provide a lower RMSE than those from 06-071-0027. Moreover, when we consider 06-065-8005 having a high correlation of NO₂ with the expensive NO₂ instrument collocated with the PA-II 7 unit, the best RMSE for all combinations using the RF model is slightly lower than that based on the MLR method.

In the case of 06-065-8005, RF models using distant NO₂ resulted in lower, but insignificant, RMSE values, compared to MLR models using distant NO₂. From these results, we draw the conclusion that the use of NO₂ collected from distant instruments with a high correlation with a collocated NO₂ site of PA-II units can improve the PA-II unit's calibration performance. Furthermore, both MLR and RF models can be good calibration models when distant NO₂ is considered. This is different from the conclusion that calibration performance of RF models is better than MLR models (Zimmerman et al., 2018).

3.3 Applicability of other PA-II units

We evaluated PA-II 8's calibration performance under the following three cases:

Case 1: Calibration model is learned with the measurements collected from the PA-II 8 in 2018 and calibration performance for the trained model is evaluated using data measured from the PA-II 8 in 2019.

Case 2: This is similar to Case 1, except that the calibration model is trained with the data measured from the PA-II 7 in 2018.

445 Case 3: The measurement data from the PA-II 8 with collocated NO₂ concentration in 2018 is used as a training dataset, while the data collected from the PA-II 8 with either collocated NO₂ or distant NO₂ concentration in 2019 is used as a test dataset.

In Case 1, we evaluated the calibration model's performance with a test dataset consisting of measurement data from the PA-II 8 in 2019. The calibration model is trained with data collected from the same PA-II 8 in 2018. Table 9 shows the calibration
450 results of the PA-II 8 using an MLR method under two different cases: with and without NO₂. We selected the same feature vectors as defined in Table 4. We observed that NO₂ can enhance calibration performance because all MLR models using NO₂, except combinations #10 and #11, yield lower errors and larger R^2 than those without NO₂. This observation aligns with the results shown in Table 5. Additionally, compared to the calibration performance for PA-II 7 shown in Table 5, PA-II 8 shows slightly larger RMSE and MAE, but similar R^2 .

455 In Case 2, we evaluated the calibration model's performance using a training dataset collected from PA-II 7 in 2018, and a test dataset collected from PA-II 8 in 2019. Table 10 shows calibration results for PA-II 8 using the MLR method under two different conditions, such as with and without NO₂. As with the observation in Table 9, NO₂ is the key factor enhancing calibration performance. With the exceptions of #10 and #11, all MLR models using NO₂ yield lower errors and larger R^2 than those without NO₂. It is important to compare this result with that shown in Table 5, as we used different test datasets. It
460 could be expected that the much worse performance for all feature combinations listed in Table 10 is achieved than for every corresponding feature vector in Table 5, since the calibration model considered in Table 10 is tested with the data measured from the PA-II 8, whereas it is trained with the measurement data collected from the PA-II 7. R^2 values of all feature vectors in Table 10 are similar to those for each corresponding feature vector in Table 5. Unlike R^2 , we observe larger RMSE and MAE when we populate the training dataset with measurements from PA-II 8 rather than PA-II 7. The maximum differences
465 of RMSE and MAE for each feature vector in Tables 10 and 5 are 0.177 $\mu\text{g}/\text{m}^3$ and 0.196 $\mu\text{g}/\text{m}^3$, respectively.

The results shown in Table 9 and Table 10 support our conclusion that reliable and consistent PA-II units, which contain two PMS 5003 sensors with high correlation, demonstrate similar calibration performance. This implies that a proposed calibration method can be applied to reliable and consistent PA-II units generally.

Lastly, in Case 3, we evaluated the effect of collocated and distant NO₂ on PA-II 8 unit's calibration performance. Table
470 11 shows the results of MLR-based calibration model for the PA-II 8 when it is verified with the test data considering either collocated or distant NO₂. As we explained in Section 3.2, we considered two monitoring sites measuring NO₂ near the Rubidoux site. One site (ID 06-065-8005) had NO₂ measurements that are much more highly correlated with the Rubidoux site than those from the other site (ID 06-071-00247). We refer to the NO₂ concentrations measured from these two sites as "distant NO₂". Three columns, describing the values of R^2 , RMSE, and MAE, of collocated NO₂ in Table 11 are exactly the

475 same as those of NO_2 included (i.e., collocated NO_2) in Table 9. In the case of site 06-065-8005 with high correlation with
the Rubidoux site, the consideration of the distant NO_2 facilitates improvement of the calibration performance, since all MLR-
based calibration models using distant NO_2 , except combinations #10 and 11, produce lower errors and larger R^2 than those
without NO_2 . This result is similar to when we consider the collocated NO_2 . However, we observe that adding distant NO_2
480 to the test dataset, which is not highly correlated to the NO_2 measurement from the reference site, deteriorates the calibration
performance. This is likely because all combinations from #10 to #21 yield lower R^2 and greater errors than all combinations
excluding NO_2 , as shown in Table 9. This result is the same as the observation of the PA-II 7 unit's calibration results in Table
8.

Hence, these results we draw from Table 11 support the same conclusions we drew from Tables 9 and 10. Reliable and
consistent PA-II units achieve similar calibration performance, and our proposed calibration model can be applied to these
485 units generally.

3.4 Effect of training period

We evaluated the effect of the training period on calibration performances. We consider four different training periods (i.e., 3,
6, 9, and 12 months), and each training set is constructed as follows: The training sets all end at the close of 2018. Their start
points are set in reverse order based on training periods. For example, for 3 months, the training set is from Oct. 2018 to Dec.
490 2018. Table S4 shows PA-II 7's calibration results using the MLR method for all four training periods. The 3-month training
period has the worst performance. The 6- and 9- month training periods generated better performances than the 12-month
training period. From a viewpoint of using NO_2 , NO_2 can improve calibration performance in all four cases, compared to
using only temperature and relative humidity. As the length of the training period increases, calibration performance improves.

3.5 Uncertainty analysis

495 We performed an uncertainty analysis of the MLR-based calibration model by using a bootstrapping technique on a test
dataset. Table 12 shows statistics of uncertainty analysis for each feature vector and t-values between two feature vectors
whose difference is the existence of NO_2 . We selected 8 feature vectors with various independent variables to verify whether
the addition of NO_2 affects the performance of our calibration model. The 4 feature vectors we considered are $\text{PM}_{2.5}$, $\text{PM}_{2.5}$, T,
 $\text{PM}_{2.5}$, RH, and $\text{PM}_{2.5}$, T, RH. We also added NO_2 to create four other feature vectors, $\text{PM}_{2.5}$, NO_2 , $\text{PM}_{2.5}$, T, NO_2 , $\text{PM}_{2.5}$, RH,
500 NO_2 , and $\text{PM}_{2.5}$, T, RH, NO_2 . We generated 1,000 test sets using a bootstrapping technique with replacement. We evaluated
mean and standard deviation values of RSME calculated over 1,000 test sets for each feature vector. In addition, we applied a
t-test to verify the effectiveness of adding NO_2 to each feature vector. Consideration of NO_2 additionally reduces mean values
of RMSE for all 4 feature vectors. Contrary to mean value, standard deviation of RMSE for every feature vector increases
slightly with the addition of NO_2 . We evaluated t-value for the mean values of RMSE for two feature vectors, with and without
505 NO_2 ; for example, the t-value between $\text{PM}_{2.5}$ and $\text{PM}_{2.5}$, NO_2 . Hence, we can evaluate 4 t-values. The Degree of Freedom
(DoF) is 1,998, so the relevant p-values are much less than 0.00001. Therefore, the difference in the mean RMSE values of

the PM_{2.5}-included and PM_{2.5}-excluded groups is significant. From these results, we can conclude that the performance of the MLR-based calibration model can be enhanced with consideration of PM_{2.5} concentrations.

4 Conclusions

510 The factors, directly affecting the performance of a low-cost PM sensor, including temperature, relative humidity, and particle composition, have been scrutinized for their impact on sensors' performance enhancement. Additionally, this study investigated the potential of NO₂, a precursor gas that gives rise to PM_{2.5} through atmospheric chemical reactions, to improve performance of the calibration model. To this end, we used the PurpleAir PA-II unit, which contains two Plantower PMS 5003 sensors, as a low-cost PM_{2.5} sensor. The PA-II units need to be typically installed close to reference monitoring sites measuring PM_{2.5}
515 concentrations and other pollutants, such as NO₂, in order to analyze their calibration. We identified a EPA-certified monitoring instrument whose deployed location is within close proximity to the installed location of 14 PA-II units, which satisfied the condition for co-location with a reference monitoring site. The monitoring site is located in Rubidoux, CA, USA. A study period of two years, i.e., from Jan. 2018 to Dec. 2019, was selected to include all seasons. Two units among 14 PA-II units were selected based on the availability of 23 months or more of measurement data from each PA-II unit, as well as its low
520 intra-model variability through correlation analysis.

One of the two selected PA-II units was compared to FRM and BAM-1020 instruments based on daily and hourly PM_{2.5} measurements. A comparison of the BAM-1020 instrument with the FRM instrument was also conducted on a daily PM_{2.5} measurement basis to evaluate the performance of the BAM-1020. The BAM-1020 instrument had a slope of 0.923, an intercept of 0.741, and a R^2 of 0.896 against the FRM instrument, which implies it provides acceptable performance as a reference mon-
525 itor for the calibration of low-cost PM_{2.5} sensors. For a PA-II unit, the Pearson correlation coefficient against the BAM-1020 instrument was shown to be 0.928 on an hourly basis. The per-month analysis was conducted on hourly PM_{2.5} measurements of the PA-II unit against the BAM-1020. Results showed the PA-II unit has a good correlation during the winter season, i.e., Nov., Dec., and Jan., with an R^2 value between 0.819 and 0.906, but a lower correlation during other months. The performance of the PA-II units was not notably affected by temperature or relative humidity (RH) during the winter months. Temperature and/or RH were found to improve R^2 during June and July 2018, but this effect in 2019 was not the same as in 2018.
530

A per-month analysis showed that NO₂ is a key factor that increased the value of R^2 during Sep. 2018, and Aug. and Sep. 2019. The effect of the addition of NO₂ for calibration of PA-II units was much larger when RH and temperature were considered together. In particular, NO₂ was shown to have more effect during months when the performance of PA-II units is moderate. It is expected that NO₂ can be used to improve the performance of low-cost PM_{2.5} sensors, but the effect of NO₂
535 should be further investigated for various ambient conditions.

Two methods for calibrating PA-II units, the Multiple Linear Regression (MLR) and Random Forest (RF), were evaluated on a test set of one year of data. We considered additive and multiplicative terms in two calibration methods. The RF method yielded better performance than the MLR method because it provides a larger R^2 as well as smaller RMSE, and MAE when NO₂, called collocated NO₂, measured from the collocated monitoring site was not used for calibration. However, when

540 collocated NO₂ is considered, MLR models showed similar performance to RF models. When several features, such as PM_{2.5},
temperature, RH, NO₂, and their multiplicative terms, are considered together to calibrate PM_{2.5} measurement data using
the MLR method, the calibration performance was shown to increase remarkably compared to cases where only PM_{2.5} are
considered. For instance, the RMSE value decreased from 4.513 µg/m³ to 3.912 µg/m³. In RF models with collocated NO₂,
545 inclusion of temperature and RH improved R^2 , RMSE, and MAE by an increase of 0.018, a decrease of 0.172 µg/m³, and 0.119
µg/m³, respectively, compared to the best RF models without NO₂. Contrary to the MLR model, multiplicative interaction
terms do not affect calibration performance with a certain direction, compared to those without NO₂; some combinations of
features provide slight enhancement, while the others cause worse performance.

We showed that NO₂ data could improve calibration performance in both MLR and RF models. The NO₂ data we referred
to was measured from an expensive reference monitor and is very reliable. However, it is not always feasible to have an NO₂
550 instrument with high accuracy collocated with a low-cost PM sensor. An alternatives is to use low-cost NO₂ sensors. However,
their performance remains questionable. To solve this issue, we investigated the effectiveness of using NO₂ measurements
collected from distant reliable NO₂ monitoring sites, called distant NO₂, whose location is not that far from a low-cost PM_{2.5}
sensor. It was demonstrated that distant NO₂ is effective for calibration models based on the MLR and RF algorithms when
distant NO₂ has a high correlation with collocated NO₂. Furthermore, we showed that MLR method can achieve a similar
555 calibration performance to the RF method when reliable distant NO₂ is considered.

We performed an evaluation of different PA-II units and found that incorporating NO₂ significantly enhanced calibration
performance across different PA-II units. This consistency held even when using models trained with different sensors at the
same location, reinforcing the reliability of generating consistent data across these units. Additionally, the uncertainty analysis
underscored a substantial performance boost by including NO₂ in the MLR method, showing a marked difference compared
560 to its omission.

Data availability. All data can be provided by the authors upon request.

Author contributions. TEXT

KK has designed and implemented the study and led the writing of the manuscript. SC helped analysis. All authors have
contributed to the writing process through discussion and feedback.

565 *Competing interests.* The authors declare that they have no conflict of interest.

Acknowledgements. This work was supported by the National Research Foundation of Korea(NRF) grant funded by the Korea government(MSIT) (RS-2022-00166847).

References

- Alvarado, M., Gonzalez, F., Fletcher, A., Doshi, A., Alvarado, M., Gonzalez, F., Fletcher, A., and Doshi, A.: Towards the Development of a
570 Low Cost Airborne Sensing System to Monitor Dust Particles after Blasting at Open-Pit Mine Sites, *Sensors*, 15, 19667–19687, 2015.
- Austin, E., Novosselov, I., Seto, E., and Yost, M.G.: Laboratory Evaluation of the Shinyei PPD42NS Low-Cost Particulate Matter Sensor,
Plos ONE, 10, 2015.
- Badura, M., Batog, P., Drzeniecka-Osciadacz, A., and Modzel, P.: Evaluation of low-cost sensors for ambient PM_{2.5} monitoring, *J. Sens.*
2018.
- 575 Nilson, B., Jackson, P. L., Schiller, C. L., and Parsons, M. T.: Development and evaluation of correction models for a low-cost fine particulate
matter monitor, *Atmos. Meas. Tech.*, 15, 3315-3328, 2022.
- Barkjohn, K. K., Bergin, M. H., Norris, C., Schauer, J. J., Zhang, Y., Black, M., Hu, M., and Zhang, J.: Using Lowcost sensors to Quantify
the Effects of Air Filtration on Indoor and Personal Exposure Relevant PM_{2.5} Concentrations in Beijing, China, *Aerosol Air Qual. Res.*,
20, 297–313, <https://doi.org/10.4209/aaqr.2018.11.0394>, 2020.
- 580 Barkjohn, K. K., Gantt, B., and Clements A. L.: Development and application of a United States-wide correction for PM_{2.5} data collected
with the PurpleAir sensor, *Atmos. Meas. Tech.*, 14, 4617-4637, 2021.
- Breiman, L.: Random Forests, *Machine Learning*, 45, 5-32, 2001.
- Budde, M., Müller, T., Laquai, B., Streibl, N., Schwarz, A., Schindler, G., Riedel, T., Beigl, M., and Dittler, A.: Suitability of the Low-Cost
SDS011 Particle Sensor for Urban PM-Monitoring, In Proceedings of the 3rd International Conference on Atmospheric Dust, Bari, Italy,
585 29–31 May 2018.
- Cavaliere, A., Carotenuto, F., Di Gennaro, F., Gioli, B., Gualtieri, G., Martelli, F., Matese, A., Toscano, P., Vagnoli, C., and Zaldei, A.:
Development of Low-Cost Air Quality Stations for Next Generation Monitoring Networks: Calibration and Validation of PM_{2.5} and
PM₁₀ Sensors, *Sensors*, 18, 2843, 2018.
- Crilley, L.R., Shaw, M., Pound, R., Kramer, L.J., Price, R., Young, S., Lewis, A.C., and Pope, F.D.: Evaluation of a low-cost optical particle
590 counter (Alphasense OPC-N2) for ambient air monitoring, *Atmos. Meas. Tech.*, 11, 709–720, 2018.
- Crilley, L. R., Singh, A., Kramer, L. J., Shaw, M. D., Alam M. S., Apte, J. S., Bloss, W. J., Ruiz L. H., Fu, P., Fu, W., Gani, S. , Gatari M.,
Ilyinskaya, E., Lewis, A. C., Ng'ang'a, D. 6, Sun, Y. , Whitty R. C. W. , Yue S., Young, S., and Pope F. D.: Effect of aerosol composition
on the performance of low-cost optical particle counter correction factors, *Atmos. Meas. Tech.*, 13, 1181–1193, 2020.
- Evans, J., van Donkelaar A., Martin, R. V., Burnett, R., Rainham, D. G., Birkett, N. J., and Krewski, D.: Estimates of globalmortality
595 attributable to particulate air pollution using satellite imagery, *Environ. Res. vol. 120*, 33-42, 2013.
- Feenstra, B., Papapostolou, V., Hasheminassab, S., Zhang, H., Boghossian, B. D., Cocker, D., and Polidori, A.: Perform-
ance evaluation of twelve low-cost PM_{2.5} sensors at an ambient air monitoring site, *Atmos. Environ.*, 216, 116946,
<https://doi.org/10.1016/j.atmosenv.2019.116946>, 2019.
- Feinberg, S., Williams, R., Hagler, G.S.W., Rickard, J., Brown, R., Garver, D., Harshfield, G., Ster, P., Mattson, E., Judge, R., and Garvey, S.:
600 Long-term evaluation of air sensor technology under ambient conditions in Denver, Colorado, *Atmos. Meas. Tech.*, 11, 4605–4615, 2018.
- Gao, M., Cao, J., and Seto, E.: A distributed network of low-cost continuous reading sensors to measure spatiotemporal variations of PM_{2.5}
in Xi'an, China, *Environ. Pollut.*, 199, 56–65, 2015.
- Hodan, W.H. and Barnard, W.R.:Evaluating the Contribution of PM_{2.5} Precursor Gases and Re-entrained Road Emissions to Mobile Source
PM_{2.5} Particulate Matter Emissions,MACTEC Federal Programs, 2004.

- 605 Holstius, D.M., Pillarisetti, A., Smith, K.R., and Seto, E.: Field calibrations of a low-cost aerosol sensor at a regulatory monitoring site in California, *Atmos. Meas. Tech.*, 7, 1121–1131, 2014.
- Hua, J., Zhang, Y., Foy, B., Mei, X., Shang, J., Zhang, Y., Sulaymon, I. D., and Zhou, D.: Improved PM_{2.5} concentration estimates from low-cost sensors using calibration models categorized by relative humidity, *Aerosol Science and Technology*, 55, 2021.
- Jayarathne R., Liu¹, X., Thai¹, P., Dunbabin, M., and Morawska, L.: The influence of humidity on the performance of a low-cost air particle
610 mass sensor and the effect of atmospheric fog, *Atmos. Meas. Tech.*, 11, 4883–4890, 2018.
- Jiao, W., Hagler, G., Williams, R., Sharpe, R., Brown, R., Garver, D., Judge, R., Caudill, M., Rickard, J., Davis, M., Weinstock, L., Zimmer-Dauphinee, S., and Buckley, K.: Community Air Sensor Network (CAIRSENSE) project: evaluation of low-cost sensor performance in a suburban environment in the southeastern United States, *Atmos. Meas. Tech.*, vol. 9, no. 11, 5281–5292, 2016.
- Johnson, K., Bergin, M., Russell, A., and Hagler, G.: Field Test of Several Low-Cost Particulate Matter Sensors in High and Low Concentration Urban Environments, *Aerosol Air. Qual. Res.* vol. 18, no. 3, 565–578, 2028.
615
- Kelly, K.E., Whitaker, J., Petty, A., Widmer, C., Dybwad, A., Sleeth, D., Martin, R., and Butterfield, A.: Ambient and laboratory evaluation of a low-cost particulate matter sensor, *Environ. Pollut.*, 221, 491–500, 2017.
- Liu, H.-Y., Bartonova, A., Schindler, M., Sharma, M., Behera, S.N., Katiyar, K., and Dikshit, O.: Respiratory Disease in Relation to Outdoor Air Pollution in Kanpur, India. *Arch. Environ. Occup. Health*, vol. 68, 204–217, 2013.
- 620 Liu, H.-Y., Dunea, D., Iordache, S., and Pohoata, A.: A Review of Airborne Particulate Matter Effects on Young Children’s Respiratory Symptoms and Diseases, *Atmosphere*, vol. 9, 150, 2018.
- Liu, H.-Y., Schneider, P., Haugen, R., and Vogt, M.: Performance Assessment of a Low-Cost PM_{2.5} Sensor for a near Four-Month Period in Oslo, Norway, *Atmosphere*, 10, 41, 2019.
- Magi, B. I., Cupini, C., Francis, J., Green, M., and Hauser, C.: Evaluation of PM_{2.5} measured in an urban setting using a lowcost optical
625 particle counter and a Federal Equivalent Method Beta Attenuation Monitor, *Aerosol Sci. Tech.*, 54, 147–159, 2019.
- Malings, C., Tanzer, R., Haurlyuk, A., Saha, P. K., Robinson, A. L., Presto, A. A., and Subramanian, R.: Fine particle mass monitoring with low-cost sensors: Corrections and longterm performance evaluation, *Aerosol Sci. Tech.*, 54, 160–174, 2020.
- Mansfield, E. R. and Helms, B. P.: Detecting Multicollinearity, *The American Statistician*, 36, 158–160, 1982.
- Mukherjee, A., Stanton, L.G., Graham, A.R., and Roberts, P.T. Assessing the Utility of Low-Cost Particulate Matter Sensors over a 12-Week
630 Period in the Cuyama Valley of California, *Sensors*, 17, 1805, 2017.
- Olivares, G., and Edwards, S. The Outdoor Dust Information Node (ODIN) – development and performance assessment of a low cost ambient dust sensor, *Atmos. Meas. Tech. Discuss.*, 8, 7511–7533, 2015.
- Pawar, H. and Sinha, B.: Humidity, density and inlet aspiration efficiency correction improve accuracy of a low-cost sensor during field calibration at a suburban site in the north-western Indo-Gangetic Plain (NW-IGP), *Aerosol Sci. Tech.*, 54, 685–703,
635 <https://doi.org/10.1080/02786826.2020.1719971>, 2020.
- PurpleAir Map, air quality Map [WWW Document] URL: <https://map.purpleair.org> (last access: 1 May 2020).
- Sayahi, T., Butterfield, A., and Kelly, K. E.: Long-term field evaluation of the Plantower PMS low-cost particulate matter sensors, *Environ. Pollut.*, vol. 245, 932–940, Feb. 2019.
- South Cost Air Quality Management District (SCAQMD): Field Evaluation AirBeam PM Sensor, available at:
640 <http://www.aqmd.gov/docs/default-source/aq-spec/laboratory-evaluations/airbeam—laboratory-evaluation.pdf?sfvrsn=6> (last access: 1 May 2020), 2017.

- South Cost Air Quality Management District (SCAQMD): Field Evaluation Purple Air (PA-II) PM Sensor, available at: <http://www.aqmd.gov/docs/default-source/aq-spec/field-evaluations/purple-air-pa-ii—field-evaluation.pdf?sfvrsn=2> (last access: 1 May 2020), 2017.
- 645 South Cost Air Quality Management District (SCAQMD): Field Evaluation Laser Egg PM Sensor, available at: <http://www.aqmd.gov/docs/default-source/aq-spec/field-evaluations/laser-egg—field-evaluation.pdf> (last access: 1 May 2020), 2017.
- Si, M., Xiong, Y., Du, S., and Du K.: Evaluation and calibration of a low-cost particle sensor in ambient conditions using machine-learning methods, *Atmos. Meas. Tech.*, 13, 1693–1707, 2020.
- Sousan, S., Koehler, K., Thomas, G., Park, J.H., Hillman, M., Halterman, A., and Peters, T.M.: Inter-comparison of low-cost sensors for measuring the mass concentration of occupational aerosols, *Aerosol Sci. Technol.*, 50, 462–473, 2016.
- 650 U.S. EPA: Reference and Equivalent Method Applications: Guidelines for Applicants, Sep. 2011.
- Wallace, L., Bi, J., Ott, W. R., Sarnat, J., and Liu, Y.: Calibration of low-cost PurpleAir outdoor monitors using an improved method of calculating PM_{2.5}, *Atmos. Environ.*, vol. 256, Jul. 2021.
- Wang, Y., Li, J., Jing, H., Zhang, Q., Jiang, J., and Biswas, P.: Laboratory Evaluation and Calibration of Three Low-Cost Particle Sensors for Particulate Matter Measurement, *Aerosol Sci. Technol.*, 49, 1063–1077, 2015.
- 655 Zheng, T., Bergin, M.H., Johnson, K.K., Tripathi, S.N., Shirodkar, S., Landis, M.S., Sutaria, R., and Carlson, D.E.: Field evaluation of low-cost particulate matter sensors in high and low concentration environments, *Atmos. Meas. Tech. Discuss.*, 11, 4823–4846, 2018.
- Zimmerman, N., Presto, A. A., Kumar, S. P. N., Gu, J., Hauryliuk, A., Robinson, E. S., Robinson, A. L., and R. Subramanian: A machine learning calibration model using random forests to improve sensor performance for lower-cost air quality monitoring, *Atmos. Meas. Tech.*, 11, 291–313, 2018.
- 660

Table 1. Information about 14 PA-II units, such as their ID, location (latitude and longitude), sensor name, start time of measurement, end time of measurement, and non-operating months.

ID	Latitude	Longitude	Sensor Name	Start Time of Measurement	End Time of Measurement	Non-Operating Months
1866	33.999978	-117.41676	RIVR_Co-loc1	7/10/17	4/27/20	Sep., Oct., Nov., and Dec. 2018
1854	33.999503	-117.41602	RIVR_Co-loc2	7/10/17	4/27/20	
2346	33.999978	-117.41676	RIVR_Co-loc3	7/31/17	4/27/20	
2325	33.999978	-117.41676	RIVR_Co-loc4	7/31/17	4/27/20	Sep., Oct., Nov., and Dec. 2018
2167	33.999978	-117.41676	RIVR_Co-loc5	7/17/17	4/27/20	
2155	33.999978	-117.41676	RIVR_Co-loc6	7/17/17	4/27/20	May, 2018
2612	33.999515	-117.41595	RIVR_Co-loc7	8/7/17	4/27/20	
2758	33.999978	-117.41676	RIVR_Co-loc8	8/11/17	4/27/20	Sep. 2018
3537	33.999381	-117.41601	RIVR_Co-loc9	9/20/17	4/27/20	May, Sep., Oct., Nov., and Dec. 2018
4748	33.999516	-117.41594	RIVR_Co-loc10	11/22/17	4/27/20	May, Aug., Sep., Oct., Nov., and Dec. 2018 and Jan 2019
4731	33.999504	-117.41593	RIVR_Co-loc11	11/22/17	3/1/19	Jan., Feb., Mar., Sep., Oct., Nov., and Dec. 2018
5280	33.99946	-117.41594	RIVR_Co-loc12	12/12/17	4/27/20	May, Sep., Oct., Nov., and Dec. 2018
5284	33.999451	-117.41591	RIVR_Co-loc13	12/12/17	4/27/20	May, Sep., Oct., Nov., and Dec. 2018
6806	33.999583	-117.41621	RIVR_Co-loc14	1/30/18	4/27/20	Apr., Sep., Oct., and Nov. 2018
6912	33.999482	-117.41627	RIVR_Co-loc15	1/31/18	4/27/20	Apr., Sep., Oct., and Nov. 2018
9226	33.999389	-117.41633	RIVR_Co-loc16	3/24/18	4/27/20	Apr., Sep., Oct., Nov., and Dec. 2018
9358	33.999319	-117.41638	RIVR_Co-loc17	3/25/18	4/27/20	Apr., Sep., Oct., Nov., and Dec. 2018

Table 2. R^2 , RMSE, and MAE of the PA-II unit against the BAM-1020 based on the hourly PM2.5 measurement data for each month.

	Jan-18	Feb-18	Mar-18	Apr-18	May-18	Jun-18	Jul-18	Aug-18	Sep-18	Oct-18	Nov-18	Dec-18
R^2	0.936	0.799	0.845	0.759	0.659	0.695	0.359	0.816	0.591	0.784	0.829	0.905
RMSE	4.201	3.735	2.932	3.938	3.477	4.097	5.615	3.204	4.550	3.650	3.832	3.765
MAE	3.171	2.721	2.196	3.098	2.716	3.267	3.597	2.424	3.358	2.844	2.913	2.743
Intercept	3.898	4.229	2.898	7.090	4.694	7.925	6.721	4.692	6.357	2.682	3.269	1.445
Slope	0.502	0.475	0.525	0.446	0.486	0.475	0.434	0.459	0.382	0.420	0.409	0.472
	Jan-19	Feb-19	Mar-19	Apr-19	May-19	Jun-19	Jul-19	Aug-19	Sep-19	Oct-19	Nov-19	Dec-19
R^2	0.884	0.750	0.735	0.618	0.801	0.730	0.893	0.405	0.441	0.523	0.880	0.813
RMSE	3.326	2.940	2.753	3.703	3.146	3.403	4.127	4.220	3.292	4.768	4.474	3.866
MAE	2.485	2.216	2.124	2.892	2.349	2.700	3.082	2.564	2.558	3.360	3.238	2.934
Intercept	1.961	2.190	1.881	4.065	2.525	3.225	3.070	5.649	5.312	5.088	2.976	1.165
Slope	0.397	0.354	0.427	0.385	0.418	0.383	0.575	0.428	0.511	0.483	0.497	0.572

Table 3. A summary statistics of daily and hourly PM_{2.5} measured from an FRM, BAM-1020, and PA-II 7 unit

	Daily PM _{2.5}			Hourly PM _{2.5}	
	FRM	BAM-1020	PA-II	BAM-1020	PA-II
Min (µg/m ³)	1.2	0	0.199	0	0.019
Max (µg/m ³)	66.3	68.3	129.069	159	263.062
Mean (µg/m ³)	11.69	12.13	18.247	12.171	18.367
Standard deviation (µg/m ³)	6.88	9.16	13.854	9.23	17.61

Table 4. A list of selected feature vectors in MLR methods

Feature Vector	PM _{2.5}	T	RH	NO ₂	PM _{2.5} ×T	PM _{2.5} ×RH	PM _{2.5} ×NO ₂	T×RH	T×NO ₂	RH×NO ₂	PM _{2.5} ×T×RH
1	X										
2	X	X									
3	X		X								
4	X		X			X					
5	X	X	X								
6	X	X	X			X					
7	X	X	X			X		X			
8	X	X	X		X	X		X			
9	X	X	X		X	X		X			X
10	X			X							
11	X			X			X				
12	X	X		X							
13	X	X		X			X				
14	X	X		X	X		X				
15	X	X		X	X		X		X		
16	X	X	X	X							
17	X	X	X	X		X					
18	X	X	X	X		X		X			
19	X	X	X	X	X	X		X			
20	X	X	X	X	X	X		X		X	
21	X	X	X	X	X	X		X	X	X	

Table 5. Calibration result (R^2 , RMSE ($\mu\text{g}/\text{m}^3$), and MAE ($\mu\text{g}/\text{m}^3$)) of hourly $\text{PM}_{2.5}$ concentrations using MLR for the PA-II 7 unit based on the selected combinations.

NO ₂ not included							NO ₂ included						
Feature	Training Set			Test Set			Feature	Training Set			Test Set		
Vector	R^2	RMSE	MAE	R^2	RMSE	MAE	Vector	R^2	RMSE	MAE	R^2	RMSE	MAE
1	0.803	4.272	3.279	0.731	4.513	3.418	10	0.806	4.241	3.259	0.741	4.424	3.329
2	0.814	4.150	3.185	0.755	4.305	3.194	11	0.806	4.236	3.255	0.741	4.423	3.326
3	0.813	4.160	3.203	0.760	4.263	3.165	12	0.826	4.010	3.075	0.789	3.997	2.871
4	0.820	4.087	3.109	0.763	4.232	3.132	13	0.827	3.997	3.071	0.789	3.993	2.857
5	0.816	4.125	3.174	0.763	4.234	3.129	14	0.829	3.977	3.042	0.792	3.962	2.843
6	0.821	4.069	3.093	0.765	4.211	3.100	15	0.829	3.975	3.041	0.793	3.954	2.838
7	0.822	4.054	3.098	0.772	4.154	3.043	16	0.826	4.008	3.077	0.790	3.986	2.866
8	0.824	4.040	3.086	0.772	4.151	3.023	17	0.829	3.979	3.028	0.789	3.990	2.863
9	0.825	4.022	3.075	0.771	4.161	3.012	18	0.831	3.958	3.029	0.798	3.912	2.793
							19	0.831	3.950	3.026	0.796	3.925	2.790
							20	0.832	3.945	3.025	0.797	3.920	2.782
							21	0.832	3.941	3.019	0.797	3.913	2.777

Table 6. A list of selected feature vectors in RF methods

Feature Vector	PM _{2.5}	T	RH	NO ₂	PM _{2.5} ×T	PM _{2.5} ×RH	PM _{2.5} ×NO ₂	T×RH	T×NO ₂	RH×NO ₂	PM _{2.5} ×T×RH
1	X										
2	X	X									
3	X		X								
4	X		X			X					
5	X	X	X								
6	X	X	X					X			
7	X	X	X		X			X			
8	X	X	X		X	X		X			
9	X	X	X		X	X		X			X
10	X			X							
11	X			X			X				
12	X		X	X							
13	X	X		X	X						
14	X	X		X			X		X		
15	X	X		X	X		X		X		
16	X	X	X	X							
17	X	X	X	X	X						
18	X	X	X	X		X		X			
19	X	X	X	X	X				X	X	
20	X	X	X	X	X		X		X	X	
21	X	X	X	X	X		X	X	X	X	
22	X	X	X	X	X	X	X	X	X	X	

Table 7. Calibration result (R^2 , RMSE ($\mu\text{g}/\text{m}^3$), and MAE ($\mu\text{g}/\text{m}^3$)) of hourly $\text{PM}_{2.5}$ concentrations using RF for for the PA-II 7 unit based on the selected combinations.

NO ₂ not included							NO ₂ included						
Feature	Training Set			Test Set			Feature	Training Set			Test Set		
Vector	R^2	RMSE	MAE	R^2	RMSE	MAE	Vector	R^2	RMSE	MAE	R^2	RMSE	MAE
1	0.826	4.014	3.072	0.739	4.439	3.318	10	0.820	4.080	3.116	0.740	4.434	3.300
2	0.842	3.830	2.933	0.764	4.223	3.156	11	0.821	4.074	3.109	0.738	4.451	3.305
3	0.857	3.632	2.785	0.786	4.026	2.951	12	0.861	3.588	2.748	0.791	3.972	2.925
4	0.875	3.398	2.611	0.786	4.024	2.957	13	0.885	3.269	2.522	0.794	3.945	2.861
5	0.883	3.290	2.526	0.785	4.034	2.970	14	0.885	3.262	2.519	0.797	3.918	2.887
6	0.862	3.568	2.740	0.787	4.014	2.955	15	0.886	3.250	2.505	0.793	3.957	2.875
7	0.884	3.276	2.515	0.779	4.092	2.964	16	0.893	3.154	2.427	0.805	3.842	2.836
8	0.861	3.584	2.747	0.782	4.059	2.956	17	0.920	2.720	2.092	0.797	3.918	2.840
9	0.905	2.968	2.257	0.785	4.029	2.853	18	0.920	2.722	2.095	0.805	3.840	2.831
							19	0.921	2.706	2.080	0.794	3.942	2.860
							20	0.921	2.699	2.073	0.795	3.936	2.857
							21	0.894	3.130	2.401	0.794	3.946	2.856
							22	0.915	2.800	2.121	0.798	3.912	2.850

Table 8. Calibration result (R^2 , RMSE ($\mu\text{g}/\text{m}^3$), and MAE ($\mu\text{g}/\text{m}^3$)) of hourly $\text{PM}_{2.5}$ concentrations using MLR and RF models for the PA-II 7 unit based on the selected combinations additionally with distant NO_2 .

Site ID	Feature Vector	MLR						RF						
		collocated NO_2			Distant NO_2			collocated NO_2			Distant NO_2			
		R^2	RMSE	MAE	R^2	RMSE	MAE	R^2	RMSE	MAE	R^2	RMSE	MAE	
06-05-08005	10	0.741	4.424	3.329	0.742	4.417	3.320	0.740	4.434	3.300	0.739	4.442	3.304	
	11	0.741	4.423	3.326	0.743	4.411	3.311	0.738	4.451	3.305	0.738	4.454	3.306	
	12	0.789	3.997	2.871	0.786	4.018	2.879	0.791	3.972	2.925	0.790	3.983	2.934	
	13	0.789	3.993	2.857	0.787	4.011	2.861	0.794	3.945	2.861	0.789	3.994	2.902	
	14	0.792	3.962	2.843	0.791	3.978	2.842	0.797	3.918	2.887	0.791	3.970	2.923	
	15	0.793	3.954	2.838	0.791	3.978	2.844	0.793	3.957	2.875	0.787	4.017	2.917	
	16	0.790	3.986	2.866	0.787	4.009	2.875	0.805	3.842	2.836	0.802	3.873	2.854	
	17	0.789	3.990	2.863	0.787	4.011	2.870	0.797	3.918	2.840	0.793	3.951	2.860	
	18	0.798	3.912	2.793	0.795	3.936	2.803	0.805	3.840	2.831	0.802	3.870	2.848	
	19	0.796	3.925	2.790	0.794	3.950	2.800	0.794	3.942	2.860	0.790	3.983	2.884	
	20	0.797	3.920	2.782	0.795	3.933	2.780	0.795	3.936	2.857	0.791	3.978	2.877	
	21	0.797	3.913	2.777	0.796	3.931	2.777	0.794	3.946	2.856	0.790	3.986	2.879	
								0.798	3.912	2.850	0.794	3.946	2.865	
	06-07-110	10	0.741	4.424	3.329	0.715	4.645	3.563	0.740	4.434	3.300	0.734	4.488	3.345
		11	0.741	4.423	3.326	0.715	4.641	3.549	0.738	4.451	3.305	0.729	4.525	3.367
		12	0.789	3.997	2.871	0.694	4.807	3.739	0.791	3.972	2.925	0.781	4.069	2.994
		13	0.789	3.993	2.857	0.695	4.799	3.706	0.794	3.945	2.861	0.692	4.826	3.624
		14	0.792	3.962	2.843	0.696	4.797	3.673	0.797	3.918	2.887	0.693	4.815	3.646
		15	0.793	3.954	2.838	0.682	4.906	3.778	0.793	3.957	2.875	0.689	4.850	3.648
		16	0.790	3.986	2.866	0.701	4.751	3.681	0.805	3.842	2.836	0.761	4.247	3.170
		17	0.789	3.990	2.863	0.714	4.651	3.576	0.797	3.918	2.840	0.733	4.494	3.325
18		0.798	3.912	2.793	0.720	4.602	3.531	0.805	3.840	2.831	0.746	4.381	3.289	
19		0.796	3.925	2.790	0.721	4.593	3.516	0.794	3.942	2.860	0.722	4.586	3.423	
20		0.797	3.920	2.782	0.721	4.595	3.516	0.795	3.936	2.857	0.721	4.592	3.422	
21	0.797	3.913	2.777	0.702	4.746	3.669	0.794	3.946	2.856	0.744	4.401	3.256		
							0.798	3.912	2.850	0.727	4.542	3.386		

Table 9. Calibration results of hourly PM_{2.5} concentrations measured from the PA-II 8 in 2019 using MLR-based calibration model learned with training data collected from the PA-II 8 in 2018.

NO ₂ not included							NO ₂ included						
Feature	Training Set			Test Set			Feature	Training Set			Test Set		
Vector	R ²	RMSE	MAE	R ²	RMSE	MAE	Vector	R ²	RMSE	MAE	R ²	RMSE	MAE
1	0.786	4.312	3.304	0.731	4.559	3.468	10	0.788	4.292	3.295	0.741	4.468	3.381
2	0.798	4.196	3.211	0.749	4.397	3.299	11	0.789	4.289	3.293	0.742	4.459	3.375
3	0.797	4.208	3.231	0.760	4.307	3.223	12	0.809	4.079	3.127	0.783	4.087	2.982
4	0.803	4.142	3.147	0.763	4.277	3.191	13	0.810	4.070	3.123	0.785	4.072	2.966
5	0.800	4.173	3.201	0.759	4.311	3.219	14	0.811	4.051	3.099	0.788	4.042	2.951
6	0.805	4.123	3.127	0.762	4.281	3.185	15	0.811	4.050	3.099	0.788	4.040	2.950
7	0.806	4.111	3.134	0.767	4.242	3.143	16	0.809	4.076	3.128	0.785	4.071	2.970
8	0.807	4.099	3.127	0.768	4.227	3.121	17	0.811	4.050	3.083	0.785	4.071	2.967
9	0.808	4.091	3.121	0.770	4.214	3.128	18	0.813	4.033	3.087	0.791	4.015	2.915
							19	0.814	4.028	3.084	0.791	4.019	2.911
							20	0.814	4.023	3.083	0.792	4.006	2.895
							21	0.814	4.021	3.081	0.792	4.002	2.892

Table 10. Calibration results of hourly PM_{2.5} concentrations measured from the PA-II 8 in 2019 using MLR-based calibration model learned with training data collected from the PA-II 7 in 2018.

NO ₂ not included				NO ₂ included			
Feature	Test Set			Feature	Test Set		
Vector	R ²	RMSE	MAE	Vector	R ²	RMSE	MAE
1	0.737	4.638	3.546	10	0.747	4.549	3.458
2	0.757	4.459	3.364	11	0.748	4.538	3.446
3	0.763	4.400	3.322	12	0.788	4.162	3.054
4	0.765	4.383	3.293	13	0.790	4.145	3.031
5	0.765	4.388	3.301	14	0.794	4.104	3.003
6	0.766	4.373	3.275	15	0.795	4.097	3.000
7	0.772	4.323	3.222	16	0.789	4.151	3.048
8	0.772	4.318	3.208	17	0.789	4.158	3.050
9	0.774	4.301	3.208	18	0.796	4.089	2.985
				19	0.795	4.100	2.984
				20	0.795	4.095	2.974
				21	0.796	4.090	2.970

Table 11. Calibration results of hourly PM_{2.5} concentrations measured from the PA-II 8 in 2019 using MLR-based calibration model learned with training data collected from the PA-II 8 in 2018 (Site ID indicates the monitoring sites for distant NO₂).

Site ID	Feature Vector	MLR					
		collocated NO ₂			Distant NO ₂		
		R^2	RMSE	MAE	R^2	RMSE	MAE
06-08065080027	10	0.741	4.468	3.381	0.742	4.458	3.371
	11	0.742	4.459	3.375	0.744	4.442	3.359
	12	0.783	4.087	2.982	0.783	4.089	2.976
	13	0.785	4.072	2.966	0.786	4.066	2.951
	14	0.788	4.042	2.951	0.789	4.031	2.927
	15	0.788	4.040	2.950	0.789	4.033	2.930
	16	0.785	4.071	2.970	0.785	4.075	2.966
	17	0.785	4.071	2.967	0.785	4.076	2.960
	18	0.791	4.015	2.915	0.790	4.022	2.911
	19	0.791	4.019	2.911	0.790	4.026	2.908
	20	0.792	4.006	2.895	0.793	3.998	2.877
	21	0.792	4.002	2.892	0.793	3.995	2.875
	10	0.741	4.468	3.381	0.716	4.681	3.600
	11	0.742	4.459	3.375	0.716	4.680	3.591
	12	0.783	4.087	2.982	0.684	4.937	3.887
	13	0.785	4.072	2.966	0.684	4.937	3.864
	14	0.788	4.042	2.951	0.680	4.965	3.850
	15	0.788	4.040	2.950	0.672	5.030	3.914
	16	0.785	4.071	2.970	0.693	4.870	3.816
	17	0.785	4.071	2.967	0.706	4.764	3.704
	18	0.791	4.015	2.915	0.710	4.733	3.676
19	0.791	4.019	2.911	0.713	4.705	3.646	
20	0.792	4.006	2.895	0.713	4.709	3.643	
21	0.792	4.002	2.892	0.699	4.818	3.756	

Table 12. Statistics of uncertainty analysis to selected feature vectors and t-values.

Feature Vector	Mean of RMSE	Std. Dev. of RMSE	Feature Vector	Mean of RMSE	Std. Dev. of RMSE	t-value	DoF
{PM _{2.5} }	4.5095	0.1026	{PM _{2.5} , NO ₂ }	4.4202	0.1037	19.3580	1,998
{PM _{2.5} , T}	4.3084	0.1000	{PM _{2.5} , T, NO ₂ }	3.9979	0.1173	63.7008	1,998
{PM _{2.5} , RH}	4.2598	0.0995	{PM _{2.5} , RH, NO ₂ }	4.1548	0.1074	22.6792	1,998
{PM _{2.5} , T, RH}	4.2387	0.1050	{PM _{2.5} , T, RH, NO ₂ }	3.9865	0.1156	51,0686	1,998




Figure 1. Correlation among all PMS 5003 sensors of the selected units PA-II 2, 3, 5, 6, 7, and 8. The left and right of each number on the x-axis represent PMS A and B sensors for its corresponding PA-II unit, respectively.




Figure 2. Scatter plot for daily PM_{2.5} comparison of BAM-1020 (Non-FEM) instrument with the FRM instrument.




Figure 3. Hourly PM_{2.5} concentrations measured from BAM-1020 (Non-FEM) and PA-II 7, and hourly temperature and relative humidity measured from PA-II 7 from Dec. 2018 to Feb. 2019.




Figure 4. R^2 and RMSE using MLR method for the PA-II unit with the BAM-1020 for the selected months based on the following feature vectors; 1:($\text{PM}_{2.5}$), 2:($\text{PM}_{2.5}, T$), 3:($\text{PM}_{2.5}, \text{RH}$), 4:($\text{PM}_{2.5}, \text{NO}_2$), 5:($\text{PM}_{2.5}, T, \text{RH}$), and 6:($\text{PM}_{2.5}, T, \text{NO}_2$).