

Concept, absolute calibration and validation of a new, bench-top laser imaging polar nephelometer

Alireza Moallemi¹, Robin L. Modini¹, Benjamin T. Brem¹, Barbara Bertozzi¹, Philippe Giaccari², and
5 Martin Gysel-Beer¹

¹Laboratory of Atmospheric Chemistry, Paul Scherrer Institute, Villigen PSI, 5232, Switzerland

²Micos Engineering GmbH, Dübendorf, CH-8600, Switzerland

Correspondence to: Martin Gysel-Beer (martin.gysel@psi.ch)

Abstract.

10 Polar nephelometers provide in situ measurements of aerosol angular light scattering and play an essential role in validating numerically calculated phase functions or inversion algorithms used in space-borne and land-based aerosol remote sensing. In this study, we present a prototype of a new polar nephelometer called uNeph. The instrument is designed to measure the phase function, F_{11} , and polarized phase function, $-F_{12}/F_{11}$ over the scattering range of around 5° to 175° with an angular resolution of 1° at a wavelength of 532 nm. In this work, we present details of the data processing procedures and instrument calibration
15 approaches. The uNeph was validated in a laboratory setting using mono-disperse polystyrene latex (PSL) and Di-Ethyl-Hexyl-Sebacate (DEHS) aerosol particles over a variety of sizes, ranging from 200 nm to 800 nm. An error model was developed and the level of agreement between uNeph measurements and Mie theory was found to be consistent within the uncertainties of the measurements and the uncertainties of the input parameters for the theoretical calculations. The estimated measurement errors were between 5% to 10 % (relative) for F_{11} and smaller than ~ 0.1 (absolute) for $-F_{12}/F_{11}$. Additionally, by applying the
20 Generalized Retrieval of Aerosol and Surface Properties (GRASP) inversion algorithm to the measurements conducted with broad unimodal DEHS aerosol particles, the volume concentration, size distribution and refractive index of the ensemble of aerosol particles were accurately retrieved. This paper demonstrates that the uNeph prototype can be used to conduct accurate measurements of aerosol phase function and polarized phase function and to retrieve aerosol properties through inversion algorithms.

25 1 Introduction

Atmospheric aerosol particles have a substantial impact on the Earth's radiative budget through their direct interaction with solar radiation and by affecting cloud formation processes (Boucher et al., 2013). Furthermore, aerosol particles are a significant component of air pollution which has been estimated to cause 4.2 million annual premature deaths worldwide (Cohen et al., 2017). The complex compositional, microphysical properties and high spatio-temporal variability of atmospheric
30 aerosols makes it difficult to properly characterize them and to constrain them in atmospheric global model simulations. This leads leading to large uncertainties in estimation of aerosol radiative forcing contribution compared to other prevalent climate

change drivers such as CO₂ (Myhre et al., 2013). Global scale long term measurements of various aerosol properties are essential for obtaining in-depth insights on atmospheric aerosol variability and to develop more realistic parametrization of aerosol particles in global atmospheric models. Passive satellite and ground-based remote sensing, which are the main approaches to obtain atmospheric aerosol observations on a global scale, rely on measurements of elastically-scattered solar radiation by atmospheric aerosols (Boucher, 2015b). Remote sensing instruments are typically designed to detect [radiance](#) ~~irradiance~~ of scattered light over single or multiple angles (radiometric measurements), while certain instruments can provide complementary measurements on the scattered light polarization state (polarimetric measurements) (Dubovik et al., 2019). The polarimetric and radiometric measurements contain implicit information of numerous aerosol micro-physical properties such as refractive index (RI), size and shape (Bohren & Huffman, 2004).

In remote sensing, aerosol properties are inferred from radiometric and polarimetric measurements by using inversion algorithms (e.g., Dubovik et al., 2021). Due to the ill-posed nature of the inverse problem, the inversion algorithms often use simplified forward scattering models and a priori assumptions to be able to retrieve aerosol properties. For instance, it is common to assume spherical shape and use Mie theory for the forward kernel of inversion algorithms (Holben et al., 1998; Omar et al., 2009). While such an assumption is sufficient for spherical aerosols, studies have shown that light scattering by complex aerosols, such as biomass burning aerosols, is quite distinct from light scattering described by Mie theory (Espinosa et al., 2019; Manfred et al., 2018). Hence, aerosol property retrievals from remote sensing measurements are prone to uncertainties and biases. Independent in situ validation techniques are required to identify potential biases and uncertainties in the retrieved aerosol properties from remote sensing measurements (Mishchenko et al., 2007). Direct validation of remote sensing measurements is a challenging and expensive task that often involves airborne measurements (Schafer et al., 2019). Alternatively, mimicking atmospheric remote sensing measurements with in situ instruments enables validation of remote sensing retrieval algorithms in laboratory environments (Schuster et al., 2019). This is a more cost-effective approach compared to the direct validation of atmospheric aerosol retrievals and allows for the testing of aerosol samples with well-defined properties.

Polar nephelometers are in situ instruments primarily designed for radiometric measurements. Following the Stokes formalism, polar nephelometers measure the $F_{11}(\theta)$ element of the aerosol scattering matrix $\mathbf{F}(\theta)$ over multiple scattering polar angles θ (Espinosa et al., 2017). The $F_{11}(\theta)$ element is also referred to as the phase function (PF) and describes the partial scattering coefficient of an aerosol as a function of θ for non-polarized incident light. A subset of polar nephelometers is also capable of performing polarimetric measurements, that is additionally providing the polarized phase function (PPF), $-F_{12}(\theta)/F_{11}(\theta)$, which describes the relative degree and orientation of linear polarization of scattered light as a function of θ for non-polarized incident light (e.g., Dolgos and Martins, 2014).

Polar nephelometers have a long history (e.g., Waldram, 1945) and over the years several different instrument designs have been introduced. Broadly, polar nephelometer designs can be categorized into three groups (Barkey et al., 2012). Goniometer-type nephelometers conduct radiometric and polarimetric measurements using a detector mounted on a rotatable arm (Li et al., 2018; Horvath et al., 2018; Waldram, 1945). These instruments can achieve high angular resolution at the expense of relatively

low time resolution. Multi-detector type nephelometers use sensors mounted at fixed scattering angles (Barkey et al., 2007; Dick et al., 2007; Nakagawa et al., 2016). These instruments can provide rapid radiometric and polarimetric measurements while the number of probed scattering angles remains limited. Laser imaging type nephelometers image light scattered by an ensemble of aerosol particles within a laser beam onto a charged couple device detector (CCD) such that scattering angle and position on the image are unambiguously related. These instruments can provide radiometric and polarimetric measurements with high angular and time resolution. While more sophisticated polar nephelometers capable of measuring more components of $F(\theta)$ exist (Muñoz et al. 2012, Hu et al., 2021), the majority of polar nephelometers used in atmospheric applications typically measure either PF only, or PF and PPF. Polarimetric data provided by these instruments make it possible to retrieve aerosol properties using inversion schemes similar (e.g. Espinosa et al., 2019, Schuster et al., 2019, Boiger et al., 2022).

In spite of several advantageous features, the use of laser imaging nephelometers has been quite limited. One of the earliest versions is the polarized imaging nephelometer (PI-Neph) that was introduced by Dolgos and Martins (2014). The PI-Neph provides both radiometric and polarimetric measurements at three wavelengths over an angle range of 3 to 177° with angular resolution of 1°. This instrument has been deployed in airborne field campaigns. For example, Espinosa et al. (2017) applied the Generalized Retrieval of Aerosol and Surface Properties (GRASP), which is a well-established retrieval algorithm, on measurements conducted by the PI-Neph and successfully retrieved size distribution, RI, and shape properties of aerosol particles.

Since the inception of the PI-Neph, several other laser imaging nephelometers have been introduced. For instance, Bian et al. (2017) and Manfred et al. (2018) used laser imaging nephelometers with only radiometric capabilities to measure the PF of ambient aerosol particles and biomass burning aerosols, respectively. More recently, Ahern et al. (2022) introduced a laser imaging nephelometer capable of conducting simultaneous polarimetric and radiometric measurements at two distinct wavelengths which is suitable for deployment on aircrafts.

In this study, a new laser imaging type nephelometer, referred to as uNeph, was designed and constructed jointly by Micos Engineering GmbH (Dübendorf, Switzerland) and the Laboratory of Atmospheric Chemistry at Paul Scherrer Institute (Villigen PSI, Switzerland). Considerable down-sizing was a design goal for the uNeph to facilitate deployment and operation of the instrument in different settings. This work presents the uNeph including instrument description and experimental setup (Section 2), basic data processing procedures (Section 3), signal calibration approaches and error assessment (Section 4) as well as validation and a first basic application (Section 5).

2 Instrument description and experimental setup

2.1 uNeph

The uNeph is a bench-top laser imaging polar nephelometer. It consists of an optical box with dimensions 44 × 61 × 18 cm and PC as well as a few small control boxes kept externally. A schematic of the key elements of the uNeph instrument is shown in Fig. 1. A solid-state continuous wave laser provides linearly polarized light at 532 nm (~200 mW). The laser beam is

[collimated with a Gaussian size around 1 mm and vignetted with apertures to slightly cut it down to ~0.5–1.0 mm.](#)

Exchangeable neutral density (ND) filters are utilized to reduce the laser intensity in order to adjust instrument sensitivity to different levels. [We particularly used three different optical density filters throughout the experiments \(Thorlabs, Inc., Newton, NJ, USA, models NE10A, NE15A, NE20A, with optical density levels of 1, 1.5 and 2.0, respectively\).](#)

The next elements serve to rotate the angle of the linear polarization state following the approach detailed in Dolgos and Martins (2014). It is an assembly consisting of a liquid crystal variable retarder (LCVR) and a Fresnel rhomb which acts as a quarter wavelength retarder. Subsequently, the beam is passed through multiple irises to adjust the cross-sectional area of the beam and to reduce the level of undesired background light (e.g., straylight) in the scattering chamber (not shown in Fig. 1).

The laser beam is directed through a scattering chamber twice using a rooftop reflector in between (a rooftop design that maintains the linear polarization nature of the laser beam). The input window of the scattering chamber is placed at angle and the small reflected part, not blocked by the AR-coating, is used to measure the input laser power (forward beam) with a photodetector ([Si free-space amplified photodetectors, Thorlabs, Inc., Newton, NJ, USA, model PDA100A2](#)). The power of the backward beam exiting the chamber after both passages is also measured with a second [similar](#) photodiode (an ND filter is used for protection). Since the laser intensity attenuation is expected to be minimal within the range of aerosol sample concentration tested in this study we only used the forward beam laser power measurements to compensate the measurements for laser amplitude variations.

The scattering chamber, also shown in Fig. 1, serves as the measurement cell. It is a metal chamber with a length of ~30 cm and a volume of ~2.5 L. The laser beam enters and leaves the cell through sealed windows. The chamber has a removable lid which enables access to the beam path, e.g., for performing calibration tasks. A pressure (p), temperature (T), and relative humidity (RH) sensor is mounted inside the scattering chamber. Sample inlet and outlet are installed at opposite ends of the scattering chamber to flush it continuously with gas or aerosol samples. A detection unit consisting of an optical objective and a camera is used to collect the scattered light. Light scatters inside the forward and backward beams in forward (scattering angles, $0^\circ \leq \theta \leq 90^\circ$ ~~$\theta = 0-90^\circ$~~) and backward ($90^\circ \leq \theta \leq 180^\circ$ ~~$\theta = 90-180^\circ$~~) directions, respectively, and appears as two distinct stripes on the camera image (Fig. 2a). The camera is an Argon filled, actively cooled, monochrome, charged coupled device (CCD) with a resolution of 1392×1040 pixels (9×6.7 mm sensor size and 6.45×6.45 μm pixel size) and a 16 bit ~~A/D~~ [analogue-to-digital \(A/D\)](#) converter (Trius Pro 825, Starlight Xpress Ltd.). A three-dimensional scheme of the laser path within the scattering chamber is demonstrated in Fig. S1. One design element of the uNeph is the use of a wide field of view pinhole lens in the camera objective, which enables the instrument to be downsized. [The camera objective is the Marshall V-PL25CS-12 \(discontinued\) with a pinhole size of ~ 2 mm, a focal length of 2.5 mm and aperture of F2.8. The field of view \(diameter\) is 100°. Direct connection of the CS-mount objective to the cooled CCD is not possible and, therefore, a Thorlabs re-imager optics is needed. The objective is placed at 45° from the direction of the beams with the pinhole being placed between both beams as depicted in Fig. S1.](#) Based on the geometry (Fig. S1), the pinhole together with a given laser beam define a scattering plane. The forward and backward laser beams were aligned to be parallel, and the pinhole location was adjusted such that the two scattering planes have 45° (forward beam) and 135° (backward beam) orientations relative to the yz plane as depicted in

Fig. S1b. They are chosen to be perpendicular to each other to achieve identical angle between scattering plane and orientation of linear polarization state for both forward and backward beams. Two distinct linear polarization states, e.g. with nominal orientation in parallel or perpendicular to the scattering planes, are sufficient to measure the phase function and polarized phase functions (see Section 4.1). [A polarimeter \(PAX1000VIS/M, Thorlabs Inc.\) was used, in a setup without the chamber, to verify that the forward and backward laser beams are almost fully linearly polarized. The set points of the LCVR to achieve parallel or perpendicular linear polarization relative to the scattering plane were determined by inserting an additional linear polarizer in front of the chamber. Its orientation was chosen to be perpendicular to the nominal laser polarization orientation, and then the LCVR setting was varied to achieve minimal beam transmission. Uncertainties in pinhole alignment relative to the laser beams is estimated to have larger effects on deviations from nominal linear polarization than the quality of polarization control.](#) ~~A polarimeter (PAX1000VIS/M, Thorlabs Inc.) was used to verify that the two laser beams crossing the chamber are almost fully linearly polarized. Uncertainties in pinhole alignment relative to the laser beams is estimated to have larger effect on deviations from nominal linear polarization than the quality of polarization control.~~ We use subscripts “1” and “2” to denote nominal parallel and perpendicular linear polarization operation set points, respectively. The raw data provided by the uNeph consist of a digital image of the light scattered from the forward and backward beams at a given state of polarization. When conducting measurements, a defined polarization state is first applied, followed by acquisition of an image (or multiple images) at a specified exposure time, t_{expo} . Auxiliary sensor readings (T , p , RH , and photo detectors) are simultaneously being logged. Figure 2a shows an example of a raw light scattering image for a particle-free air sample taken at polarization state 2, i.e., nominally perpendicular linear polarization. Further data processing will be explained in Section 3 onwards.

2.2 Experimental setup for uNeph calibration and validation

Some uNeph calibration measurements rely on probing pure, particle-free gases following the experimental setup shown in Fig. 3a. For this purpose, particle-free air, CO₂, Ar or He was flushed through the uNeph with a flow rate of $\sim 5 \text{ Lmin}^{-1}$. Further calibration and validation measurements were done using aerosol samples with well-defined properties using the experimental setups shown in Fig. 3b or Fig. 3c to generate quasi-monodisperse or broad unimodal aerosol samples, respectively. Initial aerosol generation steps were identical for these two types of experiments. The aerosol samples used in this study were spherical polystyrene latex size standards (PSL; see Table S1 for specifications) and Di-Ethyl-Hexyl-Sebacate (DEHS). DEHS is a non-absorbing oil-like liquid, hence also resulting in spherical aerosol particles.

PSL suspensions diluted with Milli-Q water were nebulized using a commercial atomizer aerosol generator (ATM 226, Topas GmbH, Dresden, Germany). Liquid DEHS in pure form was aerosolized using a Collison-type nebulizer (CH Technologies; Westwood, NJ, USA). After particle generation, aerosol samples were passed through a Kr-85 neutralizer to mitigate aerosol electrostatic losses in the sampling line. Subsequently, to remove water content from the nebulized aerosol particles, the sample flow was passed through a silica gel diffusion drier. A dilution stage was placed after the drier to adjust aerosol concentration to suitable levels, particularly during the tests with DEHS for which initial particle number concentrations were typically too high. Next, a quasi mono-disperse size cut was extracted from the sample flow by directing it through an aerosol aerodynamic

165 classifier (AAC, Cambustion, Ltd., Cambridge, UK; Tavakoli & Olfert, 2013). The major advantage of the AAC is that the
size selection only depends on particle aerodynamic diameter while being independent of charge, i.e., there is no interference
from larger particles carrying multiple charges. For PSL aerosol experiments, the AAC was operated at a resolution parameter
set point of 10 and the nominal diameter setpoint of the AAC was adjusted to maximise particle number concentration
170 in the suspension and possible agglomerated multiplets are removed without causing a shift in modal size of the selected PSL
particles. The AAC set point diameters which maximized PSL transmission agreed within 1 % to the calculated aerodynamic
diameters of the PSL size standards (Table S1), which validates accuracy of size selection by the AAC. For the DEHS aerosol
experiments, the AAC was operated at a resolution parameter set point of 20 to provide a quasi mono-disperse DEHS aerosol
of known size. The AAC aerodynamic diameter set points and corresponding volume equivalent diameters are included in
175 Table S2. In all the tests the sample flow through the AAC was $\sim 1 \text{ L min}^{-1}$. So far, particle generation and selection were
identical for all aerosol experiments. For the mono-disperse test experiments (Fig. 3b) the size-selected aerosol sample was
combined with a stream of particle-free compressed air which was maintained at $\sim 4 \text{ L min}^{-1}$, before being sent to the uNeph
and a condensation particle counter (CPC, TSI Inc., Shoreview, MN, USA, Model 3776) operated in parallel. This provides
phase function measurements for an aerosol of known refractive index (RI), size and number concentration.
180 The experimental setup was slightly modified to generate and probe a broad unimodal aerosol with larger but still moderate
width of the size distribution (Fig. 3c). A holding container (with volume of $\sim 100 \text{ L}$) was placed after the aerosol classification
stage. The AAC was stepped through 6 different set points ranging from 310 nm to 450 nm aerodynamic diameters while
filling the container in flow-through mode over a time period of $\sim 10 \text{ min}$. Afterwards, the outlet valves were closed for
 $\sim 18 \text{ hours}$ to allow the size distribution to become more homogeneous through mixing and coagulation. After the coagulation
185 process, the sample was slowly pushed through the outlet of the holding container by applying a particle free air flow of 0.5 L
 min^{-1} at the inlet. The extracted aerosol sample was diluted with 4 L min^{-1} particle-free air and then distributed to the aerosol
instruments. In addition to the uNeph and the CPC, a scanning mobility particle sizer (SMPS) was employed to measure the
aerosol number size distribution. The SMPS was a combination of a differential mobility analyser (DMA, TSI Inc., Shoreview,
MN, USA, Model 3082) and a CPC (TSI Inc., Shoreview, MN, USA, Model 3775).

190 **3 uNeph data processing**

Deriving calibrated phase function and polarized phase function data from the uNeph raw digital images requires many
intermediate data processing steps. The two main stages are image data reduction and angular signal processing as delineated
in flowcharts (Fig. S2) and detailed in the following.

3.1 Image data reduction

195 3.1.1 Dark signal corrections

Digital images acquired by the CCD contain some signal that is entirely unrelated to actual illumination of the CCD, hereafter referred to as *dark signal*. It can be characterized by acquiring images without illuminating the CCD (Manfred et al., 2018), i.e., with the uNeph laser turned off. The CCD also has a few *hot pixels* that possess abnormally large dark current (e.g., Fig. S3). Image processing, which follows the flowchart in Fig. S2a, starts with hot pixel identification and removal (Fig. S2a),
200 as detailed in Section A1. Subtraction of dark signal contribution as detailed in Section A2 comes next. The dark signal itself has two systematic contributions, a positive bias (constant) and dark current (proportional to exposure time), as well as superimposed random noise. Accordingly, two constants are sufficient to characterize the systematic components of the dark signal as a function of the exposure time (Eq. A2). Figure S4 demonstrates that this correction approach works well to subtract dark signal contribution with small residuals.

205 3.1.2 Scattering angle calibration

The relationship between scattering angle and image pixels is determined through the scattering angle calibration, which is described in more detail in Section A3. Briefly, this is achieved by relating an axial position inside the laser beam with both the scattering angle and the corresponding image pixel coordinates. For this purpose, a pinhead mounted on a 3D translation stage is placed at ~~several~~ 27 different positions inside the forward and the backward laser beams (Figs. ~~S5a~~ S5 and S6). An
210 image of the diffused laser light is recorded along with the corresponding coordinates of the pinhead position. Figure S5b illustrates how the laser beam axis, the pinhead position (S), and the center of the pinhole objective (P) define the scattering angle (θ). The center of the pinhole objective P is not known exactly, which introduces uncertainty into θ as explored and discussed in Section 4.3. The center of the bright spot on the image provides the corresponding pixel (Fig. S6a). This calibration step ultimately provides a list of pixel coordinates along the centerline of the forward and backward beams together with
215 corresponding scattering angles. Such calibration points are shown in Fig. 2b as yellow dots along with a second-order polynomial fit curve through them (red curve).

3.1.3 Angular signal extraction

The laser beam stripes on the image are wider than just one pixel (e.g., Fig. 2b). Limiting further data analysis steps to the centerline pixels would impose unnecessarily high statistical noise in the results (Ahern et al., 2022). Therefore, the next steps,
220 following the flowchart shown in Fig. S2a, aim at integrating the raw image signal along the beam cross-section for each angle. The image is first transformed to bring the laser beam stripes on a straight line in parallel to the new abscissa representing scattering angle. The top (backward beam) and bottom (forward beam) halves of Fig. 2b are separately transformed and stitched together with a common abscissa as shown in Fig. 2c ~~and described in more detail in Section A4~~. The blue lines in ~~these two panels~~ Figs. 2ba and 2cb illustrate the beam cross-section at $\theta=40^\circ$ before and after transformation as an example.

225 [Section A4](#) provides a more detailed description of the image transformation. A regular grid with 1° angular resolution was chosen for the extracted image. No attempts were made to extract the signal with higher angular resolution, given that the information content of measured phase functions is often limited by measurement uncertainties rather than angular resolution.

The transformed image signal is then integrated along the beam cross-section for each angle, illustrated in Fig. 2d for $\theta=40^\circ$ as an example. Integration boundaries were chosen to maximize the ratio of actual light scattering signal to interfering signal contributions such as stray light or dark current residuals. Specifically, we chose the integration limits, indicated by red lines in Figs 2c and 2d, at the pixels where the signal drops to $\sim 10\%$ of the peak signal at the beam center (see Section A4 for details). Integration at each angle provides, at the end of the flowchart shown in Fig. S2a, an angle-dependent scalar value, which represents the raw angular signal, $\Xi(\theta)$ (Figure 2Fig 2e).

3.2 Angular signal processing

3.2.1 Selection of valid signals

The initial data processing steps described in Section 3.1 serve to provide a signal that is, ideally, strictly proportional to image exposure time and without offset. Two signals $\Xi_1(\theta)$ and $\Xi_2(\theta)$ obtained when probing a stable homogeneous scattering medium with two different exposure times $t_{\text{expo},1}$ and $t_{\text{expo},2}$, respectively, are expected to fulfil:

$$\Xi_2(\theta) = \Xi_1(\theta) \times t_{\text{expo}2} / t_{\text{expo}1}. \quad (1)$$

240 In practice, this proportionality relation deteriorates at too small signals, when residuals of dark signal become relevant, or at too high signals, when saturation occurs. Therefore, the data processing contains a step to discard invalid signals outside the strictly proportional range (see flowchart in Fig. S2b). To estimate the valid signal range across which proportionality holds, we measured particle-free air for 12 different t_{expo} covering the range from 0.1 s to 464 s. We calculated the expected t_{expo} -dependence of the angular signal, denoted as $\Xi_{\text{exp}}(\theta, t_{\text{expo}})$, with assuming that Eq. 1 holds and with choosing the largest valid signal below onset of saturation as a reference point. This allows for testing proportionality to t_{expo} by comparing the actual measured signal, denoted as $\Xi_{\text{meas}}(\theta, t_{\text{expo}})$, against $\Xi_{\text{exp}}(\theta, t_{\text{expo}})$ as shown in Fig. 4a for $\theta=50^\circ$. This demonstrates that proportionality is fulfilled in principle. Only a more precise assessment done in Fig. 4b, which presents the relative deviation $\frac{\Xi_{\text{meas}} - \Xi_{\text{exp}}}{\Xi_{\text{exp}}}$, reveals the limits of proportionality. Systematic low bias due to saturation occurs for $\Xi > \sim 7 \cdot 10^5$. Bias also increases for very low signals, e.g., relative deviation exceeds 5% for $\Xi < \sim 2 \cdot 10^3$ in this example. Detailed results for a wide range of angles are shown in Fig. S8 and Fig. S9 for proportionality tests with cooled and uncooled CCD, respectively. Generally, proportionality was fulfilled down to lower values of Ξ when cooling the CCD due to smaller residuals of dark signal subtraction (Fig. S10).

250 Figure S11 shows repeated measurements of particle-free air samples taken over a wide range of exposure times. The two panels with longest exposure times (bottom right) demonstrate how the upper limit (saturation) of the A/D converter leads to capping of the signal (red line) in the centre part of the beam cross-section, which causes a systematic low bias in the integrated

value (blue marker, right axes). At all other exposure times ($t_{\text{expo}} \leq 100$ s), the integrated signal has a high precision (blue markers and error bars), though random noise does become important at $t_{\text{expo}} < \sim 1$ s (for particle-free air at the considered scattering angle).

260 Above results from testing proportionality of signals to t_{expo} were used to define lower and upper limits of raw \mathcal{E} outside which signals are discarded. We only retain signals for which the relative deviation remains below $\sim 6\%$ (lower limit) and for which no saturated pixel occurs (upper limit). Applying these strict limits approximately leaves a dynamic range of ~ 2 -3 or ~ 1 -2 orders in magnitude for cooled or uncooled CCD, respectively (Fig. S10).

265 Unfortunately, three orders in magnitude of dynamic range may not be sufficient to operate the uNeph with a fixed t_{expo} as the signal strength can vary over an even wider range depending on scattering angle, polarization setting and sample properties. For example, the signal at around 90° drops to very low values compared to forward/backward scattering when applying parallel polarization to a Rayleigh scatterer, or forward scattering can exceed backward scattering by orders in magnitude for large particles. When measuring aerosol samples, additional and considerable variation in the signal \mathcal{E} can occur due to statistical fluctuations of the number of particles present inside the laser. ~~Figures S12 Panels a-c in Figure 5 present data from signals for~~ measurements with a high product of DEHS particle number concentration and exposure time, ~~which. This results in homogeneous signals~~ smooth appearance of the light scattered out of the laser beam onto the CCD camera. ~~Here we mean homogeneous in the sense that light scattered out of the laser beam appears as smooth stripes~~ (Fig. S12a5a), and low level of random noise of repeated measurements ~~remains on a low level~~ (represented by the spread of the grey curves in Fig. S12b5b). Consequently, the histogram of repeatedly measured \mathcal{E} values for the aerosol sample (blue bars in Fig. S12e5c) has a narrow width. This implies that the mean number of particles present in the sensitive volume of the laser for this angle during image 275 exposure is similar for all repeats with little statistical fluctuations. The mean value of \mathcal{E} is much higher for the aerosol sample compared to the particle-free air sample shown for comparison (red bars), because the aerosol scattering coefficient clearly exceeds the air scattering coefficient at this particle size and concentration.

By contrast, ~~panels d-f in Fig. 5~~ Fig. S13 shows results for an identical aerosol sample measurements as in panels a-c Fig. S12, but for DEHS particle number concentration being a factor of 1000 lower. In this example, the product of particle number concentration times exposure time is small enough to cause inhomogeneous signals. The stripes in Fig. S13a-5d are not anymore smooth and instead bright spots along the laser beam become discernible. These are caused from single particles crossing the beam during image exposure. Accordingly, the signal along the beam cross-section also has high fluctuations between repeats (spread in the grey lines in Fig. S13b5e). In some repeats, the signal remains at the level of the air background (red line) across the entire beam cross-section or portions of it. The histogram of \mathcal{E} values shows that the aerosol sample signal 285 (blue bars in Fig. S13e5f) is identical to the air background (red bars) for a large fraction of the images because no particle crosses the laser beam at this angle during the exposure. A subset of the \mathcal{E} values is considerably larger, corresponding to cases when a particle is present during the exposure. For this example, the minimal and maximal signal of single images differ by a factor of ~ 3 due to random fluctuations caused by limited particle statistics. This problem can be mitigated by averaging sufficiently large number of images with identical exposure time (red dashed line in Fig. S13b5e).

290 Given this issue of sample homogeneity, probing the full phase function with small random noise may make it necessary to
measure with different exposure times and to include repeats at each of them. A condition for obtaining an unbiased average
is that all repeats taken at a given t_{expo} fall within the proper signal range for a given angle. If some repeats were falling outside
the proper signal range for a given angle, then the average would likely be biased. No bias is expected if either all images are
retained or if all invalid images are discarded at that angle. Therefore, all measurements taken at a t_{expo} with some invalid
295 signals are to be excluded from further data analysis. Discarding data outside proper signal range, is implemented as the first
data processing step starting from the integrated signals \mathcal{E} (see flowchart in Fig. S2b).

In the presence of large particles, the data filtering step may disqualify all t_{expo} . This is due to the fact that the difference in
signal with or without a particle present in the laser beam becomes increasingly large with increasing particle size. Above a
certain size this difference can exceed the proportional range of the instrument, such that it is not possible to find an exposure
300 time for which all signals are unbiased. At longer t_{expo} , images with particles present suffer from saturation. At shorter t_{expo} ,
images without particles present fall below the proper signal range.

One approach to mitigate the unresolvable t_{expo} trade-off is reducing instrument sensitivity by, e.g., reducing laser power with
a stronger ND filter. This would allow for longer t_{expo} without exceeding the saturation limit. Longer t_{expo} reduces signal
fluctuations related to particle statistics, such that also the smallest signals remain above the lower signal limit.

305 3.2.2 Normalization of the signal by exposure time and laser power

The signal \mathcal{E} is proportional to t_{expo} , a data acquisition setting, and to laser power. Hence, the next data processing step is
normalization of \mathcal{E} by t_{expo} and the laser power signal (flowchart in Fig. S2b), in order to account for variations in these
parameters. The forward beam photodetector provides a signal proportional to the laser power by probing the part reflected at
the chamber window (Fig. 1). The normalized signal \mathcal{E} is referred to as the compensated signal and is denoted as $\zeta(\theta)$. This
310 normalization step allows for averaging $\zeta(\theta)$ acquired at different laser power and with different t_{expo} .

3.2.3 Subtracting stray light interference

The compensated signal $\zeta(\theta)$ contains a contribution from stray light background (e.g., light scattered from the walls of the
sampling volume) which interferes with the light scattered by the sample. The next data processing step aims at subtracting
this interference (flowchart in Fig. S2b). In this study we determined the stray light signal contribution by sampling helium
315 (He) gas with the uNeph. The scattering coefficient of helium is more than sixty times smaller than that of air, such that $\zeta(\theta)$
measured for a helium sample typically is dominated by stray light contribution. Therefore, it is commonly used to quantify
stray light interference (Ahern et al., 2022; Manfred et al., 2018). We denote the normalized stray light signal as $\zeta_{\text{SL}}(\theta)$ and
subtract it from $\zeta(\theta)$ to obtain the sample signal, $\zeta_{\text{meas}}(\theta)$. Thus, $\zeta_{\text{meas}}(\theta)$ for an aerosol sample only contains contributions from
light scattered by the carrier gas and the particles (plus residuals from dark signal and stray light corrections). Stray light does
320 not depend on temperature or pressure. Therefore, correction of this interference is kept separate from air background
subtraction (described in Section 3.2.5).

3.2.4 Signal averaging to mitigate random signal fluctuations

For the reasons explained in Section 3.2.1, we typically acquired repeat measurements at different t_{expo} values. The data processing steps done so far (Fig. S2b) provide $\xi_{\text{meas}}(\theta)$, which can be averaged over repeated measurements without further corrections. We apply a weighted averaging in which individual $\xi_{\text{meas}}(\theta, t_{\text{expo}})$ are weighted by their corresponding t_{expo} in order to obtain $\xi_{\text{mean}}(\theta)$. The weighting is introduced as shorter measurements of the same sample are expected to have poorer signal to noise ratio. The results for the air samples presented in Fig. S11 justify this approach. First, the random noise in single measurements is negligible for exposure times of ~ 21.5 s or longer, whereas it is considerable for exposure times of ~ 2.15 s or shorter. Second, the mean results from repeated short measurements (red lines & blue markers) are consistent with the results at long exposure times.

3.2.5 Air background subtraction

For a particle-free gas sample, the preparatory data processing steps (flowchart in Fig. S2b) are complete after stray light subtraction and optional averaging, i.e., $\xi_{\text{gas}}(\theta) \equiv \xi_{\text{meas}}(\theta)$, which only leaves application of the calibration constants as a final step to follow (Section 4). In contrast for an aerosol sample, the signal $\xi_{\text{meas}}(\theta)$ is proportional to the sum of the light scattered by particles as well as the light scattered by the carrier gas (in this case air) present in the laser beam. Thus, subtraction of the air background contribution, $\xi_{\text{BG}}(\theta)$, is an additional step that is required to derive the signal, $\xi_{\text{aerosol}}(\theta)$, that is proportional to the light scattered by the aerosol particles only (flowchart in Fig. S2b). We apply the approach of regular filtered air measurements in order to obtain a reference value for the air background signal $\xi_{\text{BG}}(\theta)$. Data processing for this air background measurement follows the standard *gas branch* of the flowchart in Fig. S2b to obtain $\xi_{\text{air}}(\theta)$. An aerosol measurement is taken at a certain temperature (T) and pressure (p), whereas the air background is measured at potentially different conditions (T_{ref} , p_{ref}). Given that the scattering coefficient of air depends on temperature and pressure, the following correction is applied to ξ_{air} ;

$$\xi_{\text{BG}} = \xi_{\text{air}} \frac{T_{\text{ref}}}{T} \frac{p}{p_{\text{ref}}}. \quad (2)$$

The magnitude of air signal variability in the time window between two air background measurements determines the residual error in the aerosol signal due to imperfect background subtraction. Therefore, we conducted continuous air background measurements over an extended period of time in order to estimate this variability. The variations of ξ_{BG} relative to an arbitrarily chosen reference value ξ_{ref} are presented in Fig. S14-S12 and Fig. S15-S13 for the polarization set points 1 and 2, respectively. In these examples with a duration of ~ 3 days, the systematic drift dominates over random noise, while remaining within a few percent. Figure S16-S14 shows eight air background measurements distributed over 14 days. These results suggest an instrument stability of around $\pm 3\%$ over this duration. This means that residuals from imperfect air background subtraction contribute to random noise in $\xi_{\text{aerosol}}(\theta)$ on the level of $\sim 3\%$ of the air background signal.

4 Instrument calibration and error model

The uNeph data processing steps described in Section 3 provide a signal that is directly related to the light scattering phase function of either gas or aerosol samples. The last remaining step is to determine and apply a set of calibration constants to derive phase functions in absolute units.

4.1 Calibration equations

The measurement and calibration approach applied in the uNeph to determine the absolute phase function ($F_{11}(\theta)$; [$\text{Mm}^{-1}\text{sr}^{-1}$]) and the polarized phase function ($-F_{12}(\theta)/F_{11}(\theta)$; [-]) builds up on the work by Dolgos and Martins (2014), i.e., on taking measurements at two well-defined laser polarization states (i.e., $\xi_1(\theta)$ and $\xi_2(\theta)$). Using the Stokes formalism, the following pair of equations relates the measurements ($\xi_1(\theta)$, $\xi_2(\theta)$) to the scattering matrix elements ($F_{11}(\theta)$, $F_{12}(\theta)$) for a defined sample (either a gas or an aerosol) consisting of an ensemble of randomly oriented scatterers:

$$F_1(\theta) := G_1(\theta)\xi_1(\theta) = F_{11}(\theta) + q_1F_{12}(\theta), \quad (3a)$$

$$F_2(\theta) := G_2(\theta)\xi_2(\theta) = F_{11}(\theta) + q_2F_{12}(\theta). \quad (3b)$$

365

Here, $G_1(\theta)$ and $G_2(\theta)$ are instrument gain calibration factors for the two polarization set points, whereas q_1 and q_2 represent true fractions of linear polarization aligned with the nominal orientation of the linear polarization states. For perfect polarization control, q_1 and q_2 assume the values +1 and -1, representing 100% linear polarization parallel and perpendicular to the scattering plane, respectively. However, polarization control and geometry are not perfect such that q_1 and q_2 are expected to be smaller than +1 and larger than -1, respectively. In Eq. 3 we also define $F_i(\theta)$, which is the actual angular distribution of total scattered light in unit of $\text{Mm}^{-1}\text{sr}^{-1}$ for polarization set point i . To distinguish between the perfect and actual polarization states, we also define the terms F_{para} and F_{perp} , to refer to hypothetically perfect measurements with $q_1 = 1$ and $q_2 = -1$, respectively.

370

4.2 Radiometric calibration using gas samples

The first step in the calibration process is the evaluation of the gain calibration factors. In our analysis, we used particle free air and CO_2 as calibration gases. For the calibration gases, $F_{11}(\theta)$ and $F_{12}(\theta)$ are taken from the literature (Dolgos and Martins, 2014). Initially, we assumed perfect polarization set points, that is $q_1 = 1$ and $q_2 = -1$. Then we obtained $G_i(\theta)$ using Eq. 3 and $\xi_i(\theta)$ measured for the calibration gases. Specifically, we used the difference $\xi_{i,\text{CO}_2}(\theta) - \xi_{i,\text{air}}(\theta)$ to avoid interference from residual dark signal or stray light contributions to $\xi_i(\theta)$. The magenta lines in Fig. S17-S15 show the resulting angle-dependent gain calibration factors. Additionally, gain factors derived from single gas calibration measurements, i.e., with either using $\xi_{i,\text{CO}_2}(\theta)$ (red lines) or using $\xi_{i,\text{air}}(\theta)$ (blue lines) are also shown. The fact that all three curves are very similar demonstrate that the residual signal offset is very small, and that our gain calibration has a high precision.

380

In a next calibration step, we measured argon (Ar) gas to examine the validity of the assumed q values. Ar is a monatomic gas for which $F_{\text{para}}(\theta)$ approaches zero as θ approaches 90° (Fig. S18-S16). This makes it ideal to reveal errors in angle calibration or q_i . Indeed, the comparison of uNeph measurements and theoretical curves in Fig. S18-S16 (and its variant Fig. S19-S17 zoomed in at $\theta \approx 90^\circ$) suggests some bias. Applying the gain calibration, obtained with the assumption $q_1=1$, results in systematically greater uNeph measurement compared to the theoretical values in the angle range of $\sim 75^\circ$ to 105° . This suggests that the uNeph polarization control is not as perfect as assumed. Figures S18-S16 and S19-S17 additionally contain curves obtained with assuming a range of different q_i values. The measurements obtained by assuming $q_1 = 0.92$ (in the forward angular direction) and $q_1 = 0.95$ (in the backward angular direction) closely match the theoretical results. Therefore, we used these q_1 values for the subsequent data analyses. Furthermore, we also assumed that $|q_1| = |q_2|$. This is a simplified approach to calibrate the actual polarization states. Using Ar validation data alone is not sufficient to disambiguate the bias from q calibration and the bias from angle calibration. Therefore, we used a PSL aerosol with well-constrained properties to further optimize the uNeph calibration.

4.3 Refining calibration using PSL size standards

The radiometric calibration and the scattering angle calibration are inter-connected through Eq. 3. The q values were identified as a main source of uncertainty in radiometric calibration (Section 4.2). The exact position of the camera's pinhole is the main source of uncertainty in angle calibration (Section 3.1.2). In order to refine these calibration parameters, we use a PSL aerosol with well-constrained properties as a further calibration reference. Using PSL standards as a calibration step was previously implemented by Ahern et al (2022).

In this study, the monodisperse PSL aerosol was generated as described in Section 2.2. The expected absolute phase functions were calculated using Mie theory, with RI taken from the literature (Kasarova et al., 2007; Ma et al., 2003), and assuming a lognormal size distribution. We use the MiePython package (<https://github.com/scottprahl/miepython>) to carry out the Mie calculations. The lognormal parameters were taken from the certified diameter (600 nm), reported coefficient of variation (1.7%) and number concentration measured by the CPC.

For a given set of q values and pinhole location P , it is possible to process the uNeph data all the way to absolute phase functions. By varying the value of P for fixed q values it is possible to optimize the angle calibration by choosing the value which results in the smallest least-squares sum of residuals between expected and measured phase functions. Changing angle calibration, deteriorates agreement for the Ar data, which were used to optimize q values. Therefore, q values and pinhole location were alternately optimized in a few iterations.

Figure S20-6 illustrates the effect of optimizing the coordinates of the pinhole center. The dashed red-lines are the uNeph data processed with the final optimized q values along with the initial coordinates of the pinhole center taken from the angle calibration process (Section 3.1.2). The solid red lines are the uNeph data processed with the final optimized q values along with final optimized coordinates of the pinhole center. Comparison of these phase functions against the expected phase functions illustrates the considerably improved agreement after optimization.

The fact that good agreement is achieved at the end of this optimization process for one specific calibration aerosol, does not necessarily imply that the calibration constants are physically meaningful. Therefore, the uNeph measurement still needs to be validated using aerosol samples with well-constrained properties. Such validation results are presented in Section 5.

It can be seen in Fig. S20-6 that the final calibrated phase function data have a gap in the angle range between $\sim 85^\circ$ and $\sim 95^\circ$.

420 The angle of the camera's central axis relative to the laser beams was not optimally chosen in the uNeph prototype such that the portions of the forward and backward beams corresponding to this angle range fall outside the camera's field of view. This type of angular truncation (side angle truncation) is unfortunate, however, it does not substantially affect the retrieval of aerosol parameters from uNeph measurements, as shown in Moallemi et al. (2022).

4.4 Measurement error assessment

425 To better understand and quantify the errors in uNeph measured phase functions, we developed an instrument error model. This model contains the major uNeph error sources as well as uncertainty values for the parameters that govern each type of error. As demonstrated in Sections 4.2 and 4.3, uncertainties in the exact laser polarization state and angle calibration are important sources of uNeph measurement error. Air background subtraction and measurement precision also contribute to the measurement error. We account for these four independent sources of error and consider their contributions to the total phase
430 function measurement error (σ_{tot}) to be independent of each other. Hence, we combine them following standard error propagation for independent errors:

$$\sigma_{tot,l}(\theta) = \sqrt{\sigma_{\theta,l}^2(\theta) + \sigma_{BG,l}^2(\theta) + \sigma_{q,l}^2(\theta) + \sigma_{\epsilon,l}^2(\theta)}, \quad (4)$$

where $\sigma_{\theta,l}$, $\sigma_{BG,l}$, $\sigma_{q,l}$, and $\sigma_{\epsilon,l}$ represent individual contributions to phase function error arising from uncertainties in angle calibration, background subtraction, polarization state calibration, and signal precision, respectively. The subscript l is a
435 placeholder denoting errors of F_1 , F_2 , F_{11} or $-F_{12}/F_{11}$. A detailed description on the evaluation of different error components is provided in the section A5 of the appendix.

The total measurement error, $\sigma_{tot,l}$, obtained using the error model based on Eq. 4 and its constituting components, are shown in Fig. S-7 for the 600 nm PSL aerosol test case. The black lines in panels e-f of Fig. S-7 demonstrate that the total measurement error strongly depends on angle due to variable contributions from individual error components. In this example, estimated
440 total error remains below 10% in F_{11} for most angles, and below 0.1 (absolute) in $-F_{12}/F_{11}$ for all angles. The complexity of the uNeph measurement errors make it necessary to use an error model for precise error estimates as a function of phase function shape, aerosol concentration and scattering angle. The 600 nm PSL aerosol test case, which was used to illustrate how different components contribute to measurement error, is not a rigid test for the error model given that these measurement data were also used to refine the angle calibration (Section 4.3). Therefore, further validation of estimated error magnitudes is
445 presented for the uNeph validation experiments discussed in Section 5.2.

5 Instrument validation and example application

5.1 Validation of phase function absolute values

The calibration approach for the uNeph described in Section 4.2 is designed to provide phase matrix elements $F_{11}(\theta)$ and $F_{12}(\theta)$ in absolute units ($[\text{Mm}^{-1}\cdot\text{sr}^{-1}]$), as opposed to just providing normalized phase matrix elements $P_{11}(\theta)$ and $P_{12}(\theta)$ (unit: $[\text{sr}^{-1}]$).
450 Here, we assess the level of accuracy of the measured absolute values, which depends on the accuracy of the gain calibration factors G_i (Eq. 3), as well as the precision in compensated signal ζ . To do so, we used mono-disperse PSL size standards with diameters of 240 nm and 600 nm and the experimental setup shown in Fig. 3b. Validation was done for $F_1(\theta)$ and $F_2(\theta)$ measured by the uNeph. The size distribution parameters (modal size and width) and RI values are fixed and known for PSL aerosols, thus making measured $F_i(\theta)$ strictly proportional to particle number concentration at any angle. Therefore, it is
455 possible, using Mie theory, to infer particle number concentration directly from $F_i(\theta_j)$ measured at a single angle θ_j . This was done for all angles with valid uNeph measurements, for the two polarization set points and for the two PSL sizes. Statistics of particle number concentration values determined with this approach are provided in Table 1 (4th and 5th column). The coefficients of variation (CV) of the uNeph-derived particle number concentrations over θ were as low as 4%, 12%, 5% and 2% for the two PSL sizes and polarization set points. These results demonstrate that precision of the gain calibration for single
460 angles is very high, given that errors in PSL size distribution properties and other random noise can also influence the coefficients of variation. The relative bias of mean uNeph-derived particle number concentrations compared against independent measurements by a CPC is listed in the last column of Table 1. Agreement is excellent with bias ranging from -4.6% to $+3\%$. This is actually much better than the specified CPC uncertainty of $\sim 10\%$. Altogether we conclude that the gain calibration is very precise and that there is no evidence of systematic bias that goes beyond CPC uncertainty.

465 5.2 Validation of phase functions using quasi-monodisperse aerosol

The uNeph data presented in Fig. 5-7 do not validate its performance for measuring any type of phase function, because this experiment was also used to refine the angle calibration. Therefore, uNeph performance was further validated by probing quasi-monodisperse spherical aerosol particles with known complex RI and diameters of 200 nm, 400 nm, 600 nm and 800 nm. These validation aerosols were generated by extracting a narrow size cut from a broad unimodal DEHS aerosol by means of
470 an AAC (see Fig. 3b for experimental setup). The sampling period was ~ 60 min for each size and included- the recording of multiple repeats of data at different exposure times ($0.1 \text{ s} \leq t_{\text{expo}} \leq 100 \text{ s}$) to ensure that valid signals were collected for all angles and polarization set points (Section 3.2.1). The aerosol number concentration was quite stable and small drifts during the experiments were accounted for in the data analysis.

DEHS is a liquid, which therefore results in spherical particles when aerosolized. This makes it possible to use Mie theory to
475 calculate expected light scattering phase functions. Nevertheless, the true phase function is not exactly known, due to uncertainties in the aerosol parameters that are required for input in the Mie calculation. A complex RI of $1.455 + 0i$ at 532 nm was used based on Pettersson et al (2004). The AAC classified particles were assumed to have log-normal size distributions

with best estimates for modal diameter inferred from AAC aerodynamic diameter setpoints (see Table S3) and for number concentration taken from the CPC measurement. To account for uncertainties in the properties of the AAC-extracted aerosol, a Monte Carlo simulation was performed to calculate a range of expected phase functions by varying the size distribution parameters. The complex RI was held fixed and normally distributed errors were assumed for modal diameter and number concentration with coefficients of variation equal to 3% and 10%, respectively. The width of the lognormal size distribution, expressed as geometric standard deviation (GSD), is not exactly known. Therefore, possible GSD values were assumed to be evenly distributed between the limits 1.04 and 1.08. F_{11} and $-F_{12}/F_{11}$ were calculated for 1000 sets of randomly drawn size distribution parameters as described above, and the interquartile range of all resulting phase function values at a given angle is assumed to represent uncertainty of the true phase function as constrained by AAC and CPC.

The uNeph validation results are presented in Fig. 68. The uncertainty range of expected F_{11} and $-F_{12}/F_{11}$ is shown as grey shadings, and the measured phase functions are shown as red lines with error bars calculated with the error model presented in Section 4.4. The F_{11} measurements fall well within the uncertainty range of expected phase functions over all available angles and for all four sizes (top panels). More precisely, most data points fall into the uncertainty range of the predictions even without error allowance on the uNeph measurement. There is hardly any disagreement between expected and measured F_{11} that exceeds the measurement errors. The findings are quite equivalent for the $-F_{12}/F_{11}$ function: the measurements fall well within the uncertainty range of expected phase functions (bottom panels in Fig. 68), for most part even without allowance for measurement error. Disagreement that exceeds the estimated measurement errors only occurs at a few angles for the 800 nm particles. The uncertainties in the true phase functions, i.e. the widths of the grey shadings, are quite considerable despite using well-defined reference aerosols. This impedes stringent test of the error model, i.e. reliable identification of potentially underestimated measurement errors.

An alternative approach to validate the phase function measurements is to use them to retrieve the properties of the test aerosol size distributions, which are assumed to be of lognormal shape and described by the vector \mathbf{v}_{psd} defined in Eq. 5a with the elements geometric mean diameter (d_m), geometric standard deviation (GSD), and total particle number concentration (N_{tot}).

$$\mathbf{v}_{psd} = \begin{bmatrix} d_m \\ GSD \\ N_{tot} \end{bmatrix} \quad (5a)$$

For this purpose, a simple retrieval scheme was applied to retrieve \mathbf{v}_{psd} , optimized such that the corresponding phase functions calculated with Mie theory achieve best fit to the phase function measurement data, specifically F_1 and F_2 , according to the least square minimization given in Eq. 5b.

$$\mathbf{v}_{psd, fit} = \min_{\mathbf{v}_{psd}} \left(\frac{\sum_{\theta_1}^{\theta_2} N (\ln(F_{1,meas}(\theta_i)) - \ln(F_{1,Mie}(\theta_i, \mathbf{v}_{psd})))^2 + \sum_{\theta_1}^{\theta_2} N (\ln(F_{2,meas}(\theta_i)) - \ln(F_{2,Mie}(\theta_i, \mathbf{v}_{psd})))^2}{N} \right), \quad (5b)$$

where N is the number of measured angles. The complex refractive index of DEHS at 532 nm was again taken as $1.455 + 0i$ (Pettersson et al., 2004). The blue lines in Fig. 6-8 correspond to F_{11} and $-F_{12}/F_{11}$ calculated with the best fit v_{PSD} . These curves match the measurement data within measurement error except for very few data points. This shows that the shape of the measured phase functions is physically meaningful. The corresponding retrieved size distribution parameters are included in Table S3 along with independent data. The retrieved GSD values ranged from 1.035 to 1.065, which is very narrow but within a plausible range for the given AAC resolution parameter settings. The retrieved number concentrations agree with the CPC data to within -2% to +6%, which falls within the uncertainty range of the CPC. The retrieved diameters agree with the AAC data to within -0.5% to -3.8%, which is very good agreement, though located at the edge of expected AAC uncertainty. The best-fitted theoretical F_{11} fits (blue lines) also fall within the uncertainty range of the true phase functions as indicated by the grey shadings. However, the course of the blue lines within the grey shading areas is not random, instead they closely follow one edge of the shading for the three AAC set points shown in Panels c) to h). Additional analyses, presented in Fig. S21, demonstrate that measured phase functions are self-consistent across different angles. This suggests that the small but systematic low bias of retrieved diameters compared to AAC set points could just as plausibly be attributed to a small bias of the AAC. Figure S24 is identical to Fig. 6 with additional examples of calculated phase functions for different combinations of size distribution parameters. The shape of the measurement data is consistent with at least one of the magenta lines, which all represent slightly smaller diameter than AAC set point while differing in underlying GSD . By contrast, discrepancies between measurement data and all cyan or yellow lines go beyond estimated measurement uncertainty. These findings support the conclusion that measured phase functions are self-consistent across different angles and that they tightly constrain the retrieved diameter. The small but systematic low bias of retrieved diameters compared to AAC set points could just as plausibly be attributed to a small bias of the AAC.

These results demonstrate the successful validation of the uNeph. The magnitude of the modelled uNeph measurement errors are plausible, although drawing firm conclusions is difficult because uncertainties in the properties of the validation aerosol translate to considerable uncertainties in the predicted phase functions. In other words, the information content of the phase functions measured by the uNeph for a unimodal aerosol with known RI is so high, that aerosol properties (i.e., d_m , GSD and N_{tot}) retrieved with a suitable inversion algorithm have uncertainties that are similar or even smaller than the prior knowledge of these properties. This statement is in line with the findings of the uNeph information content analysis presented in Moallemi et al. (2022).

The uNeph measurement accuracy appears to be comparable to that of previous laser imaging type nephelometers. Dolgos and Martins. (2014) estimated errors to be on the level of $\sim 5\%$ for F_{11} and ~ 0.05 (absolute) for PPF in their laser imaging nephelometer. Ahern et al. (2022) reported a precision of $\pm 2\%$ for $F_{11}(\theta)$ and a positive bias of $\sim 30\%$ for the integrated scattering coefficient obtained by integrating $F_{11}(\theta)$ over θ .

The results for the 200 and 400 nm examples shown in Fig. 6-8 (left panels) show that the uNeph can provide phase function measurements between 5° to 175° under favourable conditions. On the other hand, a portion of the phase functions measured for the 600 nm and 800 nm cases in the \sim sub 35° angular range was discarded during the data processing step described in

Section 3.2.1. This is the consequence of operating the uNeph in an overly sensitive configuration combined with insufficient dynamic range, which potentially leads to systematic measurement bias when probing large aerosol particles (see Section 3.2.1 for extensive discussion). Unfortunately, the CCD was accidentally operated without cooling, which affected its dynamic range compared to operation in a cooled state. To investigate how relaxing criteria for filtering proper signals can affect the measurement results, we modified the data processing code to retain the image data acquired with $t_{\text{expo}} = 0.1$ s. These data, shown as black dashed lined in Fig. 68, are systematically larger than the predicted values. Therefore, it is considered important to rigorously filter the raw data following the procedure described in Section 3.2.1 to achieve high quality data.

5.3 uNeph-GRASP retrieval of aerosol properties

Retrieving aerosol properties from measured phase functions is a central application of aerosol polarimetry. Here we present a first test experiment to demonstrate feasibility of aerosol property retrieval from uNeph measurements. We used the GRASP-algorithm (Dubovik et al., 2014) to solve the inverse problem. GRASP is a versatile and well-established inversion algorithm used in a wide range of aerosol remote sensing applications (Dubovik et al., 2021). GRASP uses a multi-term least squares minimization in measurement space as the basis for solving the aerosol-light scattering inverse problem, including consideration of a priori constraints. It allows for using multiple types of measurement inputs, and to retrieve different types of aerosol properties. For our purpose, we tailored the open-source version GRASP OPEN (<https://www.grasp-open.com/>; last access: 25 May, 2022) to handle the single scattering inverse problem for uNeph phase function data, hereafter referred to as *uNeph-GRASP inversion*.

GRASP is mainly designed for atmospheric applications with relatively wide size distributions. The standard GRASP kernel is a pre-computed lookup-table with a finite resolution on diameter scale. This can result in discretization errors during the retrieval for lognormal size distributions with a *GSD* smaller than around 1.2, while such errors are negligible for wider size distributions. Therefore, to assess uNeph-GRASP inversion (i.e., to investigate retrieval of aerosol volume size distribution and RI from uNeph-measured F_{11} and $-F_{12}/F_{11}$) a sufficiently broad aerosol size distribution had to be generated. For this purpose, we generated a broad unimodal DEHS aerosol following the experimental procedure illustrated in Fig. 3c and explained in Section 2.2.

The aerosol was first generated and collected in a tank serving as holding chamber. Then the aerosol was sampled from this tank and probed by the uNeph as well as an SMPS and CPC to provide independent measurements of particle size distributions and number concentrations, respectively. Sampling from the tank led to gradual dilution and concurrent decrease of aerosol concentration during the uNeph measurement. In order to minimize systematic measurement bias, this concentration drift was accounted for in the uNeph data processing. Specifically, the time-resolved particle number concentration data measured by the CPC were used to compensate the $F_1(\theta)$ and $F_2(\theta)$ data, which were measured by the uNeph at different times (~5 min total measurement time per polarization set point with a 1 h time gap in between).

The measurement space considered in the uNeph-GRASP-inversion either consists of i) F_{11} only or ii) F_{11} and $-F_{12}/F_{11}-F_{14}/F_{12}$. The DEHS test aerosol consists of an ensemble of homogeneous spherical particles with equal RI and a unimodal size

575 distribution. Therefore, we chose the following two variants of aerosol state space representations for the uNeph-GRASP inversion: i) lognormal volume size distribution representation (state parameters: total volume concentration, V_{tot} ; geometric mean radius, r_g ; geometric standard deviation, GSD) or ii) binned size distribution representation (state parameters: volume concentrations at each size bin, $V_k = dV/d\log(r_k)$ for 22 size bins with central radii fixed at positions r_k). For both variants, the GRASP default size range covering particle radii from 0.05 μm to 15 μm was used, and particles were assumed to be spherical
580 with real and imaginary parts of RI allowed to vary in the ranges from 1.35 to 1.7 and $10^{-5}i$ to $0.2i$, respectively. The GSD was allowed to vary across the full available range from 1.2 to 3 for the lognormal size distribution representation. The parameters for imposing size distribution smoothness constraints for the binned size distribution representation were chosen based on the values used by Dubovik et al. (2011) for aerosol property retrieval over a single satellite pixel (difference order = 3 and Lagrange multiplier = 0.005 for volume size distribution). No further constraints such as, e.g., forcing size distribution tails to
585 zero were applied.

Figure 7-9 presents the F_{11} and $-F_{12}/F_{11}$ measured by the uNeph for the test sample, together with the fit results for different uNeph-GRASP inversions (*fit* refers to F_{11} and $-F_{12}/F_{11}$ calculated using the GRASP forward model and the inverted aerosol properties). All four uNeph-GRASP inversion configurations result in largely identical fit curves (blue and red colored lines in panels a and b), but for the exception in the angle range $\sim 150^\circ$ to 170° for the binned configurations, which will be addressed
590 later in this Section. The GRASP fits to the F_{11} match the measurements at the majority of angles (panel a). Discrepancies slightly beyond error margins only occur in the angle ranges from $\sim 95^\circ$ to $\sim 105^\circ$ (all retrieval settings) and from $\sim 155^\circ$ to $\sim 170^\circ$ (log-normal retrieval settings). The GRASP fits to the $-F_{12}/F_{11}$ function match the uNeph measurements reasonably well (panel b). However, noticeable differences beyond the error margins occur in the angle ranges from $\sim 60^\circ$ to $\sim 85^\circ$ and from $\sim 110^\circ$ to $\sim 160^\circ$. The exact reasons for these discrepancies for F_{11} and $-F_{12}/F_{11}$ remain elusive as multiple factors can play a
595 role: e.g., residual bias in the compensated aerosol concentration drifts, slightly underestimated measurement errors, or fine structures in the true aerosol size distribution shape that cannot be reproduced by the aerosol size distribution representations implemented in GRASP.

The uNeph-GRASP inversion results for the unimodal DEHS aerosol (i.e., retrieved aerosol volume size distribution and complex RI) are shown in Fig. 8-10. The volume particle size distributions (VPSD) retrieved with the four uNeph-GRASP
600 inversion variants are in close agreement with each other (red and blue colored lines in panel a), which explains the close match of four corresponding GRASP fits to F_{11} and $-F_{12}/F_{11}$ in Fig. 7-9. There is a small but clearly discernible difference in the retrieved VPSD from the binned inversions, which have additional minor modes to the right of the main mode. This leads to better agreement between GRASP-fit and measured F_{11} in the angle range of $150\text{-}170^\circ$ (Fig. 7-9a) compared to the other retrievals. However, substantial particle volume in this size range is not expected based on the aerosol generation process,
605 hence, it could be a result of over-fitting measurement bias. The uNeph-GRASP inversion also retrieves both the real and imaginary parts of the RI. These are in very close agreement among the four inversion variants, i.e., maximal absolute differences are as small as 0.018 and $\sim 3 \cdot 10^{-4}i$ for the real and imaginary parts, respectively (colored markers in Fig. 8-10 and Fig. 8e).

The results discussed so far show that the uNeph-GRASP inversion is robust in the sense that it leads to essentially identical retrieval results for different inversion variants applied to the unimodal test aerosol, which also reproduce the measured F_{11} and $-F_{12}/F_{11}$ reasonably well. As a last step, we compare the retrieval results with independently known or measured properties of the aerosol sample. Validation of retrieved VPSD is done against independent measurements by an SMPS. Figure 8a-10a shows that the mode and width of the size distributions from SMPS (black line) and uNeph-GRASP inversions (coloured lines) are consistent, while the magnitude of the retrieved size distribution is larger than that of the independent measurement. For a quantitative comparison, the integral properties total volume concentration (V_{tot}), geometric mean radius (r_g) and GSD were calculated from the measured and retrieved VPSDs (when not directly delivered as retrieved parameter).

The results listed in Table 2 demonstrate agreement for V_{tot} between SMPS and each retrieval result within 45% or better. This is a fair agreement, though outside the range of expected uncertainties of either approach under optimal performance. The reasons for this discrepancy remain elusive. In contrast to V_{tot} , the retrieval results for the state parameters r_g and GSD indicate that the uNeph-GRASP retrieval and the SMPS measurement agree quite well (Table 2). The binned retrievals have the largest r_g and GSD , which is caused by the minor tail in the retrieved VPSD at larger diameters. Agreement with the SMPS remains good despite this retrieval artefact. The lognormal retrieval variants provide slightly larger r_g than the SMPS (+10%; 0.275 μm instead of 0.250 μm) and slightly narrower GSD (~1.22 instead of 1.3). This is good agreement, thus validating the uNeph-GRASP inversion for retrieving VPSD width and size for unimodal aerosols.

Excellent agreement between independent knowledge and uNeph-GRASP inversion was also achieved for the RI. The literature value for n (i.e., the real part of the RI) of DEHS (1.455; Pettersson et al. 2004) is virtually identical to the retrieval results (~1.431 to 1.449), whereby the lowest retrieval value originates from the F_{11} -only/binned retrieval which has a second mode in the size distribution. The retrievals also correctly return a negligibly small value for k , i.e., the imaginary part of RI, ($1 \cdot 10^{-5}i$ to $5 \cdot 10^{-4}i$), which is perfectly consistent with the fact that DEHS is a non-absorbing liquid with k smaller than $1 \cdot 10^{-4}i$ (Verhaege et al., 2009).

The accuracy achieved in retrieving aerosol properties for the broad unimodal DEHS aerosol test case can be explained by the high information content of uNeph measurements. Recently, an information content analysis conducted by Moallemi et al. (2022) demonstrated that for unimodal DEHS aerosol test cases, even a polar nephelometer with high angular resolution basic radiometric configuration, i.e., only F_{11} measurements at a single wavelength, can already be quite informative in retrieving aerosol state parameters. Considering that the level of the uNeph measurement errors for the broad unimodal test aerosol is similar or lower to the base case in the aforementioned information content study, it is not surprising that aerosol properties can be retrieved with high accuracy. Furthermore, the missing benefit in retrieval accuracy of including $-F_{12}/F_{11}$ is not surprising, as this benefit becomes more prominent for more complex aerosols, as shown by Moallemi et al. (2022).

It should be noted that retrieval of k using light scattering measurements is generally more challenging than for other aerosol properties. Often, accurate absorption retrievals require auxiliary measurements, such as aerosol extinction or absorption (Schuster et al., 2019). The information content study by Moallemi et al. (2022) indicates that the scattering measurement has higher information content for retrieval of k when the aerosol is non-absorbing. This, in combination with limited complexity

of the probed aerosol, explains why good results were also achieved by the uNeph-GRASP inversion for retrieval of k for the non-absorbing DEHS aerosol test case.

645 Overall, the results from this experiment validate the uNeph-GRASP inversion to perform in situ polarimetric measurements and retrieve aerosol properties. Espinosa et al. (2017, 2019) have already shown the applicability of GRASP for aerosol property retrieval from laser imaging nephelometer measurements. Our results demonstrate that the GRASP inversion can also be applied to measurements from more compact single wavelength laser imaging nephelometers, such as the uNeph.

6 Conclusions

650 This paper introduces a new laser imaging nephelometer, the uNeph, which measures the scattering coefficient as a function of polar angle at two different states of polarization. These measurements are used to derive the absolute scattering phase function, F_{11} , and polarized phase function, $-F_{12}/F_{11}$. The instrument design and all key data processing steps are presented. We further discuss all calibration parameters and the calibration process relying on measuring both gases and a PSL aerosol size standard to achieve optimal calibration accuracy.

655 We constructed an error model and characterized the uncertainties of the parameters driving the overall measurement error. This makes it possible to provide quantitative estimates of measurement error, which depend on actual results such as phase function shape and absolute scattering intensities. Estimated measurement errors mostly are between 5% to 10 % for F_{11} and smaller than ~ 0.1 (absolute) for $-F_{12}/F_{11}$, while errors can become larger close to the lower limit of detection.

The uNeph instrument was validated using DEHS mono-disperse aerosol particles with aerodynamic diameters of 200 nm, 660 400 nm, 600 nm and 800 nm serving as reference. Good agreement was achieved between the measurements and theoretical phase functions predicted for the reference aerosols. The error model provides plausible measurement errors. However, it was not possible to rigorously validate the error model. This is mainly due to the fact that the estimated measurement errors translate to corresponding uncertainties in aerosol property space which are smaller than or comparable to the available independent knowledge of the reference aerosol properties. The results further demonstrate that it is possible to cover an angle range 665 between $\sim 5^\circ$ to $\sim 175^\circ$ for suitable samples. The small detection cell combined with high sensitivity imposes some limitations. For instance, a smaller sensitive volume is more susceptible to statistical fluctuations of the average particle number inside the sensitive volume which can limit temporal resolution at low particle number concentrations. Moreover, signal saturation can limit the maximum detectable particle size, above which systematic bias occurs in parts of the phase function, if operated at comparable sensitivity for homogeneous samples (e.g., particle free air).

670 Finally, we performed an experiment for testing the combination of the uNeph and the GRASP inversion algorithm for aerosol property retrievals. This was successfully achieved for a broad unimodal DEHS aerosol sample, i.e., retrieved volume concentration, modal size, size distribution width, and complex refractive index agreed within uncertainty with independent measurements and literature data. Generally, these experiments demonstrate the high information content of uNeph measurement data, i.e., radiometric calibration and measurement errors are sufficient for quantitative aerosol retrieval.

675 Overall, the uNeph was successfully validated and our results shows the applicability of a down sized laser imaging nephelometer. Downsizing the instrument posed challenges which result in slower measurements of aerosol samples containing large particles present in small number. At the same time, dimensions of the uNeph have be considerably reduced compared to previous laser imaging nephelometers. This demonstrates that operation of such instruments on, e.g., unmanned aerial vehicles is achievable. Such increased flexibility enables acquisition of in situ aerosol polarimetry data sets, which can
680 greatly benefit the remote sensing community for validating and improving existing retrieval algorithms. Future improvements to the uNeph will aim at increasing the dynamic range and the number of measurement wavelengths, with the goal to use it in laboratory and field settings for aerosol characterization, as well as for validation and optimization of polarimetric aerosol property retrievals for more complex aerosols.

685 ~~Overall, the uNeph was successfully validated. Future improvements will aim at increasing the dynamic range and the number of measurement wavelengths, with the goal to use it in laboratory and field settings for aerosol characterization, as well as for validation and optimization of polarimetric aerosol property retrievals for more complex aerosols.~~

Appendix A: Image data reduction

A1 Hot pixel removal

690 To remove hot pixels, we used the median of multiple dark image samples (3 repeats) that were acquired at an exposure time of 700 s, which is denoted as $\overline{I}_{m_{D,700s}}$. Assuming that the dark current is mostly homogeneous over the image pixels of $\overline{I}_{m_{D,700s}}$ we calculated the median and standard deviation (std) of all the pixel values of $\overline{I}_{m_{D,700s}}$ which were used in Eq. A1 to define a hot pixel signal threshold (L_{hot}):

$$L_{hot} = \text{median}(\overline{I}_{m_{D,700s}}) + 1.5 \times \text{std}(\overline{I}_{m_{D,700s}}), \quad (\text{A1})$$

695 Any pixel in $\overline{I}_{m_{D,700s}}$ with signal larger than the threshold value is considered a hot pixel and is removed from the image prior to any further signal processing steps. Figure S3 shows the hot pixels in a sample dark image. It should be noted that this process was also tested with dark image samples acquired at exposure times of 215 s and 464 s and the detected pixels were quite similar to the hot pixels obtained by sample images with exposure time of 700 s. Specifically, the 215 s and 462 s test cases detected 2 and 1 hot pixels less than the 700 s samples, respectively.

700 A2 Dark signal characterization

The CCD image data contain a dark signal contribution which also occurs in the absence of any illumination. The dark signal (DS) has two systematic components, a constant positive bias (B) and dark current proportional to exposure time. These can be described with a linear equation:

$$DS = G_{DC} \times t_{\text{expo}} + B, \quad (\text{A2})$$

705 where G_{DC} is the proportionality constant for the dark current part. The dark signal also contains superimposed random noise, which is not captured by Eq. A2. In our analysis we used two different t_{expo} , of 4.64 s and 700 s to characterize the DS . The DS may vary between pixels. Accordingly, the two constants can either be determined for each pixel, or for each angle (after integration over beam cross-section). We considered the signal from dark images collected at the short t_{expo} of 4.64 s (30 sample images) to mainly consist of B . Therefore, the mean values of dark images at t_{expo} of 4.64 s were used to calculate B .
710 Subsequently, we used the mean signal of dark sample images acquired at the long t_{expo} of 700 s to obtain G_{DC} .
To determine the robustness of the dark signal estimation we applied the dark signal correction on a series of dark image signals acquired over t_{expo} ranging from 0.1 s to 464 s. Residuals of the dark signal compensation, that is the difference between estimated dark signal and actual dark image signals, were integrated over scatter signal bounds for measured scattering angles (Section 3.1.3), and are shown over different t_{expo} in Fig. S4. The results show that the dark residual signal, $\bar{\epsilon}_{\text{DRS}}(\theta)$, are \sim
715 ± 200 a.u. for the uncooled CCD case over the scattering angles for the sample acquired at t_{expo} below 215 s, while for the dark image sample at $t_{\text{expo}} = 464$ s, $|\bar{\epsilon}_{\text{DRS}}(\theta)|$ as large as 500 a.u. were also observed. The results further suggest that for the CCD cooled case the $\bar{\epsilon}_{\text{DRS}}(\theta)$ are $\sim \pm 200$ a.u. for all the tested exposure times. Valid light scattering data near the lower limit of detection can only be achieved if the interference of dark signal residuals is small, i.e., if the contribution of light scattering to the CCD signal substantially exceeds the dark signal residuals. The uNeph data presented in this manuscript were all acquired
720 with uncooled CCD.

A3 Scattering angle calibration procedure

The relationship between scattering angle and image pixels is determined through the scattering angle calibration. To conduct the scattering angle calibration a multi-stepper motor mechanism, which is referred to as the *three-dimensional (3D) position probe*, was employed. Figure S5a shows the three-dimensional (3D) position probe which is made up of three orthogonal step
725 motors traversing stages which are intended to travel along the xyz coordinate axes according to the coordinate system depicted in Fig. S5a. A probing arm with a pinhead at the end of the arm was mounted on the traversing stage that travels along the x axis. The function of the 3D positional probe is to probe the spatial location of the optical objective pinhole and different locations along the path of forward and backward beams. The stepper motors used in the 3D positional probe traverse a distance of 0.025 mm per step.
730 To conduct the scattering angle calibration, the top cap of the scattering chamber is removed. The 3D positional probe is then mounted on top of the uNeph instrument such that the probe arm has access to the scattering chamber. The initial step for conducting the angular calibration is the identification of the objective pinhole coordinate. To do so, the [pinhole-probe](#) was moved from a reference position (the origin coordinate) and was carefully displaced with step motor movements until the probe pinhead reaches the location of the objective pinhole, which we define as point P . The steps taken by the motors from
735 the origin point was recorded and based on that the coordinates of the objective pinhole [isarc](#) specified relative to the origin point. Subsequently, a similar approach is employed to obtain the coordinates of the centre of the laser beams in the yz plane. The next step in the calibration process involves placing the probe pinhead at different locations along the laser beam central

axis and within the field of view of the objective pinhole lens. The coordinate of a point along the laser beam centre axis is denoted as point S . The polar scattering angle then becomes:

$$740 \quad \theta = \arccos\left(\frac{\overline{SP} \cdot \hat{i}}{\|\overline{SP}\|}\right) \quad (\text{A3})$$

In Eq. A3, P is the objective pinhole position, \overline{SP} is the vector connecting point S to P with vector length of $\|\overline{SP}\|$, and \hat{i} is the unit vector along the x axis (laser beam axis). Figure S5b shows an example scheme where scattering angle geometry is depicted for a case where the probe pinhead is placed at a given location S . When the probe pinhead is located at a given position (e.g., point S), the light reflected of the pinhead will be detected by the imaging unit and generating a bright spot in a sample image capture by the CCD camera. During the calibration process the probe was placed in [28-27](#) different locations along the laser beam centre axis of each beam (forward and backward beams). Once placed at each of these locations a picture was taken of the reflected spot of the probe pinhead. Figure S6a shows an example of the pinhead spot detected by the CCD at a single location and Fig. S6b shows a composite image of the combination of multiple pinhead spots images that were taken during the scattering angle calibration process. Thus far, this process provides pairs of polar angles (Eq. A3) and spots on the
745
750 CCD image with finite width. As a last step, an exact coordinate on the CCD image (i.e., single representative pixel) is assigned to each of these spots by calculating its centre of mass (red dots in Fig. S6).

While the pixel-angle information provided by the angular calibration are useful, they are quite limited and the angular difference ($\Delta\theta$) between most of the adjacent calibration pixels are larger than 1° . Therefore, further processing is required to obtain pixel-angle information with angular resolution of 1° . To refine the pixel-angle data, the angular calibration points at
755 each of the beam segments (forward/backward) were used to generate a second-degree polynomial fit which takes angles as input and returns pixel coordinates. These fits can be used to generate a list of refined coordinates corresponding to scattering angles ranging from $3-90^\circ$ (forward beam) to $90-177^\circ$ (backward beam) with angular resolution of 1° .

A4 Image transformation and signal integration limits

Sections 3.1.2 and A3 described how to obtain the red fit curve in Fig. 2b which is a parametrization of pixels coordinates as
760 a function of θ along the centre line of the beam on the CCD image. The next goal is to extract a pixel array representing laser beam image cross-sections for each polar angle (θ_i). For this purpose, lines perpendicular to this fit were obtained as:

$$y - y_{CL}(\theta_i) = -\frac{1}{m(\theta_i)}(x - x_{CL}(\theta_i)) \quad (\text{A4})$$

In Eq. A4, $y_{CL}(\theta_i)$ and $x_{CL}(\theta_i)$ are the x and y coordinates of points along the laser beam centre line chosen with an angular resolution of 1° , and $m(\theta_i)$ is the gradient of the fitted red fit curve at angle θ_i . The angles θ_i were chosen to represent a regular angular resolution of 1° . We then considered the pixels closest to these perpendicular lines to extract the beam cross sections for each (θ_i). Pixels in between these perpendicular lines were ignored. Figure S7 shows the transformed image with beam

cross sections as a function of θ . 120 pixels were extracted for each angle such that the full beam cross-section is included in the transformed image.

770 The next step involves defining the integration limits for each angle, inside which the signal is considered for further analyses. We used the CO₂ sample images under perpendicular polarization condition and chose as boundaries those pixels where the signal dropped to 10% of the maximum signal at the centre of the cross section (Fig. S7). Repetitions of boundary pixel identification performed over a period of approximately one month revealed stability of instrument geometry and optics such that we chose to use the median result for all further data analyses.

775 A5 Evaluation of uNeph measurement error components

The error component $\sigma_{BG,l}$ depends on precision of background subtraction. We estimate a relative error in ζ_{BG} of around $\pm 3\%$. This is based on observed stability and random noise in measured air background data (Figs S14S12-S16S14). The corresponding error $\sigma_{BG,l}$ is obtained as the difference between perturbed and unperturbed measurement results, where the perturbed measurement results are calculated based on perturbing ζ_{BG} by $\pm 3\%$ in the air background subtraction step. This directly provides σ_{BG,F_1} and σ_{BG,F_2} for positive and negative perturbations. The air background values of polarization states 1 and 2 shown in Fig. S14-S12 and Fig. S15S13, respectively, exhibit a high covariance. Therefore, we make the simplifying assumption that air background error is fully covariant for these two measurements. Hence, the perturbed F_{1l} is to be calculated by inserting perturbed F_1 and F_2 into Eq. 3, whereby the air background was perturbed with identical sign. $-F_{12}/F_{11}$ is perturbed equivalently to F_{11} to account for error covariance. The effect of the error component $\sigma_{BG,l}$ can be highly variable depending on the ratio of particle signal to air signal. To demonstrate this, we conducted error analysis on 250 nm mono-disperse DEHS aerosol particles that were measured at two particle number concentration levels of 3145 cm⁻³ (high) and 69 cm⁻³ (low). Figure S22-S19 shows the measured phase function with estimated total error and contributing error components for the high concentration experiment. F_1 , F_2 and F_{11} of the aerosol (red curves in panels a-c) remain well above 3% of the air background (dashed blue lines) at all angles. Therefore, the contribution of $\sigma_{BG,l}$ to error (magenta lines in panels e-h) is estimated to remain below $\sim 5\%$ for all angles. By contrast, F_1 , F_2 and F_{11} of the aerosol (red curves in Fig. S23a-S20a-c) are comparable or smaller than 3% of the air background (dashed blue lines) for backward scattering ($\theta > \sim 110^\circ$). Accordingly, the contribution of $\sigma_{BG,l}$ to error (magenta lines in Fig. S23e-S20e-h) is estimated to be large in this angle range. Indeed, the phase functions measured for the high concentration case are smooth and follow a fitted Mie curve across all scattering angles, while the measurement of the low concentration example is noisier and in poorer agreement with the fitted Mie curve at backward scattering angles (note, Mie curve fitting is discussed in Section 5.2).

790 The error component $\sigma_{BG,l}$ was further assessed on measurements acquired with 600 nm PSL aerosols and is shown in Fig. S7. F_2 for this aerosol (black line) has distinct features, i.e., it drops to very small values at, e.g., $\sim 105^\circ$ and $\sim 155^\circ$, which is typical for monodisperse aerosols in this size range. Accordingly, the error model predicts distinct peaks for σ_{BG,F_2} at these angles (magenta line in Fig. S7f). The relative error from σ_{BG,F_2} reaches up to $\sim 25\%$ and $\sim 15\%$ at 105° and 155° , respectively. This

800 is not the case for F_1 for which the aerosol signal remains clearly above air background at all angles, thus resulting in negligible error from background subtraction (magenta line in Fig. 5e7e). Propagating background subtraction errors in F_1 and F_2 to errors in F_{11} and $-F_{12}/F_{11}$ results in $\sigma_{BG,F_{11}}(\theta) \leq 3\%$ and $\sigma_{BG,-F_{12}/F_{11}}(\theta) \leq 0.03$. The peaks at angles around 105° and 155° in error of F_2 from background subtraction are heavily dampened, but they remain discernible and the dominant source of error at these angles for $-F_{12}/F_{11}$ (magenta line in panel Fig. 5h7h). Overall, the error analysis results for the low and high concentration 250 nm DEHS and the 600 nm PSL examples show that air background subtraction becomes a major source of error whenever aerosol signal becomes too small compared to air background, and further indicate that the error model plausibly reproduces this effect.

The error component $\sigma_{\theta,l}$ depends on angle calibration accuracy. Angle calibration errors mainly depend on accuracy of the measured pinhole location (Section 4.3). Therefore, to estimate angle calibration errors, we perturbed the pinhole position within its estimated uncertainty range. To obtain the angular perturbation ($\theta \pm \delta\theta$), the (optimized) optical pinhole location was perturbed by ± 0.1 mm in x, y and z, creating 27 perturbed cases. The perturbed pinhole configuration and corresponding angles with the largest differences (positive or negative) to the optimal angles were identified and employed in the error analysis. Resulting angle calibration errors, $\Delta\theta$, i.e., the maximal difference between perturbed and unperturbed angles, varied between 0.07° and 1.2° over the full angle range (Fig. S2+S18). The error component $\sigma_{\theta,l}$ was determined as the difference between the uNeph measurement results obtained using either the unperturbed or perturbed angles in the data processing chain. This calculation is somewhat more complicated than, e.g., the air BG perturbation calculation for the following reason: if perturbed angles are assumed to be *true*, then the gain calibration factors G_1 and G_2 derived with unperturbed angles are biased. Therefore, calculation of perturbed phase functions involves multiple changes to achieve a consistent assessment of errors. First, the angle scale is replaced by perturbed angles. Second, the calibration data are re-evaluated to determine perturbed gains. Third, processing of uNeph measurement data for the PSL sample is repeated using these perturbed angles and gain calibration constants. Potential errors in angle calibration are identical for both laser polarization set points. Therefore, angle perturbations $\Delta\theta$ are assumed to be fully covariant when using perturbed F_1 and F_2 measurements to calculate perturbed F_{11} and $-F_{12}/F_{11}$ (equivalently to handling covariance in BG subtraction error). The error component $\sigma_{\theta,l}$ for the 600 nm PSL aerosol test case is shown in Fig. 57. The blue lines in panels e and f demonstrate that $\sigma_{\theta,l}$ remains small for F_1 and F_2 at angles at which their gradients (i.e., derivative by θ) remain small. By contrast, regions with large phase function gradients result in considerable $\sigma_{\theta,l}$ of up to 35% in panel f. The direct relation between F_2 gradient and $\sigma_{\theta,l}$ is clearly seen near the local minima of F_2 , for example across the angle range from 135° - 170° (panels b and f). The error has two peaks at $\sim 147^\circ$ and $\sim 160^\circ$, which are separated by a sharp drop in error at the local minimum of F_2 (at 155°). Propagating σ_{θ,F_1} and σ_{θ,F_2} to $\sigma_{\theta,F_{11}}$ has a dampening effect, essentially because high gradients in F_1 and F_2 occur at different angles, such that $\sigma_{\theta,F_{11}}$ remains below 10%, except for slight exceedance in the angle range 40° to 60° . Similarly, errors in $\sigma_{\theta,-F_{12}/F_{11}}$ remain below 0.1 (absolute) for this test aerosol example (Fig. 57, panel h).

The error component $\sigma_{q,l}$ depends on uncertainty in the laser polarization state parameters q_1 and q_2 , which appear in Eq. 3. We use an uncertainty of $\Delta q \approx \pm 0.05$ for calculating $\sigma_{q,l}$. This estimate for Δq is based on the fact that perturbing q_1 and q_2 by this much results in clearly discernible systematic deviation between theoretical and measured F_1 for argon gas at scattering angles around $\theta = 90^\circ$ (Fig. S19S17). We determined σ_{q,F_1} and σ_{q,F_2} by subtracting uNeph results obtained with using perturbed q values from the unperturbed measurement results. The gain calibration constants G_1 and G_2 were also re-evaluated to ensure consistency, analogously to the error calculation for perturbed θ . Potential calibration biases in q_1 and in q_2 are expected to be independent of each other, hence, corresponding σ_{q,F_1} and σ_{q,F_2} are also expected to be independent of each other. Consequently, standard equations for error propagation of independent measurement errors were used to infer $\sigma_{q,F_{11}}$ and $\sigma_{q, \frac{-F_{12}}{F_{11}}}$ from σ_{q,F_1} and σ_{q,F_2} . Note, this step differs from the corresponding step in the calculation of $\sigma_{\theta,l}$, or $\sigma_{BG,l}$, where error covariance had to be considered. The error component $\sigma_{q,l}$ for the 600 nm PSL aerosol test case is also shown in Fig. S7. The error σ_{q,F_1} (red lines) is small for extreme forward angles (near $\theta = 0^\circ$) and backward angles (near $\theta = 180^\circ$), and is increasingly large as θ approaches 90° , where it exceeds 10%. High errors in σ_{q,F_1} near 90° are caused by the gain calibration step. Calculation of G_1 , as described in Section 4.2, is very sensitive to bias in q_1 . The cause of this effect is that gases have a much higher partial scattering cross section for perpendicular linearly polarized light than for parallel linearly polarized light at scattering angles near 90° . Conversely, the error σ_{q,F_2} remains very small at all angles for the opposite reason (panel f in Fig. S7). Panel g in Fig. 7 shows that the error $\sigma_{q,F_{11}}$ in F_{11} is similar to the error in F_1 at all angles where F_1 is much greater than F_2 . Conversely, the error in $\sigma_{q,F_{11}}$ is dampened, compared to σ_{q,F_1} , at angles where F_2 is similar or larger than F_1 . These effects are nicely seen when, e.g., comparing $\sigma_{q,F_{11}}$ at the angles $\sim 85^\circ$ and $\sim 95^\circ$ with corresponding errors σ_{q,F_1} . Also, the error in the polarized phase function caused by uncertainty of q values, $\sigma_{q, \frac{-F_{12}}{F_{11}}}$, is most pronounced at angles around 90° , where it reaches maximal absolute values of almost 0.1 (panel h in Fig. S7).

The precision component $\sigma_{p,l}$ in the error model is introduced to account for random contributions to measurement error which are not accounted for in the other error components. We already showed that compensated signals of particle free air samples, ζ_{air} , vary by about $\pm 3\%$ over a period of two weeks. This variability is attributed to variations in uNeph sensitivity and other random noise (Figs S14S12-S16S14). By assuming comparable random variability in ζ_{aerosol} , we estimate $\sim 3\%$ relative error for the precision component of measurement error in F_1 and F_2 . This is a rather low level of random noise, however, the assessment of gain calibration variability presented in Section 4.2 supports plausibility of this error estimate, at least for sufficiently high signal levels. Note, larger random noise is expected to occur when fluctuations in detected particle number become relevant as demonstrated in Fig. S13-5 (d,e,f) and discussed in Section 3.2.1. The random error components σ_{ε,F_1} and σ_{ε,F_2} are assumed to be independent of each other, hence, standard equations for propagation of independent errors were used to infer $\sigma_{\varepsilon,F_{11}}$ and $\sigma_{\varepsilon, \frac{-F_{12}}{F_{11}}}$ from σ_{ε,F_1} and σ_{ε,F_2} . The error component $\sigma_{\varepsilon,l}$ for the 600 nm PSL aerosol test case is also shown in

Fig. 57. The cyan lines in panels e and f of Fig. 5-7 reflect the fixed random noise $\sigma_{\varepsilon, F_1}$ and $\sigma_{\varepsilon, F_2}$ directly imposed on F_1 and F_2 . Error propagation leads to $3\% < \frac{\sigma_{p, F_{11}}}{F_{11}} < \sqrt{2} \cdot 3\%$, and the error $\sigma_{\varepsilon, \frac{-F_{12}}{F_{11}}}$ always remains below 0.05 (absolute).

Acknowledgements

865 ~~Financial support was received from MeteoSwiss through a science project in the framework of the Swiss contribution to the global atmosphere watch programme (GAW-CH) and from the Swiss National Science Foundation (BISAR project; SNSF grant no. 200021_204823).~~ The authors acknowledge Gergely Dolgos for his contributions to the initial study design and proposal writing and Nicolas Bukowiecki for contributions to proposal writing. The authors acknowledge Oleg Dubovik, Tatyana Lapyonok, Anton Lopatin, and David Fuertes for the support they provided for running GRASP-OPEN.

870 Data availability

The original contributions presented in this study are included in the article and attached supplementary information. The specific data from this study will be made publicly available on Zenodo if the manuscript is accepted for publication.

Author contributions

MGB conceptualized the study and raised the funding. PG lead the instrument design in exchange with MGB, AM and BTB.
875 AM and PG wrote the data acquisition and data analyses software. AM performed the experiments with advice from PG, BTB, RLM and MGB. AM performed the data analyses and theoretical calculations with advice and contributions from RLM, MGB, and BB. MGB secured the funding for this project. MGB and RLM supervised this project. The original draft was prepared by AM and MGB, and all the co-authors contributed in the writing, revision and editing of the paper.

Competing interests

880 The contact author has declared that none of the authors has any competing interests.

Financial support

Financial support was received from MeteoSwiss through a science project in the framework of the Swiss contribution to the global atmosphere watch programme (GAW-CH) and from the Swiss National Science Foundation (BISAR project; SNSF grant no. 200021_204823).

885 **Bibliography**

- Ahern, A. T., Erdesz, F., Wagner, N. L., Brock, C. A., Lyu, M., Slovacek, K., Moore, R. H., Wiggins, E. B., and Murphy, D. M.: Laser imaging nephelometer for aircraft deployment, *Atmos. Meas. Tech.*, 15, 1093–1105, <https://doi.org/10.5194/amt-15-1093-2022>, 2022.
- Barkey, B., Paulson, S. E., and Chung, A.: Genetic algorithm inversion of dual polarization polar nephelometer data to determine aerosol refractive index, *Aerosol Sci. Technol.*, 41, 751–760, <https://doi.org/10.1080/02786820701432640>, 2007.
- Barkey, B., Paulson, S., and Liou, K.-N.: Polar nephelometers for light scattering by ice crystals and aerosols: design and measurements, in: *Light Scattering Reviews*, Vol. 6, edited by: Kokhanovsky, A. A., Springer Berlin Heidelberg, Berlin, Heidelberg, 3–37, https://doi.org/10.1007/978-3-642-15531-4_1, 2012.
- Bian, Y., Zhao, C., Xu, W., Zhao, G., Tao, J., and Kuang, Y.: Development and validation of a CCD-laser aerosol detective system for measuring the ambient aerosol phase function, *Atmos. Meas. Tech.*, 10, 2313–2322, <https://doi.org/10.5194/amt-10-2313-2017>, 2017.
- Boiger, R., Modini, R. L., Moallemi, A., Degen, D., Adelman, A., and Gysel-Beer, M.: Retrieval of aerosol properties from in situ, multi-angle light scattering measurements using invertible neural networks. *J. Aerosol Sci.*, 163, 105977, <https://doi.org/10.1016/j.jaerosci.2022.105977>, 2022.
- 900 Bohren, C. F. and Huffman, D. R.: *Absorption and scattering of light by small particles*, Wiley-VCH, Weinheim, 530 pp., 2004.
- Boucher, O.: *Atmospheric Aerosols*, Springer Netherlands, Dordrecht, <https://doi.org/10.1007/978-94-017-9649-1>, 2015.
- Boucher, O., Randall, D., Artaxo, P., Bretherton, C., Feingold, G., Forster, P., Kerminen, V.-M., Kondo, Y., Liao, H., and Lohmann, U.: Clouds and aerosols, in: *Climate change 2013: the physical science basis. Contribution of Working Group I to the Fifth Assessment Report of the Intergovernmental Panel on Climate Change*, Cambridge University Press, 571–657, 2013.
- 905 Cohen, A. J., Brauer, M., Burnett, R., Anderson, H. R., Frostad, J., Estep, K., Balakrishnan, K., Brunekreef, B., Dandona, L., Dandona, R., Feigin, V., Freedman, G., Hubbell, B., Jobling, A., Kan, H., Knibbs, L., Liu, Y., Martin, R., Morawska, L., Pope, C. A., Shin, H., Straif, K., Shaddick, G., Thomas, M., van Dingenen, R., van Donkelaar, A., Vos, T., Murray, C. J. L., and Forouzanfar, M. H.: Estimates and 25-year trends of the global burden of disease attributable to ambient air pollution: an analysis of data from the Global Burden of Diseases Study 2015, *Lancet*, 389, 1907–1918, [https://doi.org/10.1016/S0140-6736\(17\)30505-6](https://doi.org/10.1016/S0140-6736(17)30505-6), 2017.

- Dick, W. D., Ziemann, P. J., and McMurry, P. H.: Multiangle light-scattering measurements of refractive index of submicron atmospheric particles, *Aerosol Sci. Technol.*, 41, 549–569, <https://doi.org/10.1080/02786820701272012>, 2007.
- Dolgos, G. and Martins, J. V.: Polarized Imaging Nephelometer for in situ airborne measurements of aerosol light scattering, *Opt. Express*, 22, 21972–21990, <https://doi.org/10.1364/OE.22.021972>, 2014.
- Dubovik, O., Herman, M., Holdak, A., Lapyonok, T., Tanré, D., Deuzé, J. L., Ducos, F., Sinyuk, A., and Lopatin, A.: Statistically optimized inversion algorithm for enhanced retrieval of aerosol properties from spectral multi-angle polarimetric satellite observations, *Atmos. Meas. Tech.*, 4, 975–1018, <https://doi.org/10.5194/amt-4-975-2011>, 2011.
- Dubovik, O., Lapyonok, T., Litvinov, P., Herman, M., Fuertes, D., Ducos, F., Torres, B., Derimian, Y., Huang, X., Lopatin, A., Chaikovsky, A., Aspetsberger, M., and Federspiel, C.: GRASP: a versatile algorithm for characterizing the atmosphere, *SPIE Newsroom*, <https://doi.org/10.1117/2.1201408.005558>, 2014.
- Dubovik, O., Li, Z., Mishchenko, M. I., Tanré, D., Karol, Y., Bojkov, B., Cairns, B., Diner, D. J., Espinosa, W. R., Goloub, P., Gu, X., Hasekamp, O., Hong, J., Hou, W., Knobelspiesse, K. D., Landgraf, J., Li, L., Litvinov, P., Liu, Y., Lopatin, A., Marbach, T., Maring, H., Martins, V., Meijer, Y., Milinevsky, G., Mukai, S., Parol, F., Qiao, Y., Remer, L., Rietjens, J., Sano, I., Stammes, P., Stammes, S., Sun, X., Tabary, P., Travis, L. D., Waquet, F., Xu, F., Yan, C., and Yin, D.: Polarimetric remote sensing of atmospheric aerosols: Instruments, methodologies, results, and perspectives, *J. Quant. Spectrosc. Radiat. Transf.*, 224, 474–511, <https://doi.org/10.1016/j.jqsrt.2018.11.024>, 2019.
- Dubovik, O., Fuertes, D., Litvinov, P., Lopatin, A., Lapyonok, T., Dubovik, I., Xu, F., Ducos, F., Chen, C., Torres, B., Derimian, Y., Li, L., Herreras-Giralda, M., Herrera, M., Karol, Y., Matar, C., Schuster, G. L., Espinosa, R., Puthukkudy, A., Li, Z., Fischer, J., Preusker, R., Cuesta, J., Kreuter, A., Cede, A., Aspetsberger, M., Marth, D., Bindreiter, L., Hangler, A., Lanzinger, V., Holter, C., and Federspiel, C.: A comprehensive description of multi-term LSM for applying multiple a priori constraints in problems of atmospheric remote sensing: GRASP Algorithm, Concept, and Applications, *Front. Remote Sens.*, 2, 706851, <https://doi.org/10.3389/frsen.2021.706851>, 2021.
- Espinosa, W. R., Remer, L. A., Dubovik, O., Ziemba, L., Beyersdorf, A., Orozco, D., Schuster, G., Lapyonok, T., Fuertes, D., and Martins, J. V.: Retrievals of aerosol optical and microphysical properties from Imaging Polar Nephelometer scattering measurements, *Atmos. Meas. Tech.*, 10, 811–824, <https://doi.org/10.5194/amt-10-811-2017>, 2017.
- Espinosa, W. R., Martins, J. V., Remer, L. A., Dubovik, O., Lapyonok, T., Fuertes, D., Puthukkudy, A., Orozco, D., Ziemba, L., Thornhill, K. L., and Levy, R.: Retrievals of aerosol size distribution, spherical fraction, and complex refractive index from airborne in situ angular light scattering and absorption measurements, *J. Geophys. Res. Atmos.*, 124, 7997–8024, <https://doi.org/10.1029/2018JD030009>, 2019.

- Holben, B. N., Eck, T. F., Slutsker, I., Tanré, D., Buis, J. P., Setzer, A., Vermote, E., Reagan, J. A., Kaufman, Y. J., Nakajima, T., Lavenu, F., Jankowiak, I., and Smirnov, A.: AERONET—a federated instrument network and data archive for aerosol characterization, *Remote Sens. Environ.*, 66, 1–16, [https://doi.org/10.1016/S0034-4257\(98\)00031-5](https://doi.org/10.1016/S0034-4257(98)00031-5), 1998.
- Horvath, H., Arboledas, L. A., and Reyes, F. J. O.: Angular scattering of the Sahara dust aerosol, *Atmos. Chem. Phys.*, 10, 945–955, 2018.
- Hu, Q., Qiu, Z., Hong, J., and Chen, D.: A polarized scanning nephelometer for measurement of light scattering of an ensemble-averaged matrix of aerosol particles, *J. Quant. Spectrosc. Radiat. Transf.*, 261, 107497, <https://doi.org/10.1016/j.jqsrt.2020.107497>, 2021.
- Kasarova, S. N., Sultanova, N. G., Ivanov, C. D., and Nikolov, I. D.: Analysis of the dispersion of optical plastic materials, *Opt. Mater.*, 29, 1481–1490, <https://doi.org/10.1016/j.optmat.2006.07.010>, 2007.
- Li, R., Tang, G., Ding, J., Logan, T., Brooks, S. D., Collins, D. R., Yang, P., and Kattawar, G. W.: Laboratory measurements of light scattering properties of kaolinite dust at 532 nm, *Aerosol Sci. Technol.*, 52, 666–678, <https://doi.org/10.1080/02786826.2018.1444729>, 2018.
- Ma, X., Lu, J. Q., Brock, R. S., Jacobs, K. M., Yang, P., and Hu, X.-H.: Determination of complex refractive index of polystyrene microspheres from 370 to 1610 nm, *Phys. Med. Biol.*, 48, 4165–4172, <https://doi.org/10.1088/0031-9155/48/24/013>, 2003.
- Manfred, K. M., Washenfelder, R. A., Wagner, N. L., Adler, G., Erdesz, F., Womack, C. C., Lamb, K. D., Schwarz, J. P., Franchin, A., Selimovic, V., Yokelson, R. J., and Murphy, D. M.: Investigating biomass burning aerosol morphology using a laser imaging nephelometer, *Atmos. Chem. Phys.*, 18, 1879–1894, <https://doi.org/10.5194/acp-18-1879-2018>, 2018.
- 960 Mishchenko, M. I., Cairns, B., Kopp, G., Schueler, C. F., Fafaul, B. A., Hansen, J. E., Hooker, R. J., Itchkawich, T., Maring, H. B., and Travis, L. D.: Accurate monitoring of terrestrial aerosols and total solar irradiance: introducing the glory mission, *Bull. Amer. Meteor. Soc.*, 88, 677–692, <https://doi.org/10.1175/BAMS-88-5-677>, 2007.
- Moallemi, A., Modini, R. L., Lapyonok, T., Lopatin, A., Fuertes, D., Dubovik, O., Giaccari, P., and Gysel-Beer, M.: Information content and aerosol property retrieval potential for different types of in situ polar nephelometer data, *Atmos. Meas. Tech.*, 15, 5619–5642, <https://doi.org/10.5194/amt-15-5619-2022>, 2022.
- [Muñoz, O., Moreno, F., Guirado, D., Dabrowska, D. D., Volten, H., and Hovenier, J. W.: The Amsterdam–Granada Light Scattering Database, *J. Quant. Spectrosc. Radiat. Transf.*, 113, 565–574, <https://doi.org/10.1016/j.jqsrt.2012.01.014>, 2012.](https://doi.org/10.1016/j.jqsrt.2012.01.014)

- Myhre, G., Shindell, D., Bréon, F. M., Collins, W., Fuglestedt, J., Huang, J., Koch, D., Lamarque, J. F., Lee, D., and Mendoza, B.: Climate change 2013: the physical science basis. Contribution of Working Group I to the Fifth Assessment Report of the Intergovernmental Panel on Climate Change, 659–740, 2013.
- 970 Nakagawa, M., Nakayama, T., Sasago, H., Ueda, S., Venables, D. S., and Matsumi, Y.: Design and characterization of a novel single-particle polar nephelometer, *Aerosol Sci. Technol.*, 50, 392–404, <https://doi.org/10.1080/02786826.2016.1155105>, 2016.
- Omar, A. H., Winker, D. M., Vaughan, M. A., Hu, Y., Treppe, C. R., Ferrare, R. A., Lee, K.-P., Hostetler, C. A., Kittaka, C., Rogers, R. R., Kuehn, R. E., and Liu, Z.: The CALIPSO automated aerosol classification and lidar ratio selection algorithm, *J. Atmos. Ocean. Technol.*, 26, 1994–2014, <https://doi.org/10.1175/2009JTECHA1231.1>, 2009.
- 975 Pettersson, A., Lovejoy, E. R., Brock, C. A., Brown, S. S., and Ravishankara, A. R.: Measurement of aerosol optical extinction at with pulsed cavity ring down spectroscopy, *J. Aerosol Sci.*, 35, 995–1011, <https://doi.org/10.1016/j.jaerosci.2004.02.008>, 2004.
- 980 Schafer, J. S., Eck, T. F., Holben, B. N., Thornhill, K. L., Ziemba, L. D., Sawamura, P., Moore, R. H., Slutsker, I., Anderson, B. E., Sinyuk, A., Giles, D. M., Smirnov, A., Beyersdorf, A. J., and Winstead, E. L.: Intercomparison of aerosol volume size distributions derived from AERONET ground-based remote sensing and LARGE in situ aircraft profiles during the 2011–2014 DRAGON and DISCOVER-AQ experiments, *Atmos. Meas. Tech.*, 12, 5289–5301, <https://doi.org/10.5194/amt-12-5289-2019>, 2019.
- 985 Schuster, G. L., Espinosa, W. R., Ziemba, L. D., Beyersdorf, A. J., Rocha-Lima, A., Anderson, B. E., Martins, J. V., Dubovik, O., Ducos, F., Fuertes, D., Lapyonok, T., Shook, M., Derimian, Y., and Moore, R. H.: A laboratory experiment for the statistical evaluation of aerosol retrieval (STEAR) algorithms, *Remote Sens.*, 11, 498, <https://doi.org/10.3390/rs11050498>, 2019.
- Tavakoli, F. and Olfert, J. S.: An instrument for the classification of aerosols by particle relaxation time: theoretical models of the aerodynamic aerosol classifier, *Aerosol Sci. Technol.*, 47, 916–926, <https://doi.org/10.1080/02786826.2013.802761>, 2013.
- 990 Verhaege, C., Shcherbakov, V., and Personne, P.: Retrieval of complex refractive index and size distribution of spherical particles from Dual-Polarization Polar Nephelometer data, *J. Quant. Spectrosc. Radiat. Transf.*, 110, 1690–1697, <https://doi.org/10.1016/j.jqsrt.2009.01.004>, 2009.
- Waldram, J. M.: Measurement of the photometric properties of the upper atmosphere, *Trans. Illum. Eng. Soc.*, 10, 147–187, <https://doi.org/10.1177/147715354501000801>, 1945.

Tables

Table 1 Validating absolute values of measured phase functions using PSL size standards and independent number concentration measurement from a CPC.

PSL diameter	Nominal Polarization Set point	Particle number concentration			
		CPC*	uNeph-derived ** median (q25 Q1, q75 Q3)***	uNeph-derived estimated CV****	Bias of uNeph-derived (uNeph - CPC)/CPC
nm	-	cm ⁻³	cm ⁻³	%	%
600	F_1 (parallel)	103 ± 12	98 (96, 102)	4%	-4.6%
600	F_2 (perpendicular)	110 ± 12	113 (105, 123)	12%	2.7%
240	F_1 (parallel)	1587 ± 36	1596 (1557, 1671)	5%	0.6%
240	F_2 (perpendicular)	1628 ± 39	1608 (1589, 1635)	2%	-1.2%

*Mean CPC measurements with standard deviation as error values (CPC uncertainty: $\sim\pm 10\%$)

**Requires Mie theory constrained with reported diameter and CV of PSL size standards. The number concentration was independently derived from the data points at each angle. Here we report statistics of the results from all measured angles.

*** Q1 and Q3 refer to the lower and upper quartiles, respectively.

****The coefficient of variation is estimated using the formula $(0.741 \times \text{inter-quartile range}) / \text{median}$, which is insensitive to outliers.

Table 2 Size distribution parameters (geometric mean radius, r_g ; geometric standard deviation, GSD ; total volume concentration, V_{tot}) of the broad aerosol test case obtained from SMPS measurements and retrieved from uNeph-GRASP inversions.

		r_g (μm)	GSD (-)	V_{tot} ($\mu\text{m}^3/\text{cc}$)
SMPS measurement (mean)		0.25	1.30	19.7
uNeph-GRASP inversions	Binned with F_{11}	0.30	1.41	28.9
	Binned with F_{11} & - F_{12}/F_{11}	0.32	1.58	28.1
	Lognormal with F_{11}	0.28	1.23	26.3
	Lognormal with F_{11} & - F_{12}/F_{11}	0.28	1.22	25.4

1000

1005

Figures






Figure 1. (a) Side view and (b) top view schematics of the uNeph instrument (the drawing is not to scale).

1010




1015 **Figure 2. (a)** Example of a raw sample image of particle-free air acquired at perpendicular polarization with exposure time of 215 s. **(b)** the angle calibration fit lines (red) and an example cross-section line (blue) indicating pixels corresponding to scattering angle $\theta = 40^\circ$. **(c)** Light scattering image after transformation to angle-pixel coordinates. **(d)** Cross-sectional signal for the example $\theta = 40^\circ$. **(e)** The integrated light scattering signal, Ξ , over measured scattering angles, θ , for a particle-free air sample.



1020

Figure 3. Schematics of experimental setup for probing (a) gases, (b) quasi mono-disperse aerosol, (c) broad unimodal aerosol. The aerosol generation step (solid lines) and the sampling step (dashed lines) shown in panel c) were performed in sequence as described in the main text.



1025

Figure 4. (a) shows $\Xi(\theta=50^\circ)$ vs t_{exp0} for air sample at polarization state 1 collected over different exposure times with CCD cooling turned on. (b) shows the error of Ξ_{meas} relative to Ξ_{exp} as a function of Ξ_{exp} .




Figure 5 (a) and (d) are examples of a single frame of raw data captured by uNeph at $t_{\text{exp0}} = 4.64$ s for 200 nm DEHS aerosol samples with particle number concentration of 1000 #/cc and 1 #/cc, respectively. (b) and (e) are the signals in cross-sectional pixel at scattering angle of 25° for the 200 nm DEHS aerosol samples with particle number concentration of 1000 #/cc and 1 #/cc, respectively. The grey lines are the aerosol samples in each sample frame (64 sample frames for (b) and 47 sample frames for (e)), the blue line is the mean of aerosol signals, and the red line is the mean of background (BG) air (60 sample frames, with $t_{\text{exp0}} = 4.64$ s). (c) and (f) are the histogram of the integrated signal Ξ_{air} (red) and Ξ_{aerosol} (blue) over the sampled frames for 200 nm DEHS aerosol samples with particle number concentration of 1000 #/cc and 1 #/cc, respectively.

— Expected phase function (GMD = 600.0 nm, GSD = 1.017, $N_{CPC} = 106 \text{ \#/cc}$)
 ···· uneph aerosol measurement (initial angle calibration)
 — uneph aerosol measurement (refined angle calibration)





Figure 6 Expected and measured results for a PSL size standard with geometric mean diameter (GMD)=600 nm for (a) F_1 , (b) F_2 , (c) F_{11} and (d) $-F_{12}/F_{11}$. The expected curves are based on Mie theory for homogeneous spheres constrained with certified mean particle diameter (600 nm), geometric standard deviation ($GSD = 1.017$) inferred from reported coefficient of variation and particles number concentration measured by a CPC (N_{CPC}). The grey shading corresponds to the uncertainty range of expected phase functions if the certified GMD is perturbed by the reported uncertainty (± 9 nm). The uNeph measurement is processed with finally calibrated g -values and two different angle calibration curves: the red dashed line represents initial angle calibration, whereas the red solid line represents the refined angle calibration based on optimized pinhole center coordinates.



1050 **Figure 75** (a, b, c, d) are the angular light scattering measurements with total errors for the 600 nm PSL aerosol particles (red lines). The black lines are the expected angular measurements that were calculated using Mie theory according to the description in Section 3.54.3. ~~Air background phase functions are also included (blue lines).~~ (e, f, g,) present estimated measurement errors in relative terms for F_1 , F_2 , and F_{11} , respectively. (h) presents estimated contributions to error of $-F_{12}/F_{11}$ in absolute terms. **The contribution of air background uncertainty to measurement error remains small (magenta lines in panels to the right) unless the aerosol phase function (red lines in panels to the left) approaches values only slightly above 3% of the air background (blue dashed lines). This effect is nicely seen in panels (b) and (f) at angles $\sim 105^\circ$ and $\sim 155^\circ$.**

1055




Figure 6. Phase function (F_{11}) and polarized phase function ($-F_{12}/F_{13}$) measurements for different mono-disperse DEHS aerosol test cases. The nominal aerodynamic diameters are 200 nm (panels a and b), 400 nm (panels c and d), 600 nm (panels e and f), and 800 nm (panels g and h), and the respective particle number concentrations measured by a CPC were, $190 \pm 8 \text{ cm}^{-3}$, $494 \pm 11 \text{ cm}^{-3}$, $50 \pm 4 \text{ cm}^{-3}$, and $70 \pm 5 \text{ cm}^{-3}$. The volume equivalent diameter for each sample is reported in the titles. The grey shadings indicate the uncertainty range of independently constrained phase functions considering uncertainties of input parameters for the Mie calculations (see Section 5.2 for details). The Blue lines in the figure are the best-fit phase functions obtained with a least square minimization of residuals of calculated Mie curves compared with the measurement.






Figure 68. Phase function (F_{11}) and polarized phase function ($-F_{12}/F_{11}$) measurements for different mono-disperse DEHS aerosol test cases. The nominal aerodynamic diameters are 200 nm (panels a and b), 400 nm (panels c and d), 600 nm (panels e and f), and 800 nm (panels g and h), and the respective particle number concentrations measured by a CPC were, $190 \pm 8 \text{ cm}^{-3}$, $494 \pm 11 \text{ cm}^{-3}$, $59 \pm 4 \text{ cm}^{-3}$, and $70 \pm 5 \text{ cm}^{-3}$. The volume equivalent diameter, D_{ve} , for each sample is reported in the titles. The grey shadings indicate the uncertainty range of independently constrained phase functions considering uncertainties of input parameters for the Mie calculations (see Section 5.2 for details). The Blue lines in the figure are the best fit phase functions obtained with a least square minimization of residuals of calculated Mie curves compared with the measurement.



1065 **Figure 7-9** Phase function (F_{11}) and polarized phase function ($-F_{12}/F_{11}$) measurements by the uNeph and GRASP retrieval results for the broad unimodal DEHS test case.





Figure 8 (a) volume size distribution measurement by SMPS and retrievals provided by applying GRASP on uNeph measurements. (b, c) Retrieval of real (n) and imaginary (k) parts of refractive index (RI) by applying GRASP on uNeph measurements. Figure 10 Retrieval results obtained by applying GRASP to uNeph data (a) volume size distribution compared against independent SMPS measurements. (b) Real (n) and imaginary (k) parts of the complex refractive index (RI) compared against literature data.

1070

1075

Supplementary Information (SI): Concept, absolute calibration and validation of a new, benchtop laser imaging polar nephelometer

Supplementary Section S1:

5 Supplementary tables:

Table S1. size specification of the PSL samples. A material density of 1050 kg m^{-3} for PSL was used for diameter conversion.

Certified geometric mean diameter (nm)	Specified coefficient of variation	AAC Aerodynamic diameter set point (nm)	Equivalent V volume equivalent diameter (D_{ve}) of AAC aerodynamic diameter set point (nm)
600 ± 9	0.07 1.7%	620	603
240 ± 5	1.5% 0.05	250	242

Table S2. AAC aerodynamic diameter set point and corresponding equivalent volume diameter for tested DEHS aerosols. A material density of 900 kg m^{-3} for DEHS was used for diameter conversion.

AAC aerodynamic diameter set point (nm)	Equivalent volume diameter of sample, d_{ev} D_{ve} (nm)
200	215
250	267
400	425
600	636
800	847

Table S3. Parameters of the experiments performed with AAC selected mono-disperse DEHS aerosol samples: AAC set point (aerodynamic diameter), corresponding ~~volume equivalent diameter~~ ($d_{\text{ret}}D_{\text{ve}}$) (converted using material density of 900 kg m^{-3} and assuming spherical shape), particle number concentration measured by the CPC (N_{CPC}). The following columns present size distribution parameters retrieved from phase functions measured by the uNeph using the least square minimization scheme: geometric mean diameter ($d_{\text{ret}}D_{\text{ret}}$), geometric standard deviation (GSD_{ret}), and particle number concentration (N_{ret}). The last two columns present the relative deviation of retrieved parameters from corresponding independently determined values.

AAC set point	AAC $d_{\text{ret}}D_{\text{ve}}$	N_{CPC} mean \pm SD	$d_{\text{ret}}D_{\text{ret}}$	GSD_{ret}	N_{ret}	$(d_{\text{ret}}D_{\text{ve}} - d_{\text{ret}}D_{\text{ret}}) / d_{\text{ret}}D_{\text{ve}}$	$(N_{\text{CPC}} - N_{\text{ret}}) / N_{\text{CPC}}$
(nm)	(nm)	(cm^{-3})	(nm)	(-)	(cm^{-3})	(%)	(%)
200	215	189.5 ± 8	214	1.035	185.7	-0.5%	+2.0%
400	425	494.3 ± 11	413	1.065	514	-2.8%	-4.0%
600	636	58.8 ± 4	612	1.05	61.8	-3.8%	-5.0%
800	847	70.2 ± 5	817	1.06	73.7	-3.5%	-6.0%

Supplementary figures:





Figure S1. Scattering geometry in the uNeph

(a)

(Image data reduction)



(b)

(Angular signal processing)





Figure S2. Flowchart of all the data processing steps. Calibration and external auxiliary data are shown as red boxes and text. Methods and inputs with red lines are the external data used in the data processing.

(a)

(Image data reduction)



(b)

(Angular signal processing)




Figure S2. [Flowchart of all the data processing steps. Red color indicates calibration or auxiliary data.](#)







Figure S3. Dark image in angle-pixel coordinate with hot pixels specified. To better visualize the hot pixels, the dark image shown corresponds to a dark image obtained with exposure time of 900 s.






45 **Figure S4. Residual signal from dark signal correction on the data for the cooled (Panel a) and uncooled CCD (Panel b). The constants $G_{DS}(\theta)$ and $\mathcal{E}_{PB}(\theta)$ were not perfectly optimized for the backward angle range (90°-180°) in the uncooled case leaving some systematic exposure time dependence in the residuals. Nevertheless, the residuals of ~ 200 a.u. or less for exposure times shorter than 215 s are sufficiently small given that the upper limit of valid signals (saturation) occurs at around $2 \cdot 10^5$ to $2 \cdot 10^6$ a.u. (Fig. S12).**



50 **Figure S5. (a) A picture of the 3D positional probe mounted on top the uNeph. (b) a 3D scheme of the positional probe, probing an example point S along the laser beam of the uNeph.**



55 **Figure S6. (a) An example of angular calibration image data at a point corresponding to the scattering angle of approximately 68° . (b) a composite image of all calibration data. The red dots in the images corresponds to the pixel of the centre of mass of the given spots with is considered as the pixel corresponding to the location of the 3D positional probe's pinhead. It should be noted that the spots closer to the detection units appear larger while the spots further from the detection unit appear smaller.**




Figure S7. The signal integration limits over scattering angles obtained from the CO₂ measurements collected in different days.




Figure S8. Relative error vs Ξ_{exp} for particle free air sample signals over different scattering angles for both polarization states and for a cooled CCD condition.





Figure S9. Relative error vs $\bar{\epsilon}_{\text{exp}}$ for particle free air sample signals over different scattering angles for both polarization states and for an uncooled CCD condition.



75 Figure S10. Lower and upper limits for valid signal for cooled (solid line) and uncooled (dashed lines) CCD cases based on air measurements taken over a broad range of t_{expo} . The upper limit is the signal above which at least one saturated pixel exists and the lower limit is the lowest signal below which the relative difference between Ξ_{exp} and Ξ_{meas} increases above 6%.






Figure S11. Cross section signal (normalized by t_{expo} and laser power signal) and integrated signal (normalized by t_{expo} and laser power signal) behaviour over a range of exposure times (see plot titles) for filtered air samples at a scattering angle of 30° . Samples taken at polarization state 2. The mean integrated signal, ζ , with corresponding standard error of the mean as error bars in blue, plotted against right ordinate scale. The horizontal dashed line indicates $\pm 5\%$ variation around the mean ζ at $t_{\text{expo}} = 100$ s.



95

Figure S12. Results for 200 nm DEHS aerosol test case with $t_{\text{expo}} = 4.64$ s, particle number concentration: 1000 #/cc. (a) is an example of a single frame of raw data captured by uNeph for this aerosol sample, (b) are the signals in cross-sectional pixel at scattering angle of 25°. The grey lines are the aerosol samples in each sample frame (64 sample frames), the blue line is the mean of aerosol signals, and the red line is the mean of background (BG) air (60 sample frames, with $t_{\text{expo}} = 4.64$ s). (c) is the histogram of the integrated signal Ξ_{air} (red) and Ξ_{aerosol} (blue) over the sampled frames.



100 **Figure S13. 200-nm DEHS test case with $t_{\text{expo}} = 4.64$ s and concentration: 1 #/cc (47 sample frames). The panels (a) to (c) are to the same as in Fig. S12.**




Figure S14S12. Relative difference of air background (ξ_{BG}) to an arbitrary air reference point (ξ_{ref} , sample taken on the 2021-08-10 17:43 - 17:47) during a multiday measurement. This is shown for measurements of polarization state 1 at different scattering angles (5 min averages). The horizontal dashed lines indicate $\pm 3\%$ around the arbitrarily chosen reference point ξ_{ref} .




Figure S15S13. Same as Fig. S124 for presenting polarization state 2 results with the arbitrary air reference point taken on the 2021-08-10 18:21 - 18:26).





Figure S16S14. Day to day particle-free air signal variability. The reference air signal assumed to calculate the relative difference was the sample taken on 2021-08-18.

120



125




Figure S47S15. (a) and (b) respectively shows $G_1(\theta)$ and $G_2(\theta)$ values derived from difference in CO₂ and air measurements (diff) and from individual gas (CO₂ and air) measurements.






Figure S18S16. (a) parallel phase function, F_{para} , (b) perpendicular phase function, F_{perp} , (c) phase function F_{11} , and (d) polarized phase function, $-F_{11}/F_{12}$, for uNeph measurements of Ar gas, given different q value assumptions. The blue line corresponds to the theoretical angular light scattering of Ar based on Rayleigh scattering.



130

Figure S19S17. Same as Fig. S18-S16 but for zooming in to angle range from 55° to 125° . We selected $|q_i| = 0.92$ and $|q_i| = 0.95$ as appropriate values for forward and backward beams, respectively.



135 **Figure S20. Expected and measured results for a PSL size standard with geometric mean diameter (GMD)=600 nm for (a) F_1 , (b) F_2 , (c) F_{11} and (d) $-F_{12}/F_{11}$. The expected curves are based on Mic theory for homogeneous spheres constrained with certified mean particle diameter (600 nm), geometric standard deviation ($GSD = 1.017$) inferred from reported coefficient of variation and particles number concentration measured by a CPC (N_{CPC}). The uNeph measurement is processed with finally calibrated q values and two different angle calibration curves: the red dashed line represents initial angle calibration, whereas the red solid line represents refined angle calibration based on optimized pinhole center coordinates.**

140




Figure S24S18. Angle difference between perturbed and unperturbed case. The unperturbed case is based on the best pinhole estimate.




Figure S22S19. (a, b, c, d) angular light scattering measurements with total errors for 250 nm (AAC aerodynamic set point) DEHS aerosol particles at high concentration (3010 #/cc). (e, f, g,) relative contributions of considered error factors for F_1 , F_2 , and F_{11} . (h) Absolute contributions of considered error factors for $-F_{12}/F_{11}$




Figure S23S20. (a, b, c, d) angular light scattering measurements with total errors for 250 nm (AAC aerodynamic set point) DEHS aerosol particles at low concentration (70 #/cc). (e, f, g) relative contributions of considered error factors for F_1 , F_2 , and F_{11} . (h) Absolute contributions of considered error factors for $-F_{12}/F_{11}$.

|

160




Figure S24. Same as Figure 6 but for additional Mie curves for each test cases. The additional Mie curve were simulated at three different d_m corresponding to the AAC set point, $D_{w,5}$ AAC (cyan lines) and 0.97DVE, AAC (magenta lines) and 1.03DVE, AAC (yellow lines) which are respectively the high and low extremes of DVE, AAC within the assumed error range of 3%. Mie curve at each median size case were simulated for three different GSD cases of 1.04, 1.06 and 1.08 (demonstrated at different line styles) and cover the extremes and mid values within the expected variation range of the GSD. The shape of the measurement data is consistent with at least one of the magenta lines which all represent slightly smaller diameter than AAC set point while differing in underlying GSD. By contrast, discrepancies between measurement data and all even or yellow lines go beyond estimated measurement uncertainty. These findings support the contention that measured phase functions are self-consistent across different angles and that they tightly constrain the retrieved diameter. The small but systematic low bias of retrieved diameters compared to AAC set points could just as plausibly be attributed to a small bias of the AAC.




Figure S21. Same as Fig. 8 but for additional Mie curves for each test cases. The additional Mie curve were simulated at three different d_m corresponding to the AAC set point, $D_{ve,AAC}$ (cyan lines), and 0.97 $D_{ve,AAC}$ (magenta lines) and 1.03 $D_{ve,AAC}$ (yellow lines) which are respectively the high and low extremes of $D_{ve,AAC}$ within the assumed error range of 3%. Mie curve at each median size case were simulated for three different GSD cases of 1.04, 1.06 and 1.08 (demonstrated at different line styles) and cover the extremes and mid values within the expected variation range of the GSD. The shape of the measurement data is consistent with at least one of the magenta lines, which all represent slightly smaller diameter than AAC set point while differing in underlying GSD. By contrast, discrepancies between measurement data and all cyan or yellow lines go beyond estimated measurement uncertainty. These findings support the conclusion that measured phase functions are self-consistent across different angles and that they tightly constrain the retrieved diameter. The small but systematic low bias of retrieved diameters compared to AAC set points could just as plausibly be attributed to a small bias of the AAC.