Measurement report:Evaluation of the TOF-ACSM-CV for PM_{1.0} and PM_{2.5} measurements during the RITA-2021 field campaign

Xinya Liu¹, Bas Henzing², Arjan Hensen², Jan Mulder¹, Peng Yao¹, Danielle van Dinther², Jerry van Bronckhorst³, Rujin Huang⁴, Ulrike Dusek^{*1}

¹Centre for Isotope Research (CIO), Energy and Sustainability Research Institute Groningen (ESRIG), University of Groningen, Groningen, 9747 AG, the Netherlands
 ²Department of Climate, Air and Sustainability, TNO, Utrecht, 3584 CB, the Netherlands.
 ³ Metrohm Process Analytics, Schiedam, 3125AE, The Netherlands
 ⁴ State Key Laboratory of Loess and Quaternary Geology (SKLLQG), Center for Excellence in Quaternary Science and Global
 Change, Institute of Earth Environment, Chinese Academy of Sciences, Xi'an 710061, China

Correspondence to: Ulrike Dusek (u.dusek@rug.nl)

Abstract

The recently developed time of flight-aerosol chemical speciation monitor with the capture vaporizer and a $PM_{2.5}$ aerodynamic lens (TOF-ACSM-CV-PM_{2.5}) aims to improve the collection efficiency and chemical characterization of aerosol particles with

- 15 a diameter smaller than 2.5µm. In this study, comprehensive cross-comparisons were performed between real-time online measurements and offline filter analysis with 24-hour collection time. The goal was to evaluate the capabilities of the TOF-ACSM-CV-PM_{2.5} lens, as well as the accuracy of the TOF-ACSM-CV-PM_{2.5}. The experiments were conducted at Cabauw Experimental Site for Atmospheric Research (CESAR) during the RITA-2021 campaign. The non-refractory fine particulate matter PM_{1.0} and PM_{2.5} were measured by two co-located TOF-ACSM-CV-PM_{2.5} by placing them behind a PM_{2.5} and PM_{1.0}
- 20 inlet, respectively. A comparison between the ACSMs and $PM_{2.5}$ and $PM_{1.0}$ filter samples showed a much better accuracy than ±30% less given in the previous reports, with average differences less than ± 10% for all inorganic chemical species. In addition, the ACSMs were compared to a Monitoring Instrument for Aerosol and Gas (MARGA) (slope between 0.78 - 0.97 for inorganic compounds, $R^2 \ge 0.93$), and a Mobility Particle Size Spectrometer (MPSS) measuring the particle size distribution from around 10 to 800 nm (slope was around 1.00, $R^2 = 0.91$). The intercomparison of the online measurements and the
- 25 comparison between the online and offline measurements indicated a low bias (< 10% for inorganic compounds) and demonstrated the high accuracy and stability of the TOF-ACSM-CV-PM_{2.5} lens for the atmospheric observations of particle matters. The two ACSMs exhibited an excellent agreement, with differences less than 7%, which allowed a quantitative estimate of PM_{1.0} vs PM_{2.5} chemical composition. The result showed that the PM_{1.0} accounted for about 70-80% of the PM_{2.5} on average. The NO₃ mass fraction increased but the OC mass fraction decreased from PM_{1.0} to PM_{2.5}, indicating the size-
- 30 dependence on chemical composition.

1 Introduction

Aerosols play an important role in climate change and have been intensely studied for their effects on the global radiation balance. Direct effects include absorption and scattering of solar radiation and indirect effects refer to changes of cloud properties by aerosols acting as cloud condensation nuclei (Intergovernmental Panel on Climate Change, 2014) (Fan et al.,

- 35 2016). Furthermore, air pollution is considered the biggest environmental health threat in Europe (European Environment Agency, 2020), causing considerable morbidity and mortality (Gurjar et al., 2010; Ostro et al., 2015; Southerland et al., 2022). Approximately 7.0 million premature deaths each year are caused by long-term air pollution exposure worldwide (WHO,2019). In particular, fine aerosol particles with diameters below 2.5 µm are able to penetrate deep into the lungs, possibly causing more than 3.5 million premature deaths each year (Lelieveld et al., 2015). In the Netherlands, particulate matter is usually
- 40 dominated by secondary inorganic aerosols (SIA) due to emissions from the intensive agriculture and traffic emissions, which has been become a serious problem to the local governments and globally (Brunekreef et al., 2009; Janssen et al., 2013; Gu et al., 2021).

Long-term monitoring of chemical composition and concentration is obviously important for controlling these emissions and improve the air quality. A lot of the measurement techniques and platforms have been developed and studied over the decades

45 with the aim of the long-term measurements of aerosols. The aerosol chemical species monitor (ACSM) has been developed for monitoring aerosol chemical composition, based on the aerosol mass spectrometer (AMS) (Ng et al., 2011). Compared to the AMS, the ACSM is portable, economical and relatively easy to operate.

ACSMs have been widely applied since 2011 and are continuously being improved (Wang et al., 2019). The initial design of ACSM, which has been used in most reported papers to date, was equipped with an aerodynamic lens, a standard hot vaporizer

- 50 and a lower cost residual gas analyser (RGA) quadrupole mass spectrometer (Q-ACSM) detector (Wang et al., 2019). Then the Time-of-Flight ACSM (TOF-ACSM) was developed (Fröhlich et al., 2013), which has faster response time and higher sensitivity and is used increasingly in recent years. The recent equipment of the ACSM with a capture vaporizer (Jayne and Worsnop, 2016) and a PM_{2.5} lens (Xu et al., 2017) has opened avenues for quantitative study of the chemical composition of PM_{2.5}. One potential application is monitoring the chemical differences between PM_{1.0} and PM_{2.5}, which have been studied
- 55 intensively for air quality monitoring. However, most previous studies comparing PM_{1.0} and PM_{2.5} have some limitations: most often, the comparisons were based on offline filter samples, which lack high temporal resolution (Sarti et al., 2015; Zhang et al., 2018; Giugliano et al., 2005; Vecchi et al., 2004; Perrone et al., 2013). For online approaches, the measurements usually switched between PM_{1.0} and PM_{2.5} by changing the size cut-off of the sampler inlet, making the comparison not exactly synchronous (Zheng et al., 2020; Sun et al., 2020) or observations of PM_{1.0} and PM_{2.5} were based on different instruments, thus
- 60 observed differences might result from different measurement approaches (Rodríguez et al., 2008; Budisulistiorini et al., 2014). However, to use the ACSM for such intercomparing studies requires higher accuracy than the \pm 30% cited by the manufacturer, based on the standard setup with PM_{1.0} lens. In this study we want to investigate if the introduction of the capture vaporizer (CV) and the PM_{2.5} lens sufficiently improved the accuracy and precision of the TOF-ACSM for quantitative PM_{2.5} monitoring.

In the ACSM instrument particles are converged into a narrow beam in the aerodynamic lens and then collide with the

- 65 vaporizer. The generated vapour is detected with a time of flight or quadrupole mass spectrometer after ionization (Ng et al., 2011). The ACSM equipped with the standard vaporizer (SV) have been most frequently used to date and has been evaluated in several previous studies (Zhang et al., 2017; Pieber et al., 2016; Xu et al., 2017; Canagaratna et al., 2015; Xu et al., 2017). The SV has an inverted cone structure with a porous tungsten surface, which causes particle bounce and therefore reduced collection efficiency. To reduce the particle bounce associated with the SV, the "capture vaporizer" (CV) was introduced in 2016 (Jayne
- 70 and Worsnop, 2016). The CV is made of solid molybdenum and constructed with a narrow entry "cage" and an internal structure that facilitates repeated internal bounces. This increases the residence time of the particles in contact with the thermal evaporator surfaces and therefore reduces the proportion of particles that bounce without evaporating (Hu et al., 2017). It has been reported the CV can achieve a collection efficiency (CE) of 1 for ambient aerosols (Hu et al., 2017), whereas the CE of the SV is only typically ~0.5 for ambient aerosols, and even lower for laboratory aerosols (Matthew et al., 2008; Robinson et al., 2008;
- 75 al., 2017; Liao et al., 2017; Middlebrook et al., 2012).
 - Further, the ACSM initially measured particles with aerodynamic diameters below 1.0 µm, due to the low transmission efficiency of the aerodynamic lens for the larger particles (Xu et al., 2017). The high-pressure aerodynamic lens (HPL) was developed and used for the transmission of larger particles. However, the HPL requires very high precision in the machining, which makes it difficult to reproduce consistently during manufacture (Williams et al., 2013; Xu et al., 2017). To overcome
- these limitations, Peck et al., (2016) built a new intermediate pressure lens (IPL) (3.8 Torr) and it clearly improved the transmission efficiency of particles from 1 μ m to 2.5 μ m (Xu et al., 2017; Peck et al., 2016). For a typical ambient PM_{2.5} size distribution, the PM_{2.5} aerodynamic lens system on a Q-AMS detected a higher percentage of non-refractory mass compared to the old PM_{1.0} aerodynamic lens system. Specifically, the new system detected 89% of the non-refractory mass, while the old system only detected 65% (Xu et al., 2017). A few articles reported the application of this new PM_{2.5} inlet system (Zhang et
- 85 al., 2017), but a comprehensive assessment is still missing. In this study, two identically configured and collocated TOF-ACSM-CV with both PM_{2.5} aerodynamic lens were deployed to measure the NR-PM_{1.0} and NR-PM_{2.5} during "Ruisdael Land-Atmosphere Interactions Intensive Trace-gas and Aerosol measurement campaign" (RITA-2021) at the CESAR site in the Netherlands. Other online instruments such as the Monitoring Instrument for Aerosol and Gas (MARGA), and Multi-Angle Absorption Photometer (MAAP), as well as a Mobility Particle
- 90 Size Spectrometer (MPSS) were applied for auxiliary measurements. Offline filters were collected and analysed to evaluate the TOF-ACSM-CV-PM_{2.5} lens. Cross-comparisons between online and online, online and offline were conducted to investigate the capacity of TOF-ACSM-CV-PM_{2.5} in long-term field measurements, and to give insights into the local chemical composition of the NR-PM_{1.0} and NR-PM_{2.5}.

2.1 Site and campaign description

A series of comprehensive aerosol in-situ measurements were performed during the 2021 RITA (Ruisdael land-atmosphere interactions Intensive Trace-gas and Aerosol) campaign, at the CESAR site in the Netherlands (51.97° N, 4.93° E). The Cabauw Experimental Site for Atmospheric Research (CESAR) is part of the ACTRIS¹ (Aerosol, Clouds and Trace Gases Infra-Structure) and ICOS² (Integrated Carbon Observation System) and is one of the core observation sites for the Ruisdael Observatory³ in European and global climate networks (Knoop et al., 2021). The site is located between the northeast of Rotterdam and the southwest of Utrecht and the air masses are mostly influenced by the continental and marine environments depending on the wind direction. Previous studies showed that clean air masses are often received from the North Sea or Scandinavia. In contrast, polluted air masses generally originate from Southern Europe (Mamali et al., 2018). Continuous observations of aerosol physiochemical properties were conducted during the RITA-2021 campaign from May 11th to May 24th, and from Sep16th to Oct 12th 2021, additional measurements such as meteorological data from the 213 m high mast of Cabauw tower at 10-minute time resolution were available via the KNMI Data Platform⁴.

2.2 Aerosol physical properties

Ground-based observations of aerosol physical properties were performed in the Cabauw main building using an inlet that

- 110 samples air from 4.5 meter above the ground through the roof. Every inlet consisted of 3 parts (a) a PM₁₀ size selector, (b) a wide diameter Nafion drying system to dry the ambient aerosol to below 40% RH, (c) a manifold to split the aerosol flow to the multiple instruments. The inlet systems were vertically oriented to avoid deposition losses. To minimize the electrostatic diffusional-losses all tubing was stainless steel. The measurements used in this study included: (1) A multi-angle absorption photometer (MAAP model 5012, Thermo Fisher Scientific Inc, Franklin, MA) measuring at a single nominal wavelength of 637 nm with a 5-minute time resolution to quantify the aerosol absorption coefficient (Petzold and Schönlinner, 2004). The
- 115 637 nm with a 5-minute time resolution to quantify the aerosol absorption coefficient (Petzold and Schönlinner, 2004). The mass concentration of equivalent Black Carbon (eBC) was calculated based on the optical absorbance at two different angles using a constant <u>mass absorption cross section (MAC) scattering cross section</u>-value (6.6 m²/g) (2) A Mobility Particle Size Spectrometer (MPSS, TROPOS) consisting of a bipolar particle charger (KR-85), a differential mobility analyzer (DMA, model Vienna-DMA medium), and a condensation particle counter (CPC 3750 TSI). Particle number size distributions in the
- 120 diameter range between approximately 8 and 800 nm were recorded with a time resolution of 5 min. The inversion of the raw data was performed by a custom evaluation software (DMPS-Inversion-2.13.exe), described in (Wiedensohler et al., 2012).

¹http://actris.net/ (last access: 20 July 2022)

² https://www.icos-cp.eu/ (last access: 20 July 2022)

³ https://ruisdael-observatory.nl/ (last access: 20 July 2022)

⁴ https://dataplatform.knmi.nl(last access: 20 July 2022)

2.3 Aerosol chemical composition measurement

2.3.1 Online measurements by TOF-ACSM and MARGA

- The nonrefractory (NR) chemical compositions of $PM_{1.0}$ and $PM_{2.5}$ were measured continuously during the RITA-2021 campaign with a time resolution of 5 min, including ammonium (NH₄⁺), nitrate (NO₃⁻), sulphate (SO₄²⁻), chloride (Cl⁻), and organics (OA), using two TOF-ACSMs (Aerodyne Research Inc., Billerica, MA) (Fröhlich et al., 2013), both equipped with CV and PM_{2.5} aerodynamic lens (Xu et al., 2017). The two TOF-ACSMs were installed side by side in a trailer which was around 200m away from the above-mentioned main measurement site. Teflon Coated Aluminium cyclones (URG 2000-30ED) were installed at the head of the inlet with a downward entry direction to avoid external effects such as rain. Flow rates of 2.3
- 130 L min⁻¹ and 5.0 L min⁻¹ were applied to select $PM_{2.5}$ and $PM_{1.0}$, respectively. Then a multi-tube Nafion dryer (Perma Pure, New Jersey) was used to dry the particles. It should be emphasized that the size selection occurred at ambient conditions, thus the upper limit of the dry particle size depends on humidity. The working principle of the TOF-ACSM is based on the Aerodyne aerosol mass spectrometer (AMS), and can be briefly described as follows: the particles are focused and drawn into the instrument through an aerodynamic lens, then the non-refractory constituents are evaporated rapidly by the capture vaporizer
- 135 (T = 600 °C) and subsequently ionized by electron impact. The ions are identified by their mass to charge ratio in the time of flight mass spectrometer. In the end, the electrical signal is converted to a digital signal by the signal detector and recorded (Fröhlich et al., 2013). Several calibrations need to be performed regularly to ensure the accuracy of the instruments, including the lens calibration, flow rate calibration and the tuning of the heater bias (HB) voltage, as well as the Ionization efficiency (IE) and the relative ionization efficiency (RIE) calibrations. The standard procedure of the calibration details can be found in
- 140 previous publications (Fröhlich et al., 2013; Canagaratna et al., 2007). The IE and RIE calibration were performed before the RITA campaign and the parameters used in this paper are summarized in Table 1. The data analysis was produced by Tofware (v3.2.4, Tofwerk AG, Thun, Switzerland) based on the Igor Pro 8.

Table 1: The setup details for two of the TOF-ACSMs and the corresponding IE and RIE calibration values for each species used in145this study

	TOF-ACSM PM _{1.0}	TOF-ACSM PM _{2.5}
Sampling inlet setup	URG 2000-30ED, flow rate 5.0 LPM	URG 2000-30ED, flow rate 2.3 LPM
Sampling dryer setup	Nafion dryer (Perma Pure, New Jersdy) connected to ARI sample line Flow controller (S/N fcb-03 and Greater)	Nafion dryer (Perma Pure, New Jersdy) connected to ARI sample line Flow controller (S/N fcb-03 and Greater)
Vaporizer	Capture	Capture
IE NO ₃ (pg s ⁻¹)	114.50	258.20
RIE NH4	3.25	3.51
RIE SO ₄	1.26	1.33
RIE Org	1.40	1.40
RIE Chl	1.30	1.30
AB (E+5 ions s^{-1})	2.26	4.55
Flow (cm 3 s^{-1})	1.33	1.46

The Monitor for AeRosols and Gases in ambient Air (MARGA 2060, Metrohm Applikon B.V., the Netherlands) was used during the September part of the campaign to measure the water-soluble inorganic components based on ion chromatography (IC), including hydrochloric acid (HCl), Nitric acid (HNO₃), Nitrous acid (HONO), Sulfur dioxide (SO₂), Ammonia (NH₃) in

- 150 the gas phase, and Chloride (Cl⁻), Nitrate (NO₃⁻), Sulfate (SO₄²⁻), Ammonium (NH₄⁺), Potassium (K⁺), Calcium (Ca²⁺), Magnesium (Mg²⁺) in the aerosol phase. A "MARGA-sizer" introduced by ten Brink (2007, 2009) was used to control the size of the particles (e.g. PM_{1.0}, PM_{2.5} or PM₁₀) entering the instrument. We applied the PM_{1.0} sizer in the first stage (from September 5th to September 30th, 2021) and PM_{2.5} sizer in the later stage (from October 3th to October 16th, 2021) of the campaign. The ambient air was drawn into the instrument at a constant flow rate of 16.7 LPM through a short (0.2 m) length of Teflon tubing
- 155 with an outer diameter of 25.4 mm, via a vacuum pump. Then the water-soluble gases were absorbed by a Wet Rotating Denuder (WRD) device (Wyers et al., 1993; Keuken et al., 1988) and the water-soluble aerosols were extracted in a Steam-Jet Aerosol Collector (SJAC) (Khlystov et al., 1995; Slanina et al., 2001). Eventually, the liquid of the samples was collected continuously in separate syringes and then analysed by IC at one-hour resolution. Rumsey et al provided the operational, calibration and data analysis procedures in details (Rumsey and Walker, 2016; Rumsey et al., 2014).

160 **2.3.2** Offline Filters measurements and analysis

24-hour $PM_{1.0}$ and $PM_{2.5}$ filters were collected simultaneously from midnight to midnight, according to the reference method described in the European Standards (EN12341: 1998 and EN14907: 2005). The SEQ47/50 (Leckel GmbH, Germany) instrument with a sequential low-volume system (LVS) of 2.3 m³ h⁻¹ was used for the sampling. Two polytetrafluoroethylenes (PTFE) filters (diameter 47 mm, pore size 3 µm, Millipore) and two quartz fibre filters (diameter 47 mm, Pallflex) were placed in the samplers for a paired measurement of PM_{1.0} and PM_{2.5}. All filter samples were stored at -20 ± 0.5 °C and cooled by ice

165 in the samplers for a paired measurement of $PM_{1.0}$ and $PM_{2.5}$. All filter samples were stored at -20 ± 0.5 °C and cooled by ice packs during transportation. The gravimetric mass of the PTFE filters was obtained by triple weighing before and after sampling. The weighing was performed All the filters were protected and stored under a condition of a temperature of 20.0 ± 0.5 °C and relative humidity of $50 \pm 2\%$ during the storage and transport. The detailed information about the logistic and operational (QA/QC, weighing) procedures, as well as the data acquisition was described in Schaap et al (2010).

- 170 The PM_{1.0} and PM_{2.5} quartz fibre filters were used to performing the ions analysis ion concentration analysis was done by the Ion chromatography, including the 3 inorganic anions (NO₃⁻, Cl⁻, SO₄²⁻) and the 5 cations (Na⁺, K⁺, Mg²⁺, Ca²⁺, NH₄⁺). Aliquots of the filter samples (cations: 3.28 cm²; anions: 3.0 cm²) were extracted by the 2.0 mL of the 30 mM methane sulfonic acid (MSA, cations) or 2.0 mL of extra pure water (anions) for 40 minutes under ultrasonic agitation. The determination of the concentrations was performed by the ICS-1100 and AQUION instruments (both Thermo Scientific) combined with an
- 175 autosampler AS-DV and ion exchange columns (cations: CS16; anions: AS22). The description of the equipment used can be found in Samek et al (2020). The organic carbon (OC) and elemental carbon (EC) of the PM_{1.0} and PM_{2.5} were analyzed by a Sunset thermal-optical analyzer (TOA, Sunset Laboratory Inc.). The EUSAAR2 protocol (Cavalli et al., 2010) was used to distinguish the OC and EC, using the laser transmittance signal. The details of the operation procedure can be found in Yao (2022).

180 3 Results and Discussion

3.1 Intercomparison results

The comparison between online chemical composition measurements (TOF - ACSM + MAAP) and filter measurements for daily average concentration of each species is presented in section 3.1.1. In addition, the volume concentrations derived from chemical composition measurements and the particle number size distribution (PNSD) are compared in section 3.1.2 with

185 hourly time resolution. Total online NR-PM_{1.0} and NR- PM_{2.5} mass concentration were calculated by adding the eBC to the total TOF-ACSM mass concentration (the sum of nitrate, sulfate, ammonium, organic, and chloride mass concentrations). The total mass concentrations of PM_{1.0} and PM_{2.5} filters are also calculated by the sum of the inorganic anions (NO₃, Cl, SO₄), NH₄, OC and EC concentrations.



190

Figure 1 The linear regression fitting correlations between Comparison of the online (ACSM and MAAP) and offline (Filters) daily average mass concentrations of various chemical components. PM_{1.0} is indicated in green and PM_{2.5} in orange. The shaded area represents the 95% confidential interval of the best fit line.

Figure 1 shows the comparison of each component between the online and offline measurements for PM_{1.0} (green dots) and 195 $PM_{2.5}$ (orange triangles) including a linear least-squares regression line. The uncertainties of the slope correspond to the standard error. The daily fluctuations in the online measurements are shown in Figure S7. We illustrate this by presenting the standard deviation of the daily measurements, which are taken at 10-minute intervals. These variations are marked as error bars for each individual day. Over the intensive measurement period, the daily average NO₃ mass concentrations measured by the TOF-ACSM-CV and by filters showed a high correlation with $R^2 = 0.98$ for $PM_{1.0}$ and $R^2 = 0.97$ for $PM_{2.5}$ and the 200 corresponding slopes are 0.94 ± 0.09 and 0.88 ± 0.10 , respectively. The results showed that NO₃ concentrations on the filters were slightly higher than TOF-ACSM-CV measurements. Paired t-tests were performed to investigate the significance of the difference between the online and offline measurements, and the results are shown in the Table S1 and S2. It shows a significant difference between the ACSM measured NO₃ and filters measured NO₃ (p-values are 2.38E-5 for PM_{1.0} and 5.08E-7 for PM_{2.5}). Since the filter concentrations were higher, the difference cannot be explained by the A well known bias between online and

- 205 offline measurements of nitrate is caused by evaporation of ammonium nitrate collected on the filter, which is a well know sampling artifact and the loss grows with increasing of temperature and decreasing of humidity (Malaguti et al., 2015; Kuokka et al., 2007; Chow et al., 2008; Pakkanen and Hillamo, 2002). Previous studies have shown that the loss grows with increasing of temperature and decreasing of humidity and that it this evaporation loss from quartz filters can exceed 80% up to complete evaporation when the temperature exceeds 25°C (Schaap et al., 2004; Allan et al., 2003; Pandolfi et al., 2014). Therefore, the
- 210 monitoring of humidity and the controlling of storage temperature have a crucial impact on the results. During the measurement in this study, the RH was $81.16\% \pm 14.17\%$ and temperature was 15.94 ± 4.20 °C, which should largely prevent this evaporation loss. Consequently, we observe slightly higher concentrations on the filter samples. Likely reasons for this difference are: (i) The higher offline concentration of nitrate may also be caused by the absorption of gas-phase nitric acid (HNO3) on the filter (Chow, 1995). Bhowmik et al (2022) also observed higher nitrate concentrations on filter samples with an even lower slope of
- 215 0.49 between the online AMS and offline filters NO₃ measurements. (ii) For ACSM measurements, the absolute concentration of the nitrate is highly depending on the IE calibration, which needs to be performed carefully and regularly. The calibration parameters used in this study are listed in table 1. If they are slightly biased, the ACSM concentration could be too low. But the differences are in general less than 10%, which are much better than the previous \pm 30% given in the manufactory for the with a SV and PM_{1.0} lens.
- For sulfate, the on-line and off-line measurements also showed a high correlation, although lower than for ammonium nitrate. The slope and coefficient of determination are 0.90 ± 0.16 and R² is 0.93 for sulfate PM_{1.0}, and the slope is nearly 1 (0.99 ± 0.24) and the R² is 0.87 for sulfate PM_{2.5}. The relatively lower R² is potentially due to the low sulfate mass concentration (0.67 and 0.84 µg cm⁻³ on average for PM_{1.0} and PM_{2.5}) during the measurements. Similar to nitrate, the ACSM sulfate measurements are influenced by the IE and RIE calibrations. Apart from that, higher offline values of the sulfate may also be caused by some
- refractory sulfates such as potassium sulfate, calcium sulfate and sodium sulfate, which cannot be detected by TOF-ACSM (Poulain et al., 2020). Or it can also be due to the positive sampling artifacts, for example, the absorption of SO₂ by alkaline particles in the filter membrane or by the reaction of gas-phase ammonia with sulfate aerosols to form ammonium sulfate or ammonium bisulfate (Nicolás et al., 2009; Nie et al., 2010). This is less likely to occur in the Netherlands as sulfate is usually completely neutralized by excess ammonia already in the ambient atmosphere.
- For ammonium, the coefficients of determination were $R^2 = 0.98$ in PM_{1.0} and $R^2 = 0.94$ in PM_{2.5} with slopes of 1.09 ± 0.10 and 0.96 ± 0.15 , respectively. As the ammonium measured by the ACSM mainly corresponds to ammonium nitrate and ammonium sulfate, the small deviation of the online and offline data is reasonable. However, it is worth noting that the ammonia concentrations in Europe as a whole is usually sufficient to neutralize nitric and sulphuric acid (Wichink Kruit et al., 2017). In particular, an excess of ammonium (ammonium concentrations higher than those explained by the formation of
- inorganic ammonium salts) has been observed a lot in the Netherlands in past reports (Schlag et al., 2017). Table S4 and S5 show the molar mass concentration of cation (NH₄) and anions (NO₃ and SO₄) from the filter samples and ACSM measurements. Anions are observed to be 7% higher than cation in $PM_{1,0}$ filter samples, indicated a light underestimation of

 NH_4 in Filter $PM_{1.0}$. But on the whole, the average differences between the ACSM and filter samples are less than $\pm 10\%$ for all inorganic chemical species, shows a good accuracy of the ACSM with the CV and $PM_{2.5}$ lens in the field measurements.

- 240 Regarding the measurement of the organic aerosol (<u>OA</u>) fraction, the ACSM measures <u>OA</u>-organic aerosol (OA), under the assumption that all mass, which cannot be explained by known inorganic components must be organic (Allan et al., 2004). Thus, the quantification of the OA concentration is determined by how to interpret and assign fragmentation signals. On the other hand, the offline measurement of the organics is normally done by thermal-optical analysis, which is usually detects only the carbon element of the organic compounds and is therefore referred to as organic carbon (OC). OC concentrations usually
 - 245 depend on the calculation methods and measuring protocols (Cavalli et al., 2010; Chiappini et al., 2014; Zanatta et al., 2016). As a result of the different quantification, the correlation between <u>organic matter (OM)</u> and OC is much lower than for inorganic compounds ($R^2 = 0.55$ in PM_{1.0} and $R^2 = 0.80$ in PM_{2.5}). Because OM also includes associated hydrogen, oxygen and other elements, OM is significantly higher than OC, indicated by a slope of from 2.21 ± 1.27 for PM_{1.0} and 2.27 ± 0.92 for PM_{2.5}. On average the OM/OC ratios were 1.58 ± 0.54 for PM_{1.0} and 1.97 ± 0.59 for PM_{2.5} in this study, which are common
 - 250 ratios of OM/OC observed in the organic aerosol. The lower ratio for PM_{1.0} indicates more hydrocarbon-like aerosol at smaller particles and the higher ratio for PM_{2.5} more oxidized aerosol in larger particles. Several effects could lead to inaccurate OM/OC ratios and lower correlation coefficients in the data. Volatile and semi-volatile organic compounds (VOC and SVOC) cause positive and negative artifacts in the estimation of OC (Cheng et al., 2019; Turpin et al., 1994; Cheng et al., 2011). The positive artefact results from the adsorption of VOC and SVOC on quartz filters, leading to an overestimation of OC mass and
 - 255 thus underestimated OM/OC ratios. Based on previous measurements at the same location we estimate the upper limit of the positive artifact on the order 20-30% (Dusek, unpublished data). Sometimes studies found higher artifacts up to a factor of 2, but this would lead to unrealistically high OM/OC ratios in our case. Negative artifacts arise from the evaporation of SVOC collected on the filter during sampling and potentially during storage. In order to mitigate the latter artifact, we conducted the OC-EC analysis promptly after the campaign and stored the filters in the freezer. Regarding the ACSM data, Besides, a critical
 - 260 factor is called as 'the Pieber effect', which observed that the inorganic salts can thermally decompose and release carbonaceous material already present in the instrument, leading to the formation of CO_2^+ (m/z 44) ions that are not related to the organic aerosol (Freney et al., 2019; Pieber et al., 2016). Data showed that the degree of interference was highly variable between instruments and over time and CO_2^+ was overestimated by 0.4% to 10.2%. This would lead to an overestimate of OM/OC ratios by up to 10%. In the Netherlands, values towards the upper limit are more likely due to the inorganic
 - 265 <u>concentrations in the Netherlands, especially of ammonium nitrate.</u> Specifically, NH_4NO_3 resulted in a median CO^{2+} overestimate that was 3.4% higher compared to HNO₃. The level of interference caused by other semi-refractory nitrate salts was 2-10 times higher than that caused by NH_4NO_3 . In contrast, $(NH_4)_2SO_4$ induced interference that was 3-10 times lower than NH_4NO_3 . Those may have an impact on the interpretation of the OA concentration. This artifact has a more pronounced effect in aerosol environments with high inorganic salt fractions (> 50%). High inorganic concentrations in the Netherlands
 - 270 may have likely aggravated this effect. Apart from this, a constant RIE of 1.4 was assumed for OA during the study based on the recommendation by Aerodyne, which can contribute to uncertainties in OA quantification, since this RIE can change for

different instruments and different OA composition and concentration. Although there are some studies that attempted to convert the ACSM f_{44} signal to O : C ratios and to derive OM : OC ratios, the large variability of the f_{44} signal itself causes a large uncertainty in the O : C ratio (Crenn et al., 2015; Canagaratna et al., 2015; Aiken et al., 2008; Rollins et al., 2010; Poulain

et al., 2020). Thus, this approach was not attempted in this study. <u>In summary, OM/OC ratios in this study have considerable uncertainties, but are within the range of typical values found in the literature (Aiken et al., 2008; Poulain et al., 2020; Sun et al., 2011; Zhao et al., 2020). On average, the OM/OC ratios were 1.58 ± 0.54 for PM_{1.0} and 1.97 ± 0.59 for PM_{2.5} in this study, which are common ratios of OM/OC observed in the organic aerosol.
</u>

The eBC was measured online by using the MAAP with a PM₁₀ inlet, whereas the EC was collected on the filters using PM_{2.5}

- 280 inlet and then analyzed offline by the sunset analyzer. Figure 1 shows the comparison of eBC and the $PM_{2.5}$ EC with a good correlation (R² of 0.83). The slope was 1.55 ± 0.44 , mainly reflecting reflecting the difference in size cutoff. Moreover, it is worth noting that the MAAP instrument measures eBC at 637 nm, encompassing both BC and other light-absorbing species that share the same absorption wavelength, such as brown carbon potentially leading to overestimation of the eBC measurements (Cheng et al., 2019). Additional uncertainties are related to filter loading and multiple scattering effects (Petzold
- 285 and Schönlinner, 2004; Petzold et al., 2005). The measured eBC is based on a MAC_Secondly, the eBC measured by the MAAP is based on a constant scattering cross section 6.6 m²/g of black carbon (Petzold et al., 2002) for converting the absorption to the mass concentration of eBC. In reality, this MAC value can vary widely among different environments. On the other hand, EC measurements by thermal-optical analysis (TOA) also have significant uncertainties. Previous studies show that EC can be both overestimated or underestimated by TOA depending on the thermal protocol, optical correction method,
- 290 and filter loading (Yang and Yu, 2002; Schmid et al., 2001; Panteliadis et al., 2015; Cadle et al., 1980; Zenker et al., 2020), which can introduce addition uncertainties when comparing eBC and EC measurements. A recent comparison between the MAAP and OC/EC analysis shows differences of 20% for an urban site and 70% for a regional site (Karanasiou et al., 2020). The 55% differences found in our studies with different size cutoffs shows a reasonable result.

To sum up, the comparison between the online and off-line measurements of the PM_{1.0} and PM_{2.5} showed consistent results, especially for the SIA with slopes between 0.88 - 1.09 and the R² values greater than 0.87. The OA vs OC and eBC vs EC comparisons showed results in line with previous studies. Overall, the data were fairly accurate and reliable for further study. Especially, the configuration of TOF-ACSM-CV-PM_{2.5} lens showed a high stability and accuracy. With suitable inlets it can perform both NR-PM_{1.0} and NR-PM_{2.5} measurements for the purpose of long-term field observation.

3.1.2 Comparison of chemically derived volume concentration and PNSD derived volume concentration



300

Figure 2 (a)The time series of the ACSM and MAAP volume concentrations (red line) compared with the MPSS volume concentration (blue area). (b)The correlation of the ACSM and MAAP volume concentration with the MPSS derived volume concentration. Scatters colored by the relative humidity (%).

- The total TOF-ACSM volume concentration was also compared and validated by the particle volume concentration derived from the PNSD. The aerosol particle size distribution with a range of around 8 - 800 nm in electromobility diameter was obtained by the MPSS during the RITA-2021 campaigns (in May and Sep) to further validate the chemical measurements. Simply put, the volume concentration from ACSM was calculated as the mass concentrations of individual species divided by the corresponding density. The density of each species using in this study was 1.75 g cm⁻³ for the inorganics (Haynes, 1942), 1.2 g cm⁻³ for the organics (Turpin and Lim, 2001), 1.52 g cm⁻³ for chloride (Haynes, 1942), and 1.77 g cm⁻³ for eBC (Park et al., 2004; Poulain et al., 2014). The MPSS volume concentration was estimated by converting the PNSD to the particle volume distribution. The total volume concentration of the MPSS is the integral of the particle volume distribution over all the size bins. Figure 2(a) shows the time series of the volume concentrations derived from ACSM + MAAP measurements and the
- MPSS derived volume concentration. The agreement was good over the whole measurements period indicating a stable condition of the instrument and satisfactory quality. The correlation of volume concentrations is displayed in Figure 2(b) with
- 315 data points colored by the RH. The slope was nearly 1 (± 0.02) with R² = 0.91, which was comparable with previous studies (Poulain et al., 2020; Pokorná et al., 2022). However, it demonstrates that the linear correlation between the two variables is significantly influenced by relative humidity. Higher relative humidity led to a lower size cut-off diameter, resulting in a lower mass concentration measured by ACSM. As also reported in the previous studies, the aerosol hygroscopic growth has a great impact on the size cut off in terms of dry particle size (Chen et al., 2018) when the ambient RH is high. It has been pointed out
- that the difference between ambient and dry cut-off size is approximately 10% and 20% for PM_{1.0} and PM_{2.5} in the European background, and even larger in marine or coastal stations with up to 43% and 62% for PM_{1.0} and PM_{2.5} (Poulain et al., 2020). The upper cut-off for the ACSM inlet is ~2.5 μ m (ambient, aerodynamic) and ~ 0.8 μ m (dry, electrodynamic equivalent) for the MPSS. Nevertheless, the dry, electrodynamic equivalent cut-off size of the ACSM inlet will be larger than 0.8 um. Therefore, the ACSM volume concentrations were expected to be higher and it is surprising that the agreement is so close.
- 325 However, the ACSM only measures non-refractory material, whereas the MPSS derived volume concentration also includes non-refractory material. This indicates that there is considerable contribution from non-refractory material other than BC. The

filter analysis also supports this conclusion, as seen in Figure S5, which shows approximately 21% of the $PM_{2.5}$ mass was not detected by the ion analysis. Thus, the slope of 1.00 is probably a coincidence, where missing volume from the MPSS cut-off and missing mass from the ACSM roughly cancel out. On the whole, the high R^2 values give confidence in the stability and

330 accurateness of the ACSM instrument in the long-term observations. A comparison between ACSM and MPSS volume concentrations is highly recommended as a regular quality control strategy.

3.2 Chemical composition of the PM_{1.0} and PM_{2.5}.

335

Based on the good agreement between the online and offline measurements, ACSM accurately measured both $PM_{1.0}$ and $PM_{2.5}$ concentrations. Therefore, it is possible to further quantify the $PM_{1.0}$ vs $PM_{2.5}$ chemical composition and investigate the differences.



3.2.1 Comparison of NR-PM_{1.0} and NR-PM_{2.5} species measured by TOF-ACSM

Figure 3 Time series of the NR-PM_{1.0} and NR-PM_{2.5} chemical species and the total mass concentration measured by TOF-ACSM; and corresponding linear regression fitting correlations.

- As mentioned, two identically configured TOF-ACSMs with PM_{2.5} aerodynamic lens were collocated and set up to measure the NR-PM_{1.0} and NR-PM_{2.5} during the RITA-2021 spring campaign. At the start of the campaign, both instruments were intercompared by measuring the NR-PM_{2.5}. The results shown in Figure S1 demonstrate good comparability, with the R² ranging from 0.91 to 1.0 and slopes ranging 0.94 to 0.99 for the SIA compounds SO₄, NO₃, and NH₄. The slopes were not significantly different from 1 at the 95% confidence level. For chloride, the correlation was not as good as for other species,
- because ammonium chloride had a very low concentration during the whole measurement period. Therefore, it will not be discussed in the following. The correlation of $PM_{1.0}$ and $PM_{2.5}$ OA concentrations was also reasonable with a slope of 0.93 ± 0.13 ($R^2 = 0.80$). Overall, the two co-located TOF-ACSMs compared well and can be used to compare $PM_{1.0}$ and $PM_{2.5}$ chemical composition.

Figure 4 shows the total mass concentration time series of the NR-PM_{1.0} and NR-PM_{2.5} as well as the concentration of individual

- 350 chemical species, and the corresponding scatter plots with regression lines. The mass concentration of NR-PM_{2.5} was on average $5.27 \pm 3.98 \ \mu g \ m^{-3}$ with a range from 4.84 $\mu g \ m^{-3}$ to 22.25 $\mu g \ m^{-3}$. This concentration was below the WHO PM_{2.5} annual limit values (10 $\mu g \ m^{-3}$) (World Health Organization, 2021) and also lower than previously reported concentrations in this region of 14.4 \pm 2.1 $\mu g \ m^{-3}$ (Schlag et al., 2016; Mensah et al., 2012; Mamali et al., 2018). The PM_{1.0} and PM_{2.5} mass concentrations of each species were highly correlated over the whole measurements period with R² \geq 0.98. In general, the
- 355 $PM_{1.0}$ SIA accounted for approximately 75% 85% of the $PM_{2.5}$ SIA on average, with individual contributions ranging from 82% ± 1% for ammonium, 79% ± 1%, for nitrate, to 76% ± 1% for sulphate. For organics the $PM_{1.0}$ accounted for a higher fraction of $PM_{2.5}$, with around 85% ± 1%. Similar results were also found in the filter samples as displayed in Figure S2. In addition, EC-PM_{1.0} accounted for 74% ± 14% of the EC-PM_{2.5}. In general, the PM_{1.0} mass concentration explained 80% ± 1% of the PM_{2.5} on average, and this ratio ranged from 45.21% - 94.78% throughout the campaigns. However, there was still a
- 360 substantial proportion (~21%) of unexplained mass in the PM_{2.5} as shown in FigureS5. In addition, the chemical mass fractions of PM_{1.0} and PM_{2.5} displayed in Figures S3-S4, revealed that there were some slight differences in the chemical composition of the PM_{1.0} and PM_{2.5}. Figure S3 showed the average hourly mass fraction measured by the ACSM for the NR-PM_{1.0} and NR-PM_{2.5}. The OA accounted for similar proportions namely 34.4% of NR-PM_{1.0} and 33.0% of NR-PM_{2.5}. NO₃ contributed 27.8% to NR-PM_{1.0} with a slight increase to 31.5% in NR-PM_{2.5}. Figure S4 and S5 show
- 365 the daily and the average mass fractions for $PM_{1.0}$ and $PM_{2.5}$ species from the filter samples, with a higher NO₃ fraction in $PM_{2.5}$ and a lower OC fraction in $PM_{1.0}$ for the whole period. Specifically, the NO₃ fraction increased from 38.3 % in $PM_{1.0}$ filter samples to 45.5 % in $PM_{2.5}$ filter samples, whereas the OC fraction decreased from 19.9 % to 15.1%. The difference between the ACSM OA mass fractions (similar in $PM_{1.0}$ and in- $PM_{2.5}$) and the OC mass fraction on the filters (higher in $PM_{1.0}$ than in $PM_{2.5}$) is the result of higher OM/OC ratios for larger particles. As discussed in section 3.1.1 this is likely due to the fact that
- 370 pure hydrocarbons that are often contained in primary emissions are more abundant in the smaller particle size range and the OC mass fraction on the filters (higher in $PM_{1,0}$ than in $PM_{2,5}$) could indicate that organic compounds were more abundant in the $PM_{1,0}$ $PM_{2,5}$ size range. This change in chemical composition with particle size suggests that different types of particles may dominate in different size ranges and potentially indicating a non-internal aerosol mixing state during the measurements.

The differences of OC mass fraction in $PM_{1.0}$ and in $PM_{2.5}$ also further explain the stronger correlation of ACSM OA and Filter 375 OC in $PM_{1.0}$ compared to in $PM_{2.5}$ shown in Figure 1.



3.2.2 Comparison of the SIA-PM_{1.0} by MARGA and SIA-PM_{2.5} by TOF-ACSM

Figure 4 Time series of TOF-ACSM measured SIA-PM_{2.5} during the whole period. MARGA measured PM_{1.0} from September 5th to September 30th in 2021, and PM_{2.5} from October 3rd to October 15th in 2021. The corresponding linear regression fitting correlations between MARGA PM_{1.0} and TOF-ACSM PM_{2.5}.



Figure 5 The linear regression fitting correlations between MARGA PM_{2.5} and TOF-ACSM PM_{2.5} with points colored by the relative humidity (%). The blue shaded area represents the 95% confidence interval.

- The Comparison of the MARGA and ACSM measurements were carried out for further evaluation and validation. Figure 4 displayed the time series of the MARGA measurements and TOF-ACSM measurements. Figure 5 compares MARGA and TOF-ACSM data during time periods when both instruments measured PM_{2.5}. Strong correlations with R² ranging from 0.93 to 0.97 and small intercepts demonstrated a good reliability of the two methods. However, the linear regression slopes displayed some discrepancies for individual species. The NO₃ measured by the ACSM and MARGA showed an excellent agreement, with a difference below 3% (Slope = 0.97 ± 0.03 , R² = 0.97). For the NH₄ and SO₄, the MARGA mass concentrations were lower than the ACSM mass concentrations. The slope for NH₄ was 0.83 ± 0.04 and for SO₄ it was 0.78 ± 0.02 . The analysis
- also revealed a dependence on the RH for the correlation between the two measurements. As illustrated in Figure 5, the ACSM tended to measure more mass than the MARGA under lower RH conditions. The hygroscopic growth of the aerosol particles at higher RH resulted in lower dry cut-off sizes and the different inlets of the MARGA and ACSM might lead to differences in the detected mass. Combined with Figure S6, it shows a slight bias between ACSM and MARGA at higher concentrations,
- 395 particularly when pollution originates from the south or southeast. Since the ACSM mass also includes contributions from organic nitrates, organic sulphates, and organic ammonium salts to the observed NH₄, NO₃, and SO₄ concentrations, this could also lead to higher concentrations observed by the ACSM. However, given that the validation of the TOF-ACSM against filter samples showed excellent agreement for these ions as discussed in the Section3.3.1 and listed in the Table S1-S3, the ACSM results are more likely to be closer to the true values compared with MARGA. Most previous comparisons of ACSM/AMS
- 400 and MARGA showed that the MARGA gave higher concentrations when the ACSM/AMS used the PM_{1.0} lens (Zhao et al., 2020; Stieger et al., 2018; Heikkinen et al., 2020). To the best of our knowledge, this is the first comparison between the PM_{2.5} lens on a TOF-ACSM-CV and a MARGA. We observed that a higher concentration can be achieved by using the CV and PM_{2.5} lens of the TOF-ACSM, which further verify the its capability in measuring non-refractory PM_{2.5} and PM_{1.0} concentrations quantitatively. Because of the very high correlation between MARGA and ACSM concentration, the slight bias
- 405 between the instruments can be corrected using the regression coefficients in Figure 6. Figure 4 also displayed the linear regression correlations between the MARGA measured PM_{1.0} and TOF-ACSM measured PM_{2.5} inorganic chemical species after this correction. The correlation between ACSM PM_{2.5} and MARGA PM_{1.0} all showed R² values greater than 0.85, and the slopes are 0.76 ± 0.03 for NH₄, 0.74 ± 0.03 for SO₄, 0.70 ± 0.02 for NO₃, very comparable to the slopes achieved in the spring campaign, using two different ACSMs. In summary, the local concentrations of both PM_{1.0} or PM_{2.5} were relatively low
- 410 throughout the observation period. The $PM_{1.0}$ and $PM_{2.5}$ studied by using the several different instruments have demonstrated that the $PM_{1.0}$ mass concentrations accounted for 70%-80% of the $PM_{2.5}$ concentrations for various non-refractory species.

4 Conclusions

415

This study evaluated the performance of the newly developed Time of Flight-Aerosol Chemical Species Monitor-Capture Vaporizer (TOF-ACSM-CV) with a PM_{2.5} aerodynamic lens, in comparison to other offline and online methods. Additionally, we investigated the chemical compositions of PM_{1.0} and PM_{2.5} using two co-located and identically configured TOF-ACSM-

CVs. Measurements were carried out during the RITA-2021 campaign with two intensive observation periods in Spring and Fall at CESAR (the Cabauw Experimental Site for Atmospheric Research) in the Netherlands. $PM_{1.0}$ and $PM_{2.5}$ were also collected on filters for offline analysis. We observed excellent agreement (R^2 from 0.87-0.99) between the online and offline measurements with the differences of all secondary inorganic aerosols smaller than 10%. This level of accuracy is significantly

- 420 higher than the nominal specification of \pm 30%, indicating the reliability of the ACSM with CV and PM_{2.5} lens in accurately measuring atmospheric aerosols. The integrated volume size distribution obtained from the MPSS showed a strong correlation with the summed volume concentration calculated from ACSM and MAAP measurements (slope = 1.0, R² = 0.91). The bias among the multiple online measurements (ACSM, MPSS and MARGA) was dependent on RH, which could be due to the different inlet systems (cyclones vs impactors). However, the good agreements (with all R² > 0.9) enable further quantification
- 425 of PM_{1.0} and PM_{2.5} mass concentrations with the ACSM. The average mass concentration of non-refractory (NR) compounds was $4.11 \pm 3.32 \ \mu g \ m^{-3}$ for PM_{1.0} and $5.27 \pm 3.98 \ \mu g \ m^{-3}$ for PM_{2.5}. The NR-PM_{1.0} fraction accounted for approximately 70% -80% of the NR-PM_{2.5} mass concentration, with both dominated by organics (>33%), followed by nitrate (>27%), sulphate (~18%) and ammonium (~17%). However, the mass fraction of nitrate tended to increase by 7.2% (from 38.3% to 45.5%) while the OC mass fraction tended to decrease 4.8% (from 19.9% to 15.1%) from the PM_{1.0} to PM_{2.5}. This change reveals the
- 430 size-dependence on chemical composition. In conclusion, the introduction of the CV and $PM_{2.5}$ lens significantly improved the collection and detection efficiency, enabling the TOF-ACSM to measure the $PM_{1.0}$ and $PM_{2.5}$ substance quantitively with good calibration.

Data availability

The data involved in this study is part of the Ruisdael Observatory (<u>https://ruisdael-observatory.nl</u>) project and available upon request. The meteorological data is available at the KNMI Data Platform (<u>https://dataplatform.knmi.nl</u>).

Author contribution

XL, BH, and UD designed this study. AH, JM, PY, DD, JB, UD and XL implemented the experiment and sample analysis. XL analysed the data and wrote the manuscript. All co-authors proofread and commented on the paper.

Competing interests

440 The authors declare that they have no conflict of interest.

Acknowledgements

The Chinese Scholarship Council (No.201906350118) is acknowledged for the financial support for the author X. Liu. In the project we make use of the Ruisdael observatory infrastructure, funded by the Dutch Science foundation NWO (grant number 184.034.015). The authors would like to thank Jan Pieter Lollinga and Anneliese Kasper-Giebl for their support in data collection.

References

445

465

- Aiken, A. C., Decarlo, P. F., Kroll, J. H., Worsnop, D. R., Huffman, J. A., Docherty, K. S., Ulbrich, I. M., Mohr, C., Kimmel, J. R., Sueper, D., Sun, Y., Zhang, Q., Trimborn, A., Northway, M., Ziemann, P. J., Canagaratna, M. R., Onasch, T. B., Alfarra, M. R., Prevot, A. S. H., Dommen, J., Duplissy, J., Metzger, A., Baltensperger, U., and Jimenez, J. L.: O/C and OM/OC ratios
- of primary, secondary, and ambient organic aerosols with high-resolution time-of-flight aerosol mass spectrometry, Environ. Sci. Technol., 42, 4478–4485, https://doi.org/10.1021/es703009q, 2008.
 Allan, J. D., Jimenez, J. L., Williams, P. I., Alfarra, M. R., Jayne, J. T., Coe, H., and Worsnop, D. R.: Quantitative sampling using an Aerodyne aerosol mass spectrometer 1. Techniques of data interpretation and error analysis, J. Geophys. Res. Atmos., 108, 1–10, https://doi.org/10.1029/2002jd002358, 2003.
- 455 Allan, J. D., Delia, A. E., Coe, H., Bower, K. N., Alfarra, M. R., Jimenez, J. L., Middlebrook, A. M., Drewnick, F., Onasch, T. B., Canagaratna, M. R., Javne, J. T., and Worsnop, D. R.: A generalised method for the extraction of chemically resolved J. Sci., 35. 909-922. mass spectra from Aerodyne aerosol mass spectrometer data, Aerosol https://doi.org/10.1016/j.jaerosci.2004.02.007, 2004.
- Bhowmik, H. S., Shukla, A., Lalchandani, V., Dave, J., Rastogi, N., Kumar, M., Singh, V., and Tripathi, S. N.: Intercomparison of online and offline methods for measuring ambient heavy and trace elements and water-soluble inorganic ions (NO3-, SO42-, NH4+, and Cl-) in PM2.5over a heavily polluted megacity, Delhi, https://doi.org/10.5194/amt-15-2667-2022, 2022.

in the Netherlands: the NLCS-AIR study., Res. Rep. Health. Eff. Inst., 5–89, 2009.
Budisulistiorini, S. H., Canagaratna, M. R., Croteau, P. L., Baumann, K., Edgerton, E. S., Kollman, M. S., Ng, N. L., Verma, V., Shaw, S. L., Knipping, E. M., Worsnop, D. R., Jayne, J. T., Weber, R. J., and Surratt, J. D.: Intercomparison of an Aerosol

ten Brink, H., Otjes, R., Jongejan, P., and Slanina, S.: An instrument for semi-continuous monitoring of the size-distribution of nitrate, ammonium, sulphate and chloride in aerosol, Atmos. Environ., 41, 2768–2779, https://doi.org/10.1016/j.atmosenv.2006.11.041, 2007.

ten Brink, H., Otjes, R., Jongejan, P., and Kos, G.: Monitoring of the ratio of nitrate to sulphate in size-segregated submicron aerosol in the Netherlands, Atmos. Res., 92, 270–276, https://doi.org/10.1016/j.atmosres.2008.12.003, 2009. Brunekreef, B., Beelen, R., Hoek, G., Schouten, L., Bausch-Goldbohm, S., Fischer, P., Armstrong, B., Hughes, E., Jerrett, M., and van den Brandt, P.: Effects of long-term exposure to traffic-related air pollution on respiratory and cardiovascular mortality

Chemical Speciation Monitor (ACSM) with ambient fine aerosol measurements in downtown Atlanta, Georgia, Atmos. Meas. Tech., 7, 1929–1941, https://doi.org/10.5194/amt-7-1929-2014, 2014.

475 Cadle, S. H., Groblicki, P. J., and Stroup, D. P.: Automated Carbon Analyzer For Particulate Samples, Anal. Chem., 52, 2201– 2206, https://doi.org/10.1021/ac50063a047, 1980.

Canagaratna, M. R., Jayne, J. T., Jimenez, J. L., Allan, J. D., Alfarra, M. R., Zhang, Q., Onasch, T. B., Drewnick, F., Coe, H., Middlebrook, A., Delia, A., Williams, L. R., Trimborn, A. M., Northway, M. J., DeCarlo, P. F., Kolb, C. E., Davidovits, P., and Worsnop, D. R.: Chemical and microphysical characterization of ambient aerosols with the aerodyne aerosol mass spectrometer, Mass Spectrom. Rev., 26, 185–222, https://doi.org/10.1002/mas.20115, 2007.

- 480 spectrometer, Mass Spectrom. Rev., 26, 185–222, https://doi.org/10.1002/mas.20115, 2007. Canagaratna, M. R., Jimenez, J. L., Kroll, J. H., Chen, Q., Kessler, S. H., Massoli, P., Hildebrandt Ruiz, L., Fortner, E., Williams, L. R., Wilson, K. R., Surratt, J. D., Donahue, N. M., Jayne, J. T., and Worsnop, D. R.: Elemental ratio measurements of organic compounds using aerosol mass spectrometry: Characterization, improved calibration, and implications, Atmos. Chem. Phys., 15, 253–272, https://doi.org/10.5194/acp-15-253-2015, 2015.
- 485 Cavalli, F., Viana, M., Yttri, K. E., Genberg, J., and Putaud, J.-P.: Toward a standardised thermal-optical protocol for measuring atmospheric organic and elemental carbon: the EUSAAR protocol, Atmos. Meas. Tech., 3, 79–89, https://doi.org/10.5194/amt-3-79-2010, 2010.

490 matter loading and composition measurements, Atmos. Environ., 195, 141–148, https://doi.org/10.1016/j.atmosenv.2018.09.049, 2018.

Cheng, Y., He, K. bin, Duan, F. kui, Zheng, M., Du, Z. yu, Ma, Y. liang, and Tan, J. hua: Ambient organic carbon to elemental carbon ratios: Influences of the measurement methods and implications, Atmos. Environ., 45, 2060–2066, https://doi.org/10.1016/j.atmosenv.2011.01.064, 2011.

495 Cheng, Z., Atwi, K., Onyima, T., and Saleh, R.: Investigating the dependence of light-absorption properties of combustion carbonaceous aerosols on combustion conditions, Aerosol Sci. Technol., 53, 419–434, https://doi.org/10.1080/02786826.2019.1566593, 2019.

Chiappini, L., Verlhac, S., Aujay, R., Maenhaut, W., Putaud, J. P., Sciare, J., Jaffrezo, J. L., Liousse, C., Galy-Lacaux, C., Alleman, L. Y., Panteliadis, P., Leoz, E., and Favez, O.: Clues for a standardised thermal-optical protocol for the assessment of organic and elemental carbon within ambient air particulate matter, https://doi.org/10.5194/amt-7-1649-2014, 2014.

Chow, J. C.: Measurement methods to determine compliance with ambient air quality standards for suspended particles, J. Air Waste Manag. Assoc., 45, 320–382, https://doi.org/10.1080/10473289.1995.10467369, 1995.

Chow, J. C., Watson, J. G., Lowenthal, D. H., Park, K., Doraiswamy, P., Bowers, K., and Bode, R.: Continuous and filter-based measurements of PM2.5 nitrate and sulfate at the Fresno Supersite, Environ. Monit. Assess., 144, 179–189, https://doi.org/10.1007/s10661-007-9987-5, 2008.

Chen, Y., Wild, O., Wang, Y., Ran, L., Teich, M., Größ, J., Wang, L., Spindler, G., Herrmann, H., van Pinxteren, D., McFiggans, G., and Wiedensohler, A.: The influence of impactor size cut-off shift caused by hygroscopic growth on particulate

Crenn, V., Sciare, J., Croteau, P. L., Verlhac, S., Fröhlich, R., Belis, C. A., Aas, W., Äijälä, M., Alastuey, A., Artiñano, B.,

Baisnée, D., Bonnaire, N., Bressi, M., Canagaratna, M., Canonaco, F., Carbone, C., Cavalli, F., Coz, E., Cubison, M. J., Esser-Gietl, J. K., Green, D. C., Gros, V., Heikkinen, L., Herrmann, H., Lunder, C., Minguillón, M. C., Močnik, G., O'Dowd, C. D., Ovadnevaite, J., Petit, J. E., Petralia, E., Poulain, L., Priestman, M., Riffault, V., Ripoll, A., Sarda-Estève, R., Slowik, J. G.,

 Setyan, A., Wiedensohler, A., Baltensperger, U., Prévôt, A. S. H., Jayne, J. T., and Favez, O.: ACTRIS ACSM intercomparison
 Part 1: Reproducibility of concentration and fragment results from 13 individual Quadrupole Aerosol Chemical Speciation
 Monitors (Q-ACSM) and consistency with co-located instruments, Atmos. Meas. Tech., 8, 5063–5087, https://doi.org/10.5194/amt-8-5063-2015, 2015.

Fan, J., Wang, Y., Rosenfeld, D., and Liu, X.: Review of Aerosol–Cloud Interactions: Mechanisms, Significance, and 515 Challenges, J. Atmos. Sci., 73, 4221–4252, https://doi.org/10.1175/JAS-D-16-0037.1, 2016.

- Freney, E., Zhang, Y., Croteau, P., Amodeo, T., Williams, L., Truong, F., Petit, J. E., Sciare, J., Sarda-Esteve, R., Bonnaire, N., Arumae, T., Aurela, M., Bougiatioti, A., Mihalopoulos, N., Coz, E., Artinano, B., Crenn, V., Elste, T., Heikkinen, L., Poulain, L., Wiedensohler, A., Herrmann, H., Priestman, M., Alastuey, A., Stavroulas, I., Tobler, A., Vasilescu, J., Zanca, N., Canagaratna, M., Carbone, C., Flentje, H., Green, D., Maasikmets, M., Marmureanu, L., Minguillon, M. C., Prevot, A. S. H.,
- 520 Gros, V., Jayne, J., and Favez, O.: The second ACTRIS inter-comparison (2016) for Aerosol Chemical Speciation Monitors (ACSM): Calibration protocols and instrument performance evaluations, Aerosol Sci. Technol., 53, 830–842, https://doi.org/10.1080/02786826.2019.1608901, 2019.
 - Fröhlich, R., Cubison, M. J., Slowik, J. G., Bukowiecki, N., Prévôt, A. S. H., Baltensperger, U., Schneider, J., Kimmel, J. R., Gonin, M., Rohner, U., Worsnop, D. R., and Jayne, J. T.: The ToF-ACSM: A portable aerosol chemical speciation monitor
- with TOFMS detection, Atmos. Meas. Tech., 6, 3225–3241, https://doi.org/10.5194/amt-6-3225-2013, 2013.
 Giugliano, M., Lonati, G., Butelli, P., Romele, L., Tardivo, R., and Grosso, M.: Fine particulate (PM2.5-PM1) at urban sites with different traffic exposure, Atmos. Environ., 39, 2421–2431, https://doi.org/10.1016/j.atmosenv.2004.06.050, 2005.
 Gu, B., Zhang, L., Dingenen, R. Van, Vieno, M., Grinsven, H. J. Van, Zhang, X., Zhang, S., Chen, Y., Wang, S., Ren, C., Rao, S., Holland, M., Winiwarter, W., Chen, D., Xu, J., and Sutton, M. A.: Abating ammonia is more cost-effective than nitrogen
- oxides for mitigating PM2.5 air pollution, Science (80-.)., 374, 758–762, https://doi.org/10.1126/science.abf8623, 2021.
 Gurjar, B. R., Molina, L. T., and Ojha, C. S. P.: Air pollution: Health and environmental impacts, 1–519 pp., https://doi.org/10.1201/EBK1439809624, 2010.
 Haynes, W. M.: Edition 95, 3–488 pp., 1942.

Heikkinen, L., Äijälä, M., Riva, M., Luoma, K., Dällenbach, K., Aalto, J., Aalto, P., Aliaga, D., Aurela, M., Keskinen, H.,

535 Makkonen, U., Rantala, P., Kulmala, M., Petäjä, T., Worsnop, D., and Ehn, M.: Long-term sub-micrometer aerosol chemical composition in the boreal forest: Inter- And intra-annual variability, Atmos. Chem. Phys., 20, 3151–3180, https://doi.org/10.5194/acp-20-3151-2020, 2020.

Hu, W., Campuzano-Jost, P., Day, D. A., Croteau, P., Canagaratna, M. R., Jayne, J. T., Worsnop, D. R., and Jimenez, J. L.: Evaluation of the new capture vaporizer for aerosol mass spectrometers (AMS) through field studies of inorganic species,

540 Aerosol Sci. Technol., 51, 735–754, https://doi.org/10.1080/02786826.2017.1296104, 2017.

Intergovernmental Panel on Climate Change (Ed.): Summary for Policymakers, in: Climate Change 2013 – The Physical Science Basis: Working Group I Contribution to the Fifth Assessment Report of the Intergovernmental Panel on Climate Change, Cambridge University Press, Cambridge, 1–30, https://doi.org/DOI: 10.1017/CBO9781107415324.004, 2014. Janssen, N. A. H., Fischer, P., Marra, M., Ameling, C., and Cassee, F. R.: Short-term effects of PM2.5, PM10 and PM2.5–10

- 545 on dailv mortality in the Netherlands. Sci. Total Environ.. 463-464. 20 - 26. https://doi.org/https://doi.org/10.1016/j.scitotenv.2013.05.062, 2013. Javne, J. T. and Worsnop, D. R.: Particle capture device, Aerodyne Res. Inc., 2016. Karanasiou, A., Panteliadis, P., Perez, N., Minguillón, M. C., Pandolfi, M., Titos, G., Viana, M., Moreno, T., Querol, X., and Alastuey, A.: Evaluation of the Semi-Continuous OCEC analyzer performance with the EUSAAR2 protocol, Sci. Total
- Environ., 747, https://doi.org/10.1016/j.scitotenv.2020.141266, 2020.
 Keuken, M. P., Schoonebeek, C. A. M., van Wensveen-Louter, A., and Slanina, J.: Simultaneous sampling of NH3, HNO3, HC1, SO2 and H2O2 in ambient air by a wet annular denuder system, Atmos. Environ., 22, 2541–2548, https://doi.org/10.1016/0004-6981(88)90486-6, 1988.

Khlystov, A., Wyers, G. P., and Slanina, J.: The steam-jet aerosol collector, Atmos. Environ., 29, 2229–2234, https://doi.org/10.1016/1352-2310(95)00180-7, 1995.

Knoop, S., Bosveld, F. C., De Haij, M. J., and Apituley, A.: A 2-year intercomparison of continuous-wave focusing wind lidar and tall mast wind measurements at Cabauw, Atmos. Meas. Tech., 14, 2219–2235, https://doi.org/10.5194/amt-14-2219-2021, 2021.

Kuokka, S., Teinilä, K., Saarnio, K., Aurela, M., Sillanpää, M., Hillamo, R., Kerminen, V. M., Pyy, K., Vartiainen, E.,

560 Kulmala, M., Skorokhod, A. I., Elansky, N. F., and Belikov, I. B.: Using a moving measurement platform for determining the chemical composition of atmospheric aerosols between Moscow and Vladivostok, Atmos. Chem. Phys., 7, 4793–4805, https://doi.org/10.5194/acp-7-4793-2007, 2007.

Lelieveld, J., Evans, J. S., Fnais, M., Giannadaki, D., and Pozzer, A.: The contribution of outdoor air pollution sources to premature mortality on a global scale, Nature, 525, 367–371, https://doi.org/10.1038/nature15371, 2015.

- 565 Liao, J., Brock, C. A., Murphy, D. M., Sueper, D. T., Welti, A., and Middlebrook, A. M.: Single-particle measurements of bouncing particles and in situ collection efficiency from an airborne aerosol mass spectrometer (AMS) with light-scattering detection, Atmos. Meas. Tech., 10, 3801–3820, https://doi.org/10.5194/amt-10-3801-2017, 2017. Malaguti, A., Mircea, M., La Torretta, T. M. G., Telloli, C., Petralia, E., Stracquadanio, M., and Berico, M.: Comparison of
- online and offline methods for measuring fine secondary inorganic ions and carbonaceous aerosols in the central mediterranean 570 area, Aerosol Air Qual. Res., 15, 2641–2653, https://doi.org/10.4209/aaqr.2015.04.0240, 2015.

Mamali, D., Mikkilä, J., Henzing, B., Spoor, R., Ehn, M., Petäjä, T., Russchenberg, H., and Biskos, G.: Long-term observations of the background aerosol at Cabauw, The Netherlands, Sci. Total Environ., 625, 752–761, https://doi.org/10.1016/j.scitotenv.2017.12.136, 2018.

Matthew, B. M., Middlebrook, A. M., and Onasch, T. B.: Collection Efficiencies in an Aerodyne Aerosol Mass Spectrometer

- as a Function of Particle Phase for Laboratory Generated Aerosols, Aerosol Sci. Technol., 42, 884–898, https://doi.org/10.1080/02786820802356797, 2008.
 Mensah, A. A., Holzinger, R., Otjes, R., Trimborn, A., Mentel, T. F., Ten Brink, H., Henzing, B., and Kiendler-Scharr, A.: Aerosol chemical composition at Cabauw, the Netherlands as observed in two intensive periods in May 2008 and March 2009,
- 580 Middlebrook, A. M., Bahreini, R., Jimenez, J. L., and Canagaratna, M. R.: Evaluation of Composition-Dependent Collection Efficiencies for the Aerodyne Aerosol Mass Spectrometer using Field Data, Aerosol Sci. Technol., 46, 258–271, https://doi.org/10.1080/02786826.2011.620041, 2012.

Atmos. Chem. Phys., 12, 4723–4742, https://doi.org/10.5194/acp-12-4723-2012, 2012.

Ng, N. L., Herndon, S. C., Trimborn, A., Canagaratna, M. R., Croteau, P. L., Onasch, T. B., Sueper, D., Worsnop, D. R., Zhang, Q., Sun, Y. L., and Jayne, J. T.: An Aerosol Chemical Speciation Monitor (ACSM) for routine monitoring of the

- 585 composition and mass concentrations of ambient aerosol, Aerosol Sci. Technol., 45, 780–794, https://doi.org/10.1080/02786826.2011.560211, 2011. Nicolás, J. F., Galindo, N., Yubero, E., Pastor, C., Esclapez, R., and Crespo, J.: Aerosol inorganic ions in a semiarid region on
 - the Southeastern Spanish mediterranean coast, Water. Air. Soil Pollut., 201, 149–159, https://doi.org/10.1007/s11270-008-9934-2, 2009.
- 590 Nie, W., Wang, T., Gao, X., Pathak, R. K., Wang, X., Gao, R., Zhang, Q., Yang, L., and Wang, W.: Comparison among filterbased, impactor-based and continuous techniques for measuring atmospheric fine sulfate and nitrate, Atmos. Environ., 44, 4396–4403, https://doi.org/10.1016/j.atmosenv.2010.07.047, 2010.

Ostro, B., Hu, J., Goldberg, D., Reynolds, P., Hertz, A., Bernstein, L., and Kleeman, M. J.: Associations of mortality with long-term exposures to fine and ultrafine particles, species and sources: Results from the California teachers study Cohort, Environ. Health Perspect., 123, 549–556, https://doi.org/10.1289/ehp.1408565, 2015.

Pakkanen, T. A. and Hillamo, R.: Comparison of sampling artifacts and ion balances for a Berner low-pressure impactor and a virtual impactor, Boreal Environ. Res., 7, 129–140, 2002.

Pandolfi, M., Querol, X., Alastuey, A., Jimenez, J. L., Jorba, O., Day, D., Ortega, A., Cubison, M. J., Comerón, A., Sicard, M., Mohr, C., Prévôt, A. S. H., Minguillón, M. C., Pey, J., Baldasano, J. M., Burkhart, J. F., Seco, R., Peñuelas, J., van Drooge,

- B. L., Artiñano, B., Di Marco, C., Nemitz, E., Schallhart, S., Metzger, A., Hansel, A., Lorente, J., Ng, S., Jayne, J., and Szidat, S.: Effects of sources and meteorology on particulate matter in the Western Mediterranean Basin: An overview of the DAURE campaign, J. Geophys. Res. Atmos., 119, 4978–5010, https://doi.org/https://doi.org/10.1002/2013JD021079, 2014.
 Panteliadis, P., Hafkenscheid, T., Cary, B., Diapouli, E., Fischer, A., Favez, O., Quincey, P., Viana, M., Hitzenberger, R., Vecchi, R., Saraga, D., Sciare, J., Jaffrezo, J. L., John, A., Schwarz, J., Giannoni, M., Novak, J., Karanasiou, A., Fermo, P.,
- and Maenhaut, W.: ECOC comparison exercise with identical thermal protocols after temperature offset correction -Instrument diagnostics by in-depth evaluation of operational parameters, Atmos. Meas. Tech., 8, 779–792, https://doi.org/10.5194/amt-8-779-2015, 2015.

Park, K., Kittelson, D. B., Zachariah, M. R., and McMurry, P. H.: Measurement of inherent material density of nanoparticle

agglomerates, J. Nanoparticle Res., 6, 267–272, https://doi.org/10.1023/b:nano.0000034657.71309.e6, 2004.

Peck, J., Gonzalez, L. A., Williams, L. R., Xu, W., Croteau, P. L., Timko, M. T., Jayne, J. T., Worsnop, D. R., Miake-Lye, R. C., and Smith, K. A.: Development of an aerosol mass spectrometer lens system for PM2.5, Aerosol Sci. Technol., 50, 781–789, https://doi.org/10.1080/02786826.2016.1190444, 2016.

Perrone, M. R., Becagli, S., Garcia Orza, J. A., Vecchi, R., Dinoi, A., Udisti, R., and Cabello, M.: The impact of long-range-transport on PM1 and PM2.5 at a Central Mediterranean site, Atmos. Environ., 71, 176–186, https://doi.org/10.1016/i.atmosenv.2013.02.006, 2013.

- Petzold, A. and Schönlinner, M.: Multi-angle absorption photometry—a new method for the measurement of aerosol light absorption and atmospheric black carbon, J. Aerosol Sci., 35, 421–441, https://doi.org/https://doi.org/10.1016/j.jaerosci.2003.09.005, 2004.
- Petzold, A., Kramer, H., and Schonlinner, M.: Continuous measurement of atmospheric black carbon using a multi-angle absorption photometer, Environ. Sci. Pollut. Res., 9, 78–82, 2002.
- Petzold, A., Schloesser, H., Sheridan, P. J., Arnott, W. P., Ogren, J. A., and Virkkula, A.: Evaluation of multiangle absorption photometry for measuring aerosol light absorption, Aerosol Sci. Technol., 39, 40–51, https://doi.org/10.1080/027868290901945, 2005.
- Pieber, S. M., El Haddad, I., Slowik, J. G., Canagaratna, M. R., Jayne, J. T., Platt, S. M., Bozzetti, C., Daellenbach, K. R.,
 Fröhlich, R., Vlachou, A., Klein, F., Dommen, J., Miljevic, B., Jiménez, J. L., Worsnop, D. R., Baltensperger, U., and Prévôt,
 A. S. H.: Inorganic Salt Interference on CO2+ in Aerodyne AMS and ACSM Organic Aerosol Composition Studies, Environ.
 Sci. Technol., 50, 10494–10503, https://doi.org/10.1021/acs.est.6b01035, 2016.
 Pokorná, P., Zíková, N., Vodička, P., Lhotka, R., Mbengue, S., Holubová Šmejkalová, A., Riffault, V., Ondráček, J., Schwarz,

J., and Ždímalmal, V.: Chemically speciated mass size distribution, particle density, shape and origin of non-refractory

630 PM1measured at a rural background site in central Europe, Atmos. Chem. Phys., 22, 5829–5858, https://doi.org/10.5194/acp-22-5829-2022, 2022.

Poulain, L., Birmili, W., Canonaco, F., Crippa, M., Wu, Z. J., Nordmann, S., Spindler, G., Prévôt, A. S. H., Wiedensohler, A., and Herrmann, H.: Chemical mass balance of 300 °c non-volatile particles at the tropospheric research site Melpitz, Germany, Atmos. Chem. Phys., 14, 10145–10162, https://doi.org/10.5194/acp-14-10145-2014, 2014.

635 Poulain, L., Spindler, G., Grüner, A., Tuch, T., Stieger, B., Pinxteren, D. Van, Petit, J. E., Favez, O., Herrmann, H., and Wiedensohler, A.: Multi-year ACSM measurements at the central European research station Melpitz (Germany)-Part 1: Instrument robustness, quality assurance, and impact of upper size cutoff diameter, Atmos. Meas. Tech., 13, 4973–4994, https://doi.org/10.5194/amt-13-4973-2020, 2020.

Robinson, E. S., Onasch, T. B., Worsnop, D., and Donahue, N. M.: Collection efficiency of α -pinene secondary organic

640 aerosol particles explored via light-scattering single-particle aerosol mass spectrometry, Atmos. Meas. Tech., 10, 1139–1154, https://doi.org/10.5194/amt-10-1139-2017, 2017.

Rodríguez, S., Cuevas, E., González, Y., Ramos, R., Romero, P. M., Pérez, N., Querol, X., and Alastuey, A.: Influence of sea

breeze circulation and road traffic emissions on the relationship between particle number, black carbon, PM1, PM2.5 and

PM2.5-10 concentrations in a coastal city, Atmos. Environ., 42, 6523-6534, https://doi.org/10.1016/j.atmosenv.2008.04.022,

645 2008.

655

Rollins, A. W., Fry, J. L., Hunter, J. F., Kroll, J. H., Worsnop, D. R., Singaram, S. W., and Cohen, R. C.: Elemental analysis of aerosol organic nitrates with electron ionization high-resolution mass spectrometry, Atmos. Meas. Tech., 3, 301–310, https://doi.org/10.5194/amt-3-301-2010, 2010.

Rumsey, I. C. and Walker, J. T.: Application of an online ion-chromatography-based instrument for gradient flux
measurements of speciated nitrogen and sulfur, Atmos. Meas. Tech., 9, 2581–2592, https://doi.org/10.5194/amt-9-2581-2016, 2016.

Rumsey, I. C., Cowen, K. A., Walker, J. T., Kelly, T. J., Hanft, E. A., Mishoe, K., Rogers, C., Proost, R., Beachley, G. M., Lear, G., Frelink, T., and Otjes, R. P.: An assessment of the performance of the Monitor for AeRosols and GAses in ambient air (MARGA): A semi-continuous method for soluble compounds, Atmos. Chem. Phys., 14, 5639–5658, https://doi.org/10.5194/acp-14-5639-2014, 2014.

- Samek, L., Stegowski, Z., Styszko, K., Furman, L., Zimnoch, M., Skiba, A., Kistler, M., Kasper-Giebl, A., Rozanski, K., and Konduracka, E.: Seasonal variations of chemical composition of PM2.5 fraction in the urban area of Krakow, Poland: PMF source attribution, Air Qual. Atmos. Heal., 13, 89–96, https://doi.org/10.1007/s11869-019-00773-x, 2020.
- Sarti, E., Pasti, L., Rossi, M., Ascanelli, M., Pagnoni, A., Trombini, M., and Remelli, M.: The composition of PM1 and PM2.5
 samples, metals and their water soluble fractions in the bologna area (Italy), Atmos. Pollut. Res., 6, 708–718, https://doi.org/10.5094/APR.2015.079, 2015.

Schaap, M., Spindler, G., Schulz, M., Acker, K., Maenhaut, W., Berner, A., Wieprecht, W., Streit, N., Müller, K., Brüggemann, E., Chi, X., Putaud, J. P., Hitzenberger, R., Puxbaum, H., Baltensperger, U., and Ten Brink, H.: Artefacts in the sampling of nitrate studied in the "iNTERCOMP" campaigns of EUROTRAC-AEROSOL, Atmos. Environ., 38, 6487–6496, https://doi.org/10.1016/j.atmosenv.2004.08.026, 2004.

Schaap, M., Weijers, E. ., Mooibroek, D., Nguyen, L., and Hoogerbrugge, R.: Composition and origin of particulate matter in the Netherlands, RIVM Rapp. ..., 69, 2010.

Schlag, P., Kiendler-Scharr, A., Johannes Blom, M., Canonaco, F., Sebastiaan Henzing, J., Moerman, M., Prévôt, A. S. H., and Holzinger, R.: Aerosol source apportionment from 1-year measurements at the CESAR tower in Cabauw, the Netherlands, Atmos. Chem. Phys., 16, 8831–8847, https://doi.org/10.5194/acp-16-8831-2016, 2016.

- Schlag, P., Rubach, F., Mentel, T. F., Reimer, D., Canonaco, F., Henzing, J. S., Moerman, M., Otjes, R., Prévôt, A. S. H., Rohrer, F., Rosati, B., Tillmann, R., Weingartner, E., and Kiendler-Scharr, A.: Ambient and laboratory observations of organic ammonium salts in PM1, Faraday Discuss., 200, 331–351, https://doi.org/10.1039/c7fd00027h, 2017.
 Schmid, H., Laskus, L., Jürgen Abraham, H., Baltensperger, U., Lavanchy, V., Bizjak, M., Burba, P., Cachier, H., Crow, D.,
- 675 Chow, J., Gnauk, T., Even, A., Ten Brink, H. M., Giesen, K. P., Hitzenberger, R., Hueglin, C., Maenhaut, W., Pio, C., Carvalho, A., Putaud, J. P., Toom-Sauntry, D., and Puxbaum, H.: Results of the "carbon conference" international aerosol carbon round

robin test stage I, Atmos. Environ., 35, 2111–2121, https://doi.org/10.1016/S1352-2310(00)00493-3, 2001.

Slanina, J., Ten Brink, H. M., Otjes, R. P., Even, A., Jongejan, P., Khlystov, A., Waijers-Ijpelaan, A., Hu, M., and Lu, Y.: The continuous analysis of nitrate and ammonium in aerosols by the steam jet aerosol collector (SJAC): extension and validation

- of the methodology, Atmos. Environ., 35, 2319–2330, https://doi.org/10.1016/S1352-2310(00)00556-2, 2001.
 Southerland, V. A., Brauer, M., Mohegh, A., Hammer, M. S., van Donkelaar, A., Martin, R. V., Apte, J. S., and Anenberg, S. C.: Global urban temporal trends in fine particulate matter (PM2·5) and attributable health burdens: estimates from global datasets, Lancet Planet. Heal., 6, e139–e146, https://doi.org/10.1016/S2542-5196(21)00350-8, 2022.
- Stieger, B., Spindler, G., Fahlbusch, B., Müller, K., Grüner, A., Poulain, L., Thöni, L., Seitler, E., Wallasch, M., and Herrmann,
 H.: Measurements of PM10 ions and trace gases with the online system MARGA at the research station Melpitz in Germany
 A five-year study, J. Atmos. Chem., 75, 33–70, https://doi.org/10.1007/s10874-017-9361-0, 2018.

Sun, Y., He, Y., Kuang, Y., Xu, W., Song, S., Ma, N., Tao, J., Cheng, P., Wu, C., Su, H., Cheng, Y., Xie, C., Chen, C., Lei, L., Qiu, Y., Fu, P., Croteau, P., and Worsnop, D. R.: Chemical Differences Between PM1 and PM2.5 in Highly Polluted Environment and Implications in Air Pollution Studies, Geophys. Res. Lett., 47, 1–10, https://doi.org/10.1029/2019GL086288, 690 2020.

Sun, Y. L., Zhang, Q., Schwab, J. J., Demerjian, K. L., Chen, W. N., Bae, M. S., Hung, H. M., Hogrefe, O., Frank, B., Rattigan, O. V., and Lin, Y. C.: Characterization of the sources and processes of organic and inorganic aerosols in New York city with a high-resolution time-of-flight aerosol mass apectrometer, Atmos. Chem. Phys., 11, 1581–1602, https://doi.org/10.5194/acp-11-1581-2011, 2011.

Turpin, B. J. and Lim, H. J.: Species contributions to pm2.5 mass concentrations: Revisiting common assumptions for estimating organic mass, Aerosol Sci. Technol., 35, 602–610, https://doi.org/10.1080/02786820119445, 2001.
Turpin, B. J., Huntzicker, J. J., and Hering, S. V.: Investigation of organic aerosol sampling artifacts in the los angeles basin, Atmos. Environ., 28, 3061–3071, https://doi.org/10.1016/1352-2310(94)00133-6, 1994.
Vecchi, R., Marcazzan, G., Valli, G., Ceriani, M., and Antoniazzi, C.: The role of atmospheric dispersion in the seasonal

700 variation of PM1 and PM2.5 concentration and composition in the urban area of Milan (Italy), Atmos. Environ., 38, 4437– 4446, https://doi.org/10.1016/j.atmosenv.2004.05.029, 2004.

Wang, Y., Wang, Q., Ye, J., Yan, M., Qin, Q., Prévôt, A. S. H., and Cao, J.: A Review of Aerosol Chemical Composition and Sources in Representative Regions of China during Wintertime, Atmosphere (Basel)., 10, https://doi.org/10.3390/atmos10050277, 2019.

705 Wichink Kruit, R. J., Aben, J., de Vries, W., Sauter, F., van der Swaluw, E., van Zanten, M. C., and van Pul, W. A. J.: Modelling trends in ammonia in the Netherlands over the period 1990–2014, Atmos. Environ., 154, 20–30, https://doi.org/10.1016/j.atmosenv.2017.01.031, 2017.

Wiedensohler, A., Birmili, W., Nowak, A., Sonntag, A., Weinhold, K., Merkel, M., Wehner, B., Tuch, T., Pfeifer, S., Fiebig,M., Fjäraa, A. M., Asmi, E., Sellegri, K., Depuy, R., Venzac, H., Villani, P., Laj, P., Aalto, P., Ogren, J. A., Swietlicki, E.,

710 Williams, P., Roldin, P., Quincey, P., Hüglin, C., Fierz-Schmidhauser, R., Gysel, M., Weingartner, E., Riccobono, F., Santos,

S., Grüning, C., Faloon, K., Beddows, D., Harrison, R., Monahan, C., Jennings, S. G., O'Dowd, C. D., Marinoni, A., Horn, H. G., Keck, L., Jiang, J., Scheckman, J., McMurry, P. H., Deng, Z., Zhao, C. S., Moerman, M., Henzing, B., De Leeuw, G., Löschau, G., and Bastian, S.: Mobility particle size spectrometers: Harmonization of technical standards and data structure to facilitate high quality long-term observations of atmospheric particle number size distributions, Atmos. Meas. Tech., 5, 657–

- 685, https://doi.org/10.5194/amt-5-657-2012, 2012.
 Williams, L. R., Gonzalez, L. A., Peck, J., Trimborn, D., McInnis, J., Farrar, M. R., Moore, K. D., Jayne, J. T., Robinson, W. A., Lewis, D. K., Onasch, T. B., Canagaratna, M. R., Trimborn, A., Timko, M. T., Magoon, G., Deng, R., Tang, D., De La Rosa Blanco, E., Prévôt, A. S. H., Smith, K. A., and Worsnop, D. R.: Characterization of an aerodynamic lens for transmitting particles greater than 1 micrometer in diameter into the Aerodyne aerosol mass spectrometer, Atmos. Meas. Tech., 6, 3271–
- 720 3280, https://doi.org/10.5194/amt-6-3271-2013, 2013.

740

World Health Organization: WHO global air quality guidelines, Coast. Estuar. Process., 1–360, 2021.

Wyers, G. P., Otjes, R. P., and Slanina, J.: A continuous-flow denuder for the measurement of ambient concentrations and surface-exchange fluxes of ammonia, Atmos. Environ. Part A, Gen. Top., 27, 2085–2090, https://doi.org/10.1016/0960-1686(93)90280-C, 1993.

Xu, W., Croteau, P., Williams, L., Canagaratna, M., Onasch, T., Cross, E., Zhang, X., Robinson, W., Worsnop, D., and Jayne, J.: Laboratory characterization of an aerosol chemical speciation monitor with PM2.5 measurement capability, Aerosol Sci. Technol., 51, 69–83, https://doi.org/10.1080/02786826.2016.1241859, 2017.

Yang, H. and Yu, J. Z.: Uncertainties in charring correction in the analysis of elemental and organic carbon in atmospheric particles by thermal/optical methods, Environ. Sci. Technol., 36, 5199–5204, https://doi.org/10.1021/es025672z, 2002.

- Yao, P., Ni, H., Paul, D., Masalaite, A., Huang, R.-J., Meijer, H. A. J., and Dusek, U.: An automated method for thermal-optical separation of aerosol organic/elemental carbon for 13C analysis at the sub-µgC level: A comprehensive assessment, Sci. Total Environ., 804, 150031, https://doi.org/https://doi.org/10.1016/j.scitotenv.2021.150031, 2022.
 Zanatta, M., Gysel, M., Bukowiecki, N., Müller, T., Weingartner, E., Areskoug, H., Fiebig, M., Yttri, K. E., Mihalopoulos, N., Kouvarakis, G., Beddows, D., Harrison, R. M., Cavalli, F., Putaud, J. P., Spindler, G., Wiedensohler, A., Alastuey, A.,
- Pandolfi, M., Sellegri, K., Swietlicki, E., Jaffrezo, J. L., Baltensperger, U., and Laj, P.: A European aerosol phenomenology5: Climatology of black carbon optical properties at 9 regional background sites across Europe, Atmos. Environ., 145, 346–
 364, https://doi.org/10.1016/j.atmosenv.2016.09.035, 2016.

Zenker, K., Sirignano, C., Riccio, A., Chianese, E., Calfapietra, C., Prati, M. V., Masalaite, A., Remeikis, V., Mook, E., Meijer,
H. A. J., and Dusek, U.: δ13C signatures of organic aerosols: Measurement method evaluation and application in a source study, J. Aerosol Sci., 145, 105534, https://doi.org/10.1016/j.jaerosci.2020.105534, 2020.

Zhang, Y., Tang, L., Croteau, P. L., Favez, O., Sun, Y., Canagaratna, M. R., Wang, Z., Couvidat, F., Albinet, A., Zhang, H.,
Sciare, J., Prévôt, A. S. H., Jayne, J. T., and Worsnop, D. R.: Field characterization of the PM2.5 Aerosol Chemical Speciation
Monitor: Insights into the composition, sources, and processes of fine particles in eastern China, Atmos. Chem. Phys., 17, 14501–14517, https://doi.org/10.5194/acp-17-14501-2017, 2017.

- Zhang, Y., Lang, J., Cheng, S., Li, S., Zhou, Y., Chen, D., Zhang, H., and Wang, H.: Chemical composition and sources of PM1 and PM2.5 in Beijing in autumn, Sci. Total Environ., 630, 72–82, https://doi.org/10.1016/j.scitotenv.2018.02.151, 2018.
 Zhao, Q., Huo, J., Yang, X., Fu, Q., Duan, Y., Liu, Y., Lin, Y., and Zhang, Q.: Chemical characterization and source identification of submicron aerosols from a year-long real-time observation at a rural site of Shanghai using an Aerosol Chemical Speciation Monitor, Atmos. Res., 246, 105154, https://doi.org/10.1016/j.atmosres.2020.105154, 2020.
- 750 Zheng, Y., Cheng, X., Liao, K., Li, Y., Jie Li, Y., Huang, R. J., Hu, W., Liu, Y., Zhu, T., Chen, S., Zeng, L., Worsnop, D. R., and Chen, Q.: Characterization of anthropogenic organic aerosols by TOF-ACSM with the new capture vaporizer, Atmos. Meas. Tech., 13, 2457–2472, https://doi.org/10.5194/amt-13-2457-2020, 2020.