



# **Robustness of atmospheric trace gas retrievals obtained from low spectral resolution Fourier-transform infrared absorption spectra**

Bavo Langerock<sup>1</sup>, Martine De Mazière<sup>1</sup>, Filip Desmet<sup>1</sup>, Pauli Heikkinen<sup>2</sup>, Rigel Kivi<sup>2</sup>, Mahesh Kumar Sha<sup>1</sup>, Corinne Vigouroux<sup>1</sup>, Minqiang Zhou<sup>3</sup>, Gopala Krishna Darbha<sup>4</sup>, and Mohmmed Talib<sup>4</sup>

<sup>1</sup>Royal Belgian institute for Space Aeronomy, Ringlaan 3, 1180 Ukkel, Belgium
 <sup>2</sup>Space and Earth Observation Centre, Finnish Meteorological Institute, Tähteläntie 62, 99600 Sodankylä, Finland
 <sup>3</sup>Institute of Atmospheric Physics, Chinese Academy of Sciences, Beijing, China
 <sup>4</sup>Indian Institute of Science Education and Research Kolkata, India

Correspondence: B. Langerock (bavo.langerock@aeronomie.be)

**Abstract.** For atmospheric trace gas columns retrievals obtained from ground based Fourier-transform interferometer spectra we study the sensitivity of the retrieval processing chain to changes in the number of points in the recorded interferograms. Shortening an interferogram will alter the leakage pattern in the associated spectrum and we demonstrate that the removal of a relatively small number of points from the interferogram edges creates a beat pattern in the difference of the associated spectra

5 obtained from the original and shortened interferogram. For low-resolution interferometers the beat pattern in the spectra may exceed the noise level and the effect on atmospheric gas column retrievals may be large. Sensitivity of the retrieval algorithm to the length of the underlying interferogram can be reduced by applying a non-trivial apodization such as Norton-Beer. A case study shows the effect on formaldehyde retrievals obtained from low-resolution spectra in Sodankylä and Kolkata.

## 1 Introduction

- 10 Fourier-transform infrared (FTIR) interferometers have a long history in measuring atmospheric trace gas concentrations (De Mazière et al., 2018). The Network for the Detection of Atmospheric Composition Change (NDACC) uses interferometers with a high spectral resolution of approximately 0.005 cm<sup>-1</sup> in order to provide vertically resolved trace gas abundances. The Total Carbon Column Observing Network (TCCON) uses measurements with a spectral resolution of 0.02 cm<sup>-1</sup> to provide vertically integrated gas concentrations or total columns (Wunch et al., 2015). During recent years it has been
- 15 demonstrated that FTIR interferometers with a low spectral resolution ranging from 0.2 cm<sup>-1</sup> to 0.5 cm<sup>-1</sup> have the ability to provide high-quality column data (Frey et al., 2019; Sha et al., 2020). The FTIR instruments record interferograms that are converted to spectra prior to the retrieval of an atmospheric trace gas concentration. This paper discusses the sensitivity of this processing chain to a (small) change in the number of points in the recorded interferogram.

The operation that removes points from the tails of a recorded interferogram typically occurs when a convolution is performed on the interferogram prior to applying the discrete Fourier transform (DFT). For example, a standard practice consists of resampling the interferogram to reduce its asymmetry and this operation uses a finite convolution with the sinc kernel (For-









(a) Difference in spectra for a high-resolution example  $(0.005 \text{ cm}^{-1})$  (

(b) Difference in spectra for a low-resolution example  $(0.2 \text{ cm}^{-1})$ 

Figure 1. Difference in spectra for two boxcar (BX) apodized interferograms obtained after removing  $\delta = 2^8$  points: (a) shows the beats in the difference where the length of the interferogram is  $N \approx 5 \times 10^6$  points (obtained with a Bruker 125HR instrument), (b) shows the difference where the interferogram contains  $N \approx 2 \times 10^5$  points (obtained with a Bruker Vertex70 instrument). Both spectra are recorded with an InSb detector. The header indicates the apodization function (APF) and the maximal optical path difference (OPD).

man et al., 1966, Eq. 13). The interferogram is typically shortened with half the size of the sinc kernel. Digital filtering uses a convolution in the time domain (Herres and Gronholz, 1984, part 3). Another example is the Forman phase correction method which also applies a convolution with a non-standard kernel to reduce the phase in the spectrum (Forman et al., 1966). Figure 1 shows the difference between two measured spectra obtained with or without the removal of a relatively small number of

25

30

retrievals.

ure 1 shows the difference between two measured spectra obtained with or without the removal of a relatively small number of points from the interferogram tails and reveals the appearance of beats in the difference of the spectra. For the low-resolution spectrum in Fig. 1(b), the amplitude of the beats is much larger than the noise level in the spectrum in the out of band region.

Section 2 contains a short discussion to quantify the size of the beats. Section 3 demonstrates how apodization can make the DFT more robust under such size perturbations of the interferogram. The final Section 4 studies the effect on formaldehyde retrievals and demonstrates that the use of an appropriate apodization for low-resolution spectra stabilizes the atmospheric gas

### 2 The Dirichlet kernel explains the beat pattern

Figure 2 shows the effect of edge removal in a theoretical example using a symmetrical double sided interferogram *I*[*t*], 0 ≤ *t* ≤ *N* that corresponds to an ideal bandpass model with width *L*. The right panel (b) shows the difference between the spectrum
that corresponds to the original interferogram and the one where δ << *N* points are removed from both edges prior to the DFT. The standard processing chain that we adhere to applies zero-padding of the interferogram to the next higher power of 2 and thus the removal of a few points from the edge of the interferogram does not change the number of points in the spectral domain. The difference between both spectra, when considered in the time domain, corresponds to the portion of the interferogram and the other spectra therefore equals the DFT of the product of the interferogram and

40 an appropriately positioned rectangle window of length  $2\delta$  as depicted in Fig. 2(a): DFT( $I[t]B_{2\delta}[t]$ ). The DFT of a rectangle







Figure 2. Example where the shortening operation is applied to a double sided interferogram shown in (a) that corresponds to an idealized bandpass model shown in (b) (only positive frequencies are shown in the spectral domain, interferogram size  $N = 2^8$ , bandpass width L = N/4). The effect in the spectral domain of removing  $\delta = 2^3$  points at each end of the interferogram is shown. No apodization is used.

with length  $\ell$  is related to the Dirichlet kernel (McClellan et al., 2016):

$$D_{\ell}(x) = \frac{\sin(\ell x/2)}{\sin(x/2)}$$
, sampled at  $x = \frac{2\pi}{N}k$  with  $k \in \mathbb{N}$ .

From the convolution theorem, we can state that the difference of the two spectra is therefore a circular convolution of the bandpass with an appropriately time shifted Dirichlet kernel  $D_{2\delta}(x)e^{i(N-1)x/2}$ .

45

As shown in Fig. 2(b), this convolution creates a beat pattern in the difference. In this example the removal of  $2\delta = 16$  points leads to an approximate 4% difference in the spectra. The amplitude of the beats is related to the energy in the difference of both spectra, which in turn is related to the energy in the portion of the interferogram that is removed, being  $I[t]B_{2\delta}[t]$  (Parseval's theorem). The beat amplitude therefore depends on the number of points removed  $\delta$  and on the distance to the center burst since the energy in the interferogram signal is inversely proportional to the distance to the center burst.



(a) Interferogram (ramp apodized and zero-padded with factor 2)



Figure 3. Simulation example where the shortening operation is applied to a single-sided interferogram shown in (a) that corresponds to an idealized bandpass model shown in (b) (only positive frequencies are shown in the spectral domain, interferogram size  $N = 2^9$ , bandpass width L = N/8). The effect in the spectral domain of removing  $\delta = 2^4$  points at the ends of the long arm of the single-sided interferogram is shown. The example uses a zero-filling factor of 2 and no additional apodization besides ramp is used.





For single-sided interferograms the ramp correction will ensure that the removal of  $\delta$  points on the left shorter arm of the interferogram has a negligible effect, see Fig. 3. In this case only the right side rectangle window  $B_{\delta}$  should be taken into account and the associated Dirichlet kernel takes a slightly different form due to a different time shift:  $D_{\delta}(x)e^{-i(N_r-(\delta+1)/2)x}$  where  $N_r$  is the number of points in the longer right arm.

# 3 Apodization



**Figure 4.** Similar as in Fig. 3, but here the interferogram is apodized with (a) the standard Norton-Beer medium window and (b) the Norton-Beer strong window. Apodizing reduces the beat amplitude with a factor of approximately 5 in (a) and 10 in (b) compared to Fig. 3.

The appearance of beats is related to a perturbation in the leakage pattern underlying the DFT. Including apodization in the DFT will therefore reduce its sensitivity to such relatively small changes in the number of points. We consider the standard three Norton-Beer apodization windows from (Norton and Beer, 1976, 1977) although a more arbitrary strength could be used (Ntokas et al., 2023). Figure 4 shows that the shape of the beats remains when applying different apodizations, however they are reduced in amplitude compared to the situation without apodization in Fig. 3. For stronger apodizations, the energy in the window  $B_{\delta}$  containing the removed points is lowered and so is the amplitude of the beats.

Increasing the strength of the apodization will therefore reduce the dependence of a spectrum on changes in the number of interferogram points and can bring this dependence to the level of the noise in the out-of-band region. Figure 5 shows that for the low-resolution spectrum in Fig. 1(b), the stronger Norton-Beer apodization reduces the beats amplitude to the noise level.

## 4 Case study: formaldehyde retrievals from low-resolution solar absorption spectra

- For this case study we first consider a single day of measurements obtained from a low-resolution Bruker Vertex70 instrument (maximum optical path difference OPD equals 4.5 cm) deployed at a clean site Sodankylä (Finland, high latitude, Kivi and Heikkinen (2016)) during the 2019 campaign in the framework of the FRM4GHG project (Sha et al., 2020) and one day when the instrument was deployed at a polluted site in Kolkata (India, tropics). The purpose of this study is to check the robustness of the formaldehyde retrieval to small changes in the number of points in the interferogram. For that purpose
- 70 formaldehyde columns are retrieved from 8 different sets of spectra obtained from the same underlying interferograms but with



90





**Figure 5.** Similar to Fig. 1(b), but here the interferogram is apodized with (a) the standard Norton-Beer medium (NBM) window and (b) the Norton-Beer strong (NBS) window.

4 different types of apodization applied (boxcar and 3 standard Norton-Beer) and with each interferogram shortened ( $\delta = 2^8$ ) or left unchanged ( $\delta = 0$ ). All retrievals follow the retrieval strategy that is currently adopted in the NDACC infrared working group (Vigouroux et al., 2018). This strategy is applied to the low-resolution measurements without further change because it was found that this gives the lowest bias when comparing to high-resolution measurements at Sodankyla (Sha et al., 2024). The

- 75 estimated systematic uncertainty on the formaldehyde columns for this strategy applied to low-resolution spectra is of the order of 15-30% (Vigouroux et al., 2018; Sha et al., 2024) depending on the pollution level. The retrievals are performed with the retrieval software package SFIT2 (Pougatchev et al., 1995) which was updated to SFIT4 (Hannigan et al., 2024) and is one of the standard retrieval algorithm implementations used in the NDACC infrared working group. The forward model component in SFIT4 takes into account the size of the interferogram via the maximal OPD to simulate the leakage in the modelled
- spectrum. This is done by a truncation in the time domain after applying the DFT on the modelled spectrum. Although the truncation index is determined from the maximal OPD, it is typically invariant under small perturbations of the OPD value because the truncation index requires a rounding to an integer. In the specific case of the Vertex70 spectra for Sodankylä, we have compared two sets of retrievals with maximal OPD equal to 4.5 cm (no modelled shortening of the IFG) and maximal OPD 4.49 cm (modelled shortening of the IFG with  $\delta = 2^8$  points out of 171130). The difference in retrieved formaldehyde
- 85 columns for the Sodankylä measurements used in Fig. 6 is less than 2% and is due to a dependence in the forward model wavenumber spacing on the maximal OPD. This difference can be considered negligible compared to the retrieval uncertainty. In the following all retrievals were done with a fixed maximal OPD value.

Panels (a) and (c) in Fig. 6 shows that the boxcar (BX) retrievals are strongly dependent on the number of points in the interferogram while the NBS apodized retrievals are (nearly) invariant. The formal dehyde retrieval windows (ranging between  $2760 \text{ cm}^{-1}$  and  $2785 \text{ cm}^{-1}$ ) coincide with a beat from the pattern in Fig. 1(b) which partly explains this strong dependence.

Panels (b) and (d) in Fig. 6 show that the column difference (shortened - unchanged) depends on the chosen strength of the apodization: for the clean site Sodankylä, the relative differences range from values up to 30% for NBW to 10% for NBM and below 5% for NBS. The Sodankylä retrieved columns from the NBS apodized spectra have a strongly reduced sensitivity to the number of points which becomes negligible when compared to the retrieval uncertainty budget. For the polluted site at







(a) Vertex70 (BX and NBS) formaldehyde columns at Sodankylä, shortened versus unchanged interferograms





(b) Relative column differences between shortened and original interferograms for the three Norton-Beer apodizations at Sodankylä



(c) Vertex70 (BX and NBS) formaldehyde columns at Kolkata, shortened versus unchanged interferograms

(d) Relative column differences between shortened and original interferograms for the three Norton-Beer apodizations at Kolkata

Figure 6. Sensitivity study in the retrieval of formaldehyde during one day, where the formaldehyde is retrieved from spectra obtained with either a boxcar (BX, red) or Norton-Beer strong (NBS, blue), medium (NBM, green) and weak (NBW, orange) apodization and where either no points ( $\delta = 0$ ) are removed or  $\delta = 2^8$  points are removed.

95 Kolkata, depicted in (c) and (d), the relative differences are much reduced but reveal the same pattern as for Sodankylä. Note also that Fig. 6 shows that both the perturbation in the number of points and the use of different apodizations propagate to different airmass dependence effects in the differences of the formaldehyde columns.

For Sodankylä the NBS apodization reduces the sensitivity on the interferogram size well below the reported uncertainty and is therefore chosen for a comparison between the Sodankylä Vertex70 retrievals and the Sodankylä Bruker 125HR retrieved

- 100 columns that are publicly available from the NDACC database. To quantify the bias and correlation with the NDACC 125HR data we have retrieved all Vertex70 FRM4GHG 2019 campaign spectra using a BX and a NBS apodization. Figure 7 uses a fifteen minute average of the NDACC retrievals and Vertex70 retrievals and only fifteen minute intervals that coincide are shown (Sha et al., 2024). The overall bias of low-resolution vs high-resolution (NDACC) retrieved columns reduces from approximately -55% for the low-resolution columns retrieved from the BX apodized spectra to -15% for the low-resolution
- 105 retrievals from the NBS apodized spectra. Similarly the correlation coefficient (Spearman, p-value < 0.01) between the low-resolution and high-resolution retrievals increases from 0.72 for the low-resolution BX retrievals to 0.93 for the low-resolution NBS retrievals.







**Figure 7.** Correlations for formaldehyde columns using (a) boxcar and (b) Norton-Beer strong apodizations for the Vertex70 2019 campaign data, compared to the NDACC 125HR observations, using coinciding measurements in 15 minutes averaged data.

The effect of apodization in the retrieval results does not contradict the conclusions from Amato et al. (1998), where it is shown that retrieval results are invariant under different apodizations if one takes into account that the measurement error in the spectral domain is no longer random noise after applying apodization. The retrieval software SFIT, typically used with high-resolution spectra, always considers the measurement error as random noise and therefore produces different retrieval results after applying apodization.

For completeness, it should be mentioned that the high-resolution NDACC retrievals are made from BX apodized spectra. A robustness test as in Fig. 6, showed no significant dependence of the 125HR retrieved columns on the interferogram size nor on the chosen apodization: all relative differences for the Sodankylä 125HR are below 0.1%.

## 5 Conclusions

We have demonstrated that a shortening of an interferogram creates a perturbation in the leakage pattern that resembles a beat pattern in the difference of the associated spectra. The amplitude of the beats is related to the energy of the signal in the truncated tail of the interferogram and is therefore larger for low-resolution interferograms. We have shown examples using idealized

- 120 bandpass models and real measurements where the beats exceed the noise level in the out-of-band region. Retrieval software packages may not be capable of modeling such small changes in the interferogram size. It was demonstrated that the retrieval of atmospheric gas columns from low-resolution measurements could be unstable if no appropriate apodization is used. In a case study using formaldehyde retrievals, we have chosen the strength of the apodizaton (e.g. Norton-Beer weak/medium/strong) such that the retrieved formaldehyde columns become stable under shortenings of the underlying interferograms in the sense
- 125 that retrieved column differences are within the estimated formaldehyde uncertainty budgets.





*Code and data availability.* SFIT4 (Hannigan et al., 2024) is a radiative transfer and atmospheric constituent profile retrieval algorithm for use with spectra recorded in solar absorption and emission across the infrared spectrum. The Sodankylä 125HR data used in this publication were obtained as part of the Network for the Detection of Atmospheric Composition Change (NDACC, https://ndacc.org, last public access: 31 August 2024). The other data used in this publication can be provided upon request to the corresponding author.

130 Author contributions. BL wrote the paper, did the analysis and sensitivity studies. CV implemented the NDACC formaldehyde strategy for the Vertex70 low-resolution spectra and the presentation of the statistics in the 2019 Sodankylä time series. MDM, MKS, RK and PH contributed with the Vertex70 spectra from the FRM4GHG campaign in Sodankylä, in addition RK and PH with the 125HR spectra from Sodankylä and MDM, MKS with the spectra from Kolkata. GKD and MT for the maintenance and support with the instrument operations at Kolkata. MKS, FDS and MZ contributed with valuable suggestions to understand the cause of the beat patterns. All authors have read and 125

135 agree to the published version of the manuscript.

Competing interests. The authors declare that they have no conflict of interest.

Acknowledgements. The authors thank Philippe Demoulin for useful discussion on the implementation details of the DFT at the Jungfraujoch station. The BELSPO ProDEx project TROVA-E2 (PEA 4000116692), SVANTE (4000132151/20/NL/FF/ab), FRM4GHG (4000117640/16/I-LG and 4000136108/21/I-DT-lr), and ACTRIS-BE projects provided funding for the observations used in this work.

140 Nicolas Kumps, Christian Hermans and Christof Petri for the operation and set up of the instrument during the Sodankylä campaign.





#### References

165

- Amato, U., Canditiis, D. D., and Serio, C.: Effect of apodization on the retrieval of geophysical parameters from Fourier-transform spectrometers, Appl. Opt., 37, 6537–6543, https://doi.org/10.1364/AO.37.006537, 1998.
- De Mazière, M., Thompson, A. M., Kurylo, M. J., Wild, J. D., Bernhard, G., Blumenstock, T., Braathen, G. O., Hannigan, J. W., Lambert,
- 145 J.-C., Leblanc, T., McGee, T. J., Nedoluha, G., Petropavlovskikh, I., Seckmeyer, G., Simon, P. C., Steinbrecht, W., and Strahan, S. E.: The Network for the Detection of Atmospheric Composition Change (NDACC): history, status and perspectives, Atmospheric Chemistry and Physics, 18, 4935–4964, https://doi.org/10.5194/acp-18-4935-2018, 2018.
  - Forman, M. L., Steel, W. H., and Vanasse, G. A.: Correction of Asymmetric Interferograms Obtained in Fourier Spectroscopy\*, J. Opt. Soc. Am., 56, 59–63, https://doi.org/10.1364/JOSA.56.000059, 1966.
- 150 Frey, M., Sha, M. K., Hase, F., Kiel, M., Blumenstock, T., Harig, R., Surawicz, G., Deutscher, N. M., Shiomi, K., Franklin, J. E., Bösch, H., Chen, J., Grutter, M., Ohyama, H., Sun, Y., Butz, A., Mengistu Tsidu, G., Ene, D., Wunch, D., Cao, Z., Garcia, O., Ramonet, M., Vogel, F., and Orphal, J.: Building the COllaborative Carbon Column Observing Network (COCCON): long-term stability and ensemble performance of the EM27/SUN Fourier transform spectrometer, Atmospheric Measurement Techniques, 12, 1513–1530, https://doi.org/10.5194/amt-12-1513-2019, 2019.
- 155 Hannigan, J., Palm, M., Jones, N., Ortega, I., Langerock, B., Mahieu, E., Zhou, M., and Smale, D.: SFIT4 Line-by-line nonlinear spectral fitting software: version 1.0.21, https://doi.org/10.18758/CLYG3EUO, 2024.

Herres, W. and Gronholz, J.: Understanding FT-IR data processing, pp. 352-356, 1984.

- Kivi, R. and Heikkinen, P.: Fourier transform spectrometer measurements of column CO<sub>2</sub> at Sodankylä, Finland, Geoscientific Instrumentation, Methods and Data Systems, 5, 271–279, https://doi.org/10.5194/gi-5-271-2016, 2016.
- 160 McClellan, J., Schafer, R., and Yoder, M.: Digital Signal Processing First, Global Edition, Always learning, Pearson, ISBN 9781292113869, 2016.
  - Norton, R. H. and Beer, R.: New apodizing functions for Fourier spectrometry, J. Opt. Soc. Am., 66, 259–264, https://doi.org/10.1364/JOSA.66.000259, 1976.

Norton, R. H. and Beer, R.: Errata: New Apodizing Functions For Fourier Spectrometry, J. Opt. Soc. Am., 67, 419–419, https://doi.org/10.1364/JOSA.67.000419, 1977.

- Ntokas, K. F. F., Ungermann, J., and Kaufmann, M.: Norton-Beer apodization and its Fourier transform, J. Opt. Soc. Am. A, 40, 2026–2033, https://doi.org/10.1364/JOSAA.501593, 2023.
  - Pougatchev, N. S., Connor, B. J., and Rinsland, C. P.: Infrared measurements of the ozone vertical distribution above Kitt Peak, Journal of Geophysical Research: Atmospheres, 100, 16 689–16 697, https://doi.org/https://doi.org/10.1029/95JD01296, 1995.
- 170 Sha, M. K., De Mazière, M., Notholt, J., Blumenstock, T., Chen, H., Dehn, A., Griffith, D. W. T., Hase, F., Heikkinen, P., Hermans, C., Hoffmann, A., Huebner, M., Jones, N., Kivi, R., Langerock, B., Petri, C., Scolas, F., Tu, Q., and Weidmann, D.: Intercomparison of lowand high-resolution infrared spectrometers for ground-based solar remote sensing measurements of total column concentrations of CO<sub>2</sub>, CH<sub>4</sub>, and CO, Atmospheric Measurement Techniques, 13, 4791–4839, https://doi.org/10.5194/amt-13-4791-2020, 2020.

Sha, M. K., De Mazière, M., Notholt, J., Blumenstock, T., Bogaert, P., Cardoen, P., Chen, H., Desmet, F., García, O., Griffith, D. W. T., Hase,

175 F., Heikkinen, P., Herkommer, B., Hermans, C., Jones, N., Kivi, R., Kumps, N., Langerock, B., Macleod, N. A., Makkor, J., Markert, W., Petri, C., Tu, Q., Vigouroux, C., Weidmann, D., and Zhou, M.: Fiducial Reference Measurement for Greenhouse Gases (FRM4GHG), submitted for publication, 2024.





Vigouroux, C., Bauer Aquino, C. A., Bauwens, M., Becker, C., Blumenstock, T., De Mazière, M., García, O., Grutter, M., Guarin, C., Hannigan, J., Hase, F., Jones, N., Kivi, R., Koshelev, D., Langerock, B., Lutsch, E., Makarova, M., Metzger, J.-M., Müller, J.-F., Notholt, J., Ortega, I., Palm, M., Paton-Walsh, C., Poberovskii, A., Rettinger, M., Robinson, J., Smale, D., Stavrakou, T., Stremme, W., Strong, K., Sussmann, R., Té, Y., and Toon, G.: NDACC harmonized formaldehyde time series from 21 FTIR stations covering a wide range of

column abundances, Atmospheric Measurement Techniques, 11, 5049–5073, https://doi.org/10.5194/amt-11-5049-2018, 2018.

Wunch, D., Toon, G. C., Sherlock, V., Deutscher, N. M., Liu, C., Feist, D. G., and Wennberg, P. O.: The Total Carbon Column Observing Network's GGG2014 Data Version, Tech. rep., Carbon Dioxide Information Analysis Center, Oak Ridge National Laboratory, Oak Ridge, Tennessee, U.S.A. doi: 10.14291/tccon.ggg2014.documentation.R0/1221662, https://doi.org/10.14291/tccon.ggg2014.documentation.R0/1221662, 2015.

185