

5 **Technical note: Comparison of simultaneously applied chamber-based gas flux measurements ~~Gas Chromatography vs. Mid-Infrared Laser Absorption Spectroscopy: A comparison of methods for measuring greenhouse gas fluxes~~ from arable soils using gas chromatography (static chamber) and mid-infrared laser absorption spectroscopy (dynamic chamber)**

Wolfgang Aumer^{*1,2,★}, Morten Möller^{*3,★}, Carolyn-Monika Görres¹, Christian Eckhardt⁴, Tobias Karl David Weber⁵, Carolina Bilibio⁵, Christian Bruns³, Andreas Gattinger⁶, Maria Renate Finckh², Claudia Kammann¹

10 ¹ Department of Applied Ecology, Hochschule Geisenheim University, Von-Lade-Str.1, D-65366 Geisenheim, Germany

² Section of Ecological Plant Protection, University of Kassel, Nordbahnhofstr. 1a, D-37213 Witzenhausen, Germany

³ Section of Organic Farming and Cropping Systems, University of Kassel, Nordbahnhofstr. 1a, D-37213 Witzenhausen, Germany

⁴ Institute for Plant Ecology, Giessen University, Heinrich-Buff-Ring 26, D-35392 Gießen, Germany

15 ⁵ Section of Soil Science, University of Kassel, Nordbahnhofstr. 1a, D-37213 Witzenhausen, Germany

⁶ Chair of Organic Farming, Institute of Agronomy and Plant Breeding II, Justus-Liebig University Giessen (JLU), Gießen, Germany

★*These authors ~~Wolfgang Aumer and Morten Möller~~ contributed equally and are considered co-first authors.

Correspondence to: Wolfgang Aumer (wolfgang.aumer@hs-gm.de) and Morten Möller (morten.moeller@uni-kassel.de)

20 **Abstract**

For the study of soil-atmosphere exchange of green-house gases, a commonly adopted method is to monitor the change of gas concentrations in closed chambers. Accurate determination of CO₂, CH₄, and N₂O concentrations is therefore essential for reliable flux estimations. This study compares two techniques to determine these gas concentrations: Gas Chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS). We compared both techniques by carrying out simultaneous chamber measurements under field conditions on two separate days covering a range of fluxes. The GC method involved syringe sampling into gas-tight vials and subsequent laboratory analysis. In contrast to that, a LAS analyzer was directly connected to the chambers (tubing system) and thus enabled real-time, high-temporal resolution data. We calculated gas fluxes based on GC- and LAS-derived concentration measurements, using seven distinct flux calculation setups, including systematic variations in chamber enclosure times (30, 20 and 10 min) for LAS data. Across both measurement days, the comparison resulted in a high level of agreement for determined CO₂ fluxes with a normalized Root Mean Square Error (nRMSE): 5.79 – 16.70 %. A high level of agreement between the methods was also observed for N₂O fluxes (nRMSE: 14.63 – 24.64 %). In

25
30

contrast, there was a comparatively low agreement between methods for CH₄ fluxes (nRMSE: 88.42 – 94.54 %). N₂O and CH₄ fluxes highlighted the superior precision of LAS, as it detected significant fluxes (> minimum detectable flux) that were not significant with GC. For CH₄ this explains the low agreement between methods regarding arable soils that are dominated by
35 (low) CH₄-consumption fluxes.

1. Introduction

Various methods are available to determine greenhouse gas (GHG) fluxes at the atmosphere-soil interface. A common and versatile approach to determine GHG fluxes like carbon dioxide (CO₂), nitrous oxide (N₂O), and methane (CH₄) in field experiments is the closed chamber method, where the soil surface is temporarily covered by a chamber attached to a collar that
40 is permanently anchored in the soil to ensure an air-tight seal when the chamber is placed on top of it (Livingston and Hutchinson, 1995). The gas flux rate is then calculated based on the measured change in gas concentration within the chamber headspace over time, using either a linear or non-linear model (Lundegardh, 1927; Hutchinson and Mosier, 1981; Yu and Yao, 2017; Maier et al., 2022). Ultimately, numerous analytical techniques are available for measuring these GHG concentrations, with decisions influenced by specific research questions and logistical feasibility. Gas concentrations required for flux
45 calculations can be determined through real-time monitoring using online analyzers, or by collecting [discrete](#) gas samples ~~with syringes~~ for subsequent analysis via gas chromatography (GC), followed by peak integration and calculation of gas concentrations to estimate fluxes. In both cases, flux calculation procedures demand variables according to the ideal gas law such as chamber volume, temperature and local atmospheric air pressure and the resulting flux rates are subsequently referenced to the covered soil surface (Livingston and Hutchinson, 1995).

50 Besides the chamber design, the choice of the gas analytical technique can be an influential factor estimating GHG fluxes in the ecosystems under investigation. For example, a comparison between cavity ring-down spectroscopy (CRDS) and GC (Christiansen et al., 2015a) showed that CRDS resulted in higher calculated CO₂ fluxes, provided comparable results for N₂O, and was significantly more sensitive for CH₄ fluxes compared to GC. The authors concluded, that both CRDS and GC were equally effective in capturing treatment effects for CO₂ and N₂O in laboratory and field settings, whereas this was not the case
55 for CH₄. Zheng et al. (2008) pointed out, that measuring N₂O concentrations using GC with an Electron Capture Detector (ECD), the common practice using nitrogen (N₂) as a carrier gas can lead to overestimated N₂O emissions, especially in the presence of CO₂ or from weak sources (< 200 µg N m⁻² h⁻¹), [even if CO₂ is separated before the ECD via column separation](#). To address this issue, they suggest using alternative methods, for instance chemical removal of CO₂. Although the GC has commonly been used in the past, new modern online analyzers based e.g. on mid-infrared laser absorption spectroscopy (LAS)
60 allow real-time concentration measurements within the chamber where analytical values are immediately displayed, as opposed to the GC analysis which does not allow this in most cases due to the time gap between taking samples and subsequently analyzing them; exceptions being a GC mounted in a van or container in the field to a multiplexer plus automated chamber system (e.g. Flessa et al., 2002; Yao et al., 2009). An immediate display of the measured concentration, e.g. with an

LAS analyzer, provides the advantage of immediate detection of methodical set-up issues, such as leaks in chambers or tubing, enabling corrections or repetitions as necessary. Furthermore, LAS ~~typically~~~~often~~ offer higher analytical precision, thus enabling the detection of smaller flux rates and more precise measurements within the low concentration range, which means that the minimum detectable flux (MDF) is reduced notably (Christiansen et al., 2015a; Nickerson, 2016).

Several components of the closed chamber method significantly affect MDF, including [effective chamber height \(ratio of volume to emitting surface\)](#)~~chamber size~~, closure duration, sampling frequency during closure (i.e., periodicity), and the analytical precision of gas analysis technique [as reported by Christiansen et al. \(2015a\) and Nickerson \(2016\)](#). Substantial advancements in analytical precision and temporal resolution have led to markedly improved accuracy of GHG flux measurements, enabling shorter chamber closure times (Brümmer et al., 2017; Johannesson et al., 2024) and thereby minimizing disturbance and flux gradient alteration (Livingston and Hutchinson, 1995; Maier et al., 2022). This study aims to serve as a pilot study on the comparability of the ~~world-wide very common~~~~classie~~ GC method and fast developing LAS analyzers. A challenge of adoption of new technology is also that it changes the inherent measurement uncertainties, thus, knowledge about the impact of the gas analytical method on the gas fluxes are required (Cowan et al., 2025; Kong et al., 2025). Thus, the objective is to assess how closely the GHG flux values from both methods align and to evaluate the limitations of each method in terms of measurement accuracy. For this purpose, we conducted closed chamber measurements using simultaneously static (GC) and dynamic (LAS) approaches under field conditions and calculated the corresponding MDF.

80 2. Material and methods

2.1. Site description

The measurements were carried out in a long-term field experiment (LTE), initiated in 2010, which is well described in Bilibio et al. (2025)~~Klicken oder tippen Sie hier, um Text einzugeben..~~. The LTE is located near Neu-Eichenberg, Hesse, Germany (223 m asl., 51°22'N 9°54'E), within an organic research station operated by the University of Kassel which is further described in Leisch et al. (2025). The geological formation (Keuper) is covered by a loess layer up to a thickness of 1.8 m. The soil, characterized as a silty loam, is classified as a Luvisol (Obalum et al., 2019). It consists of 13 % clay, 84 % silt, and 3 % sand, with an organic matter content of 2 % (Schmidt et al., 2017). Over a 30-year period (1991 to 2020), the average annual temperature recorded was 9.3 °C, accompanied by an average annual precipitation of 663 mm. The climate is categorized as warm-temperate and fully humid with warm summers (Kottek et al., 2006) i.e. Cfb according to the Köppen-Geiger climate classification.

2.2 Experimental design and flux estimation

[For the method comparison between GC and LAS, fluxes were determined on two dates \(21st September 2023, after winter wheat harvest and tillage, and 24th March 2024, in the subsequent clover-grass mixture\) in four replicated plots \(Fig. 1\) containing two treatments differing in organic fertilization. This resulted in 16 sets of comparative data. No fertilization was](#)

95 applied to the winter wheat itself. The last fertilization before the experiment was carried out at the end of the winter cover
crop season 2021/22 (19th April 2022). All plots received 100 kg N ha⁻¹ as hair meal pellets, while one treatment additionally
received green waste compost (5 t ha⁻¹ a⁻¹ dry matter) applied on the same day. Prior to the first flux measurement in September
2023, a series of soil tillage operations was conducted. These interventions included straw chopping and harrowing on 13th
100 August 2023; rotary tillage to a depth of ~4 cm on 12th September 2023; and, on 17th September 2023, loosening with a
subsoiler to a depth of 25 cm, tilling with a chisel plough, and harrowing to a depth of ~10 cm; followed by sowing clover-
grass on 18th September 2023. ~~For the method comparison between GC and LAS, fluxes were determined at two dates (21st~~
~~September 2023 after winter wheat and tillage, and on 24th March 2024 in the subsequent clover grass mixture) in plots (Fig-~~
~~4) containing two treatments, which differed also in organic fertilization (both treatments were fertilized with 100 kg N ha⁻¹ as~~
~~hair meal pellets; one treatment additionally received green waste compost at 5 t ha⁻¹ a⁻¹ (dry matter)), replicated four times.~~
105 ~~This resulted in sixteen sets of comparative data. Gas samples were taken from the chamber by LAS and with syringes and~~
~~subsequent analysis by GC. In our study, we compared both methods by carrying out simultaneous measurements with closed~~
~~chambers using the static and dynamic approach, respectively.~~

The chambers used for gas flux measurements were made from non-transparent PVC and fitted with a fan, thermometer,
vent (Hutchinson and Livingston, 2001), and a closable opening. The closable opening ~~to~~ prevents pressure pumping during
110 chamber placement, as the downward movement of the chamber can lead to capturing extra air inside the chamber. This results
in an artificial overpressure inside the chamber which can either flush stored gases out of the soil pore system or act as a
diffusive barrier at the soil surface. The closable opening prevents this chamber placement artefact. Additional captured air
will escape through this hole as it offers the path of least resistance for excess air. On average, the soil collars (50 x 50 cm)
were installed up to a soil depth of 12 cm, the chamber itself had a dimension of 50 cm x 50 cm x 50 cm and an effective
115 chamber height of 0.68 m. The total volume of the setup, including measurement cells, filters, tubes and chambers depends on
the depth of the soil collars, which was determined for each collar individually and was on average 0.1576 m³ (157.56 liters).
Tubes were made of Polytetrafluorethylen (PTFE). The enclosure time of the chambers was 30 minutes and for each chamber
measurement, the initial and final temperature (Probe thermometer, *LT-101, TFA Dostmann GmbH & Co. KG, DE*) values
were averaged for the flux calculation. Local air pressure data was obtained from a nearby (~200 m distance) weather station
120 (*ATMOS 41, METER Group, Munich, DE*). We analyzed the gases CO₂, N₂O and CH₄ simultaneously using two approaches:
(i.) in real time on the field using LAS and (ii.) by collecting discrete gas samples which were subsequently analyzed by GC.




Figure 1: Field sampling after winter wheat and tillage on 2023-09-21 (a) and in a clover grass mixture on 2024-03-24 (b).

125

~~We analyzed the gases CO_2 , N_2O and CH_4 simultaneously, (i.) in real time on the field with a MIRA Ultra $\text{N}_2\text{O}/\text{CO}_2$ and a MIRA Ultra Mobile LDS: $\text{CH}_4/\text{C}_2\text{H}_6$ analyzer using LAS (Direct Absorption) (AERIS Technologies, Inc., USA) connected with a multiplexer PRI-8600D (Pri-eco Technology Co. LTD, CHN), and (ii.) by taking gas samples with syringes at six points in time and subsequent analysis by GC. For GC analysis, the samples were transferred to pre-evacuated 12 ml glass vials with grey chlorobutyl rubber septa (Labco Limited, UK)~~ The LAS analyzers were a MIRA Ultra $\text{N}_2\text{O}/\text{CO}_2$ and a MIRA Ultra Mobile LDS: $\text{CH}_4/\text{C}_2\text{H}_6$ using LAS (Direct Absorption) (AERIS Technologies, Inc., USA) connected with a multiplexer PRI-8600D (Pri-eco Technology Co. LTD, CHN). Each analyzer contained a temperature- and pressure-stabilized multipass measurement cell (volume 60 cm^3) with an internal tubing volume of 6 cm^3 . Gas was circulated between the chambers and the analyzers via two tubes (inlet and outlet), each 20 m in length and 4 mm in inner diameter. The measurement cells were heated to $42 \text{ }^\circ\text{C}$ and operated under vacuum ($< 240 \text{ mbar}$). The analyzers applied mid-infrared solid-state lasers to detect gas-specific absorption lines, achieving an effective optical path length of 13 m. During operation, the pump voltage was set to 5 V for the CH_4 analyzer and 4 V for the $\text{N}_2\text{O}/\text{CO}_2$ analyzer, resulting in a combined flow rate of approximately $0.4\text{--}0.5 \text{ L min}^{-1}$.

130

135

For GC analysis the discrete gas samples were taken at six different timepoints and transferred to pre-evacuated 12 ml glass vials (Exetainers) with grey chlorobutyl rubber septa (Labco Limited, UK) using syringes. The time interval between the first and second samples was 215 seconds and from the second sample onward, samples were taken at regular intervals of 390 seconds up to the sixth sample for all chamber measurements. ~~The vials were filled with slight overpressure (~20 ml gas sample squeezed into the 12 ml container).~~ The GC used was a Bruker Model 450 (Bruker Corp., USA) connected to an autosampler and equipped with three separate detectors: a thermal conductivity detector (TCD, $200 \text{ }^\circ\text{C}$) for CO_2 , a flame ionization detector (FID, $300 \text{ }^\circ\text{C}$) for CH_4 , and an electron capture detector (ECD, $300 \text{ }^\circ\text{C}$) for N_2O . The column oven was maintained at $50 \text{ }^\circ\text{C}$. In the first channel, which carried both the TCD and FID, the gas first passed through a $1.0 \text{ m} \times 1/8''$ Hayesep Q (80/100) guard column, followed by a $1.5 \text{ m} \times 1/8''$ Molsieve 13X (80/100) column for separation of CO_2 and CH_4 . The second channel

140

145

dedicated to the ECD, was equipped with a 0.5 m × 1/8" Hayesep N 80/100 guard column and a 2.0 m × 1/8" Hayesep D 80/100 column for N₂O separation. No water flushing or additional pre-columns were used, and Ar/CH₄ (10 mL min⁻¹) was applied as make-up gas for the ECD. ~~Thermal Conductivity Detector (TCD) for CO₂, Flame Ionization Detector (FID) for CH₄ and Electron Capture Detector (ECD) for N₂O~~ Since neither the GC nor the autosampler has its own pump to draw the sample into the system, overpressure is used in the sample vials (~20 ml gas sample was transferred into a 12 ml vial). This ensures that the sample is pushed into the sample loop and also pushes any residues from the previous sample out of the sample loop to prevent the samples from mixing. Another important reason for overpressure is the protection of the gas sample during storage: if the septum becomes leaky, the gas will escape outward first, preventing contamination from ambient air. Before each run, four standard gases (SG) (*DEUSTE Gas Solutions GmbH, DE*) were used, in ascending concentration order, for calibration (see Table A1). In addition, a vial of standard gas three (SG 3) was measured every 43 samples as a control.

Prior to flux calculation, we removed the first and last 30 seconds from the LAS datasets to minimize potential disturbances caused by chamber closure and opening, resulting in a total of 1740 s (i.e. ~1740 data points) for the flux calculation. From here on, however, this approach is referred to as a 30-minute chamber closure time. In a first step, fluxes were calculated based on the full 30-minute chamber closure period using two different R packages: *gasfluxes* (version 0.4-4) (Fuss and Hüppi, 2024), which is commonly applied to GC data but not typically used with high-frequency LAS measurements, and *goFlux* (version 0.2.0) (Rheault et al., 2024), which is specifically designed for high-frequency ~~and data~~ and incorporates corrections for water vapor dilution (LI-COR, 2023). GC data were processed using *gasfluxes*, while LAS data were analyzed with both *gasfluxes* and *goFlux*. To reduce the influence of model selection on the results, we applied robust linear regression (provided in the *goFlux* script and accounts for the weighting of outlier data points in flux calculation) to the GC data and standard linear regression to the LAS data. For the GC data, please note that for each of the three gases studied, the robust linear algorithm of the *gasfluxes* script applied robust weighting in only 8 out of 16 flux calculations. The remaining flux calculations did not differ from those obtained using ordinary linear regression. In a second step, we calculated the LAS-based fluxes using the R package *goFlux* (Rheault et al., 2024). To evaluate whether shorter measurement intervals with the LASGA-derived fluxes could still yield results comparable to those from GC measurements, we shortened the LAS dataset to the initial 20 and 10 minutes to simulate shorter chamber enclosure times and recalculated fluxes using the *goFlux* script. This approach to use *goFlux* for LAS data, is supposed to reflect common practice, where shorter chamber measurement durations are often used. These were then compared to GC fluxes based on the full 30-minute measurement period. Subsequently, we implemented a flux selection procedure to ensure that the best-fitting model was applied for each flux calculation. Specifically, we employed the MDF approach described by Nickerson (2016) (Eq. 3) to determine the kappa max threshold following Hüppi et al. (2018) (Eq. 5). Based on whether the calculated kappa value exceeded or fell below this threshold, either a linear or a non-linear model (Hutchinson and Mosier Regression model, HMR) (Hutchinson and Mosier, 1981) was selected accordingly. As a result, these flux calculation setups facilitated the comparison of seven distinct approaches: (1) GC data, calculated with the *gasfluxes* package and robust linear regression (GC_gasf_rl), (2) GC data, calculated with the *gasfluxes* package and model selection

(GC_gasf), (3) LAS data, calculated with the *gasfluxes* package and linear regression (LAS_gasf_30_1), (4) LAS data, calculated with the *goFlux* package and linear regression (LAS_gof_30_1), (5) LAS data, 30 minute dataset, calculated with the *goFlux* package and model selection (LAS_gof_30), (6) LAS data, 20 minute dataset, calculated with the *goFlux* package and model selection (LAS_gof_20), (7) LAS data, 10 minute dataset, calculated with the *goFlux* package and model selection (LAS_gof_10).

All flux estimates were multiplied with a flux term:

$$flux.term.gasfluxes = \frac{V M P}{S R T} \quad (1)$$

where V is the total chamber volume (m^3), M is the molar mass of the measured gas (mol), P is the local atmospheric air pressure (Pa), S is the soil surface covered by the chamber (m^2), R is the universal gas constant ($8.314 \text{ J m}^3 \text{ Pa mol}^{-1} \text{ K}^{-1}$) and T is the temperature inside the chamber (K).

Additionally, the *goFlux* package corrects for the dilution effect caused by the increase of water vapor inside the chamber during the measurement (LI-COR, 2023), expressed as follows:

$$flux.term.goflux = \frac{(1-H_2O) V P}{S R T} \quad (2)$$

Please note, that in the *goFlux* script, V is given in liters (L) and therefore R in L kPa $K^{-1} \text{ mol}^{-1}$. The term for M is omitted, as fluxes in the script are calculated in mol and H_2O is the water vapor (mol mol^{-1}). Subsequently, the values were converted to the same units as in the *gasfluxes* script (CO_2 ($\text{g m}^{-2} \text{ h}^{-1}$), N_2O and CH_4 ($\text{mg}^{-2} \text{ h}^{-1}$)), using the respective values for M .

The methodical detection limit for the measured fluxes was determined based on MDF following the approach of Nickerson (2016). The MDF (CO_2 ($\text{g m}^{-2} \text{ h}^{-1}$), N_2O and CH_4 ($\text{mg}^{-2} \text{ h}^{-1}$)), was calculated as follows:

$$MDF = \frac{A_A}{t_c \sqrt{\frac{t_c}{p_s}}} \left(\frac{V P}{S R T} \right) \quad (3)$$

where A_A is the analytical precision of the instrument (ppm), t_c is the closure time of the chamber (h), and p_s is the sampling periodicity (h), whereby V and R again given in in the same units as in Eq. (1). The variables V , P , and T were averaged over the measurement period covering eight plots (chamber measurements) for each of the two measurement days. The analytical precisions (A_A) according to the manufacturer of the LAS analyzer were 200 ppb for CO_2 , 0.001 ppb for CH_4 , and 0.002 ppb for N_2O . The sampling frequency was 1 Hz and the analytical accuracy A_A for the two GC runs was calculated following Christiansen et al. (2015b):

$$A_A = 3 \cdot t_{99\%} \cdot SD \quad (4)$$

where $t_{99\%}$ is the t value at the 99 % confidence interval at $df = 4$ (4.604) and SD is the standard deviation of five samples standard gas 3 (SG 3, see Table A1) within one run. The average A_A of the two GC runs was 90.6 ppm for CO_2 , 0.4 ppb for CH_4 , and 0.1 ppb for N_2O .

In methodological reference to Hüppi et al. (2018) the kappa max threshold ($k.\text{max}$ (h^{-1})) was calculated as follows:

$$k.\text{max} = \frac{\text{linear flux}}{\text{MDF } t_c} \quad (5)$$

On the first day of measurements (2023-09-21), the data set showed gaps for LAS data after 20 minutes in two out of eight plots. For all scatterplots analyses only complete datasets were used. To avoid introducing inconsistencies in the method comparisons based on boxplot representations, the flux values for these plots were calculated based on the available 20-minute data and copied to in the 30-minute dataset. This decision was further supported by the observation that the differences between the 20- and 30-minute measurements were minimal.

2.3. Statistical analysis

To evaluate the linear relationship between the analytical approaches, we first calculated Pearson's correlation coefficient. The deviation of absolute fluxes was quantified using the Root Mean Square Error (RMSE), which measures the average magnitude of differences between paired observations. To facilitate interpretation and enable comparison across different flux magnitudes, we additionally calculated the normalized RMSE (nRMSE) by dividing RMSE by the means across the compared observations for each compared data set (five comparisons: see Table 2, 3 and 4). This approach does not assume one method as a "true" reference but rather assesses relative agreement between the two methods. The resulting nRMSE was expressed as a percentage, allowing for a standardized evaluation of method agreement independent of the absolute flux magnitude. Additionally, flux values from the seven different approaches were compared using a Kruskal-Wallis to test for significant differences, since the data was not normally distributed.

3. Results

The comparison of calculated MDF for CO_2 , N_2O , and CH_4 showed that the LAS method yields much lower MDFs than the GC method for all gases and chamber enclosure durations (Table 1). In the case of LAS, however, flux magnitude sensitivity decreases with shorter enclosure times, as indicated by increasing MDF values. Overall, CO_2 fluxes obtained from the GC and LAS methods generally agree well (Fig. 2, Tab. 2). Figure 2a shows GC-derived fluxes (*gasfluxes* script, robust linear regression) plotted against [LASGA](#)-derived fluxes for a 30-minute chamber closure time, using the two calculation scripts


235 *gasfluxes* (linear regression) and *goFlux* (linear regression). Most data points align closely with the line of equality, indicating
 a strong agreement between GC and LAS, as well as between both LAS-based calculation approaches. Figure 2b presents
 fluxes based on flux-model selection derived from GC fluxes (*gasfluxes* script), compared to LAS fluxes with different
 enclosure durations (30, 20 and 10 minutes) calculated using the *goFlux* script that were also selected using the same flux-
 model selection approach. For all CO₂-LAS fluxes calculated using the *goFlux* script, the flux selection algorithm applied in
 a second step resulted in HMR model fits for all fluxes. In contrast, flux selection for the GC fluxes (*gasfluxes* script) could
 240 only be carried out to a limited extent (31 %). This is because HMR models and the corresponding kappa values were available
 for only a minority of these fluxes, as the script's internal diagnostics considered the HMR model unsuitable in most cases.
 The data points cluster closely around the line of equality across all closure durations, with high correlation coefficients ($r \approx 1$),
 resulting in nRMSE values below 17 % (Table 2). The greatest scatter is observed for the 10-minute duration, indicating lower
 agreement. All fluxes exceeded the MDF and the median of the flux magnitudes of the investigated methodological approaches
 245 was relatively similar whereas the median of the GC fluxes was in tendency higher (Fig. 2c).

250

Table 1: Minimum detectable fluxes (MDF) for CO₂, CH₄, and N₂O, based on the approach by Nickerson (2016), using Gas Chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS). Calculations were performed for a 30-minute chamber closure duration (GC_30 and LAS_30), as well as for shorter durations of 20 and 10 minutes using the LAS (LAS_20 and LAS_10). Data represents averages from two measurement days.

Method	CO ₂ × 10 ⁻⁵ (g m ⁻² h ⁻¹)	N ₂ O × 10 ⁻⁵ (mg m ⁻² h ⁻¹)	CH ₄ × 10 ⁻⁵ (mg m ⁻² h ⁻¹)
GC_30	940	960	1,510
LAS_30	1.13	1.13	2.05
LAS_20	2.04	2.04	3.72
LAS_10	6.00	6.00	10

255





260 **Figure 2: Comparison of CO₂ fluxes derived from gas chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS)**
measurements using different chamber closure durations and calculation approaches. (a) GC derived fluxes for a 30-minute
chamber closure duration, calculated using the *gasfluxes* script with robust linear regression (GC_gasf_ri), are plotted against LAS
derived fluxes also for a 30-minute closure duration. The LAS fluxes were calculated using either the *gasfluxes* script
(LAS_gasf_30_l) or the *goFlux* script (LAS_gof_30_l), both with linear regression. (b) Model-selected GC fluxes for a 30-minute
closure duration using the *gasfluxes* script (GC_gasf_30) are compared to model-selected LAS fluxes using the *goFlux* script for 30,
20, and 10-minute closure durations (LAS_gof_30, LAS_gof_20, LAS_gof_10). In both panels (a) and (b), the black line represents
the line of equality, while coloured lines indicate regression lines for the respective approaches. For CO₂, all flux values exceeded
the minimum detectable flux (MDF). (c) Boxplots of all calculated CO₂ fluxes across methods and durations (n = 16). For the
summary of the mathematical analysis underlying the scatter plot comparisons, see Table 2.

270 **Table 2: Regression results comparing CO₂ fluxes derived from Gas Chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS) with varying chamber closure times and calculation scripts. Displayed are the correlation coefficient (r), regression equation, root mean square error (RMSE), and normalized RMSE (nRMSE).**

Method	r	Equation	RMSE (g m ⁻² h ⁻¹)	nRMSE (%)
LAS_gasf_30_1 vs. GC_gasf_rl	1.00	$y = 0.09 + 0.94x$	0.057	5.79
LAS_gof_30_1 vs GC_gasf_rl	1.00	$y = 0.08 + 0.96x$	0.057	5.80
LAS_gof_30 vs. GC_gasf	1.00	$y = 0.07 + 1.01x$	0.086	8.51
LAS_gof_20 vs. GC_gasf	1.00	$y = 0.07 + 1.02x$	0.100	9.38
LAS_gof_10 vs. GC_gasf	0.97	$y = 0.08 + 1.03x$	0.176	16.70

275 For N₂O, a generally good agreement was also observed between the flux values derived from GC and LAS flux estimation (Fig. 3, Table 3). This was observed both, without (Fig. 3a) and with the flux selection procedure (Fig. 3b). For N₂O, flux selection for LAS fluxes calculated with the *goFlux* script always resulted in HMR models. In contrast, model selection for GC fluxes using the *gasfluxes* script was only possible for a small subset of fluxes (18 %) due to the absence of HMR models and corresponding kappa values (HMR diagnostics). Notably, obtaining HMR models with GC fluxes was only possible on the first measurement day, when an N₂O emission pulse occurred. The values cover a **wide** range of flux magnitudes, including very low and **relatively higher** values. The data points align closely along the line of equality, and only a few GC measurements
280 fell below the MDF (low flux range), whereas all LAS fluxes (even small negative fluxes) were above the MDF. The median flux magnitudes were relatively similar overall, with the GC fluxes tending to be higher, as was also observed for CO₂. The nRMSE varied between 14.63 % and 24.64 %, with the highest values corresponding to the shortest LAS closure time (10 minutes), as also observed for CO₂.





285

Figure 3: Comparison of N₂O fluxes derived from gas chromatography (GC) mid-infrared laser absorption spectroscopy (LAS) measurements using different chamber closure durations and calculation approaches. (a) GC derived fluxes for a 30-minute chamber closure duration, calculated using the *gasfluxes* script with robust linear regression (GC_gasf_rl), are plotted against LAS derived fluxes also for a 30-minute closure duration. The LAS fluxes were calculated using either the *gasfluxes* script (LAS_gasf_30_l) or the *goFlux* script (LAS_gof_30_l), both with linear regression. (b) Model-selected GC fluxes for a 30-minute closure duration using the *gasfluxes* script (GC_gasf_30) are compared to model-selected LAS fluxes using the *goFlux* script for 30, 20, and 10-minute closure durations (LAS_gof_30, LAS_gof_20, LAS_gof_10). In both panels (a) and (b), the black line represents the line of equality, while coloured lines indicate regression lines for the respective approaches. For N₂O, only higher fluxes exceeded the minimum detectable flux (MDF) (points filled in black), whereas all LAS flux values exceeded the MDF (points filled in grey). (c) Boxplots of all calculated N₂O fluxes across methods and durations, n = 16. durations (n = 16). For the summary of the mathematical analysis underlying the scatter plot comparisons, see Table 3.


290

295

300 **Table 3: Regression results comparing N₂O fluxes derived from Gas Chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS) with varying chamber closure times and calculation scripts. Displayed are the correlation coefficient (r), regression equation, root mean square error (RMSE), and normalized RMSE (nRMSE).**

Method	r	Equation	RMSE (mg m ⁻² h ⁻¹)	nRMSE (%)
LAS_gasf_30_1 vs. GC_gasf_rl	1.00	y = 0 + 0.98x	0.010	15.01
LAS_gof_30_1 vs GC_gasf_rl	1.00	y = 0 + 1x	0.010	14.63
LAS_gof_30 vs. GC_gasf	0.99	y = 0 + 1.02x	0.012	18.21
LAS_gof_20 vs. GC_gasf	0.99	y = 0 + 1.04x	0.013	17.01
LAS_gof_10 vs. GC_gasf	0.99	y = 0.01 + 1.03x	0.018	24.64

305 Figure 4 illustrates the comparison of CH₄ fluxes between GC and LAS methods without (Fig. 4a) and with model selection (Fig. 4b). In contrast to CO₂ and N₂O, the data points displayed substantial scatter, and the agreement with the line of equality was visible weaker (compare to Fig. 2 and 3) and the nRMSE values remarkable higher (compared to Tables 2 and 3). For CH₄, flux selection for LAS fluxes calculated with the *goFlux* script always resulted in HMR models. In contrast, model selection for GC fluxes using the *gasfluxes* script was only possible for a minority of the fluxes (18 %) due to the absence HMR and kappa values, as a result of the script's internal HMR diagnostics. Despite some spread, the flux value distributions across GC and LAS measurements appear broadly similar (Fig. 4c), without significant differences between the approaches, although the median of the GC data tended to be higher (more negative). All GC fluxes were below the MDF, whereas nearly all LAS fluxes except for one value with 10-minute enclosure time were above it. The regression analysis confirmed the visual impression of lower consistency. The correlation coefficients (r) remained below 0.54 in all cases and the nRMSE values ranged between 88.42 % and 94.54 %, indicating relatively large relative deviations (Tab. 4).



315 **Figure 4: Comparison of CH₄ fluxes derived from gas chromatography (GC) and mid-infrared laser absorption spectroscopy (LAS)**
measurements using different chamber closure durations and calculation approaches. (a) GC derived fluxes for a 30-minute
chamber closure duration, calculated using the *gasfluxes* script with robust linear regression (GC_gasf_rl), are plotted against LAS
derived fluxes also for a 30-minute closure duration. The LAS fluxes were calculated using either the *gasfluxes* script
(LAS_gasf_30_l) or the *goFlux* script (LAS_gof_30_l), both with linear regression. (b) Model-selected GC fluxes for a 30-minute
closure duration using the *gasfluxes* script (GC_gasf_30) are compared to model-selected LAS fluxes using the *goFlux* script for 30,
320 **20, and 10-minute closure durations (LAS_gof_30, LAS_gof_20, LAS_gof_10). In both panels (a) and (b), the black line represents**
the line of equality, while coloured lines indicate regression lines for the respective approaches. For CH₄, no GC flux values exceeded
the minimum detectable flux (MDF), while almost all LAS flux values did (points filled in grey). Only one LAS flux fell below the
MDF (indicated by an empty fill, located in the top right corner). (c) Boxplots of all calculated CH₄ fluxes across methods and
durations, n = 16. For the summary of the mathematical analysis underlying the scatter plot comparisons, see Table 4.

Table 4: Regression results comparing CH₄ fluxes derived from GC and LAS methods with varying chamber closure times and calculation scripts. Displayed are the correlation coefficient (r), regression equation, root mean square error (RMSE), and normalized RMSE (nRMSE).

Method	r	Equation	RMSE (mg m ⁻² h ⁻¹)	nRMSE (%)
LAS_gasf_30_1 vs. GC_gasf_rl	0.54	$y = -0.01 + 0.55x$	0.027	89.40
LAS_gof_30_1 vs GC_gasf_rl	0.54	$y = -0.01 + 0.56x$	0.027	89.34
LAS_gof_30 vs. GC_gasf	0.52	$y = -0.01 + 0.51x$	0.028	91.36
LAS_gof_20 vs. GC_gasf	0.51	$y = -0.01 + 0.56x$	0.027	88.42
LAS_gog_10 vs. GC_gasf	0.46	$y = -0.01 + 0.55x$	0.028	94.54

4. Discussion

Overall, the measurement setup ~~proved~~ was suitable for a method comparison, ~~enabling as it enabled~~ the detection of both high and low flux magnitudes for CO₂ and N₂O. The CH₄ oxidation rates covered the common range for arable soils (Le Mer and Roger, 2001), further supporting the suitability of the dataset for the method comparison across all three gases ~~high and low flux magnitudes across all three investigated gases~~. While ~~possible~~ treatment effects (two treatments) were tested, none were ~~could have offered further insights, neither were~~ statistically significant ~~in the dataset~~ and ~~were~~ therefore not investigated in further detail. -The calculation of the MDF according to Nickerson (2016) revealed marked differences between the GC and LAS methods, which can be attributed to substantial disparities in analytical precision. As a result, the LAS method was able to detect significantly lower fluxes than the GC method, representing a clear analytical advantage. The LAS-method's high sensitivity allowed for the detection of fluxes that would have remained below the detection threshold of the GC and would therefore have been classified as not significantly different from zero. Consequently, the enclosure time could be reduced substantially (e.g., to 10 minutes), as also recommended in other studies (Brümmer et al., 2017), without leading to a strong increase in the MDF, offering additional methodological flexibility and potential reduction of measurement duration-induced disturbances.

For CO₂, a high level of agreement was observed between the GC and LAS data pairs across all chamber durations and both scripts (*gasfluxes* and *goFlux*). The MDF was exceeded in all cases for both LAS and GC, facilitating comparability. Absolute flux values did not differ significantly between approaches, and the nRMSE remained low across all comparisons (max. 17 %, see Table 2). The regression slopes were close to 1, and intercepts near zero, indicating that both systems reliably captured actual CO₂ fluxes under field conditions. In line with Cowan et al. (2025), this shows that the ~~traditional~~ common and still widely employed closed static chamber method is not necessarily inferior to closed dynamic chamber approaches in

situations where the measured fluxes are considerably larger than the analytical uncertainty. ~~Depending on the study site and the research question, the choice between GC and LAS can then rather be guided by logistical and budgetary constraints.~~

355 For N₂O, the comparison revealed clear differences between the two measurement days. On the first day, flux magnitudes were relatively high, and MDFs were exceeded by both methods. On the second day, however, flux magnitudes were low and, in some cases, negative. The fertilization in the crop rotation had taken place 17 months before the first measurement day and 23 months before the second measurement. We therefore consider it unlikely that the observed N₂O peak in our study was strongly influenced by earlier nitrogen inputs. Instead, it is more plausibly explained by mineralization pulses triggered by the series of soil tillage interventions carried out shortly before the first measurement day (see section 2.2), of course in
360 ~~combination with long-term effects of earlier nitrogen inputs. On the first day, flux magnitudes were relatively high, likely caused by a recent tillage event (initial shallow tillage was performed with a rotavator (~4 cm depth), followed by deeper rotary tillage with a rotary tiller), and MDFs were exceeded by both methods. On the second day, however, flux magnitudes were low and, in some cases, negative.~~ The capture of this N₂O pulse, along with the very low to negative fluxes, demonstrated a strong agreement between methods, with only minimal error for the 30-min enclosure time (~15 %,) and a tolerable deviation
365 for 10-min enclosure time (25 %) (see Table 3). Negative fluxes were consistently detected and quantified by the LAS method, whereas the GC measurements for these fluxes fell below the MDF, resulting in no significant difference from zero flux. This demonstrates the superior ability of LAS to measure very low N₂O fluxes and even small net uptake events. Such negative fluxes are commonly attributed to the final step of denitrification, where N₂O is reduced to N₂ (Cavigelli and Robertson, 2001; Glatzel and Stahr, 2001; Butterbach-Bahl et al., 2002). Although these processes are well documented, they are frequently
370 underestimated or omitted in flux datasets, as negative net fluxes often fall below detection thresholds. Chapuis-Lardy et al. (2007) highlight that this omission may lead to critical misinterpretations of the global N₂O budget. The high analytical sensitivity of the LAS method could help to better integrate N₂O sinks, but also very low N₂O emissions into future flux assessments (Cowan et al., 2025; Triches et al., in review, 2025).

The results for CH₄ showed considerable divergence. All measured LAS-derived CH₄ fluxes were negative, as were almost
375 all GC-derived CH₄ fluxes, as expected for well-aerated ~~arable~~^{upland} soils where CH₄ oxidation by methanotrophic bacteria predominates (Hütsch, 1998; Powlson et al., 2014). While the LAS measurements consistently exceeded the MDF, the GC fluxes did not, likely resulting in a higher degree of uncertainty in the GC dataset and thus a lower ability to detect potential treatment effects in CH₄ consumption rates. This discrepancy may have led to inflated variability in the GC results, particularly at low flux values (near zero). Even under conditions with only low changes in CH₄ concentrations ~~in the low concentration~~
380 ~~range~~, the LAS provided significant results (flux > MDF) for nearly all fluxes (with only one exception, with a 10-minute chamber enclosure time), highlighting its advantage in terms of flux sensitivity. Despite these differences and the low agreement of measured fluxes, the mean CH₄ fluxes did not differ significantly between methods (see Fig. 4c), suggesting that any potential effects induced by treatment factors might be captured in a similar manner. For verification of the assumption, both with GC and LAS, measuring a time series with different treatments and calculating cumulative fluxes is advisable.

385 However, the low correlation ($r \approx 0.5$) and high nRMSE up to 95 % indicate that more extensive datasets ~~for examples~~ with
more values exceeding MDF for GC fluxes are needed to validate this assumption. ~~Even so~~ ~~However~~, according to Le Mer and
~~Roger (2001)~~ ~~within this study~~, ~~CH₄ oxidation rates were in part comparatively high: CH₄ oxidation rates in aerated soils rarely~~
~~exceed values of 0.1 mg m⁻² h⁻¹ (Le Mer and Roger, 2001)~~, ~~CH₄ oxidation rates in aerobic upland soils rarely exceed 0.1 mg~~
~~m⁻² h⁻¹.~~ ~~but~~ ~~I~~ in our case, we observed rates up to 0.08 mg m⁻² h⁻¹. This leads us to assume that a more extensive data-sets of
390 simultaneous chamber measurements would probably not ~~necessarily~~ have provided further clarity, as our values were already
close to this typical upper limit. For further investigations, reducing effective chamber height ~~volume~~ may represent the most
effective measure to minimize MDF. However, in field campaigns, this approach is often not feasible, as it is often of interest
to include plants within the chamber (e.g. ~~Flessa et al., 2002; Zhang et al., 2013~~) and also to level out spatial variability over
capturing larger soil surfaces (e.g. heterogeneous distribution of solid manure and composts (Krauss et al., 2017)). In our study,
395 we used our standard chamber configuration, which is relatively high and generally applied in the LTE at the study site, where
tall crops are frequently grown. While our relatively large effective chamber height (0.68 m) increased the MDF, our results
show that CH₄ fluxes above the MDF could still be detected with the LAS system, whereas this was not possible with the GC-
based approach. This means that, because CH₄ consumption rates in arable soils are generally very low, measurements made
with large effective chamber height often yield fluxes below the MDF with GC method. ~~For verification of the assumption,~~
400 ~~that treatment effects can be captured in a similar manner, both with GC and LAS, measuring a time series with different~~
~~treatments and calculating cumulative fluxes is advisable. In any case, for well-aerated soils using relatively large chamber~~
~~setups, CH₄ fluxes are generally expected to fall below the MDF for the GC method. Consequently, under these conditions,~~
~~the measurements obtained using LAS may be considered more reliable for investigating CH₄ uptake and treatment effects on~~
~~CH₄ consumption in arable soils.~~

405

For chamber-based flux measurements, three types of water vapor corrections need to be considered: (i) absorption band
broadening, (ii) instrument cross-sensitivity, and (iii) dilution of the air sample by water vapor. The dry trace gas values
reported by the LAS are automatically corrected for the first two effects, which are instrument-specific and implemented in
410 the LAS software. However, the third effect, the dilution of the chamber air due to changes in water vapor concentration, is
not corrected internally and must be applied by the user. This additional correction is included in the *goFlux* script, which is
why it was applied in our data processing workflow. We calculated LAS fluxes both with and without water correction and
compared them among each other as well as to the non-water-corrected GC fluxes. The application of the *goFlux* script,
including correction for water vapor inside the chamber during the measurement (dilution effect) (LI-COR, 2023), had only
415 marginal effects on the absolute flux magnitudes and on the agreement between methods. This suggests that, in the present
dataset, absolute humidity had little impact on flux calculations, despite noticeable differences between measurement days
(average absolute humidity during each measurement day: 12.01 g m⁻³ on 21st September 2023 and 5.62 g m⁻³ on 25th March
2024). ~~This is in contrast to Kong et al. (2025), who compared GC with a LI-COR LI-7820 N₂O/H₂O trace gas analyzer (LI-~~

[COR Biosciences, USA](#)) based on optical feedback cavity-enhanced absorption spectroscopy (OF-CEAS) employing a near-infrared (NIR) tunable diode laser source: The authors applied water vapor corrections for dynamic chamber approach resulting in ~~who found~~ a significant effect ~~of water vapor corrections~~ on N₂O fluxes, especially below 50 µg N₂O-N m⁻² h⁻¹ (0.079 mg N₂O m⁻² h⁻¹, own conversion). ~~This~~ contributed ~~ing~~ to discrepancies in cumulated N₂O emission estimates in a similar chamber-based method comparison campaign on Danish arable soils. ~~They explained this discrepancy mainly by the different enclosure times. In contrast to our experimental design, enclosure times were not the same for the two different measurement modes. The dynamic chambers, used with the OF-CEAS, had much shorter enclosure times (2 to 16 min), which minimized the effect of water vapor changes, whereas static chambers used for GC sampling remained closed for 45 to 110 min and concentrations measurements were typically not corrected for water vapor at all. Additionally, there might have been a seasonal effect. The experiment of Kong et al. (2025) which was most similar in design to our experiment was conducted during the summer (June until September). The largest water vapor dilution effects occur on wet soils and/or soils with strongly transpiring plants at low CO₂ fluxes under dry and sunny conditions. The latter can lead to strong increases in chamber air temperature and subsequently in chamber humidity when a moisture source is available (Welles et al., 2001).~~

The flux selection approach was applied according to Hüppi et al. (2018) (LAS and GC). While HMR models were selected in all cases for the LAS fluxes, the *gasfluxes* script only allowed HMR calculations for a minority of the GC fluxes (CO₂: 31 %, N₂O and CH₄: 19%), as most did not meet the criteria defined by the script internal HMR diagnostics. On the first measurement day, when a high N₂O emission peak occurred, 3 out of 8 fluxes could be calculated using the HMR model, while on the second measurement day, however, none could be fitted with HMR. This illustrates that applying HMR calculations to GC-based measurements is often problematic, particularly at very low flux magnitudes near the MDF. It remains unclear whether this is primarily due to the limited number of measurement time points (only six per chamber) or the lower analytical precision of the method, or a combination of both. Although the *gasfluxes* script's internal HMR diagnostics allowed significantly more HMR model calculations for LAS compared to GC fluxes, it somewhat surprisingly still did not provide HMR models for all LAS fluxes, unlike the *goFlux* script. Nevertheless, when comparing the two methods (GC and LAS) with and without flux selection, it became apparent that there was only a small influence of the model selection procedure on our results and it did not significantly affect the relationship between the two methods over the 30-minute closure periods. In contrast, the shorter chamber closure durations had a more pronounced effect on the flux estimates. The relative error for CO₂ and N₂O was highest with the LAS when using a 10-minute closure time, compared to the GC fluxes, which were always based on a 30-minute enclosure. However, even in this case, the error remained within a moderate range for these two gases. Also, for CH₄, reducing the enclosure time to 20 and 10 minutes had only a minimal effect on the calculated LAS-derived fluxes. In addition to the only slightly reduced MDF resulting from the shortened chamber enclosure durations (see [Tab. 1 above](#)), this further highlights the LAS method's suitability in terms of operational practicability and the reduction of measurement duration-induced disturbance of the gas concentration gradient and hence gas flux dynamics between soil and chamber atmospheres. [Given the amount and quality of data that LAS analyzers can provide, a suitable alternative approach](#)

[might be the use of a generalized additive model for calculating fluxes from individual chamber measurements \(Themistokleous et al., 2024\), which is recommended for future assessments of LAS data.](#)

455 Taken together, although no treatment effects were explicitly tested in this study, both GC and LAS approaches resulted in comparable flux magnitudes for CO₂ and N₂O, suggesting that both methods would very likely provide consistent results when assessing potential treatment effects. For CH₄, flux magnitudes also did not differ significantly between methods; however, this apparent agreement is subject to considerably greater uncertainty, as indicated by the scatterplot analysis and the fact that GC fluxes were generally below the MDF and thus not significantly different from zero flux. However, for all three gases, a pattern of slightly higher fluxes, measured by GC, was observed. [This slight pattern should not be overinterpreted. For CO₂, the confidence is by far the highest, as all GC values are above the MDF, even though the median differences compared to LAS values are only slight \(Fig. 2c\). For N₂O, the differences in medians \(Fig. 3c\) are even smaller, and the confidence is lower because the small estimated GC fluxes for N₂O are less reliable \(< MDF\). For CH₄, the median deviation between GC and LAS values is largest \(Fig. 4c\), but almost all GC CH₄ values \(except one\) fall below the MDF. Regarding CO₂,](#)

460 [Christiansen et al. \(2015a\) reported that at very high CO₂ concentrations \(> 2000 ppm\), GC measurements tended to overestimate gas concentrations compared to cavity ring-down spectroscopy \(CRDS\), which they attributed to the non-linear response of the detector \(TCD\). In our dataset, however, the highest measured CO₂ concentration is only 1337 ppm, well below the levels at which \(Christiansen et al., 2015a\) clearly observed this effect. Therefore, the exact cause of the slight overestimation in our CO₂ flux data remains unclear. Nevertheless, it is conceivable that the detector-related non-linearity reported by Christiansen et al. \(2015a\) could play a role, implying that GC-based CO₂ flux measurements may slightly overestimate fluxes under certain conditions.](#)

465

470

The LAS offers several technical and analytical advantages: its low MDF allows for the detection of very small but significant CH₄ and N₂O fluxes; it permits shorter enclosure times without substantial increases in uncertainty; and it offers real-time measurements with minimal infrastructure requirements. These features enable faster, more flexible sampling and improve the resolution of short-term flux dynamics, such as during N₂O pulse-emission events. [While analyzers are already used at long-term monitoring sites \(e.g. ICOS Ecosystem Stations\) \(Pavelka et al., 2018\) and have also been applied in field experiments using manual \(e.g. Dix et al., 2024\) or rarely automated chambers \(Brümmer et al., 2017; Bond-Lamberty et al., 2020\), their limited mobility makes such measurements logistically demanding and constrains spatial coverage and therefore limits the ability to capture spatial variability. Recent advances in LAS techniques have led to the development of smaller and more portable analyzers, which now make high-frequency flux measurements less challenging and more feasible over a broader spatial scale with manual chamber applications, including at plot or ecosystem scale. Our results demonstrate that these portable LAS systems can detect very small CH₄ and N₂O fluxes that are often below the detection limit of conventional GC-based approaches. Consequently, portable LAS analyzers represent an important methodological advancement and can substantially improve the accuracy of annual and long-term GHG budget calculations, especially with respect to CH₄ and N₂O.](#)

475

480

485 ~~-Beside the methodological advantages there are~~At the same time, practical limitations of the LAS ~~include high acquisition~~
~~eosts,~~like increased vulnerability to environmental influences (e.g. dust, moisture, high temperature), and maintenance
demands, which may affect measurement logistics under field conditions. ~~From an economic perspective, the purchase of LAS~~
~~systems are~~generally reasonable, particularly when considering the operating costs of GC systems. ~~In terms of economic~~
490 ~~considerations, LAS analyzers do not necessarily more expensive than GC systems, particularly when operating costs are~~
~~considered. Institutions already operating GC systems may find it more economical to continue with GC, especially since~~
~~switching to LAS would involve significant additional costs.~~Although GC systems offer clear advantages, such as the ability
to easily conduct multiple measurement campaigns in parallel, the choice between GC and LAS ultimately depends on
budgetary constraints and the specific requirements and aims of the study.~~While GC systems offer clear advantages, such as~~
~~the ability to conduct multiple measurement campaigns in parallel, the choice between GC and LAS ultimately depends on~~
495 ~~budgetary constraints and the specific requirements of the study.~~

5. Conclusion

Biotic greenhouse gas (GHG) fluxes are often measured using the closed chamber method, but the choice of analytical
instrumentation influences flux estimations and reliability of the results. In this study, we compared gas chromatography (GC)
with mid-infrared laser absorption spectroscopy (LAS) based on simultaneous static (GC) and dynamic (LAS) chamber
500 measurements and derived the following conclusions:

- The measurement days proved to be suitable for the comparison, as they covered ~~a wide flux~~ range of fluxes
across the investigated gases (CO₂, CH₄, and N₂O) regarding arable soils ~~from well-aerated upland soils~~.
- For CO₂ fluxes, a very high level of agreement between the methods was observed. All fluxes were well above
505 the minimum detectable flux (MDF), and deviations were low (nRMSE < 17 %), confirming the reliability of
both methods for measuring CO₂.
- For N₂O, the methods showed strong agreement across both measurement days, with high correlation coefficients
($r \sim 1$) and low deviations (nRMSE < 25 %). For low fluxes GC-derived fluxes fell below the MDF and could
not be distinguished from zero flux.
- In the case of CH₄, the agreement between methods was poor (nRMSE up to 95 %). Covering the common range
510 for ~~Despite very high~~ CH₄ oxidation rates, none of the GC measurements exceeded the MDF. In contrast, almost
all values measured by LAS exceeded the MDF, which likely contributed to the weak correlation between the
two methods CH₄ fluxes.
- A central advantage of the LAS method lies in its considerably lower MDF ~~due to~~ in combination with the higher
515 measurement frequency (~1 Hz) and analytical precision, which enables the detection of statistically significant

fluxes (flux > MDF) even at a very low flux range. This is especially relevant for CH₄ and N₂O, where small flux rates occur and uptake of either gases into the soil-plant system might remain undetected.

- Due to the considerably lower MDF of the LAS method, shorter chamber enclosure times are potentially possible, even at small flux rates. Our results confirmed this advantage, which showed that it is possible to reduce the chamber enclosure time to 10 minutes while obtaining similar results.
- The ability to visualize and validate measurements in real time adds to the practical benefits of LAS, offering increased flexibility in field campaigns. However, GC systems offer clear advantages, such as the ability to conduct multiple measurement campaigns in parallel. ~~high acquisition costs and sensitivity to dust and moisture remain important limitations that must be considered when choosing the appropriate measurement technology.~~
- ~~Chamber closure time influenced the variability of LAS derived fluxes, with shorter durations leading to higher measurement uncertainty. However, no statistically significant differences were observed in the absolute flux values between the different closure times of 30, 20 and 10 minutes. This confirms that shorter chamber closure times are suitable for field measurements with LAS, offering more flexibility without compromising result validity.~~

Overall, this pilot study provides a comprehensive comparison of two widely used analytical techniques for chamber-based GHG flux measurements. Given the high agreement observed between methods for CO₂ and N₂O, we conclude that LAS is a valid and reliable alternative to the established GC approach for these gases. For quantifying CH₄ uptake rates in ~~arable~~ soils, there is considerable uncertainty regarding the consistency between the two methods; however, this is very likely due to the high MDF of the GC method. In conclusion, the LAS's analytical sensitivity and operational flexibility clearly supports its broader application in trace gas research, especially for gaining new insights into the natural variability of low soil GHG fluxes which are masked by instrumental noise of world-wide very common GC technique. ~~the traditional closed static chamber method.~~

Appendix A: Details of standard gases for gas chromatography measurements

Table A1: List of (non-isotope) standards used for testing and calibration¹.

Standard gas (SG)	CO ₂		CH ₄		N ₂ O		O ₂	
	Conc. (ppm)	Relative error (±%)	Conc. (ppm)	Relative error (±%)	Conc. (ppb)	Relative error (±%)	Conc. (vol%)	Relative error (±%)
SG1	304	1	1.02	2	248.4	3	19.01	0.5
SG2	402.3	0.5	1.81	2	321.3	3	20.97	0.5
SG3	1509.2	0.5	5.02	2	2010	3	15	0.5
SG4	3999.6	0.5	20.9	2	15100	2	10	0.5

550 ¹According to analysis certificate of the manufacturer (*DEUSTE Steininger GmbH, DE*).

Data availability

The processed data that support our findings are included in the supplementary material. For more detailed information, the input files for the flux calculations as well as the output from the calculation scripts are openly available in the repository Zenodo at <https://doi.org/10.5281/zenodo.15674498> (Aumer et al., 2025).

Competing interests

We have no competing interests to declare.

Author contributions

560 WA and MM contributed equally and took lead in writing the manuscript and analysing the data. CE provided GC raw data, CMG and CK were in charge of conceptualization. CBi provided weather data. AG, CBi, CBr, CE, CK, CMG, MRF, TKDW supported us in writing the manuscript and all authors approved the final version.

Acknowledgements

The long-term experiment (AKHWA project, URL: www.akhwa.de) and the laser gas analyzers were funded by the Hessian Ministry for Agriculture, Environment, Viticulture, Forestry, Hunting and Homeland Affairs. Furthermore, we acknowledge

565 the valuable contributions of all field technicians and student assistants involved in the long-term field experiment where this
investigation took place.

References

- Aumer, W., Möller, M., and Eckhardt, C.: Closed chamber flux dataset comparing gas chromatography and mid-infrared
laser absorption spectroscopy for greenhouse gas measurements, Zenodo [Data Set], doi:10.5281/zenodo.15674498,
570 2025.
- Bilibio, C., Weber, T. K. D., Hammer-Weis, M., Junge, S. M., Leisch-Waskoenig, S., Wack, J., Niether, W., Gattinger, A.,
Finckh, M. R., and Peth, S.: Changes in soil mechanical and hydraulic properties through regenerative cultivation
measures in long-term and farm experiments in Germany, *Soil and Tillage Research*, 246, 106345,
doi:10.1016/j.still.2024.106345, 2025.
- 575 [Bond-Lamberty, B., Christianson, D. S., Malhotra, A., Pennington, S. C., Sihi, D., AghaKouchak, A., Anjileli, H., Altaf
Arain, M., Armesto, J. J., Ashraf, S., Ataka, M., Baldocchi, D., Andrew Black, T., Buchmann, N., Carbone, M. S.,
Chang, S.-C., Crill, P., Curtis, P. S., Davidson, E. A., Desai, A. R., Drake, J. E., El-Madany, T. S., Gavazzi, M., Görres,
C.-M., Gough, C. M., Goulden, M., Gregg, J., Del Gutiérrez Arroyo, O., He, J.-S., Hirano, T., Hopple, A., Hughes, H.,
Järveoja, J., Jassal, R., Jian, J., Kan, H., Kaye, J., Kominami, Y., Liang, N., Lipson, D., Macdonald, C. A., Maseyk, K.,
580 \[Mathes, K., Mauritz, M., Mayes, M. A., McNulty, S., Miao, G., Migliavacca, M., Miller, S., Miniati, C. F., Nietz, J. G.,
Nilsson, M. B., Noormets, A., Norouzi, H., O'Connell, C. S., Osborne, B., Oyonarte, C., Pang, Z., Peichl, M., Pendall,
E., Perez-Quezada, J. F., Phillips, C. L., Phillips, R. P., Raich, J. W., Renchon, A. A., Ruehr, N. K., Sánchez-Cañete, E.
P., Saunders, M., Savage, K. E., Schrumph, M., Scott, R. L., Seibt, U., Silver, W. L., Sun, W., Szutu, D., Takagi, K.,
Takagi, M., Teramoto, M., Tjoelker, M. G., Trumbore, S., Ueyama, M., Vargas, R., Varner, R. K., Verfaillie, J., Vogel,
585 \\[C., Wang, J., Winston, G., Wood, T. E., Wu, J., Wutzler, T., Zeng, J., Zha, T., Zhang, Q., and Zou, J.: COSORE: A
community database for continuous soil respiration and other soil-atmosphere greenhouse gas flux data, *Global change
biology*, 26, 7268–7283, doi:10.1111/gcb.15353, 2020.\\]\\(#\\)\]\(#\)](#)
- Brümmer, C., Lyshede, B., Lempio, D., Delorme, J.-P., Rüffer, J. J., Fuß, R., Moffat, A. M., Hurkuck, M., Ibrom, A.,
Ambus, P., Flessa, H., and Kutsch, W. L.: Gas chromatography vs. quantum cascade laser-based N₂O flux
590 measurements using a novel chamber design, *Biogeosciences*, 14, 1365–1381, doi:10.5194/bg-14-1365-2017, 2017.
- Butterbach-Bahl, K., Willibald, G., and Papen, H.: Soil core method for direct simultaneous determination of N₂ and N₂O
emissions from forest soils, *Plant Soil*, 240, 105–116, doi:10.1023/A:1015870518723, 2002.
- Cavigelli, M. and Robertson, G.: Role of denitrifier diversity in rates of nitrous oxide consumption in a terrestrial ecosystem,
Soil Biology and Biochemistry, 33, 297–310, doi:10.1016/S0038-0717(00)00141-3, 2001.
- 595 Chapuis-Lardy, L., Wrage, N., Metay, A., Chotte, J.-L., and Berenoux, M.: Soils, a sink for N₂O? A review, *Global change
biology*, 13, 1–17, doi:10.1111/j.1365-2486.2006.01280.x, 2007.

- Christiansen, J. R., Outhwaite, J., and Smukler, S. M.: Comparison of CO₂, CH₄ and N₂O soil-atmosphere exchange measured in static chambers with cavity ring-down spectroscopy and gas chromatography, *Agricultural and Forest Meteorology*, 211-212, 48–57, doi:10.1016/j.agrformet.2015.06.004, 2015a.
- 600 Christiansen, J. R., Romero, A. J. B., Jørgensen, N. O. G., Glaring, M. A., Jørgensen, C. J., Berg, L. K., and Elberling, B.: Methane fluxes and the functional groups of methanotrophs and methanogens in a young Arctic landscape on Disko Island, West Greenland, *Biogeochemistry*, 122, 15–33, doi:10.1007/s10533-014-0026-7, 2015b.
- Cowan, N., Levy, P., Tigli, M., Toteva, G., and Drewer, J.: Characterisation of Analytical Uncertainty in Chamber Soil Flux Measurements, *European J Soil Science*, 76, doi:10.1111/ejss.70104, 2025.
- 605 [Dix, B. A., Hauschild, M. E., Niether, W., Wolf, B., and Gattinger, A.: Regulating soil microclimate and greenhouse gas emissions with rye mulch in cabbage cultivation. *Agriculture, ecosystems & environment*, 367, 108951, doi:10.1016/j.agee.2024.108951, 2024.](#)
- Flessa, H., Ruser, R., Schilling, R., Lofthfield, N., Munch, J., Kaiser, E., and Beese, F.: N₂O and CH₄ fluxes in potato fields: automated measurement, management effects and temporal variation, *Geoderma*, 105, 307–325, doi:10.1016/S0016-7061(01)00110-0, 2002.
- 610
- Fuss, R. and Hüppi, R.: gasfluxes (Version 0.4-4): Greenhouse Gas Flux Calculation from Chamber Measurements, doi:10.32614/CRAN.package.gasfluxes, 2024.
- Glatzel, S. and Stahr, K.: Methane and nitrous oxide exchange in differently fertilised grassland in southern Germany, *Plant Soil*, 231, 21–35, doi:10.1023/A:1010315416866, 2001.
- 615 Hüppi, R., Felber, R., Krauss, M., Six, J., Leifeld, J., and Fuß, R.: Restricting the nonlinearity parameter in soil greenhouse gas flux calculation for more reliable flux estimates, *PloS one*, 13, e0200876, doi:10.1371/journal.pone.0200876, 2018.
- Hutchinson, G. L. and Livingston, G. P.: Vents and seals in non-steady-state chambers used for measuring gas exchange between soil and the atmosphere, *European J Soil Science*, 52, 675–682, doi:10.1046/j.1365-2389.2001.00415.x, 2001.
- Hutchinson, G. L. and Mosier, A. R.: Improved Soil Cover Method for Field Measurement of Nitrous Oxide Fluxes, *Soil Science Society of America Journal*, 45, 311–316, doi:10.2136/sssaj1981.03615995004500020017x, 1981.
- 620 Hütsch, B. W.: Tillage and land use effects on methane oxidation rates and their vertical profiles in soil, *Biology and Fertility of Soils*, 27, 284–292, doi:10.1007/s003740050435, 1998.
- Johannesson, C.-F., Nordén, J., Lange, H., Silvennoinen, H., and Larsen, K. S.: Optimizing the closure period for improved accuracy of chamber-based greenhouse gas flux estimates, *Agricultural and Forest Meteorology*, 359, 110289, doi:10.1016/j.agrformet.2024.110289, 2024.
- 625
- Kong, M., Mitu, F. F., Petersen, S. O., Lærke, P. E., Abalos, D., Sørensen, P., Brændholt, A., Bruun, S., Eriksen, J., and Dold, C.: A comparison of chamber-based methods for measuring N₂O emissions from arable soils, *Agricultural and Forest Meteorology*, 370, 110591, doi:10.1016/j.agrformet.2025.110591, 2025.
- Kottek, M., Grieser, J., Beck, C., Rudolf, B., and Rubel, F.: World Map of the Köppen-Geiger climate classification updated, *metz*, 15, 259–263, doi:10.1127/0941-2948/2006/0130, 2006.
- 630

- Krauss, M., Ruser, R., Müller, T., Hansen, S., Mäder, P., and Gättinger, A.: Impact of reduced tillage on greenhouse gas emissions and soil carbon stocks in an organic grass-clover ley - winter wheat cropping sequence, *Agriculture, ecosystems & environment*, 239, 324–333, doi:10.1016/j.agee.2017.01.029, 2017.
- 635 Le Mer, J. and Roger, P.: Production, oxidation, emission and consumption of methane by soils: A review, *European Journal of Soil Biology*, 37, 25–50, doi:10.1016/S1164-5563(01)01067-6, 2001.
- Leisch, S., Bilibio, C., Weber, T., and Finckh, M. R.: Focus Area - Site Neu Eichenberg Research Farm [Data set], Leibniz Centre for Agricultural Landscape Research (ZALF), doi:10.4228/ZALF-R29F-BA86, 2025.
- LI-COR: The Importance of Water Vapor Measurements and Corrections, 2023.
- 640 Livingston, G. P. and Hutchinson, G. L.: Enclosure-based measurement of trace gas exchange: applications and sources of error.: In: Matson, P.A. and Harris, R.C., Eds., *Biogenic trace gases: measuring emissions from soil and water*. Blackwell Science Ltd., Oxford, UK., 1995.
- Lundegårdh, H.: Carbon dioxide evolution of soil and crop growth, *Soil Science*, 23, 417–453, doi:10.1097/00010694-192706000-00001, 1927.
- 645 Maier, M., Weber, T. K. D., Fiedler, J., Fuß, R., Glatzel, S., Huth, V., Jordan, S., Jurasinski, G., Kutzbach, L., Schäfer, K., Weymann, D., and Hagemann, U.: Introduction of a guideline for measurements of greenhouse gas fluxes from soils using non-steady-state chambers, *J. Plant Nutr. Soil Sci.*, 185, 447–461, doi:10.1002/jpln.202200199, 2022.
- Nickerson, N. R.: Evaluating gas emission measurements using Minimum Detectable Flux (MDF) ([White Paper](#)), doi: 10.13140/RG.2.1.4149.2089, 2016.
- 650 Obalum, S. E., Uteau-Puschmann, D., and Peth, S.: Reduced tillage and compost effects on soil aggregate stability of a silt-loam Luvisol using different aggregate stability tests, *Soil and Tillage Research*, 189, 217–228, doi:10.1016/j.still.2019.02.002, 2019.
- [Pavelka, M., Acosta, M., Kiese, R., Altimir, N., Brümmer, C., Crill, P., Darenova, E., Fuß, R., Gielen, B., Graf, A., Klemetsson, L., Lohila, A., Longdoz, B., Lindroth, A., Nilsson, M., Jiménez, S. M., Merbold, L., Montagnani, L., Peichl, M., Pihlatie, M., Pumpanen, J., Ortiz, P. S., Silvennoinen, H., Skiba, U., Vestin, P., Weslien, P., Janous, D., and Kutsch, W.: Standardisation of chamber technique for CO₂, N₂O and CH₄ fluxes measurements from terrestrial ecosystems. *International Agrophysics*, 32, 569–587, doi:10.1515/intag-2017-0045, 2018.](#)
- 655 Powlson, D. S., Stirling, C. M., Jat, M. L., Gerard, B. G., Palm, C. A., Sanchez, P. A., and Cassman, K. G.: Limited potential of no-till agriculture for climate change mitigation, *Nature Clim Change*, 4, 678–683, doi:10.1038/nclimate2292, 2014.
- Rheault, K., Christiansen, J. R., and Larsen, K. S.: goFlux: A user-friendly way to calculate GHG fluxes yourself, regardless of user experience, *JOSS*, 9, 6393, doi:10.21105/joss.06393, 2024.
- 660 Schmidt, J. H., Bergkvist, G., Campiglia, E., Radicetti, E., Wittwer, R. A., Finckh, M. R., and Hallmann, J.: Effect of tillage, subsidiary crops and fertilisation on plant-parasitic nematodes in a range of agro-environmental conditions within Europe, *Annals of Applied Biology*, 171, 477–489, doi:10.1111/aab.12389, 2017.

- 665 [Themistokleous, G., Savvides, A. M., Philippou, K., Ioannides, I. M., and Omirou, M.: A high-frequency greenhouse gas flux analysis tool: Insights from automated non-steady-state transparent soil chambers, *European J Soil Science*, 75, doi:10.1111/ejss.13560, 2024.](#)
- Triches, N. Y., Engel, J., Bolek, A., Vesala, T., Marushchak, M. E., Virkkala, A.-M., Heimann, M., and Göckede, M.: Advancing N₂O flux chamber measurement techniques in nutrient-poor ecosystems, *Atmos. Meas. Tech. Discuss.* [preprint], doi:10.5194/amt-2024-203, in review, 2025.
- 670 [Welles, J., Demetriades-Shah, T., and McDermitt, D.: Considerations for measuring ground CO₂ effluxes with chambers, *Chemical Geology*, 177, 3–13, doi:10.1016/S0009-2541\(00\)00388-0, 2001.](#)
- Yao, Z., Zheng, X., Xie, B., Liu, C., Mei, B., Dong, H., Butterbach-Bahl, K., and Zhu, J.: Comparison of manual and automated chambers for field measurements of N₂O, CH₄, CO₂ fluxes from cultivated land, *Atmospheric Environment*, 43, 1888–1896, doi:10.1016/j.atmosenv.2008.12.031, 2009.
- 675 Yu, C. and Yao, W.: Robust linear regression: A review and comparison, *Communications in Statistics - Simulation and Computation*, 46, 6261–6282, doi:10.1080/03610918.2016.1202271, 2017.
- [Zhang, Y., Mu, Y., Fang, S., and Liu, J.: An improved GC-ECD method for measuring atmospheric N₂O, *Journal of environmental sciences \(China\)*, 25, 547–553, doi:10.1016/S1001-0742\(12\)60090-4, 2013.](#)
- 680 Zheng, X., Mei, B., Wang, Y., Xie, B., Wang, Y., Dong, H., Xu, H., Chen, G., Cai, Z., Yue, J., Gu, J., Su, F., Zou, J., and Zhu, J.: Quantification of N₂O fluxes from soil–plant systems may be biased by the applied gas chromatograph methodology, *Plant Soil*, 311, 211–234, doi:10.1007/s11104-008-9673-6, 2008.

685