

In “Optimizing the precision of infrared measurements using the Eppley Laboratory, Inc. model PIR pyrgeometer”, Michalsky et al. assess the performance of four different equations that are used to convert raw voltage measurements of pyrgeometers into irradiances during transfer calibrations. The authors conclude that the method by Philipona et al. (1995) is the preferred option for this purpose. Additionally, some analysis on the conversion of thermistor-measured resistance to temperature and on the consistency among different calibration events are provided. The topic is well suitable for publication in ATM. However, I think, the study has more potential and should first receive a careful revision considering the comments below.

General comments

The introduction could be improved. First, a general motivation on why accurate thermal infrared radiation measurements are required is missing. Although this might be obvious, it might help putting the later results into context. **Paragraph added to introduction.** Second, please add a short explanation on the measurement principle of the pyrgeometers and why the instrument temperature needs to be corrected for. **Mostly unchanged except for emphasizing the need for careful receiver temperature measurements.** Thereby, especially readers less familiar with such instruments will better understand the components in Fig. 1 and Eq. (1) and the difference between T_B and T_R **Added notes in caption of Fig. 1.**

What were the meteorological conditions during the instrument calibration? Was the calibration performed according to WRC standards (calibration coefficient C from outdoor measurements in clear-sky night-time conditions and coefficients k_i from lab experiments)? L 166-167 only mentions the exclusion of outliers and periods of precipitation. Please provide more information on the measurement site and the conditions during calibration (temperature, humidity, clouds). **These are discussed in sections 3 and 5.** Perhaps, meteorological data could complement the irradiance time series shown in Fig. 3. I think, the study can largely benefit from such data. Regarding Fig. 7, this data could be used to more accurately filter for specific conditions, such as cloudiness, temperature or humidity regimes, and assess, how different conditions affect the transfer calibration of the pyrgeometers. I would look forward to an extended analysis on this problem in Sect. 4 as a second focus of the study. **See section 5.**

I would suggest testing the consistency of the standard PIRs' calibration (Fig. 6) prior to calibration transfer to get an estimate of the impact of potential instrument instabilities on the transfer calibration. Fig. 6 seems to reveal that PIR 32909 is less stable than the other ones, especially regarding the 2024 calibration. This probably causes the underestimation in measured irradiance that is visible in Fig. 3 also for Philipona's coefficients. However, due to the significantly stronger underestimation for Albrecht's coefficients, I assume that Fig. 6 would also show larger differences for Albrecht. Consider showing Fig. 6 for both Albrecht and Philipona in conjunction with Fig. 3 (see also more detailed comment on Fig. 3 below). The results might

imply a possible exaggeration of instrument instabilities by Albrecht, which would further substantiate the preference for Philipona's method.

I made plots like Fig. 6 for the Albrecht and Cox WRC calibrations; they are similar to boxplots of Fig. 6 using Philipona coefficients with somewhat less scatter. Not sure how to explain this.

Based on the general comments above, consider modifying the general structure of the manuscript: My suggestion would be the following:

Motivation: Different equations for conversion of raw signal (voltages) to irradiance, all depending on body and dome temperature (measured by thermistor)

Sect. 2: evaluate different methods to convert thermistor resistance to temperature → no significant impact on irradiance

Sect. 3: evaluate equations to convert voltages to irradiances

3 regularly calibrated "standard" PIRs: first assess instrument stability and consistency of calibration events (Fig. 6) → PIRs generally stable, but 32909 least stable, impact of instability larger for Albrecht → preference Philipona (see general comment 3)

Transfer calibration standard PIRs → test PIRs (Figs. 4, 5) → preference Philipona

Modified Sect. 4 "Impact of meteorological conditions on precision of transfer calibration" (see general comment 2)

Furthermore, I would like to encourage the authors to carefully revise the text. The wording frequently sounds too colloquial, is imprecise, or lacks clarity. Sometimes, the language could be more concise. Some comments are given below, but there is room for more improvement.

General structure was not modified. Other comments addressed elsewhere.

Specific comments

Title: Consider changing to something like "Optimizing the precision of infrared radiation measurements by Eppley PIR pyrgeometers"

I'll keep what I have.

L 16-17: Is the cosine response important for the study? Rather mention more important instrument characteristics, such as the thermistors that are used for temperature measurements and referred to in L 20.

The cosine response point emphasizes that the PIR is intended to make hemispherical radiation measurement ala pyranometers.

L 18: This sentence lacks clarity. Do “standards” and “field units” refer to the standard pyrgeometers calibrated at WRC and the physical unit of irradiance, respectively? Please clarify.

Change made to as requested.

L 87-88: Please justify why k_0 is dropped.

There is no physical reason to have this term. Text added to make this point. BTW, the most extensive use of the Reda et al. equation is in the ARM program, which does not include it.

L 91-92: Avoid the term Steinhart-Hart equation here since it was not defined yet. Better “various versions of the Steinhart-Hart equation that have been used to convert ...” → “various methods used to convert ...”.

Changed.

L 93: “three test PIRs” – later, the text mentions 6 instruments. Please clarify.

Changed.

Eq. (7): I think, the notation of the exponentiated logarithms is improper. I suggest using either $(\ln R)^2$ or $\ln^2(R)$. Furthermore, I would suggest including the quadratic term with coefficient c in Eq. (7) and set to 0 when necessary.

$(\ln R)^2$ used now, and c coefficient changed to respond to another referee.

L 117: Is my understanding right that the regression (Eq. 7) to derive the coefficients a , b , (c) , d is based on tabulated data of resistance and temperature provided by the manufacturer? At least, this statement is made in the conclusions section. If so, what is the benefit of performing the fit instead of simply interpolating between the tabulated values? Please add important information and be more accurate in describing what was done here and why.

The Steinhart-Hart equation is typically used to interpolate between the manufacturer’s tabulated values. This section simply looks at how well other forms or sets of coefficients for this equation perform in fitting to these tabulated values. I suspect that the fits discussed in Fig.2 may indicate slight inaccuracies in the manufacturer’s data. Many changes based on the other referee’s comments hopefully have made this discussion clearer.

L 118-123: Consider altering the argumentation here and sticking to ohms when including the quadratic term into the regression. Although kilohms could be skipped completely from this analysis, it might be worth mentioning the difference when the quadratic term is omitted in the following sentence. I like the idea of drawing attention to this discrepancy arising from the choice of the unit here and explain it mathematically. However, since this explanation is not the main scope of the study, I agree that the appendix is the right place to do that.

This was clarified (I think) by changing text based on the other referee’s comments.

L 125: Is the reference Gröber (2025) accessible to the public? It would be good to know which coefficients are used for the corresponding fit. Maybe, a table listing the coefficients for all fits considered in Fig. 2 can be added. **This table added.**

	a =	b =	c =	d =
Red (ohms)	0.001029607	0.0002390769	0.0	1.567609E-07
White (ohms)	0.001020630	0.0002416721	-2.47485E-07	1.64547E-07
Blue (kilohms)	0.002732470	0.0002618082	3.162474E-06	1.645474E-07
Gray (ohms)	0.0010295 Better 0.0010293	0.0002391	0.0	1.568E-07
Magenta ((ohms)	0.00102972	0.00023906	0.0	1.5677E-07

L 126: "... but does well over the entire range." Please specify the temperature range referred to at the beginning of the section. Furthermore, "entire range" is imprecise here and should be changed to something like "entire temperature range analyzed here".

Fixed.

L 129-130: Could you justify your choice quantitatively? What does the temperature deviation mean in terms of irradiance?

Text added to summarize effect of any of these temperature estimates.

L 148: Was the calibration of the field/test PIRs only done for period shown in Fig. 3? According to the text, Fig. 3 is just an example. Please be more accurate in describing the calibration procedure and mention, which instrument was calibrated when.

Text added to make this clear.

Fig. 3: To put the irradiance measurements into context with the concurrent meteorological conditions, my suggestion is to add time series of relevant meteorological quantities to the time series of the irradiances, if available. Furthermore, only show the mean of the three instruments in the time series. The difference between the individual instruments (e.g. with respect to the mean) could be plotted in an additional histogram (or boxplots), from which the magnitude of potential over- or underestimation can be better quantified.

Histogram added to clearly make the points in the reviewer's comments.

L 163: How was "the mean IR irradiance of the three standards" obtained? How were the three WRC calibration (2018, 2022, 2024) combined to calculate the standards' irradiance? Please provide more information.

The first question was dealt with in response to the first reviewer. Text was added to clarify. If the second comment refers to Fig. 6 then same comment applies. We did not combine all three WRC calibration results to get a combined standard if that is what is meant.

L 188-189: I think, it's fine to only plot the results of two instruments. However, maybe summarize statistics for all instruments in a table and mention some key numbers in the text to corroborate the conclusions.

If one assumes that there are no significant differences in the calculation of infrared irradiances using the Philipona et al. (1995) formula versus each of the other three methods discussed in this paper, this assumption is rejected with 95% confidence in 15 of the 18 cases studied (six calibrated PIRs and three formulae). The three cases where the null hypothesis cannot be rejected with 95% confidence are for three of the six PIRs using the Reda et al. (2002) formula. Changes have been made to the text.

L 218-219: Is it fair to argue that the results are better or worse in Fig. 5 compared to Fig. 4? I mean, you are comparing to different references (standards calculated with Albrecht vs. Philipona) but none of these references really represents the truth. I think the comparison only shows that calibrating the test PIRs with Albrecht's equation can fit better to either the Albrecht- or the Philipona-calibrated standard and that the consistent application of Albrecht's method does not assure a minimized spread.

I just removed this sentence.

Fig. 7: How consistent are the results if you randomly split the data into equally sized subsets for calibration and validation?

We divided all six data sets and found results consistent with Fig. 7 except in one case where results were more consistent with using the entire data set when using either half of data to calibrate and test. Note that the 2024 calibration run was generally noisier than the 2023 run.

Appendix: Equation numbers 8–14 in the text → A1–A7

Fixed.

Text improvements

throughout: the use of written-out “infrared” and abbreviated “IR” is inconsistent. I suggest using the acronym “TIR” for thermal infrared, since the infrared also covers parts of the solar spectrum where the pyrgeometer is not sensitive.

Change made.

throughout: “degrees K” → “Kelvin”

Change made.

throughout: “standards” → “standard PIRs”

Change made.

L 14: “The Eppley Model PIR is widely used ...” → “The Eppley Precision Infrared Radiometer (PIR) is a widely used pyrgeometer ...”

Change made.

L 17: “equations in the literature” → “equations suggested by the literature”

Change made.

L19-20: “... used to convert the resistance of the YSI 44021 thermistors used in PIRs for temperature measurements ...” → “... used to convert the resistance measurements of the thermistors to temperatures ...”

Change made.

L 20, 31: skip “aka case”

Change made.

L 29: “The Eppley model PIR pyrgeometer was developed ...” → “The Eppley Precision Infrared Radiometer (PIR) is a pyrgeometer developed ...”

Change made.

L 29: skip “longwave”

Change made.

L 30: “sky” → “atmosphere”

Change made.

L 46: skip “L is the external incoming infrared irradiance” since L was defined above

Change made.

Caption Fig. 1: “rays” → “radiation components on the thermopile surface” and “incoming infrared” → “incident atmospheric TIR irradiance”.

Change made.

L 112: “... where T is in degrees K and R is in ohms or kilohms” → “... where T is the temperature in Kelvin and R is the measured resistance in Ohms or Kiloohms.”

Change made.

L 118: “The least-squares fit to Eq. (7) ... is the red line ...” → “The least-squares fit to Eq. (7) ... is indicated by the red line ...”

Change made.

L 118-119 and 129-130: “full cubic” → “full cubic relationship”

Change made.

L 139: “we apply Eq. (2), (3), (4), and (6) to examine how well each performs ...” → “we examine the performance of Eqs. (2), (3), (4), and (6) ...”

Change made.

L 141: “Our three PIRs ...” → “The three standard PIRs ...”

Change made.

L 141-142 and throughout: be consistent with terms “PMOD” and “WRC”.

Change made.

L 229-235: Most information in this paragraph is redundant, because it is already known. Suggestion to simplify: “The regular calibration of the standard PIRs at the WRC leads to different calibration results. Here, the consistency and repeatability of those calibration events is assessed.” Also consider changing the structure (see general comments).

Change made.

L 306-307: “The three standard PIRs ... are sent biennially to be calibrated ...” → “The three standard PIRs ... are biennially calibrated ...”

Change made.

L 328: “In this paper, the World Infrared Standard Group (WISG) ...” → “The WISG ...”

Change made.

L 345-348: “The source of the difference ...” → “This difference is due to numerical reasons, which are explained in the following.”

Change made.