

Operational performance of the Vaisala CL61 ceilometer for atmospheric profiling

Viet Le¹, Ewan J. O'Connor¹, Maria Filioglou², and Ville Vakkari^{1,3}

¹Finnish Meteorological Institute, Helsinki, 00101, Finland

²Finnish Meteorological Institute, Atmospheric Research Centre of Eastern Finland, Kuopio, Finland

³Atmospheric Chemistry Research Group, Chemical Resource Beneficiation, North-West University, Potchefstroom, 2520, South Africa

Correspondence: Viet Le (viet.le@fmi.fi)

Abstract. The Vaisala CL61 is a new generation elastic backscatter lidar that extends conventional Automatic Lidar-Ceilometer capabilities by providing depolarization ratio measurements. Reliable use of these measurements, however, requires thorough evaluation and characterisation of the instrument performance and subsequent corrections applied. In this study, performance of multiple CL61 instruments across different sites over 3 years period have been assessed. Results indicate some differences between instruments, with most of these early production units exhibiting a pronounced decrease in laser power over time, accompanied by an increase in background noise likely due to weaker return signal. Normally, the instrument scales the internal calibration factor to compensate for changes in laser power and thus provide consistent attenuated backscatter coefficient values from profile to profile over time. However, for the instrument at the Lindenberg site, by performing manual calibration with atmospheric targets it is noted that once the laser power dropped below 40% there is no further compensation in the internal calibration factor.

The instrumental background noise and bias, characterized using the termination hood, are found to vary with temperature. A method has been developed for correcting the instrumental bias and for estimating the associated uncertainty. Additionally, an approach to estimate the uncertainty of volume and particle backscatter and linear depolarization ratio is presented. In a case study representing low aerosol load ~~condition~~ conditions in Finland, ~~the bias-corrected particle-correcting for the instrumental bias changes the volume~~ linear depolarization ratio ~~has been found to deviate~~ by up to 0.1 ~~from the original~~ 0.005 relative to the instrument-provided values. On the other hand, the difference between volume linear depolarization ratio ~~. This demonstrates the importance of~~ and particle linear depolarization ratio reached up to 0.1. These findings demonstrate that the CL61 firmware effectively removes the majority of the mean background bias, whereas accounting for the molecular contribution ~~in both volume backscatter and linear depolarization ratio when quantitatively interpreting aerosol measurements by the~~ is important for quantitative interpretations of aerosol measurements at CL61 ~~Automatic Lidar-Ceilometer operating at the~~ 's wavelength of 910.55 nm. Finally, signal loss in one unit has been traced to fogging of the inside surface of the window, and attributed to insufficient internal heating linked to the instrument's firmware behavior.

1 Introduction

Automatic Lidar-Ceilometer (ALC) are ground-based elastic backscatter lidars originally developed for the automated detection of cloud base height to support aviation and meteorological operations. Over the past two decades, advancements in both hardware and data processing have greatly enhanced their capabilities (Cimini et al., 2020). Notably, improvements in signal quality, particularly in signal-to-noise ratio, now enable ALC to provide profiles of attenuated backscatter coefficient with sufficient dynamic range to permit the observation of a wide range of atmospheric phenomena, including rainfall (e.g. Rocadenbosch et al., 2020; Roschke et al., 2025), fog (e.g. Haeffelin et al., 2016; Ribaud et al., 2021), icing (e.g. Hämäläinen et al., 2020), and even the quantitative estimation of aerosol concentrations (e.g. Shang et al., 2021; Filioglou et al., 2025). ALC data are also increasingly used to estimate atmospheric boundary layer height (e.g. Kotthaus et al., 2023), which is a critical parameter for numerical weather prediction and air quality models (e.g. Uzan et al., 2020; Barragán et al., 2023). Their ability to operate autonomously and reliably in challenging environmental conditions, as well as being eye-safe, makes them well-suited for long-term monitoring and integration into operational networks of national meteorological services and research institutions (e.g. Hirsikko et al., 2014). Many such organizations are actively investigating the use of ALC profile data to improve forecast model performance (e.g. Warren et al., 2018; Illingworth et al., 2019).

Vaisala has recently introduced the CL61 ceilometer which is an ALC with the capability of measuring the volume linear depolarization ratio, δ_v , which is the ratio of the perpendicular to the parallel component of the backscattered signal relative to the emitted polarization. This ratio provides insight into particle sphericity (e.g. Burton et al., 2012; Mamouri and Ansmann, 2016; Baars et al., 2017) and is essential for distinguishing various aerosol types (e.g. Illingworth et al., 2015; Nicolae et al., 2018; Floutsi et al., 2023; Filioglou et al., 2023; Le et al., 2024) such as pollen, smoke, dust, marine, and volcanic ash. It is also utilized in determining cloud phase (Sassen, 1991) and in the retrieval of cloud microphysical properties (Donovan et al., 2015).

To fully leverage the broad range of applications enabled by the CL61, including its new capability to measure δ_v , careful quality control of the backscattered signal from both polarizations is essential. Environmental conditions, such as temperature and relative humidity, can influence lidar performance (Campbell et al., 2002), making it necessary to assess their impact on data quality. For instance, Hervo et al. (2016) identified and corrected a temperature-dependent effect on the overlap function of a Lufft CHM15K ceilometer, which arises from temperature-induced changes in the laser as well as in the instrument's optical and electronic components. Similarly, the older Vaisala CL31 ceilometer has been known to have backscatter artifacts below approximately 70 m (Van der Kamp, 2008; Martucci et al., 2010; Tsaknakis et al., 2011), prompting many studies to omit these near-surface measurements. Identifying and correcting such artifacts in the CL61 would enable reliable observations extending down to heights that overlap with surface-based observations. This capability is especially valuable for studying near-surface meteorological phenomena such as fog, haze, and emissions from ground-level aerosol sources.

In this study, we present a long-term evaluation of the CL61 ceilometers conducted at four different ACTRIS cloud profiling sites in Finland and Germany. Our analysis focuses on the impact of laser power on signal quality, the temperature sensitivity of the instrumental background and methods for its correction. Additionally, we demonstrate our calculation of the uncertainty in

the volume and particle linear depolarization ratio (δ_v and δ_p respectively). Finally, we report instances of signal loss attributed to fogging of the inside surface of the window, observed in some of the CL61 units.

2 Instrument description

60 The CL61 ceilometer, manufactured by Vaisala, is a coaxial ALC system operating at a wavelength of 910.55 nm and equipped with depolarization measurement capability. Its key technical specifications are summarized in Table 1. In brief, the CL61 utilizes an InGaAs diode laser that emits linearly polarized pulses with an energy of 3.9 μ J at a repetition rate of 9.5 kHz. The instrument features a single-lens optical design and performs depolarization measurements using a single-receiver with an avalanche photodiode (APD) detector that switches between two polarising filters (one perpendicular and one parallel
65 to the polarisation of the transmitted pulses) in the same coaxial optical path. The CL61 alternates the acquisition of each polarization every 0.2 s, while using the same receiver module for both channels. Consequently, receiver sensitivity calibration is not required. The maximum detection range of CL61 is 15.4 km.

The overlap functions reported by the manufacturer are shown in Fig. S1. Full overlap between the transmitted laser beam and the receiver field of view is achieved at approximately 250 m range for the instrument at Kenttäröva. The instruments at
70 Vehmasmäki and Hyytiälä have different overlap functions, with approximately 95% overlap at 250 m range and complete overlap at approximately 550 m range. The overlap function is reported only in the new firmware ~~data and remains constant throughout the period. No overlap function was~~ (version 1.2.7), and no overlap function has been reported for the instrument at Lindenberg. Although the overlap function might change during operation, for example due to temperature variations (Hervo et al., 2016), the overlap correction applied to the data reported by the CL61 remains constant.

75 Data collected up to December 2024 from four CL61 deployed at four different ACTRIS Cloudnet sites were analyzed. Three of these instruments are situated in Finland: Hyytiälä, Vehmasmäki, and Kenttäröva, while the fourth is located in Lindenberg, Germany. The firmware versions of the instruments (i.e. 1.1.10 and 1.2.7), along with their respective periods of validity, are presented in Table 2. The data from this study were obtained from the ACTRIS Cloudnet data portal (Görsdorf et al., 2025).

The CL61 is equipped with an internal heater to stabilize the temperatures of the laser and optical components. It includes a
80 window heater and blower to maintain stable window conditions, as the window's transmission efficiency significantly affects the backscattered signal (see Sect. 4.3). The instrument's firmware continuously monitors and reports various housekeeping variables, such as internal temperature, laser temperature, window condition (calculated internally on a scale from 0% for a fully obstructed window to 100% for a clean window), and the status of the window blower heater. The quantity and type of variables reported may differ based on the firmware version.

85 3 Methods

The CL61 operates by emitting pulses of linear polarized laser light into the atmosphere and recording the backscattered signal. The received power per laser pulse is described by the lidar equation (Wandinger, 2005):

Table 1. Specifications of the CL61 (CL61 User Guide: <https://docs.vaisala.com/r/M212475EN-E/en-US>, last access: 21 July 2025).

Specification	Values
Laser wavelength	910.55 nm
Laser	InGaAs diode
Beam divergence	$\pm 0.2 \text{ mrad} \times \pm 0.35 \text{ mrad}$
Receiver field-of-view	$\pm 0.56 \text{ mrad}$
Measurement maximum range	15400 m
Reporting range resolution	4.8 m
Measurement interval	5 s
Pulse duration	160ns (at Full Width Half Maximum)
Pulse frequency	9.5 kHz
Maximum pulse energy	$4.7 \mu\text{J}$
Height of complete overlap	250 m (in Kenttäröva) and 550 m (in Vehmasmäki and Hyytiälä)

Table 2. Firmware version of the CL61 across all the site locations.

Site and coordinate	Time	Firmware version
Vehmasmäki 62.738°N, 27.543°E	2022-01-01 to 2022-06-14	not stated
	2022-06-15 to 2023-04-27	1.1.10
	2023-04-28 to 2024-12-31	1.2.7
Hyytiälä 61.844°N, 24.287°E	2022-01-01 to 2022-11-20	not stated
	2022-11-21 to 2023-11-22	1.1.10
	2023-11-23 to 2024-12-31	1.2.7
Kenttäröva 67.987°N, 24.243°E	2022-01-01 to 2022-06-20	not stated
	2022-06-21 to 2024-12-31	1.2.7
Lindenberg 52.208°N, 14.118°E	2024-03-01 to 2024-12-31	1.1.10

$$P(r) = \frac{1}{r^2} C_L O(r) \beta(r) e^{-2 \int_0^r \alpha(r') dr'} + B. \quad (1)$$

Here, $O(r)$ is the overlap function that is unique to each instrument and is provided by the manufacturer. The coefficients α and β represent atmospheric extinction and backscatter, respectively, B is the background signal, and C_L represents the lidar constant. This constant encapsulates the system-specific characteristics of the lidar, such as its receiver optics and laser properties. It is initially determined and provided by the manufacturer, but it may drift over time as the instrument ages and its performance changes. Although the internal firmware attempts to monitor and compensate for these changes, additional calibration, such as absolute calibration using stratocumulus clouds (O'Connor et al., 2004; Hopkin et al., 2019), is still necessary.

The instrument provides two main output signals, $ppol$ and $xpol$ in arbitrary unit [a.u.], which represent the estimated normalized and background-, range- and overlap-corrected parallel- and cross-polarized components of the attenuated backscattered signal ($^{\parallel} \beta'$ and $^{\perp} \beta'$ respectively). They can be defined as:

$$ppol = ^{\parallel} \frac{(P(r) - B_{\text{estimated}}) r^2}{C_L O(r)} = ^{\parallel} \beta', \quad (2)$$

$$xpol = ^{\perp} \frac{(P(r) - B_{\text{estimated}}) r^2}{C_L O(r)} = ^{\perp} \beta', \quad (3)$$

where $B_{\text{estimated}}$ is the internally estimated background signal by the instrument. The instrument also provides the total attenuated backscatter signal being the sum of the attenuated backscatter from all polarizations:

$$\beta' = ^{\parallel} \beta' + ^{\perp} \beta', \quad (4)$$

and the volume linear depolarization ratio being:

$$\delta_v = \frac{^{\perp} \beta'}{^{\parallel} \beta'}. \quad (5)$$

3.1 Residual background signal

The observed backscattered signal detected by a lidar consists of the true atmospheric signal and the background signal (Cao et al., 2013). The true atmospheric signal consists of backscattered contributions from particles (β_p) and molecular scattering (β_{mol}), whilst the background signal includes contributions from both atmospheric background signal $P_{\text{atmosphere}}$ (such as solar radiation, moonlight, or artificial light) and instrumental background $P_{\text{instrument}}$ signal (dark signal). Although the instrument already determines and performs the background correction internally, residual background components may still remain in the measured signal, such as the dark signal observed during termination hood measurements (see Sect 3.1.1). Therefore, we extend Eqs. 2 and 3 to account for the residual background signal (hereafter referred to as the background signal), B_{res} , that is not removed by the internal correction:

$$ppol = \parallel \frac{(P(r) - B_{\text{estimated}})r^2}{C_L O(r)} + \parallel \frac{B_{\text{res}}(r)r^2}{C_L O(r)} = \parallel \beta'_{\text{corrected}} + \parallel \frac{B_{\text{res}}(r)r^2}{C_L O(r)}, \quad (6)$$

$$115 \quad xpol = \perp \frac{(P(r) - B_{\text{estimated}})r^2}{C_L O(r)} + \perp \frac{B_{\text{res}}(r)r^2}{C_L O(r)} = \perp \beta'_{\text{corrected}} + \perp \frac{B_{\text{res}}(r)r^2}{C_L O(r)}. \quad (7)$$

To eliminate the range dependence from the background terms, we divide each expression by r^2 , yielding:

$$\frac{ppol}{r^2} = \frac{\parallel \beta'_{\text{corrected}}}{r^2} + \parallel \frac{B_{\text{res}}(r)}{C_L O(r)}, \quad (8)$$

$$\frac{xpol}{r^2} = \frac{\perp \beta'_{\text{corrected}}}{r^2} + \perp \frac{B_{\text{res}}(r)}{C_L O(r)}. \quad (9)$$

For simplicity, we define the background terms as:

$$120 \quad P_{\text{res}}(r) \equiv \frac{B_{\text{res}}(r)}{C_L O(r)}, \quad (10)$$

where $P_{\text{res}}(r) = P_{\text{atmosphere}}(r) + P_{\text{instrument}}(r)$.

To derive the $\beta'_{\text{corrected}}$, the background signals need to be quantified. Since the measured $ppol$ and $xpol$ signals include contributions from background signal in their respective polarization channels, it is necessary to estimate the background signal separately for each polarization ($\parallel P_{\text{res}}$ and $\perp P_{\text{res}}$). To achieve this, a background identification methodology has been developed. It identifies regions within the dataset that are free of aerosols and hydrometeors, where the $\frac{ppol}{r^2}$ and $\frac{xpol}{r^2}$ values are assumed to predominantly represent background signal.

The methodology proceeds as follows. First, the data is averaged over 5-minute intervals. The $\frac{ppol}{r^2}$ and $\frac{xpol}{r^2}$ profiles are then decomposed using the stationary wavelet transform (SWT) with the bior.1 wavelet, implemented via PyWavelets (Nason and Silverman, 1995; Lee et al., 2019). Subsequently, the background signal variance is reduced by applying a hard-threshold shrinkage function to the approximation and detail coefficients from levels 1 through 7, using minimax thresholding (Nason and Silverman, 1995). The profile is reconstructed via the inverse stationary wavelet transform. Finally, the background range gates are identified as regions with values below half of the previously computed minimax threshold. The mean and variance of the background are computed from the original, non-averaged data with the identified background range gates using 5-minute time intervals and 2 km range-bin intervals.

135 An example of this methodology applied to data in Kenttärova on 2024-03-28 is illustrated in Fig. 1. During daylight hours, solar radiation significantly increases the variance ($\sigma^2_{\parallel P_{\text{res}}}$ and $\sigma^2_{\perp P_{\text{res}}}$) of the background signal, resulting in a daytime variance much higher than that observed at night, as shown in panels (g) and (h). Above the aerosol and cloud layers, both the mean ($\mu_{\parallel P_{\text{res}}}$ and $\mu_{\perp P_{\text{res}}}$) and variance of the background signals in both polarizations remain relatively stable with range, or at least exhibit considerably smaller variations compared to the diurnal fluctuations.

140 By applying this methodology to the full dataset, we obtained the time series of P_{res} for all instruments in both polarizations. Comparing these background signals with other housekeeping parameters, such as laser power, enables us to evaluate the operational performance of each instrument. Moreover, since each background component is quantified independently, we can

correct the signal and estimate its uncertainty in both polarizations $ppol$ and $xpol$ and, consequently, in the derived linear depolarization ratio.




Figure 1. Time series from Kenttärova on 2024-03-28 showing: a) $ppol$, b) $xpol$, c) background $ppol$, and d) background $xpol$. Subplots e–h display the background signals at different range bins from 4000 – 6000 m, 6000 – 8000 m, 8000 – 10000 m, 10000 – 12000 m, 12000 – 14000 m, with color indicating the range bin: e) $\mu_{\parallel P_{res}}$, f) $\mu_{\perp P_{res}}$, g) $\sigma^2_{\parallel P_{res}}$, and h) $\sigma^2_{\perp P_{res}}$. These parameters are computed at a 10-second temporal resolution using all data within each range bin.



Figure 2. A termination hood (Vaisala CL61TERMHOOD) placed on top of the CL61 in Hyytiälä.

To obtain the instrumental background signal, a termination hood (Vaisala CL61TERMHOOD) was placed over the window of each ceilometer located in Finland. The hood was applied without modifying the instrument's internal operating parameters, ensuring that all measurements reflected normal operational conditions. The hood's material and design totally attenuate the outgoing laser beam, preventing any backscatter signal from reaching the detector. As a result, the measured signal can be attributed entirely to the instrumental background signal adjusted by the overlap correction. Although the overlap correction remains constant for each instrument, it was reverted to accurately interpret the instrumental background signal in the near range $P_{\text{instrument}}O(r)$.

Figure 3 presents example profiles of $ppol$ and $xpol$ both before and during the termination hood measurements. Before the termination hood was applied, the μ_{ppol} and μ_{xpol} profiles in the aerosol and hydrometeor-free region above 5 km closely follow the theoretical molecular attenuated backscatter coefficient β'_{mol} (Fig. 3b and c), indicating that the CL61 is sensitive to molecular scattering. The β'_{mol} profile was calculated following the method described by Bucholtz (1995) using meteorological input data from a numerical weather prediction model available from the ACTRIS Cloudnet data portal (O'Connor, 2025). At ranges above 10 km, however, the molecular signal becomes indistinguishable from noise. The variance profiles, σ^2_{ppol}/r^2 and σ^2_{xpol}/r^2 , also remain relatively constant in this region (Fig. 3f and g).

During the termination hood measurement, $\mu_{P_{\text{instrument}}O(r)}$ from both polarizations remains relatively stable above 1.2 km, but increases significantly below this range (Figs. 3d, e). This increase suggests the presence of a signal bias at lower ranges that has not been corrected by the firmware and will be investigated in this study. Additionally, a reduction in the signal variance of both polarization channels is observed during the termination hood measurement (Figs. 3h, i) compared to the previous period




Figure 3. Measurements at Kenttäröva on 5 March 2024. a) the time series of $ppol$. Panels b, c, f, g, display the averaged profiles in time at each range gate before the termination hood measurement (from 11 to 13 UTC): b) μ_{ppol} and $\parallel \beta'_{mol}$, c) μ_{xpol} and $\perp \beta'_{mol}$, f) σ_{ppol/r^2}^2 , g) σ_{xpol/r^2}^2 . Panels d, e, h, i, display the averaged profiles in time at each range gate during the termination hood measurement at 19 °C (from 14 to 15 UTC): d) $\mu_{\parallel P_{instrument} O(r)}$, e) $\mu_{\perp P_{instrument} O(r)}$, h) $\sigma_{\parallel P_{instrument} O(r)}^2$, i) $\sigma_{\perp P_{instrument} O(r)}^2$.

above the complete overlap range. This reduction is attributed to the absence of solar influence during the termination hood measurement.

In principle, the nighttime measurements could be used to estimate $P_{instrument}$. However, only the portion of the $P_{instrument}$ profile at higher ranges, where clouds and aerosols are absent, can be reliably estimated. An example is shown in Fig. S2 in the Supplement, where σ_{ppol/r^2}^2 agrees well between the hood termination and nighttime measurements only at ranges above approximately 2 km. This agreement is observed only under clear-sky conditions, as cloud layers at lower ranges can increase the measured background values.

Since the CL61 is a coaxial lidar, it is prone to the afterpulsing effect, such as those caused by internal reflections of the outgoing laser beam that reach the detector (Campbell et al., 2002; Welton and Campbell, 2002). Hervo et al. (2016) demonstrated that the overlap function for a ceilometer made by another manufacturer (Lufft CHM15k) was sensitive to the instrument internal temperature, T , mainly due to temperature-induced changes in its optical components. It is likely that the observed bias is temperature-dependent and is caused by variations in instrumental background associated with afterpulsing, dark signal, and

optical components under different temperature conditions. To quantify the effect of temperature, the termination hood measurement was repeated over a range of internal temperatures T , reported as `internal_temperature` in the instrument housekeeping data, across a two-year period. Hence, the instrumental background signal mean (bias) and its variance (noise) from both polarizations with respect to internal temperature can be determined.

180 3.1.2 Atmospheric background signal

For an accurate estimation of the background signal, the contribution from the atmospheric background signal, $P_{\text{atmosphere}}$, must be taken into account. As noted by Kotthaus et al. (2016), earlier instruments such as the CL31 incorporate a zero-bias level that compensates for temporal fluctuations in solar radiation. This is also evident in CL61 as shown in Fig. 3b and 3c, where the μ_{ppol} and μ_{xpol} values above 5 km closely follow β_{mol} , confirming the zero-mean solar radiation noise.
 185 Therefore, the firmware must have corrected the atmospheric background signal to a zero mean, leaving only its variance (noise) contributing to the background signal, which will be addressed in this study.

The solar noise variance can be seen from the diurnal pattern in Fig. 1g and 1h; and the difference between the σ^2 profiles before (Fig. 3f and 3g) and during (Fig. 3h and 3i) the termination hood measurement. Since this difference remains approximately constant at all ranges above the aerosol layer near the ground, we assume that the atmospheric background variance
 190 from both polarizations, $\sigma^2_{P_{\text{atmosphere}}}$, is also uniform down to the near range. Then, $\sigma^2_{P_{\text{atmosphere}}}$ is calculated by computing the difference between the variance of the background signal $\sigma^2_{P_{\text{res}}}$ (obtained between 10 to 12 km using the background identification methodology) and the variance of the instrumental background signal $\sigma^2_{P_{\text{instrument}}}$ (obtained at the same range and temperature from the termination hood measurement):

$$\sigma^2_{P_{\text{atmosphere}}} = \sigma^2_{P_{\text{res}}(r=10-12km)} - \sigma^2_{P_{\text{instrument}}(r=10-12km,T)}. \quad (11)$$

195 Here, $\sigma^2_{P_{\text{atmosphere}}}$ is independent of the range and is obtained separately for each polarization in each individual profile.

3.1.3 Correction for systematic bias

After measuring $P_{\text{instrument}}(r, T)$ across a range of T during the termination hood measurements, the results are stored as a lookup table containing $P_{\text{instrument}}(r, T)$ profiles at each T . The attenuated backscatter coefficient for each polarization can be corrected by applying the $P_{\text{instrument}}(r, T)$ value corresponding to the instrument's current internal temperature as follows:

$$200 \quad \parallel \beta'_{\text{corrected}} = ppol - \mu_{\parallel P_{\text{instrument}}(r,T)} r^2, \quad (12)$$

$$\perp \beta'_{\text{corrected}} = xpol - \mu_{\perp P_{\text{instrument}}(r,T)} r^2, \quad (13)$$

and their corresponding uncertainties given by:

$$\sigma^2_{\parallel \beta'_{\text{corrected}}} = (\sigma^2_{\parallel P_{\text{atmosphere}}} + \sigma^2_{\parallel P_{\text{instrument}}(r,T)}) r^2, \quad (14)$$

$$\sigma^2_{\perp \beta'_{\text{corrected}}} = (\sigma^2_{\perp P_{\text{atmosphere}}} + \sigma^2_{\perp P_{\text{instrument}}(r,T)}) r^2. \quad (15)$$

205 The bias-corrected attenuated backscatter coefficient and volume linear depolarization ratio are then calculated as:

$$\beta'_{\text{corrected}} = {}^{\parallel} \beta'_{\text{corrected}} + {}^{\perp} \beta'_{\text{corrected}}, \quad (16)$$

$$\delta_{\text{corrected}} = \frac{{}^{\perp} \beta'_{\text{corrected}}}{{}^{\parallel} \beta'_{\text{corrected}}}, \quad (17)$$

and their associated uncertainty, given by

$$\sigma^2_{\beta'_{\text{corrected}}} = \sigma^2_{{}^{\parallel} \beta'_{\text{corrected}}} + \sigma^2_{{}^{\perp} \beta'_{\text{corrected}}}, \quad (18)$$

$$210 \quad \sigma^2_{\delta_{\text{corrected}}} = \delta^2 \left(\frac{\sigma^2_{{}^{\perp} \beta'_{\text{corrected}}}}{{}^{\perp} \beta'^2_{\text{corrected}}} + \frac{\sigma^2_{{}^{\parallel} \beta'_{\text{corrected}}}}{{}^{\parallel} \beta'^2_{\text{corrected}}} \right). \quad (19)$$

3.2 Calibration

To ensure the CL61 produces a consistent and accurate backscatter signal, the ceilometer signal must be manually calibrated in addition to its internal calibration. This involves calibrating the backscatter signal from the CL61 using a reference target with known backscatter characteristics. Two commonly used calibration methods for deriving C_L are based on different types
215 of reference targets: atmospheric molecules (Rayleigh calibration) and liquid clouds (liquid cloud calibration).

The Rayleigh calibration (Fernald et al., 1972; Klett, 1985; Binietoglou et al., 2011; Wiegner and Geiß, 2012; Baars et al., 2016), often referred to as the backward inversion approach, is the standard method to derive the particle backscatter coefficient (β_p) for most research aerosol lidars due to their sensitivity to molecular signal (Wiegner et al., 2014). As illustrated in Fig. 3b, the attenuated backscatter coefficient profile from CL61 closely follows the attenuated molecular backscatter coefficient above
220 the aerosol layer at 4 km. This indicates a significant contribution of molecular scattering to the total CL61 signal, especially in aerosol- and hydrometeor-free regions. Therefore, the Rayleigh calibration can be applied to the CL61, provided a sufficiently long averaging time (more than 2 hours) is used.

The Rayleigh calibration method requires an assumed lidar ratio for aerosol particles (S_p). At sites equipped with a sun photometer, S_p can be constrained (Wiegner and Geiß, 2012); otherwise, a value of 50 sr was used. The first ~~50- μ m~~ 50 m
225 range has been discarded due to unreliable data (see Sect. 4.5), and it has been demonstrated that the resulting loss in optical depth for ignoring near range gates is negligible (Wiegner and Geiß, 2012). After the β_p profile is derived, the lidar constant C_L can be determined from the lidar equation (Wiegner and Geiß, 2012).

The liquid cloud calibration (O'Connor et al., 2004; Hopkin et al., 2019) relies on the fact that the lidar ratio for liquid water clouds at the ceilometer wavelength is known. This method involves calculating the integrated attenuated backscatter coefficient
230 for fully attenuated liquid water clouds, including the contribution from multiple scattering. The theoretical contribution from multiple scattering is computed for droplet diameters ranging from 8 to 20 μ m and with the CL61 beam divergence and receiver field of view. The integrated backscatter coefficient is then scaled to fit within the expected theoretical values and this scaling factor is the calibration factor C , which is the reciprocal of the lidar constant C_L . The primary advantage of this approach over Rayleigh calibration is the substantially higher signal-to-noise ratio of the backscatter from water clouds compared to
235 the molecular backscatter. As a result, it eliminates the need for long nighttime averaging periods required to obtain a reliable molecular signal.

A methodology was developed to identify suitable liquid cloud profiles. First, a representative liquid cloud profile containing a single liquid cloud layer and exhibiting complete signal attenuation approximately 200 m above the cloud base was selected as the reference profile. Each *ppol* profile in the dataset was then cross-correlated with this reference to quantify their similarity.

240 For each time step, the height corresponding to the maximum cross-correlation value was identified as the most similar location and treated as a pseudo in-cloud height. Next, the ratio of total in-cloud β' , calculated over a 150 m layer centered at the pseudo in-cloud height, to the integrated β' of the entire profile was computed. A profile was classified as containing a liquid cloud if this ratio exceeded 90%, ensuring that the integrated β' was not significantly influenced by strong aerosol loading or precipitation. Additionally, the cross-correlation value at the identified height was required to exceed 5×10^{-7} . This threshold

245 was selected to ensure that at least 1000 valid data points were available for each month.

To ensure robust data fitting, the cloud calibration is performed monthly using all suitable liquid cloud observations detected from the methodology. To validate the cloud calibration results, Rayleigh calibration was also computed occasionally when possible. Data from a co-located sun photometer was used to constrain the lidar ratio for Rayleigh calibration. If no sun photometer data is available, a lidar ratio of 50 sr is used.

250 The calibrated volume attenuated backscatter coefficient and its uncertainty can then be determined as follows, assuming the cloud calibration factor has an uncertainty of approximately 10% (Hopkin et al., 2019):

$$\beta'_v = \beta'_{\text{corrected}} \cdot C, \quad (20)$$

$$\sigma^2_{\beta'_v} = (\sigma^2_{\beta'_{\text{corrected}}} + 0.01 \cdot \beta'^2_{\text{corrected}}) \cdot C^2. \quad (21)$$

For those instruments where termination hood measurements are not available:

255 $\beta'_v = \beta' \cdot C, \quad (22)$

$$\sigma^2_{\beta'_v} = (\sigma^2_{P_{\text{atmosphere}}} + 0.01 \cdot \beta'^2) \cdot C^2 = (\sigma^2_{\parallel P_{\text{atmosphere}}} + \sigma^2_{\perp P_{\text{atmosphere}}} + 0.01 \cdot \beta'^2) \cdot C^2. \quad (23)$$

The calibration is assumed to affect both polarization channels equally. Hence, $\delta_{\text{corrected}}$ remains unchanged.

3.3 Aerosol particle inversion

The particle backscatter coefficient profile (β_p) is retrieved using the forward Klett solution (Klett, 1985) after applying the

260 previously derived cloud calibration factor and assuming that the instrument does not undergo significant degradation within a one-month period. Similar to the previous section, the first 50 m range has been discarded. The cloud calibration factor is estimated to have an uncertainty of approximately 10% (Hopkin et al., 2019). The forward inversion does not require extensive temporal averaging. When available, sun photometer observations are used to constrain the lidar ratio S_p ; otherwise, a constant value of 50 sr is applied to all aerosol layers in the retrieval.

265 Let $\beta_p = f(\beta'_v, C, S_p)$ represent the inversion function; then the uncertainty of β_p is calculated as

$$\sigma^2_{\beta_p} = \left(\frac{\partial f}{\partial \beta'_v} \sigma_{\beta'_v} \right)^2 + \left(\frac{\partial f}{\partial C} \sigma_C \right)^2 + \left(\frac{\partial f}{\partial S} \sigma_{S_p} \right)^2. \quad (24)$$

The partial derivatives can be approximated using central differences:

$$\frac{\partial f}{\partial \beta'_v} \sigma_{\beta'_v} \approx \frac{f(\beta'_v + \sigma_{\beta'_v}, C, S_p) - f(\beta'_v - \sigma_{\beta'_v}, C, S_p)}{2}, \quad (25)$$

$$\frac{\partial f}{\partial C} \sigma_C \approx \frac{f(\beta'_v, C_{\max}, S_p) - f(\beta'_v, C_{\min}, S_p)}{2}, \quad (26)$$

$$270 \quad \frac{\partial f}{\partial S} \sigma_{S_p} \approx \frac{f(\beta'_v, C, S_{p,\max}) - f(\beta'_v, C, S_{p,\min})}{2}. \quad (27)$$

with $C_{\min} = 0.9 \cdot C$, $C_{\max} = 1.1 \cdot C$, $S_{p,\min} = S_p - 10$ and $S_{p,\max} = S_p + 10$

Following (Biele et al., 2000; Freudenthaler et al., 2009), the particle linear depolarization ratio (δ_p) is then obtained as

$$\delta_p = \frac{(1 + \delta_m)\delta_v R - (1 + \delta_v)\delta_m}{(1 + \delta_m)R - (1 + \delta_v)}, \quad (28)$$

where the backscatter ratio R is defined as

$$275 \quad R = \frac{\beta_p + \beta_m}{\beta_m}, \quad (29)$$

and δ_v is the volume linear depolarization ratio, which is $\delta_{\text{corrected}}$ in Eq. 17, or $\frac{x_{\text{pol}}}{p_{\text{pol}}}$ for instruments lacking termination hood measurements.

The molecular depolarization ratio (δ_m) is estimated following Tomasi et al. (2005), accounting for major atmospheric gases and the influence of water vapor. The input data for these calculations are taken from a numerical weather prediction model (ECMWF IFS forecast) provided via the ACTRIS Cloudnet data portal (O'Connor, 2024).

The uncertainty of δ_p can be determined using partial derivatives

$$\sigma_{\delta_p}^2 = \left(\frac{\partial \delta_p}{\partial \delta_m} \sigma_{\delta_m} \right)^2 + \left(\frac{\partial \delta_p}{\partial \delta_v} \sigma_{\delta_v} \right)^2 + \left(\frac{\partial \delta_p}{\partial R} \sigma_R \right)^2. \quad (30)$$

Assuming σ_{β_m} and σ_{δ_m} are negligible and approximating $\sigma_R \approx \frac{\sigma_{\beta_p}}{\beta_m}$, this reduces to

$$\sigma_{\delta_p}^2 \approx \left(\frac{\partial \delta_p}{\partial \delta_v} \sigma_{\delta_v} \right)^2 + \left(\frac{\partial \delta_p}{\partial R} \frac{\sigma_{\beta_p}}{\beta_m} \right)^2. \quad (31)$$

285 The partial derivatives are calculated as:

$$\frac{\partial \delta_p}{\partial \delta_v} = \frac{(1 + \delta_m)^2 R (R - 1)}{((1 + \delta_m)R - (1 + \delta_v))^2}, \quad (32)$$

$$\frac{\partial \delta_p}{\partial R} = \frac{(1 + \delta_m)(1 + \delta_v)(\delta_m - \delta_v)}{((1 + \delta_m)R - (1 + \delta_v))^2}. \quad (33)$$

4 Results and discussion

4.1 Background noise

290 Figure 4 shows the time series of background noise (variance of the background signal), normalized by integration time assuming Poisson statistics ($\sigma_{ppol/r^2}^2 \times t_{integration}$), alongside the laser power. Most instruments (except for Vehmasmäki) exhibit a gradual decline in laser power, accompanied by corresponding increases in background noise, likely due to weaker return signal. These changes often occur in discrete steps rather than as a continuous trend. Notably, fluctuations in background noise are consistently related to changes in laser power. Periods of restored laser power and decreased background noise align
295 with documented hardware interventions, such as transmitter replacements at Vehmasmäki (Dec 2022; Fig. 4a) and Hyytiälä (Feb 2024; Fig. 4b).

At Hyytiälä and Kenttäröva, laser power gradually decreases from 100% to around 40% over two years. In contrast, the CL61 at Vehmasmäki maintains a consistently high laser power, ranging between 90% and 100% throughout the same period, with only a slight decrease during the summer, likely to regulate the internal temperature to prevent overheating. Meanwhile,
300 the CL61 at Lindenberg experienced a sharp decline in laser power, from 80% to below 10% within just one year of operation. We also found that changes in the firmware version do not appear to have any noticeable effect on background noise levels.

The relationship between background noise, normalized by integration time, and laser power across several instruments is illustrated in Fig. 5. Nighttime measurements (23:00 to 01:00 local time), shown in blue, are unaffected by solar radiation, while all-day measurements are displayed in grey. It is important to note that the selected nighttime window is arbitrary and
305 used solely for this analysis; for instance, locations such as Kenttäröva experience no true nighttime during summer. The nighttime data clearly indicate that a decrease in laser power leads to an increase in background noise across all instruments. Daytime background noise is generally higher due to solar radiation, which varies with factors such as solar angle, cloud height, and surface albedo, making direct comparisons more complex. Nevertheless, the figure suggests that lower laser power also contributes to increased noise levels during daytime conditions.

310 An increase in background noise can significantly reduce an instrument's ability to detect weak aerosol signals. This effect is clearly illustrated in Fig. S3 in the supplement. On 7 March 2024, an elevated aerosol layer was observed above 3 km in Lindenberg, indicated by enhanced attenuated backscatter coefficient values compared to those at lower altitudes between 1 – 3 km (Fig. S3a, c). However, by 2 December 2024, a reduction in laser power led to a higher noise floor, which meant that if a similar aerosol layer were present, it would no longer be distinguishable from the background noise (Fig. S3b, c).




Figure 4. Time series of nighttime background signal variance normalized by the integration time ($\sigma^2_{ppol/r^2} \times t_{integration}$), and the laser power are shown for (a) Vehmasmäki, (b) Hyytiälä, (c) Kenttäröva, and (d) Lindenberg. Shaded regions denote different firmware versions; unshaded areas represent periods with no recorded firmware version and laser power.




Figure 5. Variance of background noise as a function of laser power percentage, normalized by integration time, at the following locations: a) Vehmasmäki, b) Hyytiälä, c) Kenttäröva, and d) Lindenberg. Blue points represent data collected during nighttime hours (23:00–01:00 local time), while grey points include all available data.

In this section, we examine how the instrumental background bias ($\mu_{P_{\text{instrument}}O(r)}$) and noise ($\sigma_{P_{\text{instrument}}O(r)}^2$) with the overlap correction reverted, vary with temperature and over time. The termination hood measurement was deployed multiple times at Vehmasmäki, Hyytiälä, and Kenttäröva. Figure 6 presents the instrumental background profiles recorded during these termination hood measurement periods, which have been averaged and grouped according to the instruments' internal temperatures.

320 For the instruments located in Hyytiälä and Kenttäröva, the $\mu_{P_{\text{instrument}}O(r)}$ profiles remain relatively stable around zero from the far range down to about 2 km, then increase sharply as the range decreases. On the other hand, in Vehmasmäki, it also remains near zero at far range, but $\mu_{P_{\text{instrument}}O(r)}$ begins to increase already at around 5 km and continues to rise toward shorter ranges. Meanwhile, the $\mu_{\perp P_{\text{instrument}}O(r)}$ profiles for all three instruments remain near zero from the far range to about 2 km, after which they increase rapidly with decreasing range.

325 Overall, the internal temperature T has a more pronounced effect on the $\mu_{P_{\text{instrument}}O(r)}$ profiles below 1 km than at higher ranges. In particular, the $\mu_{P_{\text{instrument}}O(r)}$ profiles in both polarizations deviate substantially from zero at below 200 m (see Fig. 7) across all instruments. Notably, each instrument shows the same small peak in the instrumental background bias in both polarizations at approximately 150 m.

All instruments exhibit profiles of $\sigma_{P_{\text{instrument}}O(r)}^2$ in both polarizations that remain constant above 500 m, followed by
 330 a sharp increase below this range. These instrumental background noise profiles show a pronounced sensitivity to internal temperature (from 4°C to 38°C). Notably, very low temperatures are associated with elevated instrumental background noise across the entire measurement range. This can be explained by the effect of the internal heater. As the internal temperature falls below an instrument-specific threshold (15 °C for Hyytiälä and Kenttäröva), the internal heater is turned on. This leads to an increase in $\sigma_{P_{\text{instrument}}O(r)}^2$, with the effect becoming more pronounced as the temperature decreases. When the internal
 335 temperature is high and the internal heater is off, internal temperature positively correlates with $\sigma_{P_{\text{instrument}}O(r)}^2$.

Similar to Figs. 6 and 7, Figs. S4 and S5 show the same profiles obtained during the termination hood measurements, but include range correction and the instrument-provided overlap correction. These profiles of $ppol$ and $xpol$ provide a direct quantification of the impact of instrumental background bias and noise on the measurements and their dependence on temperature. For all CL61 instruments, the instrumental biases, μ_{ppol} and μ_{xpol} , remain below the order of 10^{-7} , while the noise terms, σ_{ppol}^2
 340 and σ_{xpol}^2 , remain below the order of 10^{-11} . These biases are even smaller below 5 km. Below a range of 1 km, μ_{ppol} remains below 5×10^{-9} and μ_{xpol} below 3×10^{-9} . These small biases are usually negligible comparing to the typical atmospheric signal.

However, at very clean sites such as Kenttäröva, the instrumental bias in $xpol$ can still affect the retrieved aerosol depolarization ratio, particularly under conditions of low depolarization. An example is shown in Fig. S6, which presents measurements
 345 from Kenttäröva on 2023-10-03 under weakly depolarizing aerosol conditions. During this period, the $xpol$ profile below 400 m is on the order of 10^{-9} , while $ppol$ remains on the order of 10^{-6} . As a result, the instrumental bias becomes relatively significant for $xpol$, leading to a noticeable, although quantitatively small, reduction in δ (from 0.002 to 0.001 at 200 m) after the bias correction is applied. Furthermore, the previously mentioned peak at 150 m is clearly visible in this case and is ef-




Figure 6. Termination hood profiles at various internal temperatures (color-coded) across different sites. In Vehmasmaki: a) $\mu_{\parallel P_{\text{instrument}} O(r)}$, b) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, c) $\mu_{\perp P_{\text{instrument}} O(r)}$, and d) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$. In Hyttiala: f) $\mu_{\parallel P_{\text{instrument}} O(r)}$, g) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, h) $\mu_{\perp P_{\text{instrument}} O(r)}$, and i) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$. In Kenttaro: j) $\mu_{\parallel P_{\text{instrument}} O(r)}$, k) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, l) $\mu_{\perp P_{\text{instrument}} O(r)}$, and m) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$.




Figure 7. Termination hood profiles at various internal temperatures (color-coded) across different sites only up to 1000 m range. In Vehmasmaki: a) $\mu_{\parallel P_{\text{instrument}} O(r)}$, b) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, c) $\mu_{\perp P_{\text{instrument}} O(r)}$, and d) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$. In Hyttiala: f) $\mu_{\parallel P_{\text{instrument}} O(r)}$, g) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, h) $\mu_{\perp P_{\text{instrument}} O(r)}$, and i) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$. In Kenttaro: j) $\mu_{\parallel P_{\text{instrument}} O(r)}$, k) $\sigma_{\parallel P_{\text{instrument}} O(r)}^2$, l) $\mu_{\perp P_{\text{instrument}} O(r)}$, and m) $\sigma_{\perp P_{\text{instrument}} O(r)}^2$.

350 fectively mitigated after the removal of the instrumental bias. Overall, the impact of the bias correction is expected to be most pronounced under clean atmospheric conditions, such as in this case in Finland, where the δ profile remains consistently low. The correction improves the overall shape of the profile, particularly around 150 m, by removing the artificial peak in x_{pol} at that height (Fig. S6). However, the change remains small in absolute terms.

For all instruments, the signal near the surface exhibits much more rapid fluctuations than at higher range gates. Figure S7 displays example profiles of $ppol$, $xpol$, and the instrument's internal temperature during a termination hood measurement at Kenttäröva, revealing periodic variations in the signal below 50 m. To investigate this, the Fourier transforms of these signals were computed at each range gate over time and compared to that of the laser temperature recorded simultaneously. The results show that both the signals and the laser temperature exhibit a coincident spectral peak at approximately 0.0079 Hz (about 120 s), indicating a strong influence of laser temperature on the signals. Furthermore, additional distinct peaks in the signal spectra imply the presence of other instrument-related effects. Overall, these periodic variations appear sporadically across all instruments, with no consistent pattern indicating when they appear or disappear. Given the difficulty in developing a reliable correction method, we recommend excluding measurements below 50 m.




Figure 8. Calibration profiles in Vehmasmäki at 23°C and 24°C, recorded at three different time points over a six-month period in 2024.

Figure 8 shows how the instrumental bias during the termination hood measurement at Vehmasmäki varies over time under consistent internal temperature conditions. Over a six-month period, the calibration profiles at each temperature show only mi-

nor fluctuations. Therefore, we recommend performing termination hood checks at least once across the operating temperature
365 range to determine the instrumental background bias and noise.

4.3 Calibration

Figure 9 presents the time series of calibration factor from the cloud calibration and Rayleigh calibration methods for all the ceilometers. Detailed results of the cloud calibration factors derived from liquid cloud signals at each site are presented in Figures S8–S11. Overall, the calibration factors obtained from the two methods agree well within their respective uncertainties,

370 In Lindenberg, the calibration factor from the cloud calibration exhibits an approximate threefold decline over time, reflecting notable signal degradation. During this degradation period, Rayleigh calibration could not be performed because the elevated background-noise level (see Section 4.1) obscured the molecular return. In contrast, the other instruments show only short-term fluctuations without any clear long-term trend. Two periods with notable deviations in the calibration factor were also identified: June to October 2023 and November to December 2024 in Vehmasmäki. The issue in the CL61 at Vehmasmäki is
375 due to the fogged window that attenuates the outgoing signal. This will be analyzed in more detail in the following section.

Across all instruments, the calibration factor generally stays within the range of 0.9 to 1.5 as long as laser power remains above 40%, aside from occasional deviations due to specific instrument issues. In Vehmasmäki, as the laser power remains above 90%, the calibration factor fluctuates around 1 within 10% fluctuation. In Hyytiälä, the highest changes in calibration factor is observed when the instrument's laser power is below 80% with the firmware 1.1.10 and then below 60% with the
380 firmware 1.2.7. However, the calibration factor remains relatively stable at 1.25 even when the laser power drops from 100% to 40% with the firmware 1.2.7. In Kenttäröva, laser power percentages were not recorded before June 2023, but the observed increase in the calibration factor is likely linked to a decrease in laser power. After a new transmitter was installed, the calibration factor returned to around 1. In Lindenberg, a noticeable decline in the cloud calibration factor was observed as laser power dropped from 40% to 10%. It is still to be ascertained whether this trend also appears in instruments running the newer
385 firmware version 1.2.7.

Figure S12 demonstrates the utility of cloud calibration for two cases; one with 80% laser power on 2024-03-03 and another with just 10% on 2024-11-16. This indicates that the method can still be used in situations with low laser power and since the internal calibration value may not scale with low laser power values, regular cloud calibration is necessary to understand instrument performance and continue to provide profiles that can be used quantitatively, especially for laser power values below
390 40%.




Figure 9. Time series of the calibration factor from cloud calibration, Rayleigh calibration and laser power at four sites: a) Vehmasmäki, b) Hyytiälä, c) Kenttäröva, and d) Lindenberg. The dashed line indicates a calibration factor of $c = 1$. Periods corresponding to different firmware versions are highlighted with shaded areas.

4.4 Signal correction and uncertainty estimation

After recording the termination hood profiles at different temperatures and determining the calibration factor, the correction and forward Klett inversion can be applied, and the associated uncertainty can be estimated using the method described in Sect. 3.3. Example results of this procedure are illustrated in Figs. 10 and 11 showcasing a typical Finnish low aerosol load condition in Kenttäröva. During this period, the calibration factor was close to 1. On this date (Fig. 11), aerosol was present below 1000 m within the boundary layer close to the ground, while several layers of clouds extended from 2 km up to 3.5 km. Figure 10 presents the β , δ and their associated uncertainties in the aerosol layer below 1000 m obtained from a measurement profile recorded at 13:01 UTC on the same day.

In below 1000 m range, the corrected β'_v differs only slightly from the original uncorrected β' , with the difference ranging from 1×10^{-9} to $2.5 \times 10^{-9} \text{ sr}^{-1} \text{ m}^{-1}$ (Fig. 10a). This difference is expected to be relatively smaller within the clouds, where the signal is stronger and the instrumental bias decreases with increasing range. In contrast, the particle backscatter coefficient β_p is substantially lower than β'_v by approximately $1.5 \times 10^{-7} \text{ sr}^{-1} \text{ m}^{-1}$, indicating a significant contribution from molecular scattering. The $\sigma_{\beta'_v}^2$ and $\sigma_{\beta_p}^2$ profiles are relatively similar to each other, and they both increase exponentially with height (Fig. 10b).

Similarly, the original uncorrected (δ) and the corrected (δ_v) profiles differ only slightly, by the difference is about 0.001 to 0.005 (Fig. 10c) while δ itself being around 0.07. In contrast, the particle linear depolarization ratio δ_p is substantially higher, differing from the original δ by approximately 0.1 with δ_p at around 0.1 to 0.15. For example, during pollen seasons in Finland, δ_p has been observed at 0.23 for pine and 0.26 for birch at this wavelength (Filioglou et al., 2023). The variances ($\sigma_{\delta_v}^2$, $\sigma_{\delta_p}^2$) are comparable below 200 m, but at higher ranges, $\sigma_{\delta_p}^2$ becomes markedly larger.

Overall, the pronounced differences between β'_v and β_p , as well as between δ_v and δ_p , highlight the importance of accounting for the molecular contribution when performing quantitative assessments of aerosol measurement with CL61 for low aerosol load conditions in Finland. While the difference between ~~β~~ β' and β'_v (and δ and δ_v) may be small, the termination hood remains important as it allows the estimation of measurement uncertainties. Averaging multiple profiles reduces the uncertainties, as illustrated in Fig. S13, which shows the profiles averaged over 10 minutes.

Figure 11 illustrates the time series of β'_v , δ_v , β_p , and δ_p profiles, along with their corresponding variances on the same day. Similar to Fig. 10, β'_v is higher than β_p while δ_v is lower than δ_p in this period. The profiles of $\sigma_{\beta'_v}^2$ and $\sigma_{\beta_p}^2$ vary with time during this period, exhibiting an exponential increase with range. On the other hand, the $\sigma_{\delta_v}^2$ and $\sigma_{\delta_p}^2$ profiles demonstrate a more pronounced dependence on the signal magnitude (Fig. 11 d and h), with stronger atmospheric signals resulting in smaller uncertainties.




Figure 10. Example of uncorrected and corrected profiles in Kenttäröva measured at 2024-06-04 13:01 UTC. a) β , β' (instrument provided), β'_v (bias corrected and calibrated), β_p , b) $\sigma_{\beta'_v}^2$, $\sigma_{\beta_p}^2$, c) δ (instrument provided), δ_v (bias corrected and calibrated), δ_p , d) $\sigma_{\delta_v}^2$, $\sigma_{\delta_p}^2$. Error bars represent one standard deviation (σ) of the measurements.




Figure 11. An example of corrected and calibrated profile in Kenttärova on 2024-06-04 13:00 - 17:00 UTC (same day as Fig. 10). a) β'_v , b) $\sigma_{\beta'_v}^2$, c) δ_v , d) $\sigma_{\delta_v}^2$, e) β_p , f) $\sigma_{\beta_p}^2$, g) δ_p , and h) $\sigma_{\delta_p}^2$. For particle properties in panels e-h, profiles are shown only up to 1 km range.

As previously noted, abrupt changes in the cloud calibration factor at Vehmasmäki were observed from June to October 2023 and again from November to December 2024. Similar deviations are present during the same periods as seen in the window condition housekeeping variable (Fig. S14). These patterns indicate that an obstruction, likely on the instrument window, was attenuating both outgoing and incoming laser beams.

425 Figure 12 illustrates a specific example from 19 May, 2023. On this date, a sudden drop in the attenuated backscatter signal below 2000 m occurred just before 06:00 UTC. At the same time, the window condition value sharply declined from around 100% to 70%. This coincided with the automatic deactivation of the window blower heater. Housekeeping data, including internal temperature and relative humidity, were used to estimate the internal dew point, which was then compared to external air temperature. When the window blower heater was off, the window cooled to match the outside air temperature. Once the
430 window temperature dropped below the internal dew point, condensation began to form on the inside surface of the window and attenuated the signal. This occurred despite the measured internal relative humidity remaining below 70%. This is because the measurement does not represent the relative humidity at the window surface, where temperature is closer to ambient temperature.

Since condensation affects the entire signal profile and the degree of attenuation depends on its severity, it is not possible to
435 precisely quantify the signal loss or reconstruct the original profile. This issue could be mitigated by keeping both the window blower and heater always on. However, at the time of this writing, this is not possible (personal communication with Vaisala). Additionally, hardware modifications may be required to maintain a sufficiently low internal dew point and prevent fogging.




Figure 12. Measurements from Vehmasmäki on 2023-05-19 a) β' , b) Internal temperature, c) window condition, d) Internal humidity, e) Window blower heater, f) Calculated dew point and measured outside air temperature.

5 Conclusions

In this study, we ~~quantify the temporal variability of background noise and calibration factors for the~~ investigate several
440 characteristics of CL61 ALC measurements across multiple sites over a three-year period, using methodologies adapted from
existing approaches. ~~In all instruments, a reduction in laser power leads to increased background noise-~~

First, we examine the temporal evolution of the background noise and its relationship with laser power. The background noise
generally increases as the laser power decreases, which is particularly important because laser aging can substantially reduce
the CL61's capability to detect weak aerosol signals. For most instruments, internal scaling of the calibration factor effectively
445 compensates for changes in laser power. However, for the instrument at the Lindenberg site with an older firmware version
(1.1.10), this compensation becomes insufficient when laser power drops below 40%, resulting in a drift in the calibration
factor. Therefore, regular cloud calibration is necessary to ensure quantitatively reliable profiles, especially when the laser
power falls below 40%. ~~Some~~ Additionally, some deviations in the calibration factor were linked ~~with~~ to the fogging of the
inside surface of the window.

450 ~~Termination hood measurements were performed on selected instruments and proved effective in characterizing the instrumental~~
Second, we performed termination hood measurements to characterize the instrumental residual background bias and noise pro-
files ~~and their temperature dependence. Notably, differences in the instrumental bias were observed.~~ The magnitude of the
residual background bias for all instruments is generally small. In a case study in Kenttärova, the magnitude of the residual
bias was approximately 1-2% of the molecular backscatter. Nevertheless, the termination hood measurements remain crucial,
455 as they allow the instrumental background to be quantified and subsequently used to estimate the uncertainties of the particle
backscatter coefficient (β_p) and the particle linear depolarization ratio (δ_p), as presented in this study.

Third, we found that the instrumental residual background bias and noise profiles are sensitive to temperature and exhibit
differences between parallel and perpendicular polarizations. ~~The calibration profiles also reveal periodic variations in the near~~
~~range signal below 50 m in both polarizations. As they appear sporadically, we~~ We also recommend discarding the first 50 m of
460 measurements. ~~Instrumental bias,~~ where periodic variations were frequently observed. Instrumental background profiles were
obtained across a range of internal temperatures and were subsequently used to correct the bias and estimate its associated
uncertainty. These profiles remained stable for at least one month for a given internal temperature, although small deviations
appeared over a six-month period. Consequently, we recommend making ~~termination hood~~ termination hood measurements at
least once across the operating temperature range to ~~determine the instrumental noise~~ characterize the instrumental background.

465 ~~In a representative summer case study at Kenttärova,~~ Fourth, we demonstrate that the molecular contribution to the measured
depolarization ratio can be substantial, reaching approximately 0.1, whereas the impact of residual background bias is only
around 0.005 in the ~~instrumental bias corrected attenuated backscatter coefficient and volume linear depolarization ratio β_v' and~~
 δ_v differ only slightly from their uncorrected counterparts β and δ , up to $2.5 \times 10^{-9} \text{ sr}^{-1} \text{ m}^{-1}$ and 0.005 respectively, reflecting
the relatively small instrumental bias in both polarization channels compared with the atmospheric signal. Nevertheless, the
470 ~~termination hood measurements remain crucial, as they allow the instrumental background to be quantified and subsequently~~
~~used to estimate the uncertainties of β_p and δ_p , as presented in this study. For typical~~ same case study. This highlights the

importance of accounting for the molecular contribution in CL61 depolarization ratio measurement under low aerosol load conditions in Finland, as in this study case, we found that the difference between β_p and β is approximately $1.5 \times 10^{-7} \text{ sr}^{-1} \text{ m}^{-1}$, and that between δ and δ_p is roughly 0.1 in the near surface layer.

475 ~~Additionally, some instruments experienced signal loss caused by the fogging~~ Finally, we identify window fogging as an important issue affecting CL61 data quality. Fogging of the inside surface of the window ~~, which was observed in some instruments and~~ resulted in degraded ~~data quality measurements~~. Identifying and excluding such periods is ~~therefore~~ essential for ensuring the integrity of the dataset and the reliability of subsequent analyses.

480 *Data availability.* The CL61 and model data used in this study are provided by the Aerosol, Clouds and Trace Gases Research Infrastructure (ACTRIS) and are available respectively from the ACTRIS Data Centre using the following DOIs: 10.60656/57adeb3f598243f2 and 10.60656/d2626d9dd3454006). The CL61 raw data was obtained from the landing page of the DOI.

Author contributions. VL, EOC and MF performed the termination hood calibrations. VL prepared the manuscript. EOC, MF, VV contributed remarks and revisions on the manuscript.

Competing interests. The authors declare that they have no conflict of interest.

485 *Acknowledgements.* Financial support by Magnus Ehrnrooth Foundation is gratefully acknowledged. This project was supported by the Research Council of Finland (grant no 337552, 343359). The authors also gratefully acknowledge the support of Vaikuttavuussäätiö (the Finnish Research Impact Foundation) through the Tandem Industry Academia (TIA) program. We also acknowledge ACTRIS, the Finnish Meteorological Institute, and the Lindenberg Meteorological Observatory - Richard Assmann Observatory (MOL-RAO) for providing the data set which is available for download from <https://cloudnet.fmi.fi>.

- Baars, H., Kanitz, T., Engelmann, R., Althausen, D., Heese, B., Komppula, M., Preißler, J., Tesche, M., Ansmann, A., Wandinger, U., Lim, J. H., Young Ahn, J., Stachlewska, I. S., Amiridis, V., Marinou, E., Seifert, P., Hofer, J., Skupin, A., Schneider, F., Bohlmann, S., Foth, A., Bley, S., Pfüller, A., Giannakaki, E., Lihavainen, H., Viisanen, Y., Kumar Hooda, R., Pereira, S. N., Bortoli, D., Wagner, F., Mattis, I., Janicka, L., Markowicz, K. M., Achtert, P., Artaxo, P., Pauliquevis, T., Souza, R. A., Prakesh Sharma, V., Gideon Van Zyl, P., Paul Beukes, J., Sun, J., Rohwer, E. G., Deng, R., Mamouri, R. E., and Zamorano, F.: An overview of the first decade of PollyNET: An emerging network of automated Raman-polarization lidars for continuous aerosol profiling, *Atmospheric Chemistry and Physics*, 16, 5111–5137, <https://doi.org/10.5194/acp-16-5111-2016>, 2016.
- Baars, H., Seifert, P., Engelmann, R., and Wandinger, U.: Target categorization of aerosol and clouds by continuous multiwavelength-polarization lidar measurements, *Atmospheric Measurement Techniques*, 10, 3175–3201, <https://doi.org/10.5194/amt-10-3175-2017>, 2017.
- Barragán, R., Molero, F., Salvador, P., Theobald, M. R., Vivanco, M. G., Rodríguez-Sánchez, A., Gil, V., Garrido, J. L., Pujadas, M., and Artíñano, B.: Study of the Effect of Different Atmospheric Conditions on the Temporal Evolution of the Mixing Layer over Madrid during the Year 2020 by Means of Two Different Methods: Ceilometer Signals and the ECMWF-IFS Meteorological Model, *Remote Sensing*, 15, 5583, <https://doi.org/10.3390/rs15235583>, 2023.
- Biele, J., Beyerle, G., and Baumgarten, G.: Polarization lidar: Corrections of instrumental effects, *Optics Express*, 7, 427–435, <https://doi.org/10.1364/OE.7.000427>, 2000.
- Biniotoglou, I., Amodeo, A., D’Amico, G., Giunta, A., Madonna, F., Mona, L., and Pappalardo, G.: Examination of possible synergy between lidar and ceilometer for the monitoring of atmospheric aerosols, in: *Lidar Technologies, Techniques, and Measurements for Atmospheric Remote Sensing VII*, vol. 8182, pp. 48–60, SPIE, <https://doi.org/10.1117/12.897530>, 2011.
- Bucholtz, A.: Rayleigh-scattering calculations for the terrestrial atmosphere, *Applied Optics*, 34, 2765–2773, <https://doi.org/10.1364/AO.34.002765>, 1995.
- Burton, S. P., Ferrare, R. A., Hostetler, C. A., Hair, J. W., Rogers, R. R., Obland, M. D., Butler, C. F., Cook, A. L., Harper, D. B., and Froyd, K. D.: Aerosol classification using airborne High Spectral Resolution Lidar measurements—methodology and examples, *Atmospheric Measurement Techniques*, 5, 73–98, <https://doi.org/10.5194/amt-5-73-2012>, 2012.
- Campbell, J. R., Hlavka, D. L., Welton, E. J., Flynn, C. J., Turner, D. D., Spinhirne, J. D., Scott, V. S., and Hwang, I. H.: Full-Time, Eye-Safe Cloud and Aerosol Lidar Observation at Atmospheric Radiation Measurement Program Sites: Instruments and Data Processing, *Journal of Atmospheric and Oceanic Technology*, 19, 431–442, [https://doi.org/10.1175/1520-0426\(2002\)019<0431:FTESCA>2.0.CO;2](https://doi.org/10.1175/1520-0426(2002)019<0431:FTESCA>2.0.CO;2), 2002.
- Cao, N., Zhu, C., Kai, Y., and Yan, P.: A method of background noise reduction in lidar data, *Applied Physics B*, 113, 115–123, <https://doi.org/10.1007/s00340-013-5447-9>, 2013.
- Cimini, D., Haeffelin, M., Kotthaus, S., Löhnert, U., Martinet, P., O’Connor, E., Walden, C., Coen, M. C., and Preissler, J.: Towards the profiling of the atmospheric boundary layer at European scale—introducing the COST Action PROBE, *Bulletin of Atmospheric Science and Technology*, 1, 23–42, <https://doi.org/10.1007/s42865-020-00003-8>, 2020.
- Donovan, D. P., Klein Baltink, H., Henzing, J. S., De Roode, S. R., and Siebesma, A. P.: A depolarisation lidar-based method for the determination of liquid-cloud microphysical properties, *Atmospheric Measurement Techniques*, 8, 237–266, <https://doi.org/10.5194/amt-8-237-2015>, 2015.

- Fernald, F. G., Herman, B. M., and Reagan, J. A.: Determination of Aerosol Height Distributions by Lidar, *Journal of Applied Meteorology and Climatology*, 11, 482–489, [https://doi.org/10.1175/1520-0450\(1972\)011<0482:DOAHDB>2.0.CO;2](https://doi.org/10.1175/1520-0450(1972)011<0482:DOAHDB>2.0.CO;2), 1972.
- Filioglou, M., Leskinen, A., Vakkari, V., O'Connor, E., Tuononen, M., Tuominen, P., Laukkanen, S., Toiviainen, L., Saarto, A., Shang, X., Tiitta, P., and Komppula, M.: Spectral dependence of birch and pine pollen optical properties using a synergy of lidar instruments, *Atmospheric Chemistry and Physics*, 23, 9009–9021, <https://doi.org/10.5194/acp-23-9009-2023>, 2023.
- Filioglou, M., Tiitta, P., Shang, X., Leskinen, A., Ahola, P., Pätsi, S., Saarto, A., Vakkari, V., Isopahkala, U., and Komppula, M.: Lidar estimates of birch pollen number, mass, and CCN-related concentrations, *Atmospheric Chemistry and Physics*, 25, 1639–1657, <https://doi.org/10.5194/acp-25-1639-2025>, 2025.
- Floutsi, A. A., Baars, H., Engelmann, R., Althausen, D., Ansmann, A., Bohlmann, S., Heese, B., Hofer, J., Kanitz, T., Haarig, M., Ohneiser, K., Radenz, M., Seifert, P., Skupin, A., Yin, Z., Abdullaev, S. F., Komppula, M., Filioglou, M., Giannakaki, E., Stachlewska, I. S., Janicka, L., Bortoli, D., Marinou, E., Amiridis, V., Gialitaki, A., Mamouri, R.-E., Barja, B., and Wandinger, U.: DeLiAn – a growing collection of depolarization ratio, lidar ratio and Ångström exponent for different aerosol types and mixtures from ground-based lidar observations, *Atmospheric Measurement Techniques*, 16, 2353–2379, <https://doi.org/10.5194/amt-16-2353-2023>, 2023.
- Freudenthaler, V., Esselborn, M., Wiegner, M., Heese, B., Tesche, M., Ansmann, A., Müller, D., Althausen, D., Wirth, M., Fix, A., Ehret, G., Knippertz, P., Toledano, C., Gasteiger, J., Garhammer, M., and Seefeldner, M.: Depolarization ratio profiling at several wavelengths in pure Saharan dust during SAMUM 2006, *Tellus B*, 61, 165–179, <https://doi.org/10.1111/j.1600-0889.2008.00396.x>, 2009.
- Görsdorf, U., Komppula, M., Moiseev, D., and O'Connor, E.: Custom collection of lidar data from Hyttiälä, Kenttäröva, Lindenberg, and Vehmasmäki between 1 Oct 2021 and 31 Dec 2024, <https://doi.org/10.60656/D2626D9DD3454006>, 2025.
- Haefelin, M., Laffineur, Q., Bravo-Aranda, J.-A., Drouin, M.-A., Casquero-Vera, J.-A., Dupont, J.-C., and De Backer, H.: Radiation fog formation alerts using attenuated backscatter power from automatic lidars and ceilometers, *Atmospheric Measurement Techniques*, 9, 5347–5365, <https://doi.org/10.5194/amt-9-5347-2016>, 2016.
- Hervo, M., Poltera, Y., and Haeferle, A.: An empirical method to correct for temperature-dependent variations in the overlap function of CHM15k ceilometers, *Atmospheric Measurement Techniques*, 9, 2947–2959, <https://doi.org/10.5194/amt-9-2947-2016>, 2016.
- Hirsikko, A., O'Connor, E. J., Komppula, M., Korhonen, K., Pfüller, A., Giannakaki, E., Wood, C. R., Bauer-Pfundstein, M., Poikonen, A., Karppinen, T., Lonka, H., Kurri, M., Heinonen, J., Moiseev, D., Asmi, E., Aaltonen, V., Nordbo, A., Rodriguez, E., Lihavainen, H., Laaksonen, A., Lehtinen, K. E. J., Laurila, T., Petäjä, T., Kulmala, M., and Viisanen, Y.: Observing wind, aerosol particles, cloud and precipitation: Finland's new ground-based remote-sensing network, *Atmospheric Measurement Techniques*, 7, 1351–1375, <https://doi.org/10.5194/amt-7-1351-2014>, 2014.
- Hopkin, E., Illingworth, A. J., Charlton-Perez, C., Westbrook, C. D., and Ballard, S.: A robust automated technique for operational calibration of ceilometers using the integrated backscatter from totally attenuating liquid clouds, *Atmospheric Measurement Techniques*, 12, 4131–4147, <https://doi.org/10.5194/amt-12-4131-2019>, 2019.
- Hämäläinen, K., Hirsikko, A., Leskinen, A., Komppula, M., O'Connor, E. J., and Niemelä, S.: Evaluating atmospheric icing forecasts with ground-based ceilometer profiles, *Meteorological Applications*, 27, e1964, <https://doi.org/10.1002/met.1964>, 2020.
- Illingworth, A. J., Barker, H. W., Beljaars, A., Ceccaldi, M., Chepfer, H., Clerbaux, N., Cole, J., Delanoë, J., Domenech, C., Donovan, D. P., Fukuda, S., Hirakata, M., Hogan, R. J., Huenerbein, A., Kollias, P., Kubota, T., Nakajima, T., Nakajima, T. Y., Nishizawa, T., Ohno, Y., Okamoto, H., Oki, R., Sato, K., Satoh, M., Shephard, M. W., Velázquez-Blázquez, A., Wandinger, U., Wehr, T., and Van Zadelhoff, G. J.: The earthcare satellite : The next step forward in global measurements of clouds, aerosols, precipitation, and radiation, *Bulletin of the American Meteorological Society*, 96, 1311–1332, <https://doi.org/10.1175/BAMS-D-12-00227.1>, 2015.

Illingworth, A. J., Cimini, D., Haeefe, A., Haeffelin, M., Hervo, M., Kotthaus, S., Löhnert, U., Martinet, P., Mattis, I., O'Connor, E. J., and
565 Potthast, R.: How Can Existing Ground-Based Profiling Instruments Improve European Weather Forecasts?, *Bulletin of the American
Meteorological Society*, 100, 605–619, <https://doi.org/10.1175/BAMS-D-17-0231.1>, 2019.

Klett, J. D.: Lidar inversion with variable backscatter/extinction ratios, *Applied Optics*, 24, 1638–1643,
<https://doi.org/10.1364/AO.24.001638>, 1985.

Kotthaus, S., O'Connor, E., Munkel, C., Charlton-Perez, C., Haeffelin, M., Gabey, A. M., and Grimmond, C. S. B.: Recommendations for
570 processing atmospheric attenuated backscatter profiles from Vaisala CL31 ceilometers, *Atmospheric Measurement Techniques*, 9, 3769–
3791, <https://doi.org/10.5194/amt-9-3769-2016>, 2016.

Kotthaus, S., Bravo-Aranda, J. A., Collaud Coen, M., Guerrero-Rascado, J. L., Costa, M. J., Cimini, D., O'Connor, E. J., Hervo, M.,
Alados-Arboledas, L., Jiménez-Portaz, M., Mona, L., Ruffieux, D., Illingworth, A., and Haeffelin, M.: Atmospheric boundary layer
height from ground-based remote sensing: a review of capabilities and limitations, *Atmospheric Measurement Techniques*, 16, 433–479,
575 <https://doi.org/10.5194/amt-16-433-2023>, 2023.

Le, V., Lobo, H., O'Connor, E. J., and Vakkari, V.: Long-term aerosol particle depolarization ratio measurements with HALO Photonics
Doppler lidar, *Atmospheric Measurement Techniques*, 17, 921–941, <https://doi.org/10.5194/amt-17-921-2024>, 2024.

Lee, G., Gommers, R., Waselewski, F., Wohlfahrt, K., and O'Leary, A.: PyWavelets: A Python package for wavelet analysis, *Journal of Open
Source Software*, 4, 1237, <https://doi.org/10.21105/joss.01237>, 2019.

580 Mamouri, R. E. and Ansmann, A.: Potential of polarization lidar to provide profiles of CCN-and INP-relevant aerosol parameters, *Atmo-
spheric Chemistry and Physics*, 16, 5905–5931, <https://doi.org/10.5194/acp-16-5905-2016>, 2016.

Martucci, G., Milroy, C., and O'Dowd, C. D.: Detection of Cloud-Base Height Using Jenoptik CHM15K and Vaisala CL31 Ceilometers,
Journal of Atmospheric and Oceanic Technology, 27, 305–318, <https://doi.org/10.1175/2009JTECHA1326.1>, 2010.

Nason, G. P. and Silverman, B. W.: *The Stationary Wavelet Transform and some Statistical Applications*, pp. 281–299, Springer New York,
585 New York, NY, https://doi.org/10.1007/978-1-4612-2544-7_17, 1995.

Nicolae, D., Vasilescu, J., Talianu, C., Binietoglou, I., Nicolae, V., Andrei, S., and Antonescu, B.: A neural network aerosol-typing algorithm
based on lidar data, *Atmospheric Chemistry and Physics*, 18, 14 511–14 537, <https://doi.org/10.5194/acp-18-14511-2018>, 2018.

O'Connor, E.: Model data from Kenttäröva on 5 March 2024, <https://hdl.handle.net/21.12132/1.b218f85f5ed14149>, 2024.

O'Connor, E.: Custom collection of model data from Hyytiälä, Kenttäröva, Lindenberg, and Vehmasmäki between 1 Oct 2021 and 31 Dec
590 2024, <https://doi.org/10.60656/57ADEB3F598243F2>, 2025.

O'Connor, E. J., Illingworth, A. J., and Hogan, R. J.: A technique for autocalibration of cloud lidar, *Journal of Atmospheric and Oceanic
Technology*, 21, 777–786, [https://doi.org/10.1175/1520-0426\(2004\)021<0777:ATFAOC>2.0.CO;2](https://doi.org/10.1175/1520-0426(2004)021<0777:ATFAOC>2.0.CO;2), 2004.

Ribaud, J.-F., Haeffelin, M., Dupont, J.-C., Drouin, M.-A., Toledo, F., and Kotthaus, S.: PARAFOG v2.0: a near-real-time decision tool to sup-
port nowcasting fog formation events at local scales, *Atmospheric Measurement Techniques*, 14, 7893–7907, [https://doi.org/10.5194/amt-
14-7893-2021](https://doi.org/10.5194/amt-

595 14-7893-2021), 2021.

Rocadenbosch, F., Barragán, R., Frasier, S. J., Waldinger, J., Turner, D. D., Tanamachi, R. L., and Dawson, D. T.: Ceilometer-Based Rain-Rate
Estimation: A Case-Study Comparison With S-Band Radar and Disdrometer Retrievals in the Context of VORTEX-SE, *IEEE Transactions
on Geoscience and Remote Sensing*, 58, 8268–8284, <https://doi.org/10.1109/TGRS.2020.2984458>, 2020.

Roschke, J., Witthuhn, J., Klingebiel, M., Haarig, M., Foth, A., Kötsche, A., and Kalesse-Los, H.: Discriminating between “drizzle or rain”
600 and sea salt aerosols in Cloudnet for measurements over the Barbados Cloud Observatory, *Atmospheric Measurement Techniques*, 18,
487–508, <https://doi.org/10.5194/amt-18-487-2025>, 2025.

- Sassen, K.: The Polarization Lidar Technique for Cloud Research: A Review and Current Assessment, *Bulletin of the American Meteorological Society*, 72, 1848–1866, [https://doi.org/10.1175/1520-0477\(1991\)072<1848:TPLTFC>2.0.CO;2](https://doi.org/10.1175/1520-0477(1991)072<1848:TPLTFC>2.0.CO;2), 1991.
- Shang, X., Mielonen, T., Lipponen, A., Giannakaki, E., Leskinen, A., Buchard, V., Darmenov, A. S., Kukkurainen, A., Arola, A., O'Connor, E., Hirsikko, A., and Komppula, M.: Mass concentration estimates of long-range-transported Canadian biomass burning aerosols from a multi-wavelength Raman polarization lidar and a ceilometer in Finland, *Atmospheric Measurement Techniques*, 14, 6159–6179, <https://doi.org/10.5194/amt-14-6159-2021>, 2021.
- Tomasi, C., Vitale, V., Petkov, B., Lupi, A., and Cacciari, A.: Improved algorithm for calculations of Rayleigh-scattering optical depth in standard atmospheres, *Applied Optics*, 44, 3320–3341, <https://doi.org/10.1364/AO.44.003320>, 2005.
- 610 Tsaknakis, G., Papayannis, A., Kokkalis, P., Amiridis, V., Kambezidis, H. D., Mamouri, R. E., Georgoussis, G., and Avdikos, G.: Inter-comparison of lidar and ceilometer retrievals for aerosol and Planetary Boundary Layer profiling over Athens, Greece, *Atmospheric Measurement Techniques*, 4, 1261–1273, <https://doi.org/10.5194/amt-4-1261-2011>, 2011.
- Uzan, L., Egert, S., Khain, P., Levi, Y., Vadislavsky, E., and Alpert, P.: Ceilometers as planetary boundary layer height detectors and a corrective tool for COSMO and IFS models, *Atmospheric Chemistry and Physics*, 20, 12 177–12 192, [https://doi.org/10.5194/acp-20-](https://doi.org/10.5194/acp-20-12177-2020)
- 615 12177-2020, 2020.
- Van der Kamp, D.: Ceilometer observations of Vancouver's urban boundary layer : validation and mixed-layer height estimation, Ph.D. thesis, University of British Columbia, <https://doi.org/10.14288/1.0066572>, 2008.
- Wandinger, U.: Introduction to Lidar, in: *Lidar: Range-Resolved Optical Remote Sensing of the Atmosphere*, edited by Weitkamp, C., pp. 1–18, Springer New York, New York, NY, https://doi.org/10.1007/0-387-25101-4_1, 2005.
- 620 Warren, E., Charlton-Perez, C., Kotthaus, S., Lean, H., Ballard, S., Hopkin, E., and Grimmond, S.: Evaluation of forward-modelled attenuated backscatter using an urban ceilometer network in London under clear-sky conditions, *Atmospheric Environment*, 191, 532–547, <https://doi.org/10.1016/j.atmosenv.2018.04.045>, 2018.
- Welton, E. J. and Campbell, J. R.: Micropulse Lidar Signals: Uncertainty Analysis, *Journal of Atmospheric and Oceanic Technology*, 19, 2089–2094, [https://doi.org/10.1175/1520-0426\(2002\)019<2089:MLSUA>2.0.CO;2](https://doi.org/10.1175/1520-0426(2002)019<2089:MLSUA>2.0.CO;2), 2002.
- 625 Wiegner, M. and Geiß, A.: Aerosol profiling with the Jenoptik ceilometer CHM15kx, *Atmospheric Measurement Techniques*, 5, 1953–1964, <https://doi.org/10.5194/amt-5-1953-2012>, 2012.
- Wiegner, M., Madonna, F., Biniotoglou, I., Forkel, R., Gasteiger, J., Geiß, A., Pappalardo, G., Schäfer, K., and Thomas, W.: What is the benefit of ceilometers for aerosol remote sensing? An answer from EARLINET, *Atmospheric Measurement Techniques*, 7, 1979–1997, <https://doi.org/10.5194/amt-7-1979-2014>, 2014.