

Towards harmonized measurements of condensable vapors: insights from the intercomparison of six chemical ionization mass spectrometers at a boreal forest site

Cecilia Righi¹, Mihai Ciobanu², Neha Deot^{2,1}, Tuija Jokinen², Chengfeng Liu¹, Qi Yuan^{3,1}, Lauriane L. J. Quéléver¹, Spiro Jorga⁴, Veronika Pospisilova⁴, Lisa J. Beck^{5,*}, Mario Simon⁵, Marcel Zauner-Wieczorek⁵, Andreas Kürten⁵, Ruben Soler⁶, Teresa Vera⁶, Amalia Muñoz⁶, Lauri R. Ahonen¹, Chao Yan⁷, Tuukka Petäjä¹ & Nina Sarnela¹

¹ Institute for Atmospheric and Earth System Research, University of Helsinki, Helsinki, 00014, Finland

² Climate and Atmosphere Research Center, The Cyprus Institute, Nicosia 2121, Cyprus

10 ³ Aerosol and Haze Laboratory, Advanced Innovation Centre for Soft Matter Science and Engineering, Beijing University of Chemical Technology, Beijing, 100029, China

⁴ Tofwerk AG, Thun, 3645, Switzerland

⁵ Institute for Atmospheric and Environmental Sciences, Goethe University Frankfurt, Frankfurt am Main, 60438, Germany

* now at: Deutscher Wetterdienst, Offenbach am Main, 63067, Germany

15 ⁶ EUPHORE Labs., CEAM Foundation, Paterna, 46980, Spain

⁷ Nanjing-Helsinki Institute in Atmospheric and Earth System Sciences, Nanjing University, Nanjing, 210023, China.

Correspondence to: Cecilia Righi (cecilia.righi@helsinki.fi), Nina Sarnela (nina.sarnela@helsinki.fi)

Abstract. Atmospheric new particle formation is driven by condensable vapors such as sulfuric acid and highly oxygenated organic molecules (HOMs). Measuring these gases is challenging because they are present at trace concentrations, they are easily lost through condensation onto surfaces, and they include a large number of chemically diverse species. Chemical Ionization Mass Spectrometry (CIMS) has been extensively used for their detection; however, results obtained in different studies are not always directly comparable. This limitation reflects differences in instrument designs, operating configurations, and reagent ion schemes employed. To investigate these factors, the Aerosol, Clouds and Trace Gases Research Infrastructure (ACTRIS) organized its first CIMS field intercomparison campaign (CI-FI1) during summer 2024 at a Finnish boreal forest site, the SMEAR II (Station for Measuring Forest Ecosystem-Atmosphere Relations) station. Six instruments employing different inlet designs, mass analyzers, and reagent ions were operated using their routine configurations and calibrated according to standard procedures. For sulfuric acid measured in nitrate mode, the conventional sulfuric acid calibration enabled moderately good agreement among instruments, although larger discrepancies were observed at the lower concentrations, typically observed during nighttime. When targeting higher-mass compounds such as monoterpene-derived HOM monomers (m/z 240-390) and dimers (m/z 480-630), sulfuric acid calibration alone proved insufficient to ensure measurement intercomparability and taking into consideration mass-dependent transmission differences became important to achieve consistent results. Notably, good agreement was observed also for selected compounds measured in bromide mode by different instruments. Overall, the results demonstrate that comparable field measurements of condensable vapors by different CIMS instruments are achievable when all relevant calibration and correction factors are carefully considered. The results further

35 highlight that similarities in instrument behavior are often more closely associated with the inlet design and instrument
operating conditions than with the reagent ion choice.

1 Introduction

The formation of new atmospheric aerosol particles is a complex process that has motivated extensive research efforts and promoted advances in analytical techniques. Although the species and mechanisms leading to the nucleation of aerosol
40 particles vary with geographical location and elevation, studies indicate that new particle formation (NPF) involving multiple neutral compounds generally dominates over ion-induced pathways in the continental boundary layer (Kerminen et al., 2010; Zhao et al., 2024). Such neutral precursors of atmospheric aerosols are formed through the oxidation of volatile compounds, including sulfur dioxide (SO₂) and several volatile organic compounds (VOCs). SO₂ can be oxidized to sulfuric acid (H₂SO₄) by hydroxyl radicals (OH) or Criegee intermediates, whereas VOCs, mainly monoterpenes, are oxidized by OH, ozone (O₃),
45 or nitrate radicals (NO₃; Peräkylä et al. 2014), leading to the formation of highly oxygenated organic molecules (HOMs; Bianchi et al., 2019). H₂SO₄ and monoterpene-derived HOMs are among the species referred to as condensable vapors, which are low-volatility atmospheric gaseous compounds that act as direct aerosol precursors, and they have long been recognized as key contributors to the formation and growth of new atmospheric particles in the boreal forest (Ehn et al., 2014; Jokinen et al., 2012, 2015; Kulmala et al., 2013; Schobesberger et al., 2013). These neutral compounds are challenging to measure because
50 (1) they are present at extremely low concentrations (often at parts-per-trillion to parts-per-quadrillion volume mixing ratios), (2) they readily condense onto surfaces due to their low volatility, and (3) they require charging before detection, which introduces additional complications, as different compounds may be charged and detected with different efficiencies. Adding to these challenges, (4) they comprise a large number of chemically diverse species: even when considering only the monoterpene-derived HOMs commonly targeted within this class, they cover a broad m/z range and include molecules with
55 varying numbers of carbon atoms and different functional groups and degrees of oxidation. Consequently, the detection of these NPF precursors requires analytical techniques that are both highly sensitive and capable of resolving a wide range of chemically diverse species.

From its early applications, chemical ionization mass spectrometry (CIMS) has proven to be particularly well suited for this purpose. Indeed, by selecting suitable reagent ions, it provides selective and sensitive online detection for a variety of chemical
60 compounds. Commonly used reagent ions in atmospheric NPF studies include both negative and positive ions, such as nitrate (NO₃⁻), acetate (CH₃COO⁻), bromide (Br⁻), and iodide (I⁻), as well as hydronium (H₃O⁺), ammonium (NH₄⁺), nitrosonium (NO⁺), and protonated benzene cluster cations ((C₆H₆)₂⁺). The selectivity of these and other reagent ions toward specific compounds has been investigated and reported in several studies (Y. Zhang et al., 2023; W. Zhang et al., 2024, and references therein).

65 Chemical ionization sampling has been performed both at reduced pressure and at ambient pressure. The first one offers comparatively better control over ionization conditions and is essentially unaffected by fluctuations in ambient parameters

such as relative humidity (RH). The latter, on the other hand, substantially reduces analyte wall losses and provides a higher ionization probability compared to reduced pressure ionization methods, owing to the high collision rate between reagent ions and sample molecules at atmospheric pressure. It has therefore been more frequently employed for the detection of condensable vapors in situ.

Over the years, mass spectrometers with increasingly higher mass resolution have been developed, together with chemical ionization inlets designed to provide efficient sample ionization. Some inlet designs also allow operation with multiple reagent ion schemes, thereby expanding the range of target species that can be measured with a single mass spectrometer. In addition to the choice of reagent ion, other instrumental parameters can be tailored to the detection of targeted compounds. For example, by tuning the instrument voltages, the operator can alter the fragmentation and the transmission of clusters with different mass-to-charge (m/z) ratios within the mass spectrometer, thus affecting the sensitivity of the method (Lopez-Hilfiker et al., 2016; Wang et al., 2024).

The flexibility of CIMS, which makes it a powerful technique in atmospheric studies, also represents a challenge for data interpretation. Since these instruments are generally not calibrated for every targeted analyte, different setups and operating configurations, even when the same reagent ion is employed, can introduce substantial variability in quantitative measurement outcomes and complicate direct comparisons across studies. To date, however, the extent of these differences has not been comprehensively investigated or discussed. Furthermore, instrument-specific biases often remain unnoticed in standalone deployments. Intercomparison studies are therefore valuable tools for addressing these challenges, as they allow one to assess how well measurements from different instruments compare and, eventually, help identify methods toward more harmonized instrument performance. Such harmonization is needed to achieve consistent and robust measurements of condensable vapors and represents one of the objectives of the Centre for Reactive Trace Gases In Situ Measurements (CiGas) within the Aerosol, Clouds and Trace Gases Research Infrastructure (ACTRIS). In this framework, CiGas coordinates continuous long-term measurements of nitrogen oxides, non-methane hydrocarbons, oxygenated VOCs, and condensable vapors in the atmosphere, while also promoting intercomparison activities aimed at improving the consistency and comparability of measurements across different sites.

As part of these efforts, the present work provides insights from the first CIMS field intercomparison campaign organized by ACTRIS (ACTRIS CI-FI1), conducted over two weeks in summer 2024 at a Finnish boreal forest site, the SMEAR II (Station for Measuring Forest Ecosystem-Atmosphere Relations) station. During the intercomparison workshop, six instruments were operated using their routine configurations and settings, and calibrations were performed following standard procedures applied by each group. The objectives were to assess (a) how well measurements from different instruments, employing either the same or different reagent ions, compare under field conditions, and thus how effectively commonly used calibration approaches ensure inter-instrument comparability, and (b) to what extent factors other than reagent ion choice influence the instrument response toward selected compounds.

2 Methods

100 2.1 Measurement site and timeline

The ACTRIS CI-FI1 campaign took place between 26 July and 9 August 2024 at the SMEAR II station, located at the Hyttiälä Forestry Field Station of the University of Helsinki, in southern Finland (61°51'N, 24°17'E, 181 m above sea level). The measurement station is surrounded by a boreal coniferous forest characterized by managed pine stands and limited local pollution, primarily from two sawmills 5 km to the southeast and the city of Tampere 60 km to the southwest. The rationale and primary objectives of the SMEAR II station, originally established in 1982, have been thoroughly described by Hari and Kulmala (2005).

The present study involved multiple research teams and a total of six mass spectrometers (Table 1). Five of these were Atmospheric Pressure interface Time-of-Flight (APi-ToF, ToFwerk AG) mass spectrometers, including three long Time-of-Flight (L-ToF, mass resolution 6,000–14,000 Th/Th; 1 Th = 1 Da/*e*, where *e* is the elementary charge) and two high-resolution Time-of-Flight (H-ToF, mass resolution 3,000–7,000 Th/Th) mass spectrometers. Three out of the five APi-ToFs featured Eisele-type inlets (Eisele and Tanner, 1991, 1993; Jokinen et al., 2012) operated with NO₃[−] ions, while the remaining two used Multi-scheme chemical IONization (MION) inlets, MION* (Rissanen et al., 2019) and MION2 (He et al., 2023), which were set to switch between NO₃[−] and Br[−] reagent ions, or alternatively to run in a non-ionizing mode to sample ambient ions. The sixth instrument was a Vocus Bipolar Time-of-Flight (B-ToF, ToFwerk AG) mass spectrometer coupled with a Vocus AIM Reactor (Riva et al., 2024). The instrument was operated with I[−] and (C₆H₆)₂⁺ as the conventional reagent ions, while NO₃[−] ionization was additionally implemented as a test configuration to explore the feasibility of nitrate ion chemistry in the AIM reactor. Hereafter, the APi-ToF mass spectrometers operated with chemical ionization are referred to as CI-APi-ToFs, while CIMS is used as a general term for all instruments included in the intercomparison, since this terminology also includes mass spectrometers that do not operate at atmospheric pressure, such as B-ToF.

120 **Table 1** Instruments participating in the ACTRIS CI-FI1 campaign. Details include institute affiliation, mass spectrometer type, chemical ionization inlet, and reagent ion.

Institute	Mass Spectrometer	Chemical Ionization inlet	Reagent ion	Instrument label
University of Helsinki	APi-H-ToF	Eisele-type (University of Helsinki)	NO ₃ [−]	EH1
Mediterranean Center for Environmental Studies Foundation (CEAM)	APi-H-ToF	Eisele-type (Aerodyne Research Inc.)	NO ₃ [−]	EH2
Goethe University Frankfurt	APi-L-ToF	Eisele-type (Aerodyne Research Inc.)	NO ₃ [−]	EL
University of Helsinki	APi-L-ToF	MION* (Karsa Ltd.)	NO ₃ [−] , Br [−]	ML1

The Cyprus Institute	APi-L-ToF	MION2 (Karsa Ltd.)	NO_3^- , Br^-	ML2
Tofwerk AG	B4-ToF	Vocus AIM Reactor (Tofwerk AG)	NO_3^- , I^- , $(\text{C}_6\text{H}_6)_2^+$	AB

The first week of the campaign was dedicated to instrument installation and side-by-side ambient air measurements for intercomparison, while the following week was focused on more targeted activities, including sulfuric acid calibrations, background measurements, and transmission calibrations. The instruments were distributed across three neighboring containers. To ensure simultaneous sampling of identical ambient air, all inlets faced the same direction, while the exhaust lines were directed towards the opposite side and merged into a large duct that transported waste flows away from the sampling area. A schematic timeline for the ACTRIS CI-FI1 campaign is presented in the Supporting Information (SI, Figure S1).

During the two-week campaign, the air temperature at the station ranged from 13 °C to 25 °C, with an average of ~17 °C, while the relative humidity varied between 39% and 100%, averaging 82%. Rainfall events occurred primarily during the first week, with precipitation recorded on 29 and 31 July 2024, as well as 1–4 and 9 August 2024. The meteorological data were obtained from datasets available on the SmartSMEAR platform; corresponding metadata are provided in Table S1 in the SI.

2.2 Instrumentation

The CI inlets deployed during the ACTRIS CI-FI1 campaign (Figure 1), as well as the APi-ToF and B-ToF mass spectrometers, have been described in dedicated publications, along with their various applications. Here, we report the operating settings applied during the intercomparison campaign (Table 2) and highlight the main differences among the chemical ionization inlets and between the mass analyzers. The relevant references are cited in the respective subsections below and, for the CI inlets, are also listed in Table 2.




Figure 1 Schematic representation of the chemical ionization inlets employed during the ACTRIS CI-FI1 campaign: Eisele-type inlet, AIM Reactor, MION* inlet, and MION2 inlet. The MION* inlet incorporates an added purge flow compared to the original MION design described by Rissanen et al. (2019), and is therefore denoted with an asterisk.

Table 2 Operating parameters of the CI inlets used during the ACTRIS CI-FI1 campaign, including inlet pressure, ion source type, residence time of NO₃⁻ reagent ion, and sample air flow rates.

CI inlet	Operating pressure [Pa]	Ion source [type, #]	NO ₃ ⁻ residence time [ms]	Sample air flow [L min ⁻¹]	References
Eisele-type	~1x10 ⁵	X-ray, 1	~200	~10	Eisele and Tanner (1993), Jokinen et al. (2012)
MION*	~1x10 ⁵	X-ray, 2	~175	~20	Rissanen et al. (2019)
MION2	~1x10 ⁵	X-ray, 2	~515	~20	He et al. (2023)
AIM Reactor	~5x10 ³	VUV, 2	~10	~2	Riva et al. (2024)

2.2.1 Atmospheric-pressure chemical ionization inlets

The Eisele-type inlet (Eisele and Tanner, 1993; Jokinen et al., 2012) consists of two concentric laminar-flow regions: an outer (O) region, directly exposed to the ionization source, where reagent ions are generated, and an inner (I) region, where these ions react with the sample molecules. Sample air is drawn into the I region through a 3/4" stainless steel tube at a flow rate of approximately 10 L min⁻¹. In the O region, around 20 L min⁻¹ of sheath flow (clean air) is mixed with 5-10 mL min⁻¹ of air saturated with the reagent gas. In the inlets used during the ACTRIS CI-FI1 campaign, soft X-ray ionization (Hamamatsu

L12535, 4.9 kV) is used to convert reagent molecules into reagent ions. Two coordinated voltages guide the reagent ions from the O region into the I region, while keeping neutral compounds in the O region. There, the reagent ions are electrostatically directed into the sample flow and have a residence time typically ranging from 100 to 300 ms, depending on the setup, during which they interact with the sample molecules.

The MION* inlet used in this study is an upgraded version of the MION inlet presented by Rissanen et al. (2019). It consists of an electrically grounded laminar flow tube with a 24 mm inner diameter, designed to accommodate multiple ion sources; in the present setup, two sources were employed, one for NO_3^- and one for Br^- . Approximately 20 L min^{-1} of sample air is drawn into the flow tube after passing through an upstream ion filter that removes any pre-existing ions from the sample air. A gas-phase stream of clean air, ranging from 5 to 50 mL min^{-1} depending on the reagent gas, is used to deliver the reagent gas through the dedicated ion source, where it is ionized by soft X-ray radiation (Hamamatsu L12535, 4.9 kV). The resulting reagent ions are accelerated and focused by electric fields through 5 mm orifices into the flow tube, where they interact with the laminar sample flow. Each ion source includes a purge flow, consisting of the same air stream used for the reagent flow, which (1) prevents neutral reagent gases from entering the sample flow, and (2) avoids the backflow of sample air into the ion source. The residence time of reagent ions in the flow tube depends on the distance between the ion source and the mass spectrometer pinhole, and thus differs between the two ion sources. Under the configuration used during this campaign, the residence time was ~ 175 ms for the upstream reagent (NO_3^-) and ~ 40 ms for the downstream reagent (Br^-).

The MION2 inlet (He et al., 2023) differs from MION* in one respect: it allows for the operation of up to three ion sources placed at equal distance from the pinhole. Additionally, the connecting pipe lengths can be adjusted both between the upstream source and the instrument pinhole and between the upstream and downstream sources, providing flexibility in setting the residence times of both reagent ions, unlike in MION*, where the closest source is fixed in position. In the setup used during this campaign, only two ion sources were operated, one for NO_3^- and one for Br^- . The sources were separated by ~ 350 mm, resulting in a residence time of ~ 515 ms for NO_3^- and ~ 40 ms for Br^- .

2.2.2 Low-pressure chemical ionization inlet

The Vocus AIM Reactor (Riva et al., 2024) operates at sub-atmospheric (medium) pressures, typically around 50 mbar and within a range of 20-500 mbar, and it further differs from the atmospheric-pressure CI inlets described in Section 2.2.1 in that all surfaces in contact with the sample before entering the inlet are made of Teflon® rather than stainless steel. Sample air is introduced into the reactor at a flow rate of approximately 2 L min^{-1} . Reagent ions are generated using compact vacuum ultraviolet (VUV; UV lamp krypton DC PID PKS 106, Heraeus) ion sources, with reagent gas delivered through the ion source and into the ion–molecule reaction (IMR) region at a flow rate of about 0.25 L min^{-1} . Similar to the MION inlets, the AIM Reactor can accommodate multiple ion sources, allowing up to three sources to be arranged radially. In the configuration used in this study, two sources were implemented, injecting reagent ions at an angle of approximately 45° relative to the sample flow. The reactor temperature is actively controlled and typically maintained at 50°C to ensure stable reaction conditions. Under these operating conditions, the residence time of reagent ions is on the order of ~ 10 ms.

2.2.3 Atmospheric Pressure interface Time-of-Flight mass spectrometer

The APi-ToF is a high-resolution mass spectrometer capable of detecting natural ions with minimal fragmentation, limited to weakly bound clusters. The instrument has been described in detail by Junninen et al. (2010), and a brief overview is also provided in Section S4 in the SI. For the purposes of this intercomparison, it is sufficient to highlight that the key difference between H-ToF and L-ToF mass analyzers lies in the length of the ToF chamber—665 mm and 1330 mm, respectively—which directly affects mass resolving power: a longer flight path allows for better separation of ions, thus resulting in higher resolution. According to specifications from ToFwerk AG, the theoretical mass resolution ranges from 3,000–7,000 Th/Th for H-ToF and up to 6,000–14,000 Th/Th for L-ToF.

For the NO₃[−] mode measurements presented in this study, the mass resolution at m/z = 201, calculated from the full width at half maximum (FWHM) of the corresponding peak, is reported for each CI-APi-ToF in Table 3.

Table 3 Full width at half maximum (FWHM) and corresponding mass resolution at m/z = 201 (NO₃[−] mode) for the five CI-APi-ToF mass spectrometers included in the intercomparison study.

Instrument label	Mass analyzer	FWHM (m/z 201)	Mass resolution (m/z 201)
EH1	H-ToF	0.120	1680
EH2	H-ToF	0.061	3319
EL	L-ToF	0.018	11188
ML1	L-ToF	0.029	6827
ML2	L-ToF	0.026	7648

2.2.4 Bipolar Time-of-Flight mass spectrometer

The Vocus B4 is a bipolar time-of-flight mass spectrometer in which sample ions are transmitted through a common, differentially pumped vacuum interface before entering polarity-specific ToF paths. Unlike single-analyzer instruments (e.g., APi-ToF), polarity switching in the Vocus B4 does not require switching of the high voltages of the ToF chambers or detectors. Instead, the positive- and negative-ion ToF chambers remain operated at their respective high-voltage settings, while fast switching of low voltages (those of the interface) selects the transmitted polarity or reagent ion mode. This enables the quasi-simultaneous detection of positive and negative ions, with polarity switching on the order of 50 ms. Being L-ToF analyzers, both ToF chambers achieve a theoretical mass resolution of ~10,000 Th/Th.

2.2.5 Reagent ion chemistry

All the instruments involved in the ACTRIS CI-FII campaign operated with nitric acid (HNO₃) either as the sole or as one of the reagent gases. When HNO₃ is used, the ionization predominantly generates nitrate ion clusters, (HNO₃)_xNO₃[−] (x = 0–2;

(Ehn et al., 2014; Jokinen et al., 2012). These reagent ions are highly selective toward H_2SO_4 and other gas-phase acids stronger than HNO_3 , as well as toward oxidation products of VOCs containing more than five oxygen atoms, with the binding strength of the reagent ion to the target molecule increasing almost linearly with the number of oxygen atoms (Hyytinen et al., 2018). In the first case, the acids are directly ionized and detected either as single deprotonated ions or as clusters with HNO_3 . In the second case, the detection of such highly oxygenated species relies primarily on clustering with NO_3^- via collisions with nitrate ion clusters, through a ligand exchange reaction (Hyytinen et al., 2015, 2018).

Instruments ML1 and ML2 also employed a second reagent gas, dibromomethane (CH_2Br_2), whose ionization primarily produces Br^- , with Br_3^- formed to a lesser extent through $\text{Br}_2\text{-Br}^-$ recombination (Finkenzeller et al., 2024). In this case, the analyte is charged via direct adduct formation with Br^- (Hyytinen et al., 2018). Although this ionization method is less selective than NO_3^- , it has proven suitable for studying atmospheric organic compounds that are less oxidized than those typically favored by NO_3^- (Huang et al., 2021; Rissanen et al., 2019). Moreover, Br^- ionization has been applied to the detection of OH and hydroperoxyl radicals (HO_2 ; Albrecht et al., 2019; Lambe et al., 2022; Sanchez et al., 2016), as well as halogenated compounds (Wang et al., 2021). Finally, Br^- shares with NO_3^- the capability to detect H_2SO_4 (Wang et al., 2021), which makes it quantitatively comparable to the nitrate ion chemistry.

Instrument AB employed two additional reagent gases alongside HNO_3 : methyl iodide (CH_3I) and benzene (C_6H_6), which are ionized to I^- and $(\text{C}_6\text{H}_6)_2^+$, respectively. The ionization pathways and target compound classes of I^- largely overlap with those of Br^- , although I^- forms weaker adducts with the analytes (Hyytinen et al., 2018). As for $(\text{C}_6\text{H}_6)_2^+$ mode, it predominantly targets volatile precursor species of condensable vapors, such as isoprene and terpenes (Lavi et al., 2018) or dimethyl sulfide (Kim et al., 2016). In the present work, however, only iodide-mode data are discussed for instrument AB.

2.3 Data analysis

2.3.1 Selected peak list

The target species considered in this study were selected among condensable vapors involved in the different stages of NPF, and were based on two main criteria: the measurement site (i.e., a boreal forest environment), and the selectivity of the reagent ions employed (Section 2.2.5).

The shortlist of peaks selected for NO_3^- mode (Table S2 in the SI) includes sulfuric acid, a few strong organic and inorganic acids (hereafter referred to as HA), as well as several monoterpene-derived HOMs belonging to different chemical classes, namely organic peroxy radicals, closed-shell non-nitrate monomers, closed-shell organonitrate monomers, and dimers. All selected species contain seven or more oxygen atoms, for which NO_3^- shows high selectivity.

The peak list used for Br^- mode (Table S3 in the SI) includes the same species as in NO_3^- ion mode, with NO_3^- in the clusters replaced by Br^- and the corresponding reagent ions adjusted accordingly. However, due to the higher sensitivity of Br^- toward less oxidized acids, a set of additional organic compounds containing three to six oxygen atoms were included in both the nitrate and bromide peak lists to enable a more meaningful comparison between the two reagent ion modes in instruments

operating with both. These species were classified as semi-volatile organic compounds (SVOCs) according to the parametrization proposed by Donahue et al. (2011) and later updated by Mohr et al. (2019).

245 Finally, among the peaks fitted in I⁻ mode, only six species were selected for this study, corresponding to compounds already included in the other two peak lists within the SVOC group, detected here as clusters with I⁻ (Table S4 in the SI).

2.3.2 Sulfuric acid calibration

The method for sulfuric acid calibration of chemical ionization mass spectrometers, introduced by Kürten et al. (2012), was applied to each CI-APi-ToF during the campaign using the same setup. The calibration factor, C_{SA} , was determined from the
250 slope of the linear regression between the modeled H₂SO₄ concentration and the reagent ion normalized H₂SO₄ signal, $S_{H_2SO_4}$, at different calibration steps. The modeled H₂SO₄ concentration was calculated using a dedicated MATLAB model, while the normalized H₂SO₄ signal for each calibration step in nitrate mode was calculated according to Eq. (1):

$$S_{H_2SO_4} = \frac{HSO_4 + HNO_3 \cdot HSO_4 + H_2SO_4 \cdot HSO_4}{NO_3 + HNO_3 \cdot NO_3 + (HNO_3)_2 \cdot NO_3 + H_2O \cdot NO_3} \quad (1)$$

During the ACTRIS CI-FI1 campaign, sulfuric acid calibrations were conducted in five steps, each lasting 10 minutes. For all
255 instruments, signals were averaged over 2-minute intervals. The calibration was not performed for instrument AB, since it was not expected to detect H₂SO₄ at the low concentrations typical of the measurement site.

Since both nitrate and bromide ionization detect H₂SO₄ near collision limit (Ehn et al., 2014; Wang et al., 2021), sulfuric acid is considered a suitable calibration compound also when operating in this second reagent ion mode. The equation used to calculate the Br⁻ mode normalized H₂SO₄ signal is provided in the SI (Section S8), along with a detailed description of the
260 calibration setup employed during the campaign (Section S7).

2.3.3 Transmission calibration

In cases where a direct calibration method for the target compound has not yet been established, the accurate quantification of the analyte requires knowledge of the transmission efficiency of the instrument (Alfaouri et al., 2025). In APi-ToF, the unequal transmission of ions depending on their m/z, known as mass discrimination, occurs primarily in the first two quadrupoles of
265 the APi section, the ToF orthogonal extraction unit, and the microchannel plate (MCP) detector (Heinritzi et al., 2016). When a CI source is used upstream of the APi-ToF, the transmission of an ion at a given m/z can be determined relative to that of the reagent ion.

Transmission calibrations were performed on all CI-APi-ToFs in NO₃⁻ mode, following the depletion method described by Heinritzi et al. (2016), and briefly here. A Union Tee piece of 3/4" was attached to the inlet tubes of the CI-APi-ToFs, using
270 adaptors to 1" for instruments with larger inlet tube diameter. The third outlet of the Tee was closed with a cap, in which a small quantity of the perfluorinated acids used for the calibration was placed, either in liquid or solid form, one compound at a time. The calibration experiment consisted of heating the cap with a heat gun to induce evaporation or sublimation of the

calibrant, resulting in depletion of the reagent ions. After the heating cycle, the signal from the perfluorinated acid decreased as the system cooled and the primary reagent ions recovered. The perfluorinated acids used in the experiments are listed in Table 4. During the ACTRIS CI-FI1 campaign, C5 and C7 were employed, with C7 excluded for EH2, whereas C7, C8, and C9 were used for ML2, which was calibrated post-campaign in the ACTRIS Laboratory of the University of Helsinki. The use of different calibrant sets was mainly related to time limitations during the campaign, as well as to ongoing testing and optimization of the method. Based on the experience gained during the campaign, we recommend performing at least three heating cycles for each calibrant and aiming for different degrees of reagent ion depletion, as also suggested by Heinritzi et al. (2016).

Table 4 Perfluorinated acids used for the transmission calibrations.

Compound name	Compound label	Chemical formula	Molecular mass [amu]	Supplier
Perfluorinated pentanoic acid	C5	$\text{CF}_3(\text{CF}_2)_3\text{COOH}$	264.05	Sigma-Aldrich, 97% purity
Perfluorinated heptanoic acid	C7	$\text{CF}_3(\text{CF}_2)_5\text{COOH}$	364.06	Sigma Aldrich, 99% purity
Perfluorinated octanoic acid	C8	$\text{CF}_3(\text{CF}_2)_6\text{COOH}$	414.07	Sigma Aldrich, 96% purity
Perfluorinated nonanoic acid	C9	$\text{CF}_3(\text{CF}_2)_7\text{COOH}$	464.08	Sigma Aldrich, 97% purity

Curves describing the relative transmission as a function of m/z were derived by least-squares fitting of a 2-fold Gaussian distribution (Eq. 2) to the depletion experiment data:

$$f(x) = a_1 \cdot \exp\left(-\left(\frac{x-b_1}{c_1}\right)^2\right) + a_2 \cdot \exp\left(-\left(\frac{x-b_2}{c_2}\right)^2\right), \quad (2)$$

where x corresponds to the m/z of the selected species. Following Heinritzi et al. (2016), lower limits were imposed on the two parameters in the denominator of the exponential term ($c_1, c_2 \geq 200$) to ensure a more robust representation of the measured data. Additionally, since the instruments were calibrated using H_2SO_4 , the relative transmission for all peaks contributing to the H_2SO_4 signal was constrained to unity. As all peaks associated with the selected HA species are in same mass range as sulfuric acid peaks ($m/z = 97, 160, \text{ and } 195$), we assume, as an approximation, that their relative transmission is likewise equal to 1. Therefore, no transmission correction factor, C_{trans} , is applied to these species (see Eq. 4), as it is also the case for H_2SO_4 (see Eq. 3).

Since the relative transmission curve describes the mass-dependent discrimination of analyte ions relative to that of the reagent ions, its applicability to other ionization modes must be carefully considered. The comparable binding strengths of NO_3^- and Br^- to oxidized molecules (Hyttinen et al., 2018) suggests that the corresponding clusters with HOMs are detected with similar

efficiency under the same mass spectrometer settings (Rissanen et al., 2019). Therefore, the transmission curve obtained from the calibrations performed in NO_3^- mode was applied to Br^- mode measurements as well.

2.3.4 Data processing

Ambient data, background measurement data, and sulfuric acid calibration data were processed using the Tofware analysis software package, while transmission calibration data were processed using the MATLAB-based tofTools software, as the functions used for deriving transmission curves were implemented in MATLAB.

Signals from ambient measurements were pre-averaged over 10-minute intervals for all CI-API-ToFs, regardless of their individual acquisition time. This averaging time was selected to better reflect typical data processing practices for these instruments during field measurements, where, depending on the application, averaging times of 30 or even 60 minutes are commonly used. In this campaign, instruments ML1 and ML2 alternated ionization modes every 10 minutes, which made the choice of 10-minute pre-averaging convenient for the purposes of the study. As described in Section 2.3.2, sulfuric acid calibration data were treated differently and averaged over 2-minute intervals. This choice was related to the 10-minute duration of each calibration step: since the signal is often not stable during the initial part of each step, 2-minute averaging provides enough data points per step when evaluating the time series of the sulfuric acid signal. Background measurement data were also averaged over 2-minute intervals, for consistency with the sulfuric acid calibration data. This was also motivated by the relatively short duration of the background measurements (~ 30 min), since using longer averaging times, such as 10 minutes as for the ambient data, would have resulted in too few data points.

In NO_3^- mode, H_2SO_4 and the other strong acids were quantified using Eq. (3) and Eq. (4), respectively. Selected HOM species were quantified using Eq. (5) and subsequently grouped by chemical class (Table S2-S4 in the SI).

$$[H_2SO_4] = S_{H_2SO_4} \times C_{SA}, \quad (3)$$

$$[HA] = \frac{A + HNO_3 \cdot A}{NO_3 + HNO_3 \cdot NO_3 + (HNO_3)_2 \cdot NO_3 + H_2O \cdot NO_3} \times C_{SA} \times C_{inlet}, \quad (4)$$

$$[HOM] = \frac{HOM \cdot NO_3}{NO_3 + HNO_3 \cdot NO_3 + (HNO_3)_2 \cdot NO_3 + H_2O \cdot NO_3} \times C_{SA} \times C_{inlet} \times C_{trans}. \quad (5)$$

In the equations above, C_{SA} , represents the calibration factor derived from the sulfuric acid calibration described in Section 2.3.2. The sulfuric acid concentration is derived using the same set of peaks as those used to calculate $S_{H_2SO_4}$ during the calibration. For species other than sulfuric acid, C_{inlet} accounts for diffusional losses of the sample in the inlet tube and it is calculated based on species-specific gas diffusion coefficients, estimated following the approach of Fuller et al. (1966, 1969; Section S6 in the SI). Finally, C_{trans} is the mass-dependent transmission correction factor, obtained as the inverse of the relative transmission function fitted for each instrument and resolved for each species as detailed in Section 2.3.3. For H_2SO_4 , C_{SA} intrinsically includes both inlet loss and transmission correction factors (Heinritzi et al., 2016; Kürten et al., 2012). For HAs and HOMs, C_{inlet} and C_{trans} are calculated relative to H_2SO_4 , which serves as the calibration reference.

Analogous equations were applied for Br⁻ mode (Section S8 in the SI).

The AB instrument data were treated separately. Analyte signals were processed by normalizing them to respective reagent ion count rate of 1 million counts s⁻¹, as measured at the detector (Lee et al., 2014). This normalization was applied to account for variations in instrument response caused by changes in ion source intensity and detector gain. For compounds that were not directly calibrated, concentrations were estimated using the collision-limited sensitivity of 10 ncps ppt⁻¹ to convert normalized signals to mixing ratios (Aggarwal et al., 2025)

2.3.5 Background measurements and detection limit

Background measurements in NO₃⁻ mode were performed for all CI-APi-ToFs by connecting a high-efficiency particulate air (HEPA) filter (Cytiva HEPA Capsule Versapor™ 1.2 μm HB, 1 pack) to the inlet tube and sampling for approximately 30 minutes. The use of HEPA filter for this type of measurements has been reported in previous studies (Boyer et al., 2024; Häkkinen et al., 2023) and is straightforward to apply during field measurements. Although HEPA filters are conventionally used to remove particles in particle-phase measurements, they can also be used for background measurements of gas-phase species when the compounds of interest are sufficiently low volatile and are efficiently removed by interaction with the filter material. The efficiency of this removal is, however, compound-dependent and the filter should therefore not be considered a universal gas-phase zeroing method. To demonstrate the effectiveness of the approach for the compounds targeted in this study, measurements with and without the HEPA filter are shown in Figure S3 in the SI. The signals of the investigated compounds were strongly suppressed when the filter was installed, although the signals did not decrease completely to zero and the extent of signal decrease differed among sulfuric acid, HAs, and HOMs.

The limit of detection (LOD), defined as the minimum analyte signal that can be distinguished from background noise, can be determined using different approaches. Here, following He et al. (2023), the LOD for H₂SO₄ was calculated as:

$$LOD_{SA} = (\mu + 3\sigma) \cdot C_{SA}, \quad (6)$$

where μ represents the mean value of the normalized sulfuric acid signal during the background measurement, and σ is the corresponding standard deviation. As described by He et al. (2023), for species without a direct calibration, such as HOMs, the LOD calculated using the sulfuric acid calibration factor can be applied, although this generally represents a lower limit of the actual LOD.

To further characterize differences in noise levels between instruments, we calculated a more general LOD, referred to as LOD_{CIMS} , using the same formula but a different definition of μ : for each HOM peak included in the NO₃⁻ mode peak list, a nearby region where no peak was expected was selected, and the signal in this region was fitted and taken as μ . For example, for a peak at nominal masses 308 and 494 Th, the signal at 308.5 and 494.5 Th were used, and similarly for other peaks up to a nominal mass of 620 Th. Since the HOM peaks span a broad m/z range, LOD_{CIMS} arguably provides a more general characterization of the background signal and noise level, rather than a compound-specific detection limit, such as LOD_{SA} , and can thus help in interpreting differences in measurement outcomes between instruments. It should be noted that the detection

limit of an individual compound can also be affected by the presence and intensity of nearby peaks, an effect that is not accounted for in this approach and may become increasingly relevant at higher m/z .

360 Under the relatively clean conditions of the SMEAR II station and during rainy days with limited photochemical activity, the ambient concentrations of sulfuric acid and other analytes were generally low. This made the measurements challenging for all instruments, with several data points approaching the detection limits, but the situation was particularly critical for the AB instrument, for which the lower pressure of the AIM inlet inherently affects sensitivity and increases detection limits. Consequently, the NO_3^- mode did not perform well for this instrument, possibly further affected by contamination issues in
365 the bipolar configuration during the campaign. For this reason, NO_3^- mode data from AB were excluded from the intercomparison (Section 3.3), and the focus was instead on I^- mode measurements. Accordingly, LOD_{SA} and LOD_{CIMS} were not calculated for this instrument.

3 Results and discussion

3.1 Sulfuric acid calibration factor and detection limit

370 Table 5 collects C_{SA} values resulting from the sulfuric acid calibrations (Section 2.3.2), along with the LOD_{SA} and LOD_{CIMS} values calculated based on the background measurements (Section 2.3.5). NO_3^- mode C_{SA} values generally lie in the range $1\text{--}2 \times 10^{10} \text{ cm}^{-3}$, with the exception of EL, which shows a calibration factor of $7.25 \times 10^9 \text{ cm}^{-3}$. The highest C_{SA} is that of EH2, which also exhibits the highest inlet losses, likely due to its comparatively narrower and longer inlet tube combined with a lower inlet flow rate (Y Zhang et al., 2023). Nevertheless, LOD_{SA} values calculated for all instruments are fairly similar, ranging
375 between $1\text{--}3 \times 10^5 \text{ cm}^{-3}$, with ML2 showing the lowest detection limit for H_2SO_4 . LOD_{CIMS} values, instead, are about one order of magnitude lower than LOD_{SA} values, and ML2 again shows a notably lower detection limit. Although the ranking of the instruments based on LOD_{SA} and LOD_{CIMS} is not identical, EH1 and EH2 show overall higher background noise throughout the mass spectrum, while ML2 consistently exhibits lower detection limits compared to the other instruments.

The LOD_{CIMS} values obtained here are overall comparable to the detection limit of $3.6 \times 10^4 \text{ cm}^{-3}$ reported by Jokinen et al.
380 (2012) for 15-minute averaged NO_3^- mode measurements with a CI-API-ToF. By contrast, the LOD_{SA} values obtained in this study are higher than the compound-specific sulfuric acid detection limits of $7.6 \times 10^4 \text{ cm}^{-3}$ and $8.8 \times 10^4 \text{ cm}^{-3}$ reported by He et al. (2023) and Boyer et al. (2024), respectively. However, differences in the background measurement procedure and averaging time need to be considered when comparing these values. Boyer et al. (2024) used HEPA-filter background measurements, as in the present study, and 5-minute averaged data, whereas He et al. (2023) introduced pure nitrogen or
385 synthetic air into the CI inlet and calculated the LOD from 1 h of data averaged to 1 min intervals. In addition, possible incomplete removal of sulfuric acid or sulfuric-acid-containing clusters during HEPA-filter background measurements may have contributed to an overestimation of the derived LOD_{SA} values. For these reasons, the reported LOD values should not be considered strict limits below which the measurements are necessarily unreliable.

For ML1 and ML2, C_{SA} values obtained in Br^- mode are higher than those in NO_3^- mode. This is consistent with the different ion chemistry of NO_3^- and Br^- toward H_2SO_4 , with NO_3^- providing higher sensitivity and therefore lower calibration factors. The longer residence time of the NO_3^- source may also partially contribute to this difference.

We note that for ML1 operated in Br^- mode, the sulfuric acid calibration was performed following an abrupt shift in the ML1 bromide reagent ions behavior that occurred on July 31st; consequently, the data acquired thereafter in this mode are not considered reliable for the remainder of the ACTRIS CI-FI1 campaign. The calibration factor derived from this experiment was therefore not used in the analysis. Instead, a revised calibration factor, C'_{SA} , was derived from the two neighboring calibrations obtained for this instrument with the same experimental setup, one before the campaign and one following the campaign, when the Br^- mode was operating properly. This highlights the importance of regular calibrations, which allowed this deviation from the regular instrument performance to be identified and addressed. Additional details are given in Section S13 in the SI.

Table 5 Sulfuric acid calibration factors and detection limits calculated for the five CI-API-ToFs, reported alongside selected inlet parameters. The C_{SA} value reported for ML1 operating in Br^- mode corresponds to the revised calibration factor (C'_{SA}).

Instrument label	Inlet tube length [mm]	Inlet tube inner diameter [mm]	Inlet flow [L min^{-1}]	NO_3^- mode C_{SA} [cm^{-3}]	Br^- mode C_{SA} [cm^{-3}]	NO_3^- mode LOD_{SA} [cm^{-3}]	NO_3^- mode LOD_{CIMS} [cm^{-3}]
EH1	750.0	15.30	~ 10.0	1.14×10^{10}	-	3.06×10^5	2.01×10^4
EH2	1040.0	15.30	~ 11.0	2.05×10^{10}	-	2.80×10^5	2.74×10^4
EL	550.0	15.30	~ 11.0	7.25×10^9	-	1.85×10^5	1.65×10^4
ML1	970.0	20.18	~ 23.0	1.06×10^{10}	8.58×10^{10}	2.39×10^5	1.06×10^4
ML2	730.0	20.18	~ 21.0	1.31×10^{10}	$*4.59 \times 10^{10}$	1.13×10^5	7.88×10^3

3.2 Transmission curves

Figure 2 shows the relative transmission curves for all the CI-API-ToFs, obtained from the transmission calibrations performed as described in Section 2.3.3. With the exception of EH1, the relative transmission increases above unity and reaches its maximum between m/z 400 and 500, before decreasing and approaching zero beyond m/z 800. This trend is especially pronounced for ML2, which shows a peak relative transmission more than twice that of ML1, the instrument with the second-highest maximum. Considering that the HOM clusters we target in this study have masses between m/z 240–620 in nitrate mode and m/z 260–640 in bromide mode, we expect improved detectability of both monomers and dimers in instruments exhibiting higher transmission in this mass range, in the order $\text{ML2} > \text{ML1} > \text{EL}$, followed by EH2. The curve obtained for EH1, however, shows a different behavior: the relative transmission never exceeds unity and gradually decreases starting from

about m/z 200, approaching zero already around m/z 600. Hence, based on the transmission curve alone, EH1 would be expected to show little to no signal in the mass range of interest.





Figure 2 Relative transmission curves derived for all instruments based on the transmission calibrations with perfluoro compound depletion technique (Heinritzi et al., 2016; C5 – blue, C7 – green, C8 – pink, C9 – yellow). The curves were obtained by least-squares fitting of a 2-fold Gaussian function to the depletion data. Note the different y-axis scale for ML2.

Given the peculiar relative transmission curve of EH1, the resulting C_{trans} for the selected HOMs becomes considerably large, especially for HOM dimers. This occurs because C_{trans} is calculated as $1/f(x)$, and when $f(x)$ is close to zero, the resulting correction factors are unrealistically large relative to actual HOM concentrations. For this reason, the application of the transmission correction factor is arguably unreliable when $f(x)$ is zero or near zero, as it is the case for the EH1 relative transmission function beyond m/z 400.

To assess whether the derived transmission curves are consistent with the observed mass spectra, we compared the spectra recorded by the different CI-API-ToFs in the m/z regions relevant for the species targeted in this study. Specifically, we selected three regions: (i) the range containing the main sulfuric acid signal, corresponding to the bisulfate peak (m/z 97; 85–110), (ii) the region where monoterpene-derived HOM monomers clustered with NO_3^- fall (m/z 240-390), and (iii) the region where monoterpene-derived HOM dimers clustered with NO_3^- appear (m/z 480-630). Figure 3 shows the three selected segments of the unit mass resolution (UMR) mass spectra measured by the five mass spectrometers, averaged over four daytime hours (12:00–16:00) during peak photochemical activity on 1 August 2024.



435

Figure 3 Selected segments of the unit mass resolution (UMR) mass spectra measured by the five CI-API-ToF mass spectrometers operating in nitrate mode, averaged over four daytime hours (12:00–16:00) during peak photochemical activity on 1 August 2024. Peak intensities were normalized to the reagent ion intensities. The three columns show (i) the range containing the main sulfuric acid signal, corresponding to the bisulfate peak (m/z 97; 85–110), (ii) the region where monoterpene-derived HOM monomers clustered with NO_3^- fall (m/z 240–390), and (iii) the region where monoterpene-derived HOM dimers clustered with NO_3^- appear (m/z 480–630). Each row corresponds to a different mass spectrometer. Peaks corresponding to target species included in the NO_3^- mode peak list are highlighted in black.

Overall, the figure show that the instruments do not produce identical mass spectra, even when operated with the same reagent ion. In regions (i) and (ii), the spectra measured by ML1 and ML2 show comparable signal intensities and similar peak patterns. The spectra from EH1 and EL also compare well, although the spectrum measured by EL exhibits slightly higher signal intensity, whereas EH2 shows substantially lower signals in region (ii), consistent with its lower transmission in this mass range compared to the other instruments. In region (iii), the differences between instruments become more pronounced, with EH2 and especially EH1 showing much lower signal intensities compared to ML1 and ML2. This behavior is consistent with the transmission curves shown in Figure 2, which indicate higher transmission for ML1 and ML2 at such high m/z values. Notably, EH1 still shows a clear signal in region (ii), despite the derived relative transmission curve indicating very low transmission at these m/z values. Thus, we estimated (1) to what extent the application of the transmission correction factor C_{trans} (together with the inlet loss correction C_{inlet}) improves the consistency and agreement between measurements from different instruments, and (2) to what extent, in the present case, not applying C_{trans} to the EH1 measurements improves the agreement compared to applying it. To this end, correlation analysis was performed on concentrations of different HOM classes measured by all CI-API-ToFs operating in NO_3^- and Br^- modes over selected days (1–2 August and 27–31 July, respectively). For nitrate mode, the analysis was conducted by (a) applying only C_{SA} , (b) applying C_{SA} , C_{inlet} , and C_{trans} to all instruments, and (c) applying the same corrections to all instruments except EH1. In addition to the coefficient of determination (R^2), the relative median absolute deviation (rMAD) was calculated to quantify how much the measurements differ between instruments. Table 6 summarizes the results for NO_3^- mode. Figure S4 in the SI provides a more immediate visualization of the effect of applying the different correction factors to measurements from the five CI-API-ToFs operating in nitrate mode.

Table 6 Coefficients of determination (R^2) and relative median absolute deviation (rMAD) calculated from concentrations of different HOM classes measured by five CI-API-ToF mass spectrometers on 1–2 August 2024. Values are shown for cases (a) where only C_{SA} is applied, (b) where C_{SA} , C_{inlet} , and C_{trans} are applied to all instruments, and (c) where the same corrections are applied to all instruments except EH1.

HOM class	R_a^2	R_b^2	R_c^2	$rMAD_a$ [%]	$rMAD_b$ [%]	$rMAD_c$ [%]
organic peroxy radicals	0.037	0.012	0.089	22.100	18.000	18.000
closed shell non-nitrate monomers	0.625	0.412	0.697	16.000	18.900	18.900
closed shell organonitrate monomers	0.730	0.555	0.809	20.900	19.000	19.000
dimers	0.288	0.003	0.029	26.100	48.300	19.800

As shown in Table 6, applying all correction factors to all instruments (case b) generally leads to a decrease in R^2 , indicating reduced consistency among instruments. However, excluding C_{trans} for EH1 (case c) improves the agreement compared to both case (a) and (b), yielding higher R^2 values for most HOM groups. Two cases deviate from this general behavior. First, for HOM dimers, the highest R^2 value is observed in case (a), rather than in case (c). A possible explanation for this unexpected result is discussed in Section 3.3. Nevertheless, the rMAD in case (c) is lower than in case (a) and, especially, case (b), indicating improved agreement despite the lower R^2 . Second, for HOM closed shell non-nitrate monomers, R^2 follows the expected behavior and increases from case (a) to case (c); however, the rMAD increases upon applying the full set of

corrections and does not improve when excluding C_{trans} for EH1. This suggests that the observed variability for this group is not primarily driven by the EH1 transmission correction.

These results indicate that differences in transmission among instruments are not negligible and should generally be considered when interpreting mass spectrometry data. On the other hand, they also highlight the importance of critically evaluating whether a given relative transmission curve is suitable for correction purposes and whether the transmission calibration itself provides a reliable representation of instrument behavior. In the present case, indeed, applying the transmission correction factor to all instruments without such evaluation leads to reduced agreement overall, an effect that can be attributed primarily to EH1 (Section S10 in the SI). Therefore, based on all the considerations presented in this section, C_{trans} was applied to all CI-API-ToFs except for EH1 data in the subsequent analysis.

3.3 Measurements in nitrate mode

Figure 4 shows sulfuric acid concentrations measured in nitrate mode by five mass spectrometers participating in ACTRIS CI-F11 during two days of ambient sampling (1-2 August 2024). As discussed in Section 2.3.5, NO_3^- mode data from the AB instrument were excluded from the comparison. The concentrations were calculated according to Eq. (2) and the panels in the two rows, (a-c) and (d-f), present the same analysis but account for two different limits of detection, namely LOD_{SA} and LOD_{CIMS} , respectively. For the following discussion, it should be noted that the abrupt shift in the EL time series occurring around 05:40 on 2 August (visible as an outlier peak in the sulfuric acid time series) was caused by an issue encountered while adjusting the HNO_3 flow mixed with the sheath flow. Hence, statements concerning EL refer to its behavior before this event. Hereafter, we define good agreement as $R^2 > 0.7$, moderate agreement that with $0.4 < R^2 < 0.7$, and poor agreement as $R^2 < 0.4$.




Figure 4 Sulfuric acid concentrations measured in nitrate mode by five CI-APi-ToFs during 1-2 August 2024. Upper and lower panel show the same comparison using two different limits of detection: LOD_{SA} in the upper panel (figures (a) to (c)) and LOD_{CIMS} in the lower panel (figures (d) to (f)). For each panel, the left figure ((a) and (d)) shows the time series of sulfuric acid concentrations, with instrument-specific LOD values indicated by dotted horizontal lines. The central figure ((b) and (e)) shows a box plot summarizing the distributions over the same period; boxes indicate the interquartile range, horizontal lines indicate the median, whiskers represent the non-outlier range, and empty circles denote outliers. Data points below LOD are excluded from the box plot. The right figure ((c) and (f)) shows the hourly median sulfuric acid concentration measured by each instrument against the median of the hourly median concentrations across all five instruments. Dashed lines indicate log-log linear regressions for each instrument, with the corresponding coefficient of determination (R^2) reported in the legend. Points below the corresponding LOD are shown as crosses (X), while outliers are plotted as smaller semi-transparent points; both are excluded from the linear fits.

The instruments show good agreement during periods of peak H_2SO_4 concentrations in the central part of the day, whereas larger differences are observed at lower H_2SO_4 concentrations. Specifically, in the latter case, the time series show higher signal for the instruments with Eisele-type inlets compared to those with MION inlets. This is further confirmed by the scatter plots: when the concentration measured by an individual instrument is plotted against the median concentration reported by the five instruments, the data points are arranged in a fan-like pattern, with increasing dispersion toward lower concentrations. Consistent with the time series, at lower concentrations the data points from different instruments cluster according to inlet type, with similar distributions observed for Eisele-type inlets and for MION inlets.

To explain this difference in behavior, we examined the detection limits calculated for the five instruments in NO_3^- mode. For sulfuric acid, two definitions of detection limit were considered: LOD_{SA} , a compound-specific detection limit calculated for each instrument according to Eq. (6), and LOD_{CIMS} , a background noise level derived following the procedure described in Section 2.3.5. As discussed in Section 3.1, EH1 and EH2 show higher detection limits than ML1 and ML2 in both cases,

consistent with the behavior observed in the time series. In contrast, the detection limit calculated for EL is comparatively low,
510 lying within the range of the other instruments (for LOD_{CIMS}) or even below ML1 (for LOD_{SA}), despite the higher background
signal observed in the time series. One possible explanation is that the C_{SA} value for EL is underestimated relative to the
calibration factors of the other instruments: since this parameter is included in the LOD calculation, such an underestimation
could lead to an apparently lower detection limit. However, if C_{SA} were indeed significantly underestimated, larger
discrepancies in the sulfuric acid concentrations measured by EL would also be expected at higher concentrations, which is
515 not observed.

Notably, when considering LOD_{SA} , only the sulfuric acid concentrations measured by EL remain above the detection limit over
the period considered, although they approach it during nighttime, whereas measurements from all other instruments fall
clearly below it. When applying LOD_{CIMS} , however, measurements from all instruments are above the detection limit. Since
measurements below the detection limit are typically excluded from statistical analyses, these differences can directly affect
520 the measurement outcome and interpretation. This highlights the need for a more rigorous and consistent definition of detection
limit within the CIMS community.

Similarly to Figure 4, Figure 5 shows concentrations of iodic acid, closed-shell organonitrate monomers, and dimers during
the selected days. Analogous figures for other HAs and HOM classes are provided in Section S11 the SI. Iodic acid
concentrations were calculated according to Eq. (4), whereas HOM concentrations were calculated according to Eq. (5); for
525 EH1, the C_{trans} factor was not applied in this case (Section 3.2).




Figure 5 Concentrations of (top panel) iodic acid, (central panel) closed shell organonitrate monomers, and (bottom panel) dimers measured in nitrate mode by five CI-APi-ToFs during 1-2 August 2024. For each panel, the left figure ((a), (d), and (g)) shows concentration time series, with instrument-specific LOD_{CIMS} values indicated by dotted horizontal lines. The central figure ((b), (e), and (h)) shows a box plot summarizing the distributions over the same period; boxes indicate the interquartile range, horizontal lines indicate the median, whiskers represent the non-outlier range, and empty circles denote outliers. Data points below LOD_{CIMS} are excluded from the box plot. The right figure ((c), (f), and (i)) plot shows the hourly median concentration measured by each instrument against the median of the hourly median concentrations across all five instruments. Dashed lines indicate log-log linear regressions for each instrument, with the corresponding coefficient of determination (R^2) reported in the legend. Points below the corresponding LOD_{CIMS} are shown as crosses (X), while outliers are plotted as smaller semi-transparent points; both are excluded from the linear fits.

The difference at low analyte concentrations between instruments with Eisele-type inlets and to those with MION inlets is even more pronounced in the iodic acid plots (Figure 5a-c) than in the sulfuric acid case. Most instruments report higher concentrations in close agreement, with the exception of EH1. At lower concentrations, however, a clear separation of roughly one order of magnitude is observed.

Closed-shell organonitrate monomers (Figure 5d-f), selected across the m/z range 339-387, show a similar behavior. In this case, however, agreement at peak concentrations is even stronger, while differences at lower concentrations are considerably smaller than for iodic acid.

For the dimers (Figure 5g-i), all plots clearly show the separation between instruments with Eisele-type and MION inlets. Furthermore, the box plots suggest that ML1 and ML2 more effectively distinguish this group of HOMs from the background noise, as the difference between higher and lower concentrations is larger, consistent with an improved signal-to-noise ratio in the m/z range 494–620. This is in line with their comparatively lower LODs, which are favored by reduced inlet tube losses and more efficient delivery of ionized sample to the pinhole (Finkenzeller et al., 2024), as well as higher relative transmission in the region of interest compared to the Eisele-type inlets used in this study.

Table 7 summarizes the results of the correlation analysis for the species targeted in NO₃⁻ mode. Overall, the instruments show moderate agreement for sulfuric acid, poor to moderate agreement for the selected HA species, and moderate to good agreement for closed-shell HOM monomers, including both non-nitrate and organonitrate species. In contrast, agreement is poor for the remaining HOM monomer class, namely organic peroxy radicals, as well as for HOM dimers. However, the low coefficient of determination for the former is largely driven by the behavior of EH1 (Figure S4 in the SI). For HOM dimers, the poor agreement mainly reflects the clear separation between instruments between the behavior of instruments with Eisele-type inlets and with MION inlets. For HAs, the reason for the poorer agreement could not be clearly identified. Additional correlation analysis tables are presented in Section S12 in the SI.

Table 7 Statistical parameters of the correlation analysis for the species measured in nitrate mode based on data recorded by five CI-APi-ToFs during 1-2 August 2024. Points below the corresponding *LOD_{CIMS}* are excluded from the statistical analysis. R² is the coefficient of determination, p is the probability value, and parameters a and b are the slope and the offset of the linear fit, respectively.

Species	R ²	p	a	b
sulfuric acid	0.577	<0.001	0.897	0.525
methanesulfonic acid	0.271	<0.001	0.677	1.579
malonic acid	0.561	<0.001	0.995	0.010
iodic acid	0.339	<0.001	0.959	0.140
nitrophenol	0.314	<0.001	0.666	2.068
organic peroxy radicals	0.089	<0.001	0.507	3.043
closed shell non-nitrate monomers	0.678	<0.001	0.947	0.383
closed shell organonitrate monomers	0.809	<0.001	0.974	0.184
dimers	0.029	<0.05	0.312	3.562

3.4 Measurements in bromide mode

Similar to Figure 4 and Figure 5 in Section 3.3, Figure 6 shows the concentrations of different species detected using bromide as the reagent ion, comparing ML1 and ML2. Specifically, the species shown here are SVOCs (Table S3), for which Br^- is more selective towards, and the HOM class of closed-shell non-nitrate monomers (m/z 308-374; Table S3), which are expected to cluster less efficiently with Br^- compared to NO_3^- (see Section 2.2.5).




Figure 6 Concentrations of (top panel) semi-volatile organic compounds, and (bottom panel) closed-shell non-nitrate monomers measured in bromide mode by ML1 and ML2 during 27-30 July 2024. For each panel, the left figure ((a) and (d)) shows concentration time series, while the central figure ((b) and (e)) shows a box plot summarizing the distributions over the same period. Boxes indicate the interquartile range, horizontal lines indicate the median, whiskers represent the non-outlier range, and empty circles denote outliers. The right figure ((c) and (f)) shows the hourly median concentration measured by each instrument against the mean of the corresponding hourly median concentrations from the two instruments. Dashed lines indicate log-log linear regressions for each instrument, with the corresponding coefficient of determination (R^2) reported in the legend. Outliers are plotted as smaller semi-transparent points and are excluded from the linear fits. The top panel of Figure 6 shows that the two instruments are in good agreement ($R^2 = 0.853$; see Table S11 in the SI) for most of the time when measuring SVOCs. In contrast, larger discrepancies exist for species that bind more weakly to Br^- , such as the selected closed-shell non-nitrate monomers. This is illustrated in the bottom panel of Figure 6, where noticeable differences are observed both at low and at high concentrations. With the settings used in the ACTRIS CI-F11 campaign, ML1 appears to more effectively resolve both the lowest and the peak concentrations for these species.

Despite these differences, the overall performance in bromide mode is encouraging. As reported in previous studies (He et al., 2023; Hyttinen et al., 2018), the RH conditions during the campaign (39-100%; see Section 2.1) were not optimal for operating in Br^- mode, since the detection sensitivity decreases noticeably at RH above approximately 20–40%, with the effect expected to be stronger for compounds forming less stable clusters with the reagent ion. Nevertheless, for species that are typically targeted using Br^- , the two instruments show consistent behavior, suggesting that bromide-mode measurements may remain

robust under field conditions and may be less susceptible to variations in ambient moisture than previously anticipated. However, it cannot be excluded that ML1 and ML2 were similarly affected by the variations in RH, and this possibility is further considered in the following section by comparing the Br^- and NO_3^- measurements.

3.5 Measurements in different reagent ion modes: nitrate, bromide, and iodide

Having an instrument measure in multiple reagent ion modes can be useful for identifying issues affecting the individual modes. Accordingly, since two instruments were operated in both NO_3^- and Br^- mode, we investigated whether the response toward selected compounds varied consistently with reagent ion choice in the two instruments. During ambient sampling side by side, ML1 and ML2 switched every 10 minutes among NO_3^- ionization, Br^- ionization and non-ionizing mode; in this way, both instruments and each ionization mode were exposed to the same analyte concentrations under identical ambient conditions.

In Figure 7, we selected two compounds for which both reagent ions are selective (see Section 2.2.5), namely sulfuric acid and the SVOC $\text{C}_{10}\text{H}_{16}\text{O}_6$, and plotted the concentrations measured in Br^- mode against those measured in NO_3^- mode for the two instruments. The figure highlights some differences in behavior between the instruments. For ML2, the high R^2 values in both cases suggest good agreement between the two ionization modes in detecting the selected species over a wide concentration range. The distribution of the data points further indicates that, as expected, NO_3^- is more sensitive to sulfuric acid than Br^- , whereas the opposite behavior is observed for $\text{C}_{10}\text{H}_{16}\text{O}_6$, with which Br^- forms more stable clusters than NO_3^- does (Hytinen et al., 2018). The RH bars in all panels further suggest that RH does not play a major role in driving the performance of Br^- mode relative to NO_3^- mode, despite the reduced Br^- sensitivity expected under the elevated RH experienced during the measurements, further supporting the interpretation proposed in the previous section. For ML1, however, the lower R^2 values and the larger dispersion of the data points indicate greater discrepancies in the detection of sulfuric acid at different concentration levels. In part, this behavior likely owes to the substantial fraction of the Br^- mode measurements falling close to or below the detection limit for ML1. In addition, peak fitting for sulfuric acid may be more challenging in Br^- mode than in NO_3^- mode, since the HSO_4^- peak (m/z 96.960) nearly overlaps with the Br^- water cluster, $(\text{H}_2\text{O})\cdot\text{Br}^-$ (m/z 96.929). The fact that such a discrepancy is not observed for ML2 does not exclude the possibility that peak overlap affected ML1 more strongly, since different instrument settings may lead to different fragmentation of Br^- water clusters, which can affect the intensity of $(\text{H}_2\text{O})\cdot\text{Br}^-$ and its separation from HSO_4^- , making the sulfuric acid peak more difficult to fit in ML1.




Figure 7 Scatter plots of hourly median concentrations measured in bromide mode versus nitrate mode for (a) sulfuric acid and (b) $C_{10}H_{16}O_6$ obtained with ML1 and ML2 during 27-30 July 2024. Dashed lines indicate linear regressions for each plot, with the corresponding coefficient of determination (R^2), number of data points considered (n), and probability value (p) reported. The color of the points represents the relative humidity, as indicated by the color bar in the center. Points falling below the corresponding LOD_{CIMS} are shown as crosses (X), while outliers are plotted as smaller semi-transparent points; both are excluded from the linear fits.

These results indicate that the relative performance of the two ionization modes cannot be explained solely by reagent ion chemistry and that it is also strongly instrument-dependent. In particular, the different behavior observed for ML1 and ML2 suggests that differences in instrument settings and other operating conditions may contribute substantially to the observed differences not only among the Br^- measurement themselves, but also in comparison to the NO_3^- measurement.

Finally, to further investigate how the response toward selected compounds varies with both instrument and reagent ion, we took advantage of the fact that AB also operated in I^- mode and selected a few SVOCs that were expected to be detectable, at least to some extent, by all three reagent ions (NO_3^- , Br^- , and I^-). This provided an opportunity to compare all instruments participating in the ACTRIS CI-FII campaign. Figure 8a shows the diurnal variation of $C_{10}H_{16}O_5$, while Figure 8b presents the concentrations measured by each instrument and ionization mode plotted against the median concentration reported across all of them (additional species are shown in Section S15 in the SI). We note that this species, containing only five oxygen atoms, is not expected to be detected with full sensitivity with NO_3^- (see Section 2.2.5); nevertheless, it was selected because it represents one of the most oxygenated compounds that could be detected by all six instruments in the three ionization modes of interest.




Figure 8 (a) Box plots showing the diurnal variation of $C_{10}H_{16}O_5$ based on measurements conducted on 1-2 August 2024. Data are from five instruments operating in nitrate mode (EH1, EH2, EL, ML1, ML2), one instrument operating in bromide mode (ML2), and one instrument operating in iodide mode (AB). Boxes indicate the interquartile range, whiskers represent the non-outlier range, and full circles denote outliers. For the instruments operating in nitrate mode, instrument-specific LOD_{CIMS} values are indicated by dotted horizontal lines. (b) Scatter plots of hourly median concentrations showing, for each instrument and ionization mode combination, the concentrations measured by the individual instrument-mode against the median concentration measured across all instruments-mode combinations over the same period. Dashed lines indicate linear regressions for each instrument, with the corresponding coefficient of determination (R^2) reported in the figure.

As predicted, in NO_3^- mode, instruments with an Eisele-type inlet do not detect $C_{10}H_{16}O_5$ with optimal sensitivity, especially in the case of EH1, for which the corresponding peak is not clearly distinguishable from the background noise signal. Similarly, for EH2 and EL, the signal is close to or below the detection limit. However, the two instruments with MION inlets can resolve the peak more clearly over the daily concentration range also in this ionization mode, and they show similar behavior. ML2 operating in Br^- mode, which is expected to be the configuration best suited for detecting this species, shows a response very close to the NO_3^- mode measurements from ML1 and ML2. By comparison, under the present sampling conditions and instrument configuration, I^- mode showed substantially poorer agreement with the other instruments and a limited ability to resolve variations in $C_{10}H_{16}O_5$ concentrations. This behavior is not unexpected, given the fundamental differences in operating pressure of the inlet between the bipolar instrument and the other mass spectrometers and, consequently, the higher LOD expected for AB. The substantial difference between this instrument and the others may also contribute to the overall higher concentrations reported by AB, although other factors, such as differences in data processing and concentration calculation, may also play a role. In this regard, for the purpose of the intercomparison, the different response of AB to variations in $C_{10}H_{16}O_5$ concentrations is more relevant than the difference in absolute concentration.

Overall, instruments with the same inlet configuration behave more similarly to each other compared to instruments using the same reagent ion. In particular, the way EH1, EH2, and EL detect $C_{10}H_{16}O_5$ can be clearly distinguished from the behavior of ML1 and ML2, both in NO_3^- and Br^- mode. Likewise, the different behavior observed in I^- mode appears to be more strongly related to the substantially different instrument configuration employed in this mode than to the reagent ion type itself, which would be expected to detect a compound containing five oxygen atoms in a manner broadly comparable to NO_3^- (Hytinen et al., 2018; Riva et al., 2019, 2024). These results further confirm that reagent ion choice is not the only important factor determining selectivity toward selected compounds; instrument configuration, here primarily the inlet design and its operating pressure, also plays a defining role.

4 Conclusions

Understanding the formation and growth of new atmospheric particles requires accurate measurements of key condensable vapors, including sulfuric acid and highly oxygenated organic molecules among others. Chemical ionization mass spectrometry has proven to be well suited for this purpose; however, the many factors influencing instrument performance pose challenges for achieving consistent and comparable measurements across different studies. Intercomparison campaigns are therefore important for assessing how differences in instrument configuration and operation methods influence the response toward the detection and quantification of selected compounds and, consequently, the agreement among instruments.

In this study, we compared data from six chemical ionization mass spectrometers, some of which operated using multiple reagent ions, resulting in a total of ten measurement modes, of which eight were analyzed in this work. The ambient measurements were conducted at a boreal forest site under rainy conditions, resulting in relatively low measured concentrations that were often close to the instruments' limits of detection. Although these conditions introduced some measurement limitations, the ACTRIS CI-FII campaign provided several important insights into the comparability of condensable vapor measurements obtained using different chemical ionization mass spectrometers during field operations.

(1) For sulfuric acid measured in nitrate mode, the traditional calibration based on gas-phase sulfuric acid enabled moderate agreement among the participating instruments ($R^2 = 0.577$), with the highest consistency observed during daytime concentration maxima. Larger discrepancies occurred at lower concentrations, where the measurements approached the corresponding detection limits.

(2) For higher molar-weight compounds, such as highly oxygenated organic molecules, sulfuric acid calibration alone was not sufficient to ensure consistent inter-instrument comparison. The results showed that accounting for mass-dependent transmission differences can substantially improve the agreement for HOM measurements. At the same time, the study demonstrated the importance of critically evaluating whether a derived transmission curve provides a reliable representation of instrument behavior before applying the corresponding correction. In nitrate mode, after applying the corrections discussed in this study, the instruments showed good agreement for closed-shell HOM monomers, with R^2 values of 0.678 and 0.809 for non-nitrate and organonitrate monomers, respectively (0.831 and 0.891 when excluding EH1 from the correlation analysis).

685 In contrast, agreement was poorer for organic peroxy radicals and HOM dimers; the latter showing clear differences between instruments with Eisele-type and MION inlets over the range of concentrations measured.

(3) For organic and inorganic acids of relatively low molecular mass (below m/z 250) considered in this study (HAs), the agreement among instruments was generally poorer than for sulfuric acid and HOMs. The reason for these larger differences could not be clearly identified. One possible explanation considered was the lack of correction for relative transmission differences for these species; however, additional analysis showed that applying this correction did not improve the agreement among instruments. The results therefore suggest that the measurement of these compounds remains challenging, and that further work is needed to better understand and improve their quantitative determination by CIMS.

690 (4) The bromide mode comparison yielded encouraging results for the selected SVOCs, for which ML1 and ML2 showed good agreement ($R^2 = 0.853$) despite the elevated RH conditions encountered during the campaign. The comparison of nitrate and bromide mode measurements further indicates that the elevated and variable RH conditions did not strongly affect the relative behavior of Br^- mode for the selected compounds. These observations suggest that bromide-mode measurements can remain robust under ambient conditions for compounds forming sufficiently stable clusters with Br^- ions.

700 (5) Finally, the comparison between nitrate, bromide, and iodide modes demonstrated that, although all instruments showed qualitatively similar diurnal patterns, differences in signal intensity were often more closely associated with instrument configuration than with reagent ion choice itself. Specifically, instruments sharing the same inlet design showed more similar responses toward selected compounds than instruments operating with the same reagent ion but different configurations.

The ACTRIS CI-FII campaign proves that (a) comparable measurements of condensable vapors by different CIMS instruments are achievable when all relevant calibration and correction factors are critically applied, and (b) inter-instrument comparability is strongly influenced not only by reagent ion chemistry, but also by inlet design and instrument operating conditions, including operating pressure. At the same time, it is worth acknowledging that, since these instruments are generally not calibrated for every targeted analyte, the use of a common calibration factor for groups of species remains an approximation because the sensitivity of a given reagent ion can vary between them, and the application of additional correction factors can only partially account for these differences, which may contribute to the less consistent agreement observed for some species. Hence, detailed guidelines on measurement principles and calibration procedures are urgently needed to ensure good comparability and quantification of condensable vapors across different observations, thereby improving our ability to evaluate and interpret new particle formation processes.

710

Data availability. All data are available upon reasonable request from the corresponding author (cecilia.righi@helsinki.fi). The meteorological data used in this study are available on the SmartSMEAR platform (<https://smear.avaa.csc.fi/>, last access: 13 January 2026). The MATLAB functions used to model sulfuric acid concentrations for the calibration experiment are available on GitHub (https://github.com/ceciliarighi/ACTRIS_CiGas_condensable_vapors, last access: 18 March 2026).

715

Supplement. The supplement related to this article is provided as separate file.

Author contributions. NS conceived the study idea and organized the intercomparison campaign with LA; the details were decided together with the whole intercomparison collaboration. LA, LB, MC, ND, SJ, CL, LQ, CR, NS, RS, TV, QY, and MZW contributed to the instrument installation and operation during the campaign. CR analyzed the data and wrote the paper. All co-authors participated in reviewing the paper and were involved in the scientific interpretation of the results.

Competing interests. At least one of the (co-)authors is a member of the editorial board of Atmospheric Chemistry and Physics.

Acknowledgements. We acknowledge the discussions and technical help from the Faculty of Science at the University of Helsinki and from the Institute for Atmospheric and Earth System Research (INAR), as well as the support from the staff of the Hyytiälä Forest Station. We acknowledge also the ACTRIS Topical Centre Units hosted by University of Helsinki for supporting the implementation and operation of the Units. We thank Douglas Worsnop for valuable discussions and advice on data analysis.

Financial support. This work was supported by the Research Council of Finland through the ACTRIS Central Facilities 2020–2024 programme (FIRI funding, grant no. 329274) and the Virtual Laboratory for Molecular Level Atmospheric Transformations (VILMA; grant nos. 346372 and 364230). Additional support was provided by the European Union through the European Research Council (ERC Starting Grant 2022, BAE project, grant no. 101076311) and the EMME-CARE project. Further funding was received from the ATMOBE project (PID2022-142366OB-I00) and the equipment project (EQC2018-005323-P), funded by MCIN/AEI/10.13039/501100011033 and by the European Regional Development Fund (ERDF, “A way of making Europe”). The CEAM Foundation is partially funded by the Valencian regional government (GVA). Additional support was provided through the joint research project ACTRIS-D_NF (Grant ID: 01LK2001I), funded by the Federal Ministry of Research, Technology and Space in Germany, and by the DFG within the SPP 1294 HALO (grant no. 461450583).

References

- Aggarwal, S., Bansal, P., Wang, Y., Jorga, S., Macgregor, G., Rohner, U., Bannan, T., Salter, M., Zieger, P., Mohr, C., and Lopez-Hilfiker, F.: Identifying key parameters that affect sensitivity of flow tube chemical ionization mass spectrometers, *Atmos. Meas. Tech.*, 18, 4227–4247, <https://doi.org/10.5194/amt-18-4227-2025>, 2025.
- Albrecht, S. R., Novelli, A., Hofzumahaus, A., Kang, S., Baker, Y., Mentel, T., Wahner, A., and Fuchs, H.: Measurements of hydroperoxy radicals (HO₂) at atmospheric concentrations using bromide chemical ionisation mass spectrometry, *Atmos. Meas. Tech.*, 12, 891–902, <https://doi.org/10.5194/amt-12-891-2019>, 2019.
- Alfaouri, D., Passananti, M., Sarnela, N., Kangasluoma, J., and Vehkamäki, H.: An optimization of transmission measurement of an atmospheric pressure interface time-of-flight mass spectrometer (APi-ToF MS), *Environmental Science: Atmospheres*, 5, 1341–1353, <https://doi.org/10.1039/d5ea00029g>, 2025.

- 750 Boyer, M., Aliaga, D., Quéléver, L. L. J., Bucci, S., Angot, H., Dada, L., Heutte, B., Beck, L., Duetsch, M., Stohl, A., Beck, I., Laurila, T., Sarnela, N., Thakur, R. C., Miljevic, B., Kulmala, M., Petäjä, T., Sipilä, M., Schmale, J., and Jokinen, T.: The annual cycle and sources of relevant aerosol precursor vapors in the central Arctic during the MOSAiC expedition, *Atmos. Chem. Phys.*, 24, 12595–12621, <https://doi.org/10.5194/acp-24-12595-2024>, 2024.
- Ehn, M., Thornton, J. A., Kleist, E., Sipilä, M., Junninen, H., Pullinen, I., Springer, M., Rubach, F., Tillmann, R., Lee, B.,
- 755 Lopez-Hilfiker, F., Andres, S., Acir, I. H., Rissanen, M., Jokinen, T., Schobesberger, S., Kangasluoma, J., Kontkanen, J., Nieminen, T., Kurtén, T., Nielsen, L. B., Jørgensen, S., Kjaergaard, H. G., Canagaratna, M., Maso, M. D., Berndt, T., Petäjä, T., Wahner, A., Kerminen, V. M., Kulmala, M., Worsnop, D. R., Wildt, J., and Mentel, T. F.: A large source of low-volatility secondary organic aerosol, *Nature*, 506, 476–479, <https://doi.org/10.1038/nature13032>, 2014.
- Eisele, F. L. and Tanner, D. J.: Ion-assisted tropospheric OH measurements, *J. Geophys. Res.*, 96, 9295–9308,
- 760 <https://doi.org/10.1029/91JD00198>, 1991.
- Eisele, F. L. and Tanner, D. J.: Measurement of the gas phase concentration of H₂SO₄ and methane sulfonic acid and estimates of H₂SO₄ production and loss in the atmosphere, *J. Geophys. Res.*, 98, 9001–9010, <https://doi.org/10.1029/93JD00031>, 1993.
- Finkenzeller, H., Mikkilä, J., Righi, C., Juuti, P., Sipilä, M., Rissanen, M., Worsnop, D., Shcherbinin, A., Sarnela, N., and Kangasluoma, J.: Multiphysical description of atmospheric pressure interface chemical ionisation in MION2 and Eisele type
- 765 inlets, *Atmos. Meas. Tech.*, 17, 5989–6001, <https://doi.org/10.5194/amt-17-5989-2024>, 2024.
- Fuller, E. N., Schettler, P. D., and Giddings, J. C.: A new method for prediction of binary gas-phase diffusion coefficients, *Ind. Eng. Chem.*, 19–27, <https://doi.org/10.1021/ie50677a007>, 1966.
- Fuller, E. N., Ensley, K., and Giddings, C. J.: Diffusion of Halogenated Hydrocarbons in Helium. The Effect of Structure on Collision Cross Sections, *J. Phys. Chem.*, 73, 3679–3685, <https://doi.org/10.1021/j100845a020>, 1969.
- 770 Häkkinen, E., Zhao, J., Graeffe, F., Fauré, N., Krechmer, J. E., Worsnop, D., Timonen, H., Ehn, M., and Kangasluoma, J.: Online measurement of highly oxygenated compounds from organic aerosol, *Atmos. Meas. Tech.*, 16, 1705–1721, <https://doi.org/10.5194/amt-16-1705-2023>, 2023.
- Hari, P. and Kulmala, M.: Station for Measuring Ecosystem-Atmosphere Relations (SMEAR II), *Boreal Environment Research*, 10, 315–322, 2005.
- 775 He, X. C., Shen, J., Iyer, S., Juuti, P., Zhang, J., Koirala, M., Kytökari, M. M., Worsnop, D. R., Rissanen, M., Kulmala, M., Maier, N. M., Mikkilä, J., Sipilä, M., and Kangasluoma, J.: Characterisation of gaseous iodine species detection using the multi-scheme chemical ionisation inlet 2 with bromide and nitrate chemical ionisation methods, *Atmos. Meas. Tech.*, 16, 4461–4487, <https://doi.org/10.5194/amt-16-4461-2023>, 2023.
- Heinritzi, M., Simon, M., Steiner, G., Wagner, A. C., Kürten, A., Hansel, A., and Curtius, J.: Characterization of the mass-
- 780 dependent transmission efficiency of a CIMS, *Atmos. Meas. Tech.*, 9, 1449–1460, <https://doi.org/10.5194/amt-9-1449-2016>, 2016.
- Huang, W., Li, H., Sarnela, N., Heikkinen, L., Tham, Y. J., Mikkilä, J., Thomas, S. J., Donahue, N. M., Kulmala, M., and Bianchi, F.: Measurement report: Molecular composition and volatility of gaseous organic compounds in a boreal forest -

From volatile organic compounds to highly oxygenated organic molecules, *Atmos. Chem. Phys.*, 21, 8961–8977, 2021, <https://doi.org/10.5194/acp-21-8961-2021>, 2021.

785 Hyttinen, N., Kupiainen-Määttä, O., Rissanen, M. P., Muuronen, M., Ehn, M., and Kurtén, T.: Modeling the Charging of Highly Oxidized Cyclohexene Ozonolysis Products Using Nitrate-Based Chemical Ionization, *Journal of Physical Chemistry A*, 119, 6339–6345, <https://doi.org/10.1021/acs.jpca.5b01818>, 2015.

Hyttinen, N., Otkjær, R. V., Iyer, S., Kjaergaard, H. G., Rissanen, M. P., Wennberg, P. O., and Kurtén, T.: Computational
790 Comparison of Different Reagent Ions in the Chemical Ionization of Oxidized Multifunctional Compounds, *Journal of Physical Chemistry A*, 122, 269–279, <https://doi.org/10.1021/acs.jpca.7b10015>, 2018.

Jokinen, T., Sipilä, M., Junninen, H., Ehn, M., Lönn, G., Hakala, J., Petäjä, T., Mauldin, R. L., Kulmala, M., and Worsnop, D. R.: Atmospheric sulphuric acid and neutral cluster measurements using CI-API-TOF, *Atmos. Chem. Phys.*, 12, 4117–4125, <https://doi.org/10.5194/acp-12-4117-2012>, 2012.

795 Jokinen, T., Berndt, T., Makkonen, R., Kerminen, V. M., Junninen, H., Paasonen, P., Stratmann, F., Herrmann, H., Guenther, A. B., Worsnop, D. R., Kulmala, M., Ehn, M., and Sipilä, M.: Production of extremely low volatile organic compounds from biogenic emissions: Measured yields and atmospheric implications, *Proc. Natl. Acad. Sci. U. S. A.*, 112, 7123–7128, <https://doi.org/10.1073/pnas.1423977112>, 2015.

Junninen, H., Ehn, M., Petäjä, T., Luosujärvi, L., Kotiaho, T., Kostiainen, R., Rohner, U., Gonin, M., Fuhrer, K., Kulmala, M.,
800 and Worsnop, D. R.: A high-resolution mass spectrometer to measure atmospheric ion composition, *Atmos. Meas. Tech.*, 3, 1039–1053, <https://doi.org/10.5194/amt-3-1039-2010>, 2010.

Kerminen, V. M., Petäjä, T., Manninen, H. E., Paasonen, P., Nieminen, T., Sipilä, M., Junninen, H., Ehn, M., Gagné, S., Laakso, L., Riipinen, I., Vehkamäki, H., Kurten, T., Ortega, I. K., Dal Maso, M., Brus, D., Hyvärinen, A., Lihavainen, H., Leppä, J., Lehtinen, K. E. J., Mirme, A., Mirme, S., Hörrak, U., Berndt, T., Stratmann, F., Birmili, W., Wiedensohler, A.,
805 Metzger, A., Dommen, J., Baltensperger, U., Kiendler-Scharr, A., Mentel, T. F., Wildt, J., Winkler, P. M., Wagner, P. E., Petzold, A., Minikin, A., Plass-Dülmer, C., Pöschl, U., Laaksonen, A., and Kulmala, M.: Atmospheric nucleation: Highlights of the EUCAARI project and future directions, *Atmos. Chem. Phys.*, 10, 10829–10848, <https://doi.org/10.5194/acp-10-10829-2010>, 2010.

Kim, M. J., Zoerb, M. C., Campbell, N. R., Zimmermann, K. J., Blomquist, B. W., Huebert, B. J., and Bertram, T. H.:
810 Revisiting benzene cluster cations for the chemical ionization of dimethyl sulfide and select volatile organic compounds, *Atmos. Meas. Tech.*, 9, 1473–1484, <https://doi.org/10.5194/amt-9-1473-2016>, 2016.

Kulmala, M., Kontkanen, J., Junninen, H., Lehtipalo, K., Manninen, H. E., Nieminen, T., Petäjä, T., Sipilä, M., Schobesberger, S., Rantala, P., Franchin, A., Jokinen, T., Järvinen, E., Äijälä, M., Kangasluoma, J., Hakala, J., Aalto, P. P., Paasonen, P., Mikkilä, J., Vanhanen, J., Aalto, J., Hakola, H., Makkonen, U., Ruuskanen, T., Mauldin III, R. L., Duplissy, J., Vehkamäki,
815 H., Bäck, J., Kortelainen, A., Riipinen, I., Kurtén, T., Johnston, M. V., Smith, J. N., Ehn, M., Mentel, T. F., Lehtinen, K. E. J., Laaksonen, A., Kerminen, V.-M., and Worsnop, D. R.: Direct Observations of Atmospheric Aerosol Nucleation, *Science*, 339, 943–946, <https://doi.org/10.1126/science.1227385>, 2013.

- Kürten, A., Rondo, L., Ehrhart, S., and Curtius, J.: Calibration of a chemical ionization mass spectrometer for the measurement of gaseous sulfuric acid, *Journal of Physical Chemistry A*, 116, 6375–6386, <https://doi.org/10.1021/jp212123n>, 2012.
- 820 Lambe, A. T., Avery, A. M., Bhattacharyya, N., Wang, D. S., Modi, M., Masoud, C. G., Ruiz, L. H., and Brune, W. H.: Comparison of secondary organic aerosol generated from the oxidation of laboratory precursors by hydroxyl radicals, chlorine atoms, and bromine atoms in an oxidation flow reactor, *Environmental Science: Atmospheres*, 2, 687–701, <https://doi.org/10.1039/d2ea00018k>, 2022.
- Lavi, A., Vermeuel, M. P., Novak, G. A., and Bertram, T. H.: The sensitivity of benzene cluster cation chemical ionization
825 mass spectrometry to select biogenic terpenes, *Atmos. Meas. Tech.*, 11, 3251–3262, <https://doi.org/10.5194/amt-11-3251-2018>, 2018.
- Lee, B. H., Lopez-Hilfiker, F. D., Mohr, C., Kurtén, T., Worsnop, D. R., and Thornton, J. A.: An iodide-adduct high-resolution time-of-flight chemical-ionization mass spectrometer: Application to atmospheric inorganic and organic compounds, *Environ. Sci. Technol.*, 48, 6309–6317, <https://doi.org/10.1021/es500362a>, 2014.
- 830 Lopez-Hilfiker, F. D., Iyer, S., Mohr, C., Lee, B. H., D’ambro, E. L., Kurtén, T., and Thornton, J. A.: Constraining the sensitivity of iodide adduct chemical ionization mass spectrometry to multifunctional organic molecules using the collision limit and thermodynamic stability of iodide ion adducts, *Atmos. Meas. Tech.*, 9, 1505–1512, <https://doi.org/10.5194/amt-9-1505-2016>, 2016.
- Peräkylä, O., Vogt, M., Tikkanen, O.-P., Laurila, T., Kajos, M. K., Rantala, P., Patokoski, J., Aalto, J., Yli-Juuti, taina, Ehn, M., Sipilä, M., Paasonen, P., Rissanen, M., Nieminen, T., Taipale, R., Keronen, P., Lappalainen, H. K., Ruuskanen, T., Rinne, J., Kerminen, veli-matti, Kulmala, M., Bäck, J., and Petäjä, T.: Monoterpenes’ oxidation capacity and rate over a boreal forest: temporal variation and connection to growth of newly formed particles, *Boreal Environment Research*, 19, 293–310, 2014.
- 835 Rissanen, M. P., Mikkilä, J., Iyer, S., and Hakala, J.: Multi-scheme chemical ionization inlet (MION) for fast switching of reagent ion chemistry in atmospheric pressure chemical ionization mass spectrometry (CIMS) applications, *Atmos. Meas. Tech.*, 12, 6635–6646, <https://doi.org/10.5194/amt-12-6635-2019>, 2019.
- Riva, M., Rantala, P., Krechmer, E. J., Peräkylä, O., Zhang, Y., Heikkinen, L., Garmash, O., Yan, C., Kulmala, M., Worsnop, D., and Ehn, M.: Evaluating the performance of five different chemical ionization techniques for detecting gaseous oxygenated organic species, *Atmos. Meas. Tech.*, 12, 2403–2421, <https://doi.org/10.5194/amt-12-2403-2019>, 2019.
- Riva, M., Pospisilova, V., Frege, C., Perrier, S., Bansal, P., Jorga, S., Sturm, P., Thornton, J. A., Rohner, U., and Lopez-
845 Hilfiker, F.: Evaluation of a reduced-pressure chemical ion reactor utilizing adduct ionization for the detection of gaseous organic and inorganic species, *Atmos. Meas. Tech.*, 17, 5887–5901, <https://doi.org/10.5194/amt-17-5887-2024>, 2024.
- Sanchez, J., Tanner, D. J., Chen, D., Huey, L. G., and Ng, N. L.: A new technique for the direct detection of HO₂ radicals using bromide chemical ionization mass spectrometry (Br-CIMS): Initial characterization, *Atmos. Meas. Tech.*, 9, 3851–3861, <https://doi.org/10.5194/amt-9-3851-2016>, 2016.
- 850 Schobesberger, S., Junninen, H., Bianchi, F., Lönn, G., Ehn, M., Lehtipalo, K., Dommen, J., Ehrhart, S., Ortega, I. K., Franchin, A., Nieminen, T., Riccobono, F., Hutterli, M., Duplissy, J., Almeida, J., Amorim, A., Breitenlechner, M., Downard, A. J.,

- Dunne, E. M., Flagan, R. C., Kajos, M., Keskinen, H., Kirkby, J., Kupc, A., Kürten, A., Kurtén, T., Laaksonen, A., Mathot, S., Onnela, A., Praplan, A. P., Rondo, L., Santos, F. D., Schallhart, S., Schnitzhofer, R., Sipilä, M., Tomé, A., Tsagkogeorgas, G., Vehkamäki, H., Wimmer, D., Baltensperger, U., Carslaw, K. S., Curtius, J., Hansel, A., Petäjä, T., Kulmala, M., Donahue, N. M., and Worsnop, D. R.: Molecular understanding of atmospheric particle formation from sulfuric acid and large oxidized organic molecules, *Proc. Natl. Acad. Sci. U. S. A.*, 110, 17223–17228, <https://doi.org/10.1073/pnas.1306973110>, 2013.
- Wang, H., Baker, Y., Shen, H., Wu, R., Kang, S., Zhao, D., Wahner, A., Zorn, S. R., and Mentel, T. F.: Decomposition of Clusters of Oxygenated Compounds with NO₃- by Applying Voltage Scanning to Chemical Ionization Mass Spectrometry in Steady-State Experiments, *Environ. Sci. Technol. Lett.*, 11, 694–700, <https://doi.org/10.1021/acs.estlett.4c00276>, 2024.
- Wang, M., He, X. C., Finkenzeller, H., Iyer, S., Chen, D., Shen, J., Simon, M., Hofbauer, V., Kirkby, J., Curtius, J., Maier, N., Kurtén, T., Worsnop, D. R., Kulmala, M., Rissanen, M., Volkamer, R., Tham, Y. J., Donahue, N. M., and Sipilä, M.: Measurement of iodine species and sulfuric acid using bromide chemical ionization mass spectrometers, *Atmos. Meas. Tech.*, 14, 4187–4202, <https://doi.org/10.5194/amt-14-4187-2021>, 2021.
- Zhang, W., Xu, L., and Zhang, H.: Recent advances in mass spectrometry techniques for atmospheric chemistry research on molecular-level, <https://doi.org/10.1002/mas.21857>, 1 September 2024., 2024.
- Zhang, Y., Liu, R., Yang, D., Guo, Y., Li, M., and Hou, K.: Chemical ionization mass spectrometry: Developments and applications for on-line characterization of atmospheric aerosols and trace gases, <https://doi.org/10.1016/j.trac.2023.117353>, 1 November 2023., 2023.
- Zhao, B., Donahue, N. M., Zhang, K., Mao, L., Shrivastava, M., Ma, P. L., Shen, J., Wang, S., Sun, J., Gordon, H., Tang, S., Fast, J., Wang, M., Gao, Y., Yan, C., Singh, B., Li, Z., Huang, L., Lou, S., Lin, G., Wang, H., Jiang, J., Ding, A., Nie, W., Qi, X., Chi, X., and Wang, L.: Global variability in atmospheric new particle formation mechanisms, *Nature*, 631, 98–105, <https://doi.org/10.1038/s41586-024-07547-1>, 2024.